

Trypan Blue Tissue Culture

Sample Name:	Sample Name	Created Date/Time:	2026-01-23 10:56:41
Account:	General	Passage:	
Software Version:	v4.3.4	Firmware Version:	11.66
Instrument SN:	s-07097	Record:	20260123105641
Description:			

Results

Last Processed Date/Time:	2026-01-23 10:56:41	Gating Min Diameter (um):	4
Gating Max Diameter (um):	20	Live Cell Count:	186
Dead Cell Count:	96	Total Cell Count:	282
% Viability:	65.9574	Mean Diameter (um):	14.03
Live Cells/mL:	1.09e+06	Dead Cells/mL:	5.64e+05
Total Cells/mL:	1.66e+06		



Dilution

Target Cells/mL:	N/A	Target Number of Cells:	N/A
Target Volume (mL):	0	Original Volume (mL):	0
Dilution Guidance:			





Protocol


Protocol:	Scm	DirectPipette (or Slide):	Pipette
Chamber Height:	100	Dilution Factor:	2
Protocol Min Diameter (um):	4	Protocol Max Diameter (um):	20
Live Roundness:	35	Dead Roundness:	15
Stained Threshold:	5		
Small Cell Mode:	True		
Irregular Cell Mode:	False		

Sample Images



Sample Histograms





Cluster Size