From: Taras Stecy TStecy@sbgh.mb.ca @ Subject: reusable silicone style mask

Date: April 8, 2020 at 14:23

To: Colleen Reidke Creidke@wrha.mb.ca

Hi Colleen:

One fit test was conducted today on the reusable silicone style respirator.

Attached are the results.

As well, attached is a picture of the fit test event

The fit test was successful.

Note:

Please ignore the name of the respirator in the report.

A field with an N95 was selected as there are no choices in the drop down menu for silicone reusable respirators.

Taras Stecy

St Boniface Hospital

Safety Coordinator and Biosafety Officer

Occupational Health and Safety

409 Tache Ave., Winnipeg Mb.

T: 204-237-2184 C: 204-479-9173 F: 204-237-2041 e: tstecy@sbgh.mb.ca

FIT TEST REPORT 4/8/2020

ID NUMBER	SBH		
LAST NAME	BERENA	CUSTOM1	
FIRST NAME	CHARLENE	CUSTOM2	
COMPANY		CUSTOM3	
LOCATION		CUSTOM4	
TEST DATE	4/8/2020 11:03	PORTACOUNT S/N	8038201220
DUE DATE	4/8/2022	N95 COMPANION	Υ
RESPIRATOR	AOSAFETY PLEATS PLUS N95 [100]	PROTOCOL	CSA Z94.4-2011
MANUFACTURER	AOSAFETY	PASS LEVEL	100
MODEL	PLEATS PLUS		
MASK STYLE	N95	APPROVAL	
MASK SIZE	S/M	EFFICIENCY<99%	True

EXERCISE	DURATION (sec.)	FIT FACTOR	PASS
NORMAL BREATHING	60	200+	Y
DEEP BREATHING	60	200+	Y
TURN HEAD SIDE TO SIDE	60	200+	Y
NOD HEAD UP AND DOWN	60	200+	Y
TALK OUT LOUD	60	200+	Y
BENDING OVER	60	200+	Y
NORMAL BREATHING	60	200+	Υ
		000.	

OVERALL FF		200+	т
FIT TEST OPERATOR		DATE	
	Q. R.		
NAME		DATE	
	CHARLENE BERENA		

Note:

Respirator Fit Test Card
Name: CHARLENE BERENA Test Date Test Date: 4/8/2020 ID: SBH Next Test Date: 4/8/2022

Respirator Mfg: AOSAFETY Model: PLEATS PLUS Results Overall FF: 200+ FF Pass Level: 100 Style: N95 Pass: Y Size: S/M Operator: Q. R. Protocol: CSA Z94.4-2011

Fit Test Method: QNFT using TSI PortaCount
*** Your company contact information here ***



