

**EXCELLENT***Instrumentation Services***SPECIALISED IN :**

Supply, Installation. Commissioning,
ISO9001 Calibration, Validation,
Manpower Supply & Maintance Contract

20, New Nilgiri C.H.S., L-2/3, Sector7,
Shree Nagar, Wagale Ind. Estate,
Thane (W) 400604, Maharashtra.

Mob. 9820889164 / 59 7506177332

E-mail – excellentservices44@yahoo.co.in

www.excellentinstrumentationservices.co.in

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE NO:		EIS/VI/TE/2024-25/1245	
CLIENT NAME & ADDRESS:		ARK Scientific, 5-83, Drithi Towers, Prasanthi Nagar IDA, Kukatpally, Near Naina Garden Lane, Hyderabad, Telangana 500072	
SOP NO.	EIS/06/24-25	Inst Name : Temperature Element	
INST. SR. NO.	VJSEN-42	TAG NO : VAI/TE-01	CaL Frq. : 12 Months
MAKE	VAISHNAV INST	PLANT :	
RANGE	0 To 200 Deg C	INSTRUMENT ACCEPTED ACCURACY: +/- 0.5 % F.S.D	
OR RANGE	NA	LOCATION :	
CALI RANGE	0 To 200 Deg C	CAL DATE : 23-01-2025	DUE DATE : 22-01-26
LEAST COUNT	NA	Other:	

CALIBRATION RESULTS

Sr.No	STD. Input		Converted to Output/Display	Actual Output/Display				Remarks
				Before Calibration		After Calibration		
	Actual	Deviation	Actual	Deviation				
	Deg C	Ohm	Deg C	Deg C	Deg C	Deg C		
1	35.0	113.60	35.0	35	0	-	-	OK
2	50.0	119.40	50.0	50	0	-	-	OK
3	100.0	138.50	100.0	101	1	-	-	OK
4	150.0	157.31	150.0	150	0	-	-	OK
5	200.0	175.84	200.0	200	0	-	-	OK

DETAILS OF CALIBRATION STANDARDS USED:

NO	NAME	RANGE	SR. NO	MAKE/ MODEL	CERTIFICATE NO.	DUE DATE
1	Digital Calibrator	Ohm	SE-406	Sonit	2023- 24/CFC/MUM/120/4	10-09-2026
2	Temperature Calibrator	Amb to 650 Deg C	HT07021560	Nagman 650 H	2023- 24/CFC/MUM/120/1	10-09-2026

***THE STANDARD USED TRACEABLE TO NATIONAL STANDARD CALIBRATION**

***REMARK: INSTRUMENT WORK ON OK CONDITION**

CHECKED BY
(Amit Jain)

CALIBRATED BY
(Excellent Instrumentation Services)