

HARDWARE: BWT-2508



© 2026 Brain Waves Tech. All Rights Reserved

DEVICE OVERVIEW

Technical Category	Specification / Detail
Model Designation	BWT-2508
Sensor Quantity	3
Sensor Technology	Medical-Grade Dry Electrodes
Anatomical Sensor Placement	Fixed Position, FP1 Forehead Axis (Prefrontal Cortex)
Setup Time	2 minutes
Wireless Protocol	Bluetooth 5.2 Low Energy
Power System	Internal Rechargeable Battery
Active Battery Runtime	Up to 9 Hours
Data Quality Rating	High-Fidelity Signal-to-Noise Ratio (Good/Clinical Grade)
Headband Design	Ergonomic, Adjustable Fixed Position Strap

Global Certifications & Compliance

CE Certified: Meets European health, safety, and environmental protection standards.

ISO 9001:2015 Certified: Standardized quality management for global reliability.

ISO 13485 Certified: Compliant with international quality management standards for medical devices.

RoHS Compliant: Free from hazardous substances, ensuring safe daily contact

