



Native Mathematica fitreport: Hg 404.770

File: 05_Hg_calibration_405.020nm_404.770_at_50_nm_hg_2nm_scan_REAL_FOR_CURVEFIT.dat

Fit window: {405.05, 405.42} nm



parameter	estimate	standard error
c	-0.0931865	0.130797
d	0.253533	0.605427
amp	8.9496	0.148527
mu	405.219	0.0011192
sigma	0.0613799	0.00149447
RMS residual	0.147319	
max residual	0.311692	
true Hg wavelength	404.77	
used for calibration?	True	