



# Native Mathematica fitreport: Hg 365.119 diagnostic

File: 03\_Hg\_calibration\_365.365nm\_365.12\_at\_50\_nm\_hg\_2nm\_scan\_REAL\_FOR\_CURVEFIT.dat

Fit window: {365.5, 365.85} nm



| parameter             | estimate  | standard error |
|-----------------------|-----------|----------------|
| c                     | -0.239951 | 0.145244       |
| d                     | 3.8514    | 0.57789        |
| amp                   | 8.39922   | 0.142734       |
| mu                    | 365.662   | 0.00111694     |
| sigma                 | 0.0679802 | 0.00163876     |
| RMS residual          | 0.229822  |                |
| max  residual         | 0.484286  |                |
| true Hg wavelength    | 365.119   |                |
| used for calibration? | False     |                |