



Native Mathematica fitreport: Hg 546 repeat wide

File: 07_Hg_calibration_546.250nm_562hg_2nm_scan_REAL_FOR_CURVEFIT.dat

Fit window: {546.55, 546.9} nm



parameter	estimate	standard error
c	-0.252561	0.0509003
d	1.08603	0.208745
amp	2.95771	0.0526362
mu	546.694	0.000967592
sigma	0.0759202	0.00146106
RMS residual	0.0458885	
max residual	0.124447	
true Hg wavelength	546.226	
used for calibration?	False	