30031C	Spinning Rotor Gages, below 0.1 Pa, Additional Gas	5756
30032S	Special Test of Spinning Rotor Gages, Transition Range (above 0.1 Pa)	At Cost
30034C	Ionization Gages, 10^{-4} Pa to 10^{-1} Pa, Nitrogen Gas	5459
30035C	Ionization Gages, 10^{-5} Pa to 10^{-1} Pa, Nitrogen Gas	6657
30036C	Ionization Gages, 10^{-7} Pa to 10^{-1} Pa, Nitrogen Gas	7537
30037C	Ionization Gages, Additional Filament or Gas for Above Tests	At Cost
30038C	Ionization Gages, NIST Supplied Gage Tube for Above Tests	303
30040S	Special Tests of Low-Pressure Gages	At Cost
30050S	Special Tests of Vacuum Gages	At Cost
30060S	Special Tests of Leak Artifacts (10^{-13} mol/s to 10^{-6} mol/s)	At Cost
30061C	Helium Leaks, Primary Calibration (10^{-13} mol/s to 10^{-6} mol/s)	6464
30062C	Helium Leaks, Comparison Calibration (10^{-13} mol/s to 10^{-9} mol/s)	5033
30063S	Special Tests of Low-Gas-Flow Instruments	At Cost

Fees are subject to change without notice.

NOTE: Due to the time and effort required preparing vacuum instrumentation for calibration it is particularly important that they be known to be in proper operating condition when they are submitted to NIST. Equipment will be inspected upon receipt and the customer notified of any obvious damage. If the schedule permits, we will cooperate with the customer's efforts to repair or replace damaged equipment so that the calibration of their equipment can proceed. However, concealed damage or operational deficiencies most likely will not be detected before the instrument is operating on the vacuum system or the calibration has started; in such cases, **if the equipment cannot be calibrated, we will charge 20% of the regular calibration fee for low-pressure transducers and 30% of the regular fee for spinning rotor and ionization gages.**