photobiologylnOut Version 0.2.1 User Guide

Pedro J. Aphalo

April 10, 2015

Contents

1	Introduction	
2	Examples	

1 Introduction

```
library(photobiologyg)
library(photobiologygg)
library(photobiologyInOut)
```

This package defines functions for importing spectral data from different instruments (Table 1) and simulation models (Table 2).

R function	Instrument	Program
read_ooss_file() read_oojaz_file() read_macam_file() read_licor_file()	Ocean Optics spectrometers Ocean Optics Jaz Macam LI-COR LI-1800	SpectraSuite instrument instrument PC1800 (MS-DOS)

Table 1: Functions for importing measured spectral emission data

R function	Simulation model	version
\/	TUV (Sasha Madronich) TUV (Sasha Madronich)	,

Table 2: Functions for importing simualted spectral data from models

2 Examples

Reading a file geenrated by Ocean Optics' Jaz spectrometer.

```
jaz.spct <- read_oojaz_file(file = "spectrum.JazIrrad")

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -3.2

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032</pre>
```

Photon and energy irradiances.

```
q_irrad(jaz.spct, PAR()) * 1e6 # umol m-2 s-1

## PAR
## 47.43464
## attr(,"time.unit")
## [1] "second"
## attr(,"radiation.unit")


## [1] "photon irradiance total"

e_irrad(jaz.spct, PAR()) # W m-2

## PAR
## 10.12685
## attr(,"time.unit")
## [1] "second"
## attr(,"radiation.unit")
## [1] "energy irradiance total"
```

Plotting the spectrum.

```
plot(jaz.spct)
## Warning: Removed 1 rows containing missing values (geom_path).
```




All in one statement.

```
plot(read_oojaz_file(file = "spectrum.JazIrrad"))

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -3.2

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032

## Warning: Removed 1 rows containing missing values (geom_path).
```



As above but limiting the wavelength range plotted.


```
plot(read_oojaz_file(file = "spectrum.JazIrrad"), range = c(315,900))

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -3.2

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.00062
```



Adding our custom "adaptive" smoothing.

```
plot(smooth_spct(read_oojaz_file(file = "spectrum.JazIrrad")), range = c(315,900))


## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -3.2

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032

## Warning in smooth_spct.source.spct(read_oojaz_file(file = "spectrum.JazIrrad")):
89 'bad' estimates in spectral irradiance


## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.027

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.027
```




Now a file from an Ocean Optics' Q6500? spectrometer.

```
plot(read_ooss_file(file = "spectrum.SSIrrad"))
```



Macam's single column DTA files can also be imported.

```
plot(read_macam_file(file = "spectrum.DTA"))
```



And a file generated by LI-COR's PC1800 program for the LI-1800 sepctroradiometer.

```
read_licor_file(file = "spectrum.PRN")
##
        w.length
                    s.e.irrad
##
             300 6.053193e-05
    1:
##
    2:
             301 1.333399e-04
             302 8.702762e-05
    3:
##
##
    4:
             303 1.279194e-04
             304 1.810952e-04
##
    5:
##
## 597:
             896 2.060117e-03
             897 1.571039e-03
## 598:
## 599:
             898 1.029629e-03
             899 5.312083e-04
## 600:
             900 2.383262e-04
## 601:
```

In all cases as much information as possible is decoed, and the data file headers are preserved as comments in the source.spct objects.

```
cat(comment(jaz.spct))
## Ocean Optics:
## Jaz Absolute Irradiance File
```

```
## Date: Tue Feb 03 09:44:41 2015
## User: jaz
## Dark Spectrum Present: Yes
## Processed Spectrum Present: Yes
## Spectrometers: JAZA1065
## Integration Time (usec): 193000 (JAZA1065)
## Spectra Averaged: 3 (JAZA1065)
## Boxcar Smoothing: 5 (JAZA1065)
## Correct for Electrical Dark: Yes (JAZA1065)
## Strobe/Lamp Enabled: No (JAZA1065)
## Correct for Detector Non-linearity: Yes (JAZA1065)
## Correct for Stray Light: No (JAZA1065)
## Number of Pixels in Processed Spectrum: 2048
## Fiber (micron): 3900
## Collection Area: 0.119459
## Int. Sphere: No
```

A variable with the user supplied date and time data, or the date read from the header (the text itself) not the file date as the file date may noy reflect the creation date and time.

```
jaz.spct <- read_oojaz_file(file = "spectrum.JazIrrad", date = NULL)</pre>
## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -3.2
## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032
jaz.spct
         w.length s.e.irrad
     1: 189.2849 0.00000000 2015-02-03 09:44:41
##
    2: 189.7444 0.00000000 2015-02-03 09:44:41
    3: 190.2040 0.00000000 2015-02-03 09:44:41
     4: 190.6635 0.00000000 2015-02-03 09:44:41
##
##
     5:
         191.1229 0.00000000 2015-02-03 09:44:41
##
## 2043: 1031.7194 -0.01576508 2015-02-03 09:44:41
## 2044: 1032.0767  0.00227682  2015-02-03  09:44:41
## 2045: 1032.4341 -0.03179707 2015-02-03 09:44:41
## 2046: 1032.7911 0.00224783 2015-02-03 09:44:41
## 2047: 1033.1483 -0.02682935 2015-02-03 09:44:41
```

```
jaz.spct <- read_oojaz_file(file = "spectrum.JazIrrad", date = today())

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -3.2

## Warning in range_check(x, strict.range = strict.range): Negative spectral energy
irradiance values; minimun s.e.irrad = -0.032

jaz.spct

## w.length s.e.irrad date

## 1: 189.2849 0.00000000 2015-04-10</pre>
```

```
## 2: 189.7444 0.00000000 2015-04-10

## 3: 190.2040 0.00000000 2015-04-10

## 4: 190.6635 0.00000000 2015-04-10

## 5: 191.1229 0.00000000 2015-04-10

## ---

## 2043: 1031.7194 -0.01576508 2015-04-10

## 2044: 1032.0767 0.00227682 2015-04-10

## 2045: 1032.4341 -0.03179707 2015-04-10

## 2046: 1032.7911 0.00224783 2015-04-10

## 2047: 1033.1483 -0.02682935 2015-04-10
```

The output from the TUV model can be imported either by editing it before import, or by making a simple edit to the output routine of TUV. This function is known to work with TUV version 5.0 output. The output from TUV can contain a variable number of spectra in "parallel" columns, which are *melted* into a single column, with a factor with letter as levels, a numeric variable with the zenith angle and a POSIXct column with times. A date needs to be always supplied as the output file from TUV has only time of day information.

```
tuv.spct <- read_tuv_file(file = "usrout.txt", date = ymd("2014-03-21"))</pre>
summary(subset(tuv.spct, spectrum == "A"))
## wavelength ranges from 280.5 to 761.5 nm
## largest wavelength step size is 1 nm
## spectral irradiance ranges from 3.041e-15 to 1.947 W m-2 nm-1
## energy irradiance is 636.7 W m-2
## photon irradiance is 2928 umol s-1 m-2
tuv.spct
##
         w.length spectrum s.e.irrad angle
                                                             date
          280.5 A 3.041e-15 1.829 2014-03-21 12:00:00 281.5 A 1.164e-13 1.829 2014-03-21 12:00:00
##
     1:
##
     2:
                       A 1.824e-12 1.829 2014-03-21 12:00:00
##
     3:
            282.5
##
     4:
           283.5
                      A 9.874e-12 1.829 2014-03-21 12:00:00
     5:
           284.5
                       A 1.258e-10 1.829 2014-03-21 12:00:00
##
##
## 3852:
                        H 0.000e+00 103.213 2014-03-21 19:00:00
           757.5
## 3853:
           758.5
                        H 0.000e+00 103.213 2014-03-21 19:00:00
## 3854:
           759.5
                        H 0.000e+00 103.213 2014-03-21 19:00:00
## 3855:
           760.5
                        H 0.000e+00 103.213 2014-03-21 19:00:00
## 3856:
           761.5
                     H 0.000e+00 103.213 2014-03-21 19:00:00
```

It is possible to extract individual spectra with subset, or as done here plot them in different panels.

```
plot(tuv.spct, annotations = c("colour.guide")) +
  facet_wrap(~date, ncol = 2)
```

