photobiologyLamps Version 0.1.15
Catalogue of Lamps

Pedro J. Aphalo

November 14, 2014

1 Introduction

We will plot the emission spectra of the different lamps for which data is provided in the pacakge. We plot side-by-side the lamp output as spectral energy irradiance and as spectral photon irradiance. All spectra are normalized to an area of one under the whole curve.

```
library(ggplot2)
library(photobiology)
library(photobiologyLamps)
library(photobiologygg)
```

We define a function to do the actual plotting so as to not repeat code, and to make changes easier in the future.

```
lamp.plotter <- function(lamp.name, w.low=250.0, w.high=900.0, scaled="area"){
  w.band <- waveband(c(w.low, w.high))
  object.name <- paste(lamp.name, ".spct", sep="")
a.spct <- copy(get(object.name))
# a.spct <- trim_spct(a.spct, w.band, fill = NA)
e2q(a.spct, byref=TRUE)
print(plot(a.spct, unit="energy") + labs(title=object.name) + theme_grey(10))
# print(plot(a.spct, unit="photon") + labs(title=object.name) + theme_grey(10))
}
```

2 UV-B lamp spectra

```
UVB.lamps <- c("philips.tl01.bentham", "philips.tl12.bentham")
for (lamp in UVB.lamps) {
  lamp.plotter(lamp.name=lamp)
}
```






```
UVB.M.lamps <- c("philips.tl12", "qpanel.uvb313")
for (lamp in UVB.M.lamps) {
  lamp.plotter(lamp.name=lamp)
}
```






```
fig_temp <- ggplot(data=qpanel.uvb313.temperature.dt,
                   aes(x=w.length, y=s.e.irrad, colour=factor(temperature))) +
               scale_colour_brewer(type="seq", palette="YlOrRd")
fig_temp <- fig_temp + geom_line() +
  labs(x="Wavelength (nm)", y="Spectral energy irradiance (W m-2 nm-1)",
       colour="Temperature (C)")
print(fig_temp)
```



3 UV-A lamp spectra

```
UVA.lamps <- c("qpanel.uva340")
for (lamp in UVA.lamps) {
  lamp.plotter(lamp.name=lamp)
}
```



4 Narrow spectrum VIS lamps

```
colour.lamps <- c("philips.tld36w.15", "philips.tld36w.18")
for (lamp in colour.lamps) {
  lamp.plotter(lamp.name=lamp)
}
```





5 Broad VIS lamps

```
white.lamps <- c("incandescent.60w", "osram.36w.25", "philips.pls11w.827",
                 "philips.tld36w.83", "philips.tld36w.865", "philips.tld36w.89",
                 "philips.tld36w.92", "philips.tll36w.950", "sylvania.215w.vho")
for (lamp in white.lamps) {
  lamp.plotter(lamp.name=lamp)
}
```


incandescent.60w.spct
UVA
0.0653
PAR
14
E (W m⁻²)
Spectral energy irradiance E(λ) (W m⁻² nm⁻¹)
0.00
0.05
0.10
0.15
0.20
400
600
800
Wavelength (nm)


osram.36w.25.spct
UVA
0.0502
PAR
15.1
E (W m⁻²)
Spectral energy irradiance E(λ) (W m⁻² nm⁻¹)
436
546
578
406
492
1020
812
0.00
0.05
0.10
0.15
0.20
500
750
1000
Wavelength (nm)


philips.pls11w.827.spct
UVA
0.38
PAR
11
E (W m⁻²)
Spectral energy irradiance E(λ) (W m⁻² nm⁻¹)
612
546
436
366
406
488
710
0.0
0.1
0.2
0.3
0.4
400
600
800
Wavelength (nm)


philips.tld36w.83.spct
Spectral energy irradiance E(λ) (W m⁻² nm⁻¹)
E (W m⁻²)
UVA
0.172
PAR
7.73
365
405
436
488
545
587
613
652
689
710
761
812
Wavelength (nm)


philips.tld36w.865.spct
Spectral energy irradiance E(λ) (W m⁻² nm⁻¹)
E (W m⁻²)
UVA
0.0921
PAR
7.57
365
405
436
488
546
588
613
651
688
710
812
Wavelength (nm)


philips.tld36w.89.spct
Spectral energy irradiance E(λ) (W m⁻² nm⁻¹)
E (W m⁻²)
UVA
0.0672
PAR
8.76
365
405
437
488
546
588
613
652
689
710
761
812
Wavelength (nm)







6 Calibration lamps

```
FEL.spct <- FEL_spectrum(250:900)
D2.spct <- D2_spectrum(250:900)
calibration.lamps <- c("FEL","D2")
```

```
for (lamp in calibration.lamps) {
  lamp.plotter(lamp.name=lamp, scaled=NULL)
}
```