

Supplementary Sensitivity Analysis III

TABLE S6. BLAND-ALTMAN AGREEMENT SUMMARY AT THE PATIENT LEVEL (N = 13) FOR MEAN CLINICAL ROI (CROI) VALUES OF STORAGE MODULUS (G_d) AND LOSS MODULUS (G_l). FOR EACH MODEL, THE TABLE REPORTS THE MEAN BIAS AND THE 95% LIMITS OF AGREEMENT RELATIVE TO THE REFERENCE 3D-MRE-DERIVED VALUES. SMALLER ABSOLUTE BIAS AND NARROWER LIMITS OF AGREEMENT INDICATE BETTER AGREEMENT.

Model	G_d bias (kPa)	G_d LoA (kPa)	G_d CCC [95% CI]	G_l bias (kPa)	G_l LoA (kPa)	G_l CCC [95% CI]
Random init.	+0.127	[-1.34, +1.60]	0.917 [0.793, 0.968]	+0.102	[-1.16, +1.37]	0.877 [0.760, 0.939]
AE (Full Finetuning)	+0.136	[-1.09, +1.37]	0.951 [0.849, 0.985]	+0.118	[-0.72, +0.95]	0.953 [0.874, 0.983]
AE (Warmup-PerBlockLR)	+0.114	[-1.08, +1.31]	0.952 [0.855, 0.985]	+0.089	[-0.77, +0.95]	0.949 [0.882, 0.979]
AE (Warmup-LRRreset)	+0.119	[-1.05, +1.29]	0.956 [0.863, 0.986]	+0.108	[-0.63, +0.85]	0.962 [0.906, 0.985]
AE (Fixed)	+0.109	[-1.00, +1.22]	0.960 [0.877, 0.988]	+0.088	[-0.66, +0.84]	0.962 [0.906, 0.985]
Rot (Full Finetuning)	+0.078	[-0.89, +1.05]	0.966 [0.913, 0.987]	+0.068	[-1.04, +1.18]	0.911 [0.811, 0.960]
Rot (Warmup-PerBlockLR)	+0.008	[-1.07, +1.09]	0.957 [0.901, 0.982]	+0.003	[-1.21, +1.22]	0.889 [0.791, 0.943]
Rot (Warmup-LRRreset)	+0.046	[-1.00, +1.09]	0.960 [0.904, 0.984]	+0.026	[-1.20, +1.25]	0.887 [0.783, 0.943]
Rot (Fixed)	+0.019	[-1.10, +1.14]	0.953 [0.896, 0.979]	+0.007	[-1.28, +1.29]	0.874 [0.764, 0.934]

95% CIs for the Bland-Altman LoA endpoints are provided in Supplementary Table S1. CCC is reported as Lin's concordance correlation coefficient with 95% CIs via Fisher z-transformation. Log-ratio Bland-Altman, percentage Bland-Altman, and bootstrap CCC 95% confidence intervals at both slice and patient levels are reported in Supplementary Table S3. Best per-column values are shown in bold.