

SYNTHETIC TRAINING FIXTURE - NOT A
REAL STANDARD OR CERTIFICATION
RECORD

SYN-EMC-LAB-303 ESD OBSERVATION
SHEET

SYN-PDF-SCAN-01-SEC-01
START THE SYNTHETIC ESD SEQUENCE
WITH CONTACT DISCHARGE ON CONDUCTIVE
POINTS.

SYN-PDF-SCAN-01-SEC-02
USE AIR DISCHARGE ONLY FOR
INSULATING SURFACES OR INACCESSIBLE
CONDUCTIVE POINTS.

SYN-PDF-SCAN-01-SEC-03
THE DEMO RECOVERY LOG RECORDS THE
RESET TYPE AND TIMESTAMP.

SYN-PDF-SCAN-01-SEC-04
IF BEHAVIOR CHANGES AFTER DISCHARGE,
SAVE THE WAVEFORM AND OPERATOR NOTE
BEFORE POWER CYCLING.