

JM Test Systems, LLC.

Calibration Datasheet

Work Order: 576563-002

Manufacturer: CRYSTAL ENGINEERING

Model: 2KPSIXP2I

Description: DIGITAL TEST GAUGE

Mfr S/N: 763018

Cust ID: COI

Cust S/N: N/A

Data created on: 4/8/2026 12:46:22 PM

*Calibration Procedure: I/A/W JM529 Rev. 2.0 08-20-2024
Robokal ver. 1.6*

**Range: Pressure Measure, 0 to 2000 psig, Accuracy: $\pm 0.02\%$ of Full Scale from 0 to 20% of
Range: $\pm 0.1\%$ of Reading from 20 to 110% of Range**

As-Found / As-Left Report

Upscale Test

Standard Read Value	Low Limit	UUT Read Value	High Limit	Deviation	Result
0.000	-0.4	0.0	0.4	0.0	Pass
400.000	399.6	400.0	400.4	0.0	Pass
500.001	499.5	500.0	500.5	0.0	Pass
1,000.001	999.0	1,000.2	1,001.0	0.2	Pass
1,500.002	1,498.5	1,500.2	1,501.5	0.2	Pass
2,000.005	1,998.0	2,000.1	2,002.0	0.1	Pass

All values are reported in **psig** unless otherwise stated