

Data created on: 04/15/2026 03:42:44 PM
Calibration Procedure: Fluke 789 Processmeter
Calsystem ver. 0.5.0

AC Voltage
Tolerance: $\pm 0.7\%$ rdg + ± 4 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
100.0 mV @ 60 Hz	98.9	100.0	101.1	0.000	Pass
300.0 mV @ 60 Hz	297.5	300.0	302.5	0.000	Pass
1.0 V @ 60 Hz	1.0	1.0	1.0	0.000	Pass
2.0 V @ 60 Hz	2.0	2.0	2.0	0.000	Pass
3.0 V @ 60 Hz	3.0	3.0	3.0	0.000	Pass
10.0 V @ 60 Hz	9.9	10.0	10.1	0.000	Pass
30.0 V @ 60 Hz	29.8	30.0	30.2	0.000	Pass
100.0 V @ 60 Hz	99.1	100.0	100.9	0.000	Pass
300.0 V @ 60 Hz	297.7	300.0	302.3	0.000	Pass
100.0 V @ 60 Hz	97.0	100.0	103.0	0.000	Pass
800.0 V @ 60 Hz	792.0	800.0	808.0	0.000	Pass

All values in mV unless otherwise stated

Frequency
Tolerance: $\pm 0.005\%$ rdg + ± 1 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
5.0 V @ 100 Hz	5.0	100.0	5.0	95.000	Pass
5.0 V @ 1000 Hz	4.9	1000.0	5.1	995.000	Pass
5.0 V @ 10 kHz	5.0	10.0	5.0	5.000	Pass

All values in V unless otherwise stated

DC Voltage
Tolerance: $\pm 0.1\%$ rdg + ± 1 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
1.0 V	1.0	1.0	1.0	0.000	Pass
3.0 V	3.0	3.0	3.0	0.000	Pass
10.0 V	10.0	10.0	10.0	0.000	Pass
30.0 V	30.0	30.0	30.0	0.000	Pass
100.0 V	99.8	100.0	100.2	0.000	Pass
300.0 V	299.6	300.0	300.4	0.000	Pass
100.0 V	99.0	100.0	101.0	0.000	Pass
800.0 V	798.0	800.0	802.0	0.000	Pass

All values in V unless otherwise stated

DC Millivolts

Tolerance: $\pm 0.1\%$ rdg + ± 1 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
100.0 mV	99.8	100.0	100.2	0.000	Pass
300.0 mV	299.6	300.0	300.4	0.000	Pass

All values in mV unless otherwise stated

Resistance

Tolerance: $\pm 0.2\%$ rdg + ± 2 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
120.0 Ohm	119.6	120.0	120.4	0.000	Pass
300.0 Ohm	299.2	300.0	300.8	0.000	Pass
1.2 kOhm	1.2	1.2	1.2	0.000	Pass
3.0 kOhm	3.0	3.0	3.0	0.000	Pass
12.0 kOhm	12.0	12.0	12.0	0.000	Pass
30.0 kOhm	29.9	30.0	30.1	0.000	Pass
120.0 kOhm	119.7	120.0	120.3	0.000	Pass
300.0 kOhm	299.3	300.0	300.7	0.000	Pass
1.2 MOhm	1.2	1.2	1.2	0.000	Pass
3.0 MOhm	3.0	3.0	3.0	0.000	Pass
12.0 MOhm	11.7	12.0	12.3	0.000	Pass
30.0 MOhm	29.2	30.0	30.8	0.000	Pass

All values in Ohm unless otherwise stated

Diode

Tolerance: $\pm 2\%$ rdg + ± 1 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
2.0 V	2.0	2.0	2.0	0.000	Pass

All values in V unless otherwise stated

DC Milliamps

Tolerance: $\pm 0.05\%$ rdg + ± 2 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
4.0 mA	4.0	4.0	4.0	0.000	Pass
12.0 mA	12.0	12.0	12.0	0.000	Pass
20.0 mA	20.0	20.0	20.0	0.000	Pass

All values in mA unless otherwise stated

DC Amps

Tolerance: $\pm 0.2\%$ rdg + ± 2 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
0.1 A	0.1	0.1	0.1	0.000	Pass
0.4 A	0.4	0.4	0.4	0.000	Pass

All values in A unless otherwise stated

AC Current

Tolerance: $\pm 1\%$ rdg + ± 2 dgt

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
0.1 mA @ 60 Hz	0.1	0.1	0.1	0.000	Pass
0.4 mA @ 60 Hz	0.4	0.4	0.4	0.000	Pass

All values in mA unless otherwise stated

Current Source

Tolerance: $\pm 0.05\%$ span

Nominal Value	Low Limit	UUT Reading	High Limit	Deviation	Result
4.0 mA	4.0	4.0	4.0	0.000	Pass
12.0 mA	12.0	12.0	12.0	0.000	Pass
20.0 mA	20.0	20.0	20.0	0.000	Pass

All values in mA unless otherwise stated