
ICRM NEWSLETTER

Issue 13 - January 1999



International Committee for Radionuclide Metrology

Editor: Dietmar F.G. Reher



EUROPEAN COMMISSION

JOINT RESEARCH CENTRE
Institute for Reference Materials and Measurements
IRMM
Geel, Belgium



LABORATORY

National Institute of Standards and Technology

NAMES

R. Collé

ACTIVITY

Determination of Calibration Factors for the Nondestructive Assay of Pure Beta-Emitting Brachytherapy Sources

RESULTS

Ionization-chamber calibration factors have been determined for two commercially-manufactured intravascular brachytherapy sources: a TiNi-encapsulated 32P source developed by Guidant Intravascular Intervention (Houston, TX), and a stainless-steel encapsulated 90Sr-90Y developed by Bebig Isopentechnic und Umweltdiagnostik GmbH (Berlin, Germany) in collaboration with the Novoste Corp. (Norcross, GA). The calibration factor for the former was derived from ionization current measurements with a Capintec CRC-12 ("dose calibrator") which is the nuclearmedicine community's de facto standard instrument, followed by very quantitative, destructive assays of the ³²P content in the sources. Similarly, for the former sources, NIST "4πγ chamber A" calibration factors were determined for both bare ceramic and SS-steel encapsulated source configurations. The assay results used to establish these calibration factors at the same time, resulted in establishing the requisite calibration factors for subsequent nondestructive measurements by their manufacturers, e.g., for in-house quality control.

OTHER RELATED PUBLICATIONS

R. Collé, B. E. Zimmerman, C. G. Soares, and B. M. Coursey, "Determination of a Calibration Factor for the Nondestructive Assay of Guidant ³²P Brachytherapy Sources," *in press, Appl. Rad. Iso.* (1998).

ADDRESS-

NATL INST STDS AND TECH 100 Bureau Drive STOP 8462 Gaithersburg, MD 20899-8462 USA

Tel: 1-301-975-5527

Fax:1-301-926-7416 ronald.colle@nist.gov

CONTACT

R. Collé

LABORATORY

National Institute of Standards and Technology

NAMES

R. Collé

ACTIVITY

Development of Procedures for the Chemical Digestion and Radionuclidic Assay of Encapsulated, Pure Beta-Emitting, Intravascular Brachytherapy Sources

RESULTS

Very quantitative radiochemical procedures for the destructive assay of encapsulated sources containing pure beta-emitting nuclides have been devised. These types of sources are intended for use in the prophylactic treatment of restenosis following balloon angioplasty in heart-disease patients. The developed methods have been applied to two very different types of sources: a polymer-based, ³²P-containing source with TiNi encapsulation; and a ceramic-based 90Sr-90Y source with stainless-steel encapsulation. The internal compositions for both types of sources had constituents that were chemically impervious. The assays involved partial dissolutions (or extractions) followed by 4πβ liquid scintillation (LS) spectrometry (with ³H-standard efficiency tracing) of the resulting solutions. The procedures for the former included provisions for accounting for all possible losses of ³²P in the digestion procedure (based on radiochemical tracing experiments), for any unrecovered activity in the remaining source material, and for any residual activity in the solution- and source-handling tools. The procedures for the latter source consisted of extracting a fraction of the 90Sr activity from the ceramic-like material for LS assay, and determining the fraction of unextracted activity by before and after ionization current measurements on the extracted source material. The uncertainties in the assays were typically 2% to 4% for two standard uncertainty intervals. These destructive assays were required for relating radiochromic-film measurements of the absorbed dose spatial distributions for the sources to theoretic dose modeling, and for establishing calibration factors for subsequent non-destructive radionuclidic measurements on the sources. The generalized protocols developed for this work have been extended and deployed for the assay of ³²P-ion-implanted coronary stents, and will soon be used to perform assays on 32Pcontaining balloon catheters.

IN PREPARATION

R. Collé, "On the Radioanalytical Methods Used to Assay Stainless-Steel-Encapsulated, Ceramic-Based ⁹⁰Sr-⁹⁰Y Intravascular Brachytherapy Sources," *submitted, Appl. Rad. Iso.* (1998).

OTHER RELATED

PUBLICATIONS

R. Collé, "Chemical Digestion and Radionuclidic Assay of TiNi-Encapsulated ³²P Intravascular Brachytherapy Sources," *in press*,

Appl. Rad. Iso. (1998).

ADDRESS

NATL INST STDS AND TECH

100 Bureau Drive STOP 8462

Gaithersburg, MD 20899-8462 USA

Tel: 1-301-975-5527 Fax:1-301-926-7416 ronald.colle@nist.gov

CONTACT

R. Collé

LABORATORY

National Institute of Standards and Technology

NAMES

R. Collé

ACTIVITY

New "Primary" Standardizations of ²²⁶Ra and ²²²Rn

IN PROGRESS

All extant activity standards and calibrations for ²²⁶Ra and ²²²Rn are based on comparative measurements against an artifact radium mass standard (namely the international 1935 Hönigschmid standards and its derivatives). Efforts are currently underway to remove this dependence, and to develop and perform

underway to remove this dependence, and to develop and perform primary standardizations for both $^{226}\mathrm{Ra}$ and the $^{222}\mathrm{Rn}$ subseries based on $4\pi\alpha\beta$ LS spectrometry with $^3\mathrm{H}\text{-standard}$ efficiency tracing. Work has been completed on successfully employing this new standardization technique to $^{222}\mathrm{Rn}$, such as for the calibration

of the NIST radon-in-water standard generator and for

measurements of the polyethylene-encapsulated ²²⁶Ra-solution emanation standards (SRM 4968). The techniques are now being extended to the entire ²²⁶Ra decay series. Preliminary findings have demonstrated that the method will be able to adequately resolve the ²¹⁰Pb subseries (i.e., from ²²⁶Ra and the ²²²Rn

subseries) in even aged radium solutions. The ability to perform

such a resolution is significant in that it will allow direct

intercomparisons between the various ²²⁶Ra standards issued by NIST over the past 50 years, irrespective of their present degree

of radioactive equilibrium between ²²⁶Ra and ²¹⁰Pb.

ADDRESS

NATL INST STDS AND TECH 100 Bureau Drive STOP 8462

Gaithersburg, MD 20899-8462 USA

Tel: 1-301-975-5527 Fax:1-301-926-7416 ronald.colle@nist.gov

CONTACT

R. Collé