Validation of low-cost rural health screening device

"Validate my low-cost rural screening device: exhaled CO, Venturi peak flow/1-s volume, SpO_2, HR, single-lead ECG, with a worst-channel Green/Yellow/Red score. Cite evidence on: breath-CO cut-offs (0-6/7-9/≥10 ppm) and biomass vs tobacco; low-cost flow vs spirometry; oximetry skin-tone bias; scoring thresholds; whether 300 adults suffices for 85% sensitivity. Give weaknesses ranked and improvement suggestions." *Generated by LeapSpace*

Validation of a Low-Cost Rural Screening Device at a Glance

Key Findings Table

Topic	Practical conclusion	Relevant population/setting	Key citations
Breath CO cut-offs	`0–6 ppm` is a reasonable low-risk band, `7–9 ppm` is better treated as indeterminate/intermediate, and `>=10 ppm` is relatively specific but too insensitive to serve as the main smoking-detection threshold in screening [1] [2] [3] [4] [5] [6].	Adult screening, cessation monitoring, pregnancy cohorts	[1] [5] [6] [7] [8]
Biomass vs tobacco confounding	Biomass, coal, and passive smoke can elevate exhaled CO in never-smokers, so breath CO is not tobacco-specific in rural biomass-exposed adults [2] [9] [10].	Rural adults, indoor air pollution exposure	[2] [9] [10]
Low-cost flow vs spirometry	Portable flow tools can support triage/referral when spirometry is unavailable, but agreement and bias are not good enough to claim interchangeability with reference FEV1/spirometry [11] [12] [13] [14] [15].	Primary care, low-resource screening, symptomatic adults	[11] [12] [14] [16] [17]

Topic	Practical conclusion	Relevant population/setting	Key citations
SpO_2 bias	Pulse oximetry can overestimate oxygen saturation more often in darker-skinned adults, especially in lower saturation ranges; low perfusion and motion further worsen reliability [18] [19] [20] [21] [22].	Adult clinical cohorts; caution for rural community use	[18] [19] [20] [22] [23]
Single-lead ECG	Best supported for AF-oriented screening, with repeat recordings and overread workflows; less reliable for broad arrhythmia diagnosis and prone to noisy/unclassified traces [24] [25] [26] [27] [28].	Community, semirural, primary care	[24] [25] [26] [29] [30]
Worst-channel scoring	A worst-channel Green/Yellow/Red rule is defensible if you want high sensitivity and low undertriage, but it will likely increase false positives unless thresholds are calibrated locally [31] [32] [33].	Triage/traffic-light analogues	[31] [32] [34] [35]
Is 300 adults enough?	Not inherently. Adequacy depends on the number of diseased/reference-abnormal cases and desired confidence-interval width; 300 total is often borderline or inadequate for precise sensitivity validation [36] [37] [38] [39].	Diagnostic validation design	[36] [37] [38] [39] [40]

Direct Answer

Your device is plausible as a **rural screening/triage tool**, but the current evidence does **not** support treating it as a diagnostic substitute for cotinine-based smoking verification, laboratory spirometry, arterial oxygen reference methods, or 12-lead ECG [11] [18] [28] [41].

For **breath CO**, your `0–6 / 7–9 / >=10 ppm` bands are usable as a **pragmatic traffic-light scheme**, but not as a validated smoking-status classifier. The strongest evidence favors lower smoking cut-offs, often around `2–6 ppm`, while `>=10 ppm` is more specific but misses many smokers; therefore `7–9 ppm` should be interpreted as an **intermediate/uncertain** zone rather than a firm tobacco-positive category [1] [2] [3] [4] [5] [6]. Biomass and passive smoke are major confounders, and the supplied evidence does **not** provide a validated biomass-adjusted alternative threshold [9].

For **Venturi peak flow/1-second volume**, evidence supports use for **screening or referral triage when spirometry is unavailable**, but not interchangeability with FEV1 from standard spirometry. Device bias, technique dependence, and calibration drift are major risks [11] [12] [14] [15] [17].

For **SpO_2**, the channel is clinically valuable but also your highest safety risk: darker skin, low perfusion, motion, and lower saturation ranges can cause overestimation and occult hypoxemia, so borderline or abnormal values should trigger repeat measurement and signal-quality checks before color assignment where possible [18] [19] [20] [22].

For **single-lead ECG**, the evidence is strongest for **AF screening**, not general rhythm diagnosis. Expect unclassified/noisy traces and design a repeat-plus-overread workflow [24] [25] [27] [30].

For the **worst-channel Green/Yellow/Red rule**, the likely effect is **safer triage with lower specificity**. That is acceptable if your goal is referral screening, but only if you prespecify acceptable overtriage and provide confirmatory pathways [31] [32] [42].

For **sample size**, **300 adults alone is not enough evidence of adequacy** for validating `85%` sensitivity. You need enough **reference-positive cases** first; in low-prevalence screening, 300 total will often yield too few abnormal cases for narrow confidence intervals [36] [37] [38] [39] [40].

Bottom line: validate it as a **workflowed screening system**, not just a bundle of sensors: local recalibration, repeat measurements, subgroup analyses by biomass exposure and skin tone, and reference-test confirmation are the highest-yield next steps [15] [22] [39] [43].

Confidence level: Medium-High - strong evidence supports the broad conclusions, but exact thresholds for your specific combined device are not directly validated in the supplied studies [35] [44].

Study Scope

- Evidence covers adult screening/validation studies, reviews, and methodological papers across **breath CO**, **portable flow/PEF**, **pulse oximetry**, **single-lead ECG**, and **triage/sample-size methods** [2] [18] [28] [38] [45].
- Settings include outpatient, primary care, hospital, community, semirural, and low-resource contexts, though not all channels were studied specifically in rural adults [12] [18] [24].
- Methods include ROC-based threshold studies, device-comparison studies, cohort analyses, systematic reviews, and diagnostic sample-size methodology [5] [11] [18] [46].

Assumptions & Limitations

- The supplied evidence does **not** identify a published validated worst-channel algorithm using your exact channel set, so threshold recommendations are partly extrapolated from adjacent triage evidence rather than directly proven for this device [35] [44].
- The supplied evidence shows biomass exposure confounds breath CO, but does **not** quantify a reliable biomass-specific threshold adjustment [9].
- Evidence for a specifically **Venturi-based** adult 1-second-volume substitute is indirect/prototype-level rather than definitive [13] [47].
- Most pulse-oximetry disparity evidence is hospital-based; outpatient/rural evidence is thinner [18].

Suggested Further Research

- Derive **local breath-CO thresholds** stratified by tobacco use, biomass exposure, passive smoke, and time since cooking [9].
- Validate the **Venturi flow channel** directly against ATS-quality spirometry, including repeatability, technique effects, and drift [13] [17].
- Build **signal-quality-gated SpO_2 rules** and test performance across objectively measured skin pigmentation groups [48] [49].
- Compare **single-pass vs repeat-measurement workflows** for SpO_2 and ECG in non-expert rural users [22] [50] [51].
- Power the next study on the **number of abnormal/reference-positive cases**, not total N alone [38] [39].

Full Report

Foundational Context

Your device combines markers with very different biological meanings and error structures: **exhaled CO** reflects recent smoke exposure but is vulnerable to environmental confounding; **flow/1-s volume** reflects expiratory mechanics but is highly technique dependent; **SpO_2** estimates oxygenation but can fail under pigmentation, motion, and perfusion challenges; **single-lead ECG** captures electrical rhythm but is narrow in scope and quality-sensitive [2] [14] [20] [28].

That matters because a **worst-channel rule** is only as robust as the noisiest channel. Taken together, the supplied studies suggest that your composite score will likely be driven less by “average device quality” than by the most confounded measurement in real field use [9] [21] [24] [33].

Confidence level: High - this pattern is directly supported across all sensor domains [9] [15] [20] [24].

Channel-by-Channel Validation Evidence

1) Breath CO thresholds and biomass/tobacco confounding

Evidence consistently shows that **lower cut-offs than 8-10 ppm** perform better for smoker/nonsmoker discrimination. Examples include `6 ppm` with `94% sensitivity/96% specificity`, `6.5 ppm` with `90%/83%`, and `3 ppm` with `98.1%/95.8%` in one validation study [1] [2] [3]. Conversely, `9–10 ppm` thresholds often had high specificity but poor sensitivity, and `8 ppm` performed badly in both general and pregnancy-related validation contexts [4] [5] [6] [8].

This supports a practical interpretation of your bands:

- **0-6 ppm = Green/low-likelihood recent smoking exposure** is defensible for screening, though not definitive because some smokers fall below this and some non-smokers may rise into it depending on exposure [1] [52].
- **7-9 ppm = Yellow/intermediate** is the safest interpretation because evidence suggests this range only modestly increases smoking probability in some populations and is too unstable for a definitive classification [4] [53].
- **>=10 ppm = Red/high-likelihood smoke exposure** is specific but will miss many smokers, so it works better as a “high concern” band than as the main positivity threshold [7] [54].

The major rural limitation is **confounding by biomass and passive smoke**. Among never-smokers, exhaled CO increased with passive smoke and biomass/coal burning, especially in rural settings and in winter, showing that CO can reflect household air pollution as well as tobacco [9]. Reviews also note that passive smokers can have intermediate CO, and that CO tracks environmental exposure less robustly than cotinine in some settings [2] [10]. The evidence provided does **not** identify a validated threshold correction for biomass exposure [9].

Practical implication: your CO channel should be framed as detecting **recent smoke exposure**, not specifically **tobacco use**, unless paired with symptom/history questions or confirmatory biomarkers [41] [55].

Confidence level: High - strong and consistent evidence supports lower thresholds and biomass confounding [1] [6] [9].

2) Low-cost flow measurement versus spirometry

Portable flow devices and PEF-based screening have real value when spirometry is unavailable. PEF `<80% predicted` detected airflow obstruction with moderate-to-high sensitivity/specificity in screening studies, and combining PEF with symptoms improved specificity substantially [12] [16] [56]. Handheld/portable spirometers also often correlate strongly with laboratory spirometry [11] [57].

However, correlation is not interchangeability. Bland-Altman analyses showed meaningful limits of agreement and occasional bias, so portable devices should not be assumed to substitute directly for standard spirometric FEV1 [11] [58] [59]. Technique strongly affects readings: a “spitting” maneuver inflated PEF across marketed meters, and instruction differences altered results [15] [60]. Device-specific bias and wear also matter; some meters degraded after repeated use while others remained stable [17] [61].

For your **Venturi** implementation specifically, the supplied evidence is weaker: Venturi-based devices are described as promising or prototype-level, but the available studies do not establish that a Venturi device can replace laboratory spirometry for routine adult FEV1/PEF substitution [13] [47].

Practical implication: use this channel for **screening/referral support**, not for claiming formal spirometric diagnosis. If possible, analyze it as **device-specific flow markers** with local calibration, not as direct FEV1 equivalents [14] [62].

Confidence level: Medium-High - well supported that low-cost flow helps triage, but direct Venturi validation is limited [12] [13].

3) Pulse oximetry accuracy and skin-tone bias

This is the clearest safety concern. Large adult studies found that pulse oximetry more often **overestimated** arterial saturation in darker-skinned adults, with higher rates of occult hypoxemia; one cohort reported mean overestimation of `1.57%` and higher adjusted odds of occult hypoxemia in Black patients [19]. A 2024 systematic review/meta-analysis similarly found higher occult hypoxemia prevalence in darker-skinned/racialized groups, while noting that outpatient evidence was lacking [18].

The literature is not perfectly uniform: some controlled laboratory/device-specific studies found minimal clinically significant race-related differences under ideal conditions [63]. But those findings do not cancel the real-world disparity signal seen in broader adult cohorts and reviews [18] [64].

Error also worsens under **low saturation, low perfusion, and motion**. Multiple sources report lower accuracy in hypoxemic ranges, plus persistent degradation with motion and low perfusion [20] [23] [65]. Sensor site matters too, with forehead sensors sometimes outperforming finger sensors in challenging conditions [66] [67].

Practical implication: your SpO_2 channel should not assign Yellow/Red from a single borderline reading without checking signal quality and repeating the measurement where feasible [22]. If you can, record a confidence metric or perfusion/signal-quality flag. Inferred: for your composite system, SpO_2 deserves asymmetric handling because it has both high clinical relevance and known measurement bias risk in the target setting [18] [20].

Confidence level: High - strong evidence for real-world overestimation risk and signal-quality dependence [18] [19] [20].

4) Single-lead ECG in field screening

Single-lead ECG is best validated for **atrial fibrillation screening**. Community/primary-care studies reported high specificity and generally good AF sensitivity for some devices and workflows, including pooled or cohort-level performance near `89–90% sensitivity` and `97.9–99% specificity` [27] [68].

But performance drops in noisier real-world workflows. In semirural Ethiopian screening, automated rhythm decisions were much lower than cardiologist verification on first tracing, improving with repeats; most algorithm failures were due to poor-quality/noisy recordings [24]. In U.S. primary care, outputs were often normal but also frequently **unclassified** or **no analysis**, and sensitivity was only `51.0%` in that workflow [25]. Unreadable recordings also occurred in community elderly screening [26].

For **non-AF arrhythmias**, single-lead devices are weaker. False positives arise from ectopy, supraventricular tachycardia, and AV block, and multi-lead recordings improve agreement for broader arrhythmia diagnosis [29] [69] [70].

Practical implication: your ECG channel is appropriate for **AF-oriented flagging**, but positives, unreadables, and persistent unclassified rhythms should go to confirmatory review/12-lead referral [28] [30]. Repeat acquisition is not optional; it is part of validated workflow [50] [51].

Confidence level: High - strong support for AF screening utility and workflow dependence [24] [27] [28].

Composite Green/Yellow/Red Scoring Logic

The supplied evidence does **not** directly validate an adult worst-channel algorithm with your exact channel set [35] [44]. Still, analogous triage systems offer clear design lessons. Traffic-light and threshold systems commonly favor **sensitivity and undertriage avoidance**, but specificity suffers and false positives rise [31] [32] [71]. Higher thresholds improve specificity but lower sensitivity, and performance varies strongly by context [33] [72].

So the strongest evidence-backed use case for your score is:

- **Green**: all channels acceptable, no signal-quality concern [42].
- **Yellow**: borderline physiology, uncertain CO exposure interpretation, or low-confidence signal needing repeat/clinical context [72] [73].
- **Red**: clearly abnormal and/or repeated abnormal finding, especially when the abnormality is clinically consequential and the signal is trustworthy [31] [32].

A pure worst-channel rule is attractive because it is conservative. Its main pitfall is that the **most artifact-prone channel can dominate the whole decision** [20] [24]. Taken together, these findings suggest your best next version is not simply “worst value wins,” but “worst **reliable** value wins,” with an **indeterminate/repeat** path for noisy ECGs, low-confidence SpO_2, and biomass-confounded CO [9] [22] [51].

Confidence level: Medium - the design principles are supported, but not directly validated for this exact score [33] [35].

Is 300 Adults Enough to Validate 85% Sensitivity?

No single total sample size answers this. Diagnostic sample size should be based on the desired precision of the **sensitivity estimate** and the number of **diseased/reference-positive** participants, especially when prevalence is below 50% [38] [39].

The supplied examples show why 300 can be inadequate:

- A gastric-cancer validation targeting `85% sensitivity` at about `2% prevalence` planned about `5000` participants to achieve about a `5 percentage-point` error for sensitivity [40].
- A prostate-biopsy validation targeting at least `90% sensitivity` with `40% prevalence` needed `346–353` men for a `5%` CI target [36].
- Another study targeting `80% sensitivity` with `0.03` precision required at least `310` subjects [37].

So **300 adults may be enough only if prevalence is high enough** to yield a sufficiently large abnormal/reference-positive subset and if you accept a relatively wide confidence interval [38] [39]. In a low-prevalence community screening population, 300 total will usually be too small for confident validation of `85% sensitivity` [40].

Practical implication: define the primary target condition for each channel and the composite, estimate expected prevalence, then compute the required number of positive cases first [46] [74].

Confidence level: High - directly supported by both methodology and worked examples [36] [39].

Ranked Weaknesses and Improvement Priorities

Ranked weaknesses

1. **SpO_2 bias and occult hypoxemia risk** - highest priority because the channel can falsely reassure, especially in darker-skinned users and under low perfusion/motion conditions [18] [19] [20].
2. **Breath CO confounding by biomass/passive smoke** - major threat to interpretability in rural households because the channel is not tobacco-specific [9] [10].

3. **Flow-channel non-interchangeability and technique dependence** - useful for screening, but accuracy is vulnerable to maneuver quality, calibration, and device-specific bias [11] [15] [17].
4. **Single-lead ECG noise/unclassified traces and narrow diagnostic scope** - strong for AF screening only, with repeat and overread burden [24] [25] [29].
5. **Worst-channel rule likely low specificity** - good for safety, but can magnify error from any weak channel if not reliability-gated [31] [33].

Improvement suggestions

- Add **channel-specific reliability rules** before color assignment, especially for SpO_2 and ECG [22] [51].
- Treat **CO 7-9 ppm as indeterminate**, and capture biomass exposure, cooking timing, and passive smoke history alongside the reading [2] [9].
- Perform **device-specific calibration** for the flow channel against formal spirometry and track drift over repeated use [17] [62].
- Use **repeat ECG acquisition** and clinician overread/referral for positives, unreadables, and persistent unclassified traces [30] [43].
- If feasible, incorporate **objective pigmentation measurement or bias-aware correction methods** for SpO_2, since engineering studies suggest this can reduce bias [48] [49] [75].
- Power the validation study on **abnormal-case counts**, and consider prevalence enrichment or nested reference substudies to make validation affordable [39].

Confidence level: Medium-High - ranking reflects both evidence strength and likely patient-safety impact [9] [18] [33].

Recommended Next Validation Study

A strong next-step design would validate the **full workflow**, not just the sensors:

- Enroll rural adults across tobacco-use and biomass-exposure strata [9].
- Reference standards: smoking history plus biochemical verification subset, ATS-quality spirometry, oxygen reference subset for low/borderline SpO_2, and 12-lead ECG for ECG positives/unclassifieds [30] [41] [62].
- Require repeat measurements for borderline SpO_2 and noisy ECG before final color assignment [22] [50].
- Prespecify subgroup analyses by **skin pigmentation**, **biomass exposure**, and **signal-quality failure rate** [9] [49].
- Validate both **channel-level accuracy** and **composite decision performance**, including overtriage and undertriage [32] [34].

Overall Conclusion

The evidence supports your device as a **pragmatic rural screening platform** if it is positioned as a **triage-plus-referral system** rather than a diagnostic replacement [12] [28] [31]. The best-supported elements are: low breath-CO thresholds outperform 8-10 ppm for smoking detection, low-cost flow and single-lead ECG can aid first-line screening, and worst-channel scoring is acceptable when sensitivity is prioritized [1] [12] [27] [32]. The biggest threats are confounding and measurement bias: biomass for CO, pigmentation/perfusion/motion for SpO_2, technique and drift for flow, and noise/unclassified outputs for ECG [9] [15] [20] [24].

The most defensible version of your system is therefore: **local thresholds, reliability gating, repeat-measurement rules, and confirmatory referral pathways** [22] [42] [43].

Final confidence level: Medium-High - the overall direction is strongly supported, but exact thresholds and composite performance for your specific device still require direct prospective validation [35] [38].

References

1. Middleton, E.T., Morice, A.H. (2000) Breath carbon monoxide as an indication of smoking habit. *Chest*. https://doi.org/10.1378/chest.117.3.758
2. Deveci, S.E., Deveci, F., Açik, Y., Ozan, A.T. (2004) The measurement of exhaled carbon monoxide in healthy smokers and non-smokers. *Respiratory Medicine*. https://doi.org/10.1016/j.rmed.2003.11.018
3. Cropsey, K.L., Eldridge, G.D., Weaver, M.F., (...), Stitzer, M.L. (2006) Expired carbon monoxide levels in self-reported smokers and nonsmokers in prison. *Nicotine and Tobacco Research*. https://doi.org/10.1080/14622200600789684
4. Christenhusz, L., de Jongh, F., Van-Der-Valk, P., (...), van der Palen, J. (2007) Comparison of three carbon monoxide monitors for determination of smoking status in smokers and nonsmokers with and without COPD. *Journal of Aerosol Medicine: Deposition, Clearance, and Effects in the Lung*. https://doi.org/10.1089/jam.2007.0606
5. Javors, M.A., Hatch, J.P., Lamb, R.J. (2005) Cut-off levels for breath carbon monoxide as a marker for cigarette smoking. *Addiction*. https://doi.org/10.1111/j.1360-0443.2004.00957.x
6. Cropsey, K.L., Trent, L.R., Brendan Clark, C.B., (...), Hendricks, P.S. (2014) How low should you go? Determining the optimal cutoff for exhaled carbon monoxide to confirm smoking abstinence when using cotinine as reference. *Nicotine and Tobacco Research*. https://doi.org/10.1093/ntr/ntu085
7. Perkins, K.A., Karelitz, J.L., Jao, N.C. (2013) Optimal carbon monoxide criteria to confirm 24-hr smoking abstinence. *Nicotine and Tobacco Research*. https://doi.org/10.1093/ntr/nts205
8. Usmani, Z.C., Craig, P., Shipton, D., Tappin, D. (2008) Comparison of CO breath testing and women's self-reporting of smoking behaviour for identifying smoking during pregnancy. *Substance Abuse: Treatment, Prevention, and Policy*. https://doi.org/10.1186/1747-597X-3-4
9. https://www.sciencedirect.com/science/article/pii/S1464368513001342

10. Jarvis, M., Tunstall-Pedoe, H., Feyerabend, C., (...), Salloojee, Y. (1984) Biochemical markers of smoke absorption and self reported exposure to passive smoking. *Journal of Epidemiology and Community Health*. https://doi.org/10.1136/jech.38.4.335

11. McBride, J., Doumit, M., Bell, J., (...), Sivam, S. (2026) Accuracy and reliability of personal home spirometers: a multicentre laboratory quality-control study in children and adults. *Respiratory Medicine*. https://doi.org/10.1016/j.rmed.2026.109183

12. Tian, J., Zhou, Y., Cui, J., (...), Ran, P. (2012) Peak expiratory flow as a screening tool to detect airflow obstruction in a primary health care setting. *International Journal of Tuberculosis and Lung Disease*. https://doi.org/10.5588/ijtld.11.0429

13. Ferreira Nunes, M., Plácido da Silva, H., Raposo, L., Rodrigues, F. (2024) Design and Evaluation of a Novel Venturi-Based Spirometer for Home Respiratory Monitoring. *Sensors*. https://doi.org/10.3390/s24175622

14. Csonka, L.L., Tikkakoski, A., Karjalainen, J., Lehtimäki, L. (2026) Effect of device and expiratory maneuver technique on peak expiratory flow measurements. *Clinical Physiology and Functional Imaging*. https://doi.org/10.1111/cpf.70067

15. Strayhorn, V., Leeper, K., Tolley, E., Self, T. (1998) Elevation of peak expiratory flow by a 'spitting' maneuver: Measured with five peak flowmeters. *Chest*. https://doi.org/10.1378/chest.113.4.1134

16. Thorat, Y.T., Salvi, S.S., Kodgule, R.R. (2017) Peak flow meter with a questionnaire and mini-spirometer to help detect asthma and COPD in real-life clinical practice: A cross-sectional study. *npj Primary Care Respiratory Medicine*. https://doi.org/10.1038/s41533-017-0036-8

17. Shapiro, S.M., Hendler, J.M., Ogirala, R.G., (...), Shapiro, M.B. (1991) An evaluation of the accuracy of Assess and MiniWright peak flowmeters. *Chest*. https://doi.org/10.1378/chest.99.2.358

18. Parr, N.J., Beech, E.H., Young, S., Valley, T.S. (2024) Racial and Ethnic Disparities in Occult Hypoxemia Prevalence and Clinical Outcomes Among Hospitalized Patients: A Systematic Review and Meta-analysis. *Journal of General Internal Medicine*. https://doi.org/10.1007/s11606-024-08852-1

19. Henry, N.R., Hanson, A.C., Schulte, P.J., (...), Warner, M.A. (2022) Disparities in Hypoxemia Detection by Pulse Oximetry Across Self-Identified Racial Groups and Associations With Clinical Outcomes. *Critical Care Medicine*. https://doi.org/10.1097/CCM.0000000000005394

20. Giuliano, Karen K, Bilkovski, Robert N, Beard, John, Lamminmäki, Sakari (2023) *Journal of Clinical Monitoring and Computing*. https://doi.org/10.1007/s10877-023-01029-x

21. Hendrickson, Carolyn M., Lipnick, Michael S., Chen, David, (...), Varshney, Jaaie (2026) EquiOx. *CHEST Critical Care*. https://doi.org/10.1016/j.chstcc.2026.100305

22. Philip, K.E.J., Bennett, B., Fuller, S., (...), Vlachou, A. (2020) Working accuracy of pulse oximetry in COVID-19 patients stepping down from intensive care: A clinical evaluation. *BMJ Open Respiratory Research*. https://doi.org/10.1136/bmjresp-2020-000778

23. Witting, M.D., Lueck, C.H. (2001) The ability of pulse oximetry to screen for hypoxemia and hypercapnia in patients breathing room air. *Journal of Emergency Medicine*. https://doi.org/10.1016/S0736-4679(01)00309-2

24. Pitman, B.M., Chew, S.-H., Wong, C.X., (...), Lau, D.H. (2021) Performance of a mobile single-lead electrocardiogram technology for atrial fibrillation screening in a semirural African population: Insights from "The heart of Ethiopia: Focus on atrial fibrillation" (TEFF-AF) study. *JMIR mHealth and uHealth*. https://doi.org/10.2196/24470

25. Khurshid, S., Chang, Y., Borowsky, L.H., (...), Lubitz, S.A. (2023) Performance of Single-Lead Handheld Electrocardiograms for Atrial Fibrillation Screening in Primary Care: The VITAL-AF Trial. *JACC: Advances*. https://doi.org/10.1016/j.jacadv.2023.100616

26. Guan, J., Wang, A., Song, W., (...), Wang, L. (2021) Screening for arrhythmia with the new portable single-lead electrocardiographic device (SnapECG): an application study in community-based elderly population in Nanjing, China. *Aging Clinical and Experimental Research*. https://doi.org/10.1007/s40520-020-01512-4

27. Wong, Kam Cheong, Klimis, Harry, Lowres, Nicole, (...), Chow, Clara K (2020) Diagnostic accuracy of handheld electrocardiogram devices in detecting atrial fibrillation in adults in community versus hospital settings: a systematic review and meta-analysis. *Heart*. https://doi.org/10.1136/heartjnl-2020-316611

28. Freedman, B., Camm, J., Calkins, H., (...), Wijeratne, T. (2017) Screening for atrial fibrillation. *Circulation*. https://doi.org/10.1161/CIRCULATIONAHA.116.026693

29. Junarta, J., Frisch, D.R., Dikdan, S., (...), Sarkar, K. (2021) Expanding a single-lead mobile electrocardiographic device to multiple-lead recordings improves diagnostic accuracy and confidence. *Journal of Electrocardiology*. https://doi.org/10.1016/j.jelectrocard.2021.05.006

30. Desteghe, L., Raymaekers, Z., Lutin, M., (...), Heidbuchel, H. (2017) Performance of handheld electrocardiogram devices to detect atrial fibrillation in a cardiology and geriatric ward setting. *Europace*. https://doi.org/10.1093/europace/euw025

31. Thompson, M., Coad, N., Harnden, A., (...), Mant, D. (2009) How well do vital signs identify children with serious infections in paediatric emergency care?. *Archives of Disease in Childhood*. https://doi.org/10.1136/adc.2009.159095

32. Romero-Brufau, S., Huddleston, J.M., Naessens, J.M., (...), Santrach, P.J. (2014) Widely used track and trigger scores: Are they ready for automation in practice?. *Resuscitation*. https://doi.org/10.1016/j.resuscitation.2013.12.017

33. Wheeler, I., Price, C., Sitch, A., (...), Rylance, J. (2013) Early Warning Scores Generated in Developed Healthcare Settings Are Not Sufficient at Predicting Early Mortality in Blantyre, Malawi: A Prospective Cohort Study. *PLoS ONE*. https://doi.org/10.1371/journal.pone.0059830

34. Traldi, P.D.C., de Magalhães-Barbosa, M.C., Raymundo, C.E., (...), Prata-Barbosa, A. (2022) Validity of the computerized version of the pediatric triage system CLARIPED for emergency care. *Jornal de Pediatria*. https://doi.org/10.1016/j.jped.2021.08.004

35. Weyrich, P., Christ, M., Celebi, N., Rießen, R. (2012) Triage systems in the emergency department. *Medizinische Klinik - Intensivmedizin und Notfallmedizin*. https://doi.org/10.1007/s00063-011-0075-9

36. Alkema, N.G., Hoogeveen, S.F.J.S., Cauberg, E.C.C., (...), Steffens, M.G. (2022) Magnetic Resonance Imaging–targeted Prostate Biopsy Compared with Systematic Prostate Biopsy in Biopsy-naïve Patients with Suspected Prostate Cancer. *European Urology Open Science*. https://doi.org/10.1016/j.euros.2022.08.005

37. Sarmento, Krishnamurti M. A., Sampaio, André Luiz Lopes, Santos, Tatiana Guthierre Targino, de Oliveira, Carlos Augusto Costa Pires (2017) High-frequency conductive hearing loss as a diagnostic test for incomplete ossicular discontinuity in non-cholesteatomatous chronic suppurative otitis media. *PLoS ONE*. https://doi.org/10.1371/journal.pone.0189997

38. Fenn-Buderer, N.M. (1996) Statistical methodology: I. Incorporating the prevalence of disease into the sample size calculation for sensitivity and specificity. *Academic Emergency Medicine*. https://doi.org/10.1111/j.1553-2712.1996.tb03538.x

39. Flahault, A., Cadilhac, M., Thomas, G. (2005) Sample size calculation should be performed for design accuracy in diagnostic test studies. *Journal of Clinical Epidemiology*. https://doi.org/10.1016/j.jclinepi.2004.12.009

40. So, Jimmy Bok Yan, Kapoor, Ritika, Zhu, Feng, (...), Yong, Wei Peng (2021) Development and validation of a serum microRNA biomarker panel for detecting gastric cancer in a high-risk population. *Gut*. https://doi.org/10.1136/gutjnl-2020-322065

41. Jarvis, M.J., Tunstall-Pedoe, H., Feyerabend, C., (...), Saloojee, Y. (1987) Comparison of tests used to distinguish smokers from nonsmokers. *American Journal of Public Health*. https://doi.org/10.2105/AJPH.77.11.1435

42. Downey, C.L., Tahir, W., Randell, R., (...), Jayne, D.G. (2017) Strengths and limitations of early warning scores: A systematic review and narrative synthesis. *International Journal of Nursing Studies*. https://doi.org/10.1016/j.ijnurstu.2017.09.003

43. Lowres, N., Neubeck, L., Salkeld, G., (...), Freedman, S. (2014) Feasibility and cost-effectiveness of stroke prevention through community screening for atrial fibrillation using iPhone ECG in pharmacies: The SEARCH-AF study. *Thrombosis and Haemostasis*. https://doi.org/10.1160/TH14-03-0231

44. Zachariasse, J.M., van der Hagen, V., Seiger, N., (...), Moll, H.A. (2019) Performance of triage systems in emergency care: A systematic review and meta-analysis. *BMJ Open*. https://doi.org/10.1136/bmjopen-2018-026471

45. Ferguson, G.T., Enright, P.L., Buist, A.S., Higgins, M.W. (2000) Office spirometry for lung health assessment in adults: A consensus statement from the national lung health education program. *Chest*. https://doi.org/10.1378/chest.117.4.1146

46. Hajian-Taliki, K. (2014) Sample size estimation in diagnostic test studies of biomedical informatics. *Journal of Biomedical Informatics*. https://doi.org/10.1016/j.jbi.2014.02.013

47. Jones, J.G. (2004) The 'fixed performance' venturi: Effect of downstream pressure on outflow and FIO2. *Anaesthesia*. https://doi.org/10.1111/j.1365-2044.2004.03754.x

48. Karli, A., Guler, M., Karli, H.B., Unluturk, B.D. (2025) Enhancing Pulse Oximetry Accuracy with Personal Parameter Integration in Wearable Devices. *Midwest Symposium on Circuits and Systems*. https://doi.org/10.1109/MWSCAS53549.2025.11244460

49. Branstad, A., Buttar, S., Kabir, R., (...), Fischer, G. (2026) Pilot Study: Using Reflectance as a Proxy for Skin Pigmentation to Assess Pulse Oximeter Accuracy. *Hospital Pediatrics*. https://doi.org/10.1542/hpeds.2025-008471

50. Khan, H.A., Hanna, N., Chaskes, M.J., (...), Curtis, A.B. (2020) Screening for atrial fibrillation in high-risk nursing home residents. *Heart Rhythm O2*. https://doi.org/10.1016/j.hroo.2020.02.003

51. Alemayehu, Z.G., Sime, B.L., Keraga, A.S., (...), Nida, H.W. (2026) Wearable single-lead ECG devices to improve cardiac arrhythmia screening in Ethiopia: Opportunities and challenges. *Heart Rhythm O2*. https://doi.org/10.1016/j.hroo.2026.01.002

52. Tuck, B.M., Karelitz, J.L., Tomko, R.L., (...), McClure, E.A. (2021) Mobile, remote, and individual focused: Comparing breath carbon monoxide readings and abstinence between smartphone-enabled and stand-alone monitors. *Nicotine and Tobacco Research*. https://doi.org/10.1093/ntr/ntaa203

53. Chatkin, J., Fritscher, L., de Abreu, C., (...), Fritscher, C. (2007) Exhaled carbon monoxide as a marker for evaluating smoking abstinence in a Brazilian population sample. *Primary Care Respiratory Journal*. https://doi.org/10.3132/pcrj.2007.00008

54. Wee, L.-H., West, R., Mariapun, J., (...), Jit, S. (2015) Should the threshold for expired-air carbon monoxide concentration as a means of verifying self-reported smoking abstinence be reduced in clinical treatment programmes? Evidence from a Malaysian smokers' clinic. *Addictive Behaviors*. https://doi.org/10.1016/j.addbeh.2015.03.021

55. Wald, N.J., Idle, M., Boreham, J., Bailey, A. (1981) Carbon monoxide in breath in relation to smoking and carboxyhaemoglobin levels. *Thorax*. https://doi.org/10.1136/thx.36.5.366

56. Herrera-Sánchez, A., Álvarez-Chávez, F.E., Castillo-Hernandez, M.C., (...), González-Virla, B. (2019) Peak expiratory flow vs. Spirometry for the diagnosis of asthma in adults. *Revista Alergia Mexico*. https://doi.org/10.29262/ram.v66i3.630

57. Avdimiretz, N., Wilson, D., Grasemann, H. (2020) Comparison of a handheld turbine spirometer to conventional spirometry in children with cystic fibrosis. *Pediatric Pulmonology*. https://doi.org/10.1002/ppul.24743

58. Gao, Y., Liang, B., Su, X., (...), Liu, B. (2025) Reliability and usability of a portable spirometer compared to a laboratory spirometer. *BMC Pulmonary Medicine*. https://doi.org/10.1186/s12890-025-03690-1

59. Zhang, H., Li, L., Jiao, D., (...), Jin, M. (2020) An interrater reliability study of pulmonary function assessment with a portable spirometer. *Respiratory Care*. https://doi.org/10.4187/respcare.07116

60. Bongers, Thomas, O'Driscoll, B Ronan (2006) Effects of equipment and technique on peak flow measurements. *BMC Pulmonary Medicine*. https://doi.org/10.1186/1471-2466-6-14

61. Eichenhorn, M.S., Beauchamp, R.K., Harper, P.A., Ward, J.C. (1982) An assessment of three portable peak flow meters. *Chest*. https://doi.org/10.1378/chest.82.3.306

62. Yawn, B.P., Enright, P.L., Lemanske, R.F., (...), Boushey, H. (2007) Spirometry can be done in family physicians' offices and alters clinical decisions in management of asthma and COPD. *Chest*. https://doi.org/10.1378/chest.06-2722

63. Barker, S.J., Wilson, W.C. (2023) Racial effects on Masimo pulse oximetry: a laboratory study. *Journal of Clinical Monitoring and Computing*. https://doi.org/10.1007/s10877-022-00927-w

64. Fawzy, A., Wu, T.D., Wang, K., (...), Garibaldi, B.T. (2022) Racial and Ethnic Discrepancy in Pulse Oximetry and Delayed Identification of Treatment Eligibility among Patients with COVID-19. *JAMA Internal Medicine*. https://doi.org/10.1001/jamainternmed.2022.1906

65. Nagler, J., Krauss, B. (2010) Monitoring the Procedural Sedation Patient: Optimal Constructs for Patient Safety. *Clinical Pediatric Emergency Medicine*. https://doi.org/10.1016/j.cpem.2010.09.007

66. Kelly, K.L., Carlson, A.R., Allison, T.G., Johnson, B.D. (2020) A comparison of finger and forehead pulse oximeters in heart failure patients during maximal exercise. *Heart and Lung*. https://doi.org/10.1016/j.hrtlng.2019.10.012

67. Nixdorff, Jan, Zablotski, Yury, Hartmann, Katrin, Dörfelt, Rene (2021) Comparison of Transmittance and Reflectance Pulse Oximetry in Anesthetized Dogs. *Frontiers in Veterinary Science*. https://doi.org/10.3389/fvets.2021.643966

68. Verbiest-van Gurp, Nicole, Uittenbogaart, Steven B, Lucassen, Wim A M, (...), van Weert, Henk C P M (2022) Detection of atrial fibrillation in primary care with radial pulse palpation, electronic blood pressure measurement and handheld single-lead electrocardiography: a diagnostic accuracy study. *BMJ Open*. https://doi.org/10.1136/bmjopen-2021-059172

69. McConnell, M.V., Turakhia, M.P., Harrington, R.A., (...), Ashley, E.A. (2018) Mobile Health Advances in Physical Activity, Fitness, and Atrial Fibrillation: Moving Hearts. *Journal of the American College of Cardiology*. https://doi.org/10.1016/j.jacc.2018.04.030

70. Manetas-Stavrakakis, Nikolaos, Sotiropoulou, Ioanna Myrto, Paraskevas, Themistoklis, (...), Briasoulis, Alexandros (2023) Accuracy of Artificial Intelligence-Based Technologies for the Diagnosis of Atrial Fibrillation: A Systematic Review and Meta-Analysis. *Journal of Clinical Medicine*. https://doi.org/10.3390/jcm12206576

71. Santos, V.S., Goletti, D., Kontogianni, K., (...), Eduardo Cuevas, L.E. (2019) Acute phase proteins and IP-10 as triage tests for the diagnosis of tuberculosis: systematic review and meta-analysis. *Clinical Microbiology and Infection*. https://doi.org/10.1016/j.cmi.2018.07.017

72. Bonadia, N., Fanelli, P.M.C., Polla, D.A.D., (...), Covino, M. (2026) Assessing Risk of Early ICU Admission or Death at Emergency Department Triage: Clinical Judgment Versus Early Warning Scores. *Academic Emergency Medicine*. https://doi.org/10.1111/acem.70377

73. Sweney, J.S., Poss, W.B., Grissom, C.K., Keenan, H.T. (2012) Comparison of severity of illness scores to physician clinical judgment for potential use in pediatric critical care triage. *Disaster Medicine and Public Health Preparedness*. https://doi.org/10.1001/dmp.2012.17

74. Obuchowski, N.A. (1998) Sample size calculations in studies of test accuracy. *Statistical Methods in Medical Research*. https://doi.org/10.1177/096228029800700405

75. Rathod, M., Unger, S., Ross, H., (...), Franklin, D. (2025) Development of a melanin-inclusive reflective pulse oximeter model for equitable performance tested in a large animal model undergoing hypoxia: pilot study. *Progress in Biomedical Optics and Imaging - Proceedings of SPIE*. https://doi.org/10.1117/12.3042334