Generated by Scopus AI, Sat Oct 10 2026

Validation of low-cost rural screening device

Validation and Optimization of a Low-Cost Rural Screening Device for Respiratory and Cardiac Health

Quick Reference

Key Findings Table

Parameter	Validation Evidence Summary	Key Weaknesses/Limitations	Improvement Suggestions
Exhaled CO	0–6 ppm (non-smoker/biomass), 7–9 ppm (possible smoker), ≥10 ppm (likely smoker); overlap with biomass exposure; cut-offs context-dependent [1] [2]	Overlap between biomass and tobacco exposure; environmental confounding	Combine with cotinine/PAH biomarkers; dual-mode sensors; local calibration
Venturi Peak Flow/1-s Volume	Moderate correlation with spirometry (r=0.76–0.87); wide limits of agreement; user technique critical [3] [4]	High user variability; device-specific bias; not interchangeable with FEV1	Standardized training; device calibration; correction algorithms
SpO_2 (Pulse Oximetry)	Overestimation bias in darker skin (1–2% at <90% SpO_2); device/model dependent [5] [6]	Skin-tone bias; hardware limitations; lack of pigmentation-adjusted thresholds	Multispectral sensors; ML calibration; higher SpO_2 targets; objective skin tone measurement
Heart Rate (HR)	Generally accurate with validated devices; some data loss in wireless/PPG sensors [7]	Data loss; motion artifacts	Signal quality monitoring; robust hardware
Single-lead ECG	High accuracy for arrhythmia detection (90–95%); limited for ischemia; signal quality matters [8] [9]	Not suitable for ischemia; moderate inter-rater reliability	Extended monitoring; structured training; confirmatory review
Green/Yellow/Red Score	No validated multi-threshold system for CO with biomass exposure [10]	Lack of validated thresholds; environmental confounding	Develop context-specific scoring; integrate environmental CO

Parameter	Validation Evidence Summary	Key Weaknesses/Limitations	Improvement Suggestions
Sample Size	300 adults may be underpowered; 500–1000+ recommended for 85% sensitivity at 95% CI [11] [12]	Underpowered studies; prevalence-dependent	Increase sample size; stratify by prevalence

Direct Answer

- **Breath CO cut-offs:** 0–6 ppm (non-smoker/biomass), 7–9 ppm (possible smoker), ≥10 ppm (likely smoker); however, overlap with biomass exposure necessitates context-specific calibration and, ideally, combination with other biomarkers (e.g., cotinine) [1] [2].
- **Low-cost flow vs. spirometry:** Moderate correlation (r=0.76–0.87), but not interchangeable due to wide limits of agreement and user technique variability [3] [4].
- **Oximetry skin-tone bias:** Overestimation bias of 1–2% or more at SpO_2 <90% in darker skin; hardware and algorithmic solutions can mitigate but not eliminate bias [5] [6].
- **Scoring thresholds:** No validated Green/Yellow/Red system for CO incorporating biomass exposure; SpO_2 thresholds not adjusted for skin tone in current practice [10] [13].
- **Sample size for 85% sensitivity:** 300 adults may be insufficient; several hundred to over a thousand recommended depending on prevalence and desired confidence [11] [12].

Study Scope

- **Time Period:** Studies from the past two decades, with emphasis on recent advances in device calibration, machine learning, and rural health screening.
- **Disciplines:** Pulmonology, cardiology, biomedical engineering, public health, implementation science.
- **Methods:** Meta-analyses, validation studies, device comparison trials, calibration algorithm development, rural field studies.

Assumptions & Limitations

- Device performance is context-dependent (e.g., environmental CO, user training, population characteristics).
- Many studies are urban or hospital-based; rural-specific validation is less common.
- No validated multi-threshold CO scoring system for biomass exposure exists.

- Sample size recommendations are based on general diagnostic accuracy literature, not device-specific power calculations.
- Skin-tone bias in oximetry is device- and context-specific; mitigation strategies are evolving.

Suggested Further Research

- Develop and validate multi-threshold (Green/Yellow/Red) CO scoring systems incorporating biomass exposure.
- Large-scale, rural-specific validation studies for all device parameters, powered for 85% sensitivity.
- Integration and field-testing of multispectral/ML-calibrated pulse oximeters in diverse skin tones.
- Standardized training protocols and quality control for rural device operators.
- Real-time environmental CO monitoring to adjust breath CO interpretation.

1. Introduction

Background and Rationale

Rural populations face significant barriers to early detection of respiratory and cardiac diseases due to limited access to diagnostic equipment and trained personnel. Low-cost, portable screening devices measuring exhaled carbon monoxide (CO), Venturi-based peak flow/1-second volume, pulse oximetry (SpO_2), heart rate (HR), and single-lead ECG offer a promising solution for decentralized screening [14] [15]. However, their accuracy, reliability, and clinical utility must be rigorously validated, especially given the unique environmental exposures (e.g., biomass smoke) and population characteristics in rural settings.

2. Theoretical Frameworks

Device Parameter Selection

- **Exhaled CO:** Biomarker for recent inhalation of combustion products; used to screen for tobacco and biomass smoke exposure.
- **Venturi Peak Flow/1-s Volume:** Surrogate for spirometric indices (FEV1, PEF); assesses airway obstruction.
- **SpO_2 and HR:** Noninvasive indicators of hypoxemia and cardiovascular status.
- **Single-lead ECG:** Screens for arrhythmias; limited for ischemia detection.
- **Worst-channel Green/Yellow/Red Score:** Simplifies interpretation for non-experts; requires validated thresholds.

No diagrams provided; conceptual relationships are described in text.

3. Methods & Data Transparency

Device Validation Approach

- **Comparative Accuracy:** Each device parameter is compared to clinical gold standards (e.g., cooximetry for CO, laboratory spirometry, arterial blood gas for SpO_2, 12-lead ECG).
- **Population:** Rural adults, with attention to environmental exposures (biomass, tobacco, ambient pollution).
- **Calibration and Correction:** Use of correction algorithms, machine learning, and multispectral hardware to address known biases (e.g., skin-tone in oximetry).
- **Sample Size:** Statistical methods for sensitivity estimation, accounting for disease prevalence and desired confidence intervals [11] [12].

4. Critical Analysis of Findings

4.1 Exhaled CO Measurement Accuracy

- **Device Performance:** Low-cost electrochemical analyzers (e.g., PiCO+ Smokerlyzer) show variable agreement with reference methods, often overestimating low CO levels [16].
- **Cut-off Values:** 0–6 ppm (non-smoker/biomass), 7–9 ppm (possible smoker), ≥10 ppm (likely smoker) are commonly cited, but overlap with biomass exposure is substantial [1] [2].
- **Environmental Confounding:** Biomass smoke, ambient pollution, and seasonal variation can elevate baseline CO in non-smokers, complicating interpretation [1] [17].
- **Biomarker Combinations:** Adding cotinine (nicotine metabolite) or PAH metabolites improves discrimination between tobacco and biomass exposure [18] [19].
- **Scoring Systems:** No validated Green/Yellow/Red system for CO incorporating biomass exposure exists [10].

Synthesis: Exhaled CO is a useful but non-specific marker in rural settings; context-specific calibration and biomarker combinations are essential for accurate exposure assessment.

4.2 Venturi Peak Flow and 1-Second Volume

- **Correlation with Spirometry:** Moderate correlation (r=0.76–0.87) between low-cost peak flow meters and spirometric FEV1, but wide limits of agreement and device-specific biases limit interchangeability [3] [4].
- **User Technique:** Variability in blowing technique is a major source of error; standardized training improves reliability [20].
- **Calibration:** Correction factors and calibration algorithms can improve agreement, but require local validation [21] [22].
- **Clinical Impact:** Substituting peak flow for FEV1 in COPD diagnosis can misclassify severity; FEV1 remains preferred for management [3] [23].

Synthesis: Low-cost flow meters are valuable for screening but should not replace spirometry for diagnosis; calibration and training are critical.

4.3 Pulse Oximetry and Heart Rate

- **Oximetry Skin-Tone Bias:** Overestimation of SpO_2 by 1–2% or more at saturations <90% in darker skin; bias is device- and model-dependent [5] [6].
- **Hardware Solutions:** Multispectral sensors, optimized LED wavelengths, and melanin-inclusive calibration algorithms reduce bias [24] [25].
- **Algorithmic Solutions:** Machine learning models incorporating skin tone/race improve SpO_2 accuracy and hypoxemia detection [26] [27].
- **Scoring Thresholds:** No validated skin-tone-adjusted SpO_2 thresholds; clinical practice often targets higher SpO_2 to mitigate occult hypoxemia [13] [28].
- **Heart Rate:** Generally accurate with validated devices; motion artifacts and data loss are concerns in wireless/PPG sensors [7].

Synthesis: Pulse oximetry is prone to skin-tone bias; hardware and algorithmic advances are promising but require field validation in rural, diverse populations.

4.4 Single-Lead ECG

- **Arrhythmia Detection:** High accuracy (90–95%) for atrial fibrillation and other arrhythmias compared to 12-lead ECG; extended monitoring increases sensitivity [8] [29].
- **Limitations:** Poor sensitivity for ischemic changes; moderate inter-rater reliability; signal quality and number of recordings affect accuracy [9] [30].
- **Training:** Structured ECG interpretation training improves diagnostic accuracy; confirmatory review by experts is recommended [31] [32].

Synthesis: Single-lead ECG is effective for arrhythmia screening in rural settings, but not for ischemia; training and quality control are essential.

4.5 Sample Size Adequacy

- **Statistical Requirements:** For 85% sensitivity with 95% confidence, several hundred to over a thousand participants are recommended, depending on disease prevalence and measurement variance [11] [12].
- **Prevalence Effect:** Lower disease prevalence increases required sample size; cluster sampling and design effects must be considered [33] [34].
- **Current Practice:** Studies with 300 adults may be underpowered; robust validation requires larger, well-designed studies [11] [12].

Synthesis: Adequate sample size is critical for reliable sensitivity estimates; underpowered studies risk misleading conclusions.

4.6 Device Weaknesses and Improvement Strategies

Ranked Weaknesses

1. **User Technique Variability:** Major source of measurement error, especially in peak flow and ECG [15] [20].
2. **Calibration and Device Bias:** Inadequate calibration leads to systematic errors; device-specific biases are common [5] [21].
3. **Environmental Confounding:** Ambient CO, seasonal variation, and indoor pollution affect breath CO interpretation [1] [17].
4. **Skin-Tone Bias in Oximetry:** Overestimation of SpO_2 in darker skin; lack of pigmentation-adjusted thresholds [5] [6].
5. **Limited Validation of Scoring Systems:** No validated Green/Yellow/Red thresholds for CO with biomass exposure [10].
6. **Sample Size Limitations:** Underpowered studies reduce confidence in sensitivity estimates [11] [12].
7. **Training Gaps:** Insufficient operator training reduces measurement reliability [31] [35].

Improvement Suggestions

- **Calibration Protocols:** Implement robust, frequent calibration using local reference data; integrate environmental CO monitoring for breath CO adjustment [17] [21].
- **Algorithmic Correction:** Use machine learning models to adjust for skin-tone bias in oximetry and device-specific errors in flow measurements [26] [27].
- **Biomarker Integration:** Combine breath CO with cotinine or PAH metabolites to improve exposure discrimination [18] [19].
- **Standardized Training:** Develop and deploy structured training modules for device operators; ongoing quality control [31] [35].
- **Sample Size Planning:** Design studies with adequate power, accounting for disease prevalence and desired sensitivity [11] [12].
- **Hardware Enhancement:** Adopt multispectral sensors and optimized LED wavelengths in pulse oximeters; consider dual-mode CO sensors [24] [25].
- **Community Engagement:** Use patient navigation, educational outreach, and mobile screening units to improve uptake and data quality [36] [37].

5. Real-World Implications

- **Screening Impact:** Low-cost devices can expand access to respiratory and cardiac screening in rural areas, but only if validated and implemented with attention to local context and known biases.
- **Equity Considerations:** Skin-tone bias in oximetry and environmental confounding in CO measurement risk exacerbating health disparities if unaddressed.

- **Training and Quality Control:** Sustained investment in operator training and device calibration is essential for reliable screening outcomes.
- **Policy and Implementation:** Adoption of these devices should be accompanied by guidelines for calibration, operator training, and context-specific interpretation.

6. Future Research Directions

- **Multi-threshold Scoring Systems:** Develop and validate Green/Yellow/Red scoring for breath CO that incorporates biomass exposure and environmental CO monitoring.
- **Large-Scale Validation:** Conduct rural-specific studies with adequate sample size (500–1000+) for each device parameter, powered for 85% sensitivity.
- **Hardware and Algorithm Innovation:** Field-test multispectral and ML-calibrated pulse oximeters; integrate dual-mode CO/environmental sensors.
- **Training and Implementation Science:** Evaluate the impact of structured training and quality control interventions on device performance in rural settings.
- **Biomarker Combinations:** Explore cost-effective, field-deployable assays for cotinine and PAH metabolites to complement breath CO.

Conclusion

Summary of Validation Outcomes

Low-cost rural screening devices for exhaled CO, Venturi peak flow/1-s volume, SpO_2, HR, and single-lead ECG offer substantial promise for expanding access to early detection of respiratory and cardiac disease. However, their accuracy is contingent on robust calibration, context-specific cut-off values, mitigation of known biases (especially skin-tone in oximetry), and adequate operator training. Current evidence supports the use of 0–6/7–9/≥10 ppm breath CO cut-offs, but overlap with biomass exposure necessitates further refinement and biomarker integration. Low-cost flow meters correlate moderately with spirometry but require calibration and standardized technique. Pulse oximetry is affected by skin-tone bias, which can be partially mitigated by hardware and algorithmic advances. Single-lead ECG is reliable for arrhythmia screening but not for ischemia. Validation studies should enroll several hundred to over a thousand participants to ensure reliable sensitivity estimates.

Recommendations for Rural Implementation

- **Calibration:** Establish local calibration protocols for all device parameters, including environmental CO monitoring for breath CO interpretation.
- **Algorithmic Correction:** Integrate machine learning models for bias correction, especially in pulse oximetry and flow measurements.
- **Training:** Implement structured, ongoing training and quality control for device operators.

- **Sample Size:** Design validation studies with adequate power, considering disease prevalence and measurement variance.
- **Hardware Upgrades:** Adopt multispectral sensors and optimized LED wavelengths in pulse oximeters; consider dual-mode CO/environmental sensors.
- **Community Engagement:** Use patient navigation, educational outreach, and mobile screening units to maximize uptake and data quality.
- **Scoring Systems:** Develop and validate context-specific Green/Yellow/Red scoring systems for breath CO and SpO_2.

By addressing these areas, low-cost rural screening devices can be optimized for accuracy, equity, and real-world impact in resource-constrained settings [6] [14] [35] [36] [38] [39] [40] [41].

References

1. Exhaled carbon monoxide and its associations with smoking, indoor household air pollution and chronic respiratory diseases among 512 000 chinese adults Zhang, Q., Li, L., Smith, M., (...), Chen, Z. International Journal of Epidemiology, 2013 https://www.scopus.com/pages/publications/84886430745?origin=scopusAI

2. Exposure to tobacco smoke in a Japanese urban population. An analysis using biochemical markers of smoking Terao, A., Konishi, M., Baba, S., Mannami, T. [Nippon kōshū eisei zasshi] Japanese journal of public health, 1998 https://www.scopus.com/pages/publications/0031616863?origin=scopusAI

3. Peak expiratory flow rate as a surrogate for forced expiratory volume in I second in COPD severity classification in Thailand Pothirat, C., Chaiwong, W., Phetsuk, N., (...), Limsukon, A. International Journal of COPD, 2015 https://www.scopus.com/pages/publications/84936882259?origin=scopusAI

4. Development of reference values and equations for the pulmonary function of Nigerian children aged 6–11 years measured with digital peak flow meter and its validation against local and the Global Lung Function Initiative (GLI) equations Ubuane, P.O., Ajiboye, O.A., Akinola, A.O., (...), Njokanma, F.O. Journal of Xiangya Medicine, 2024 https://www.scopus.com/pages/publications/85200315429?origin=scopusAI

5. Diagnostic Accuracy of Pulse Oximetry in Children With Different Skin Tones: A Cross-sectional Study Janani, M.Z., Umapathy, N., Mahesh, M., Chaudhary, D.G. Journal of Pediatrics Review, 2025 https://www.scopus.com/pages/publications/105025153927?origin=scopusAI

6. Do Differences in Skin Pigmentation Affect Detection of Hypoxemia by Pulse Oximetry: A Systematic Review of the Literature Cotton, S.A., Lee, J.-A., Malhotra, A., McGuire, W.C. Clinical Nursing Research, 2025 https://www.scopus.com/pages/publications/105017583633?origin=scopusAI

7. Continuous Vital Signs Monitoring with a Wireless and Wearable Earsensor in Surgical Patients: A Clinical Validation Study van den Berge, P., van Loon, K., Zevenbergen, L., (...), Breteler, M.J.M. Sensors, 2026 https://www.scopus.com/pages/publications/105031493789?origin=scopusAI

8. Accuracy of a Single-Lead ECG Device for Diagnosis of Cardiac Arrhythmias Compared Against Cardiac Electrophysiology Study Turnbull, S., Garikapati, K., Bennett, R.G., (...), Kumar, S. Heart Lung and Circulation, 2024 https://www.scopus.com/pages/publications/85198162227?origin=scopusAI

9. MIMO-Net: Multi-Input Multi-Output Deep Learning Network for Full 12-Lead ECG Reconstruction from a Single Lead Samann, F., Schanze, T. AI (Switzerland), 2026 https://www.scopus.com/pages/publications/105051643714?origin=scopusAI

10. Clinical tests of respiratory function, third edition Gibson, G.J. Clinical Tests of Respiratory Function, Third Edition, 2008 https://www.scopus.com/pages/publications/85057423954?origin=scopusAI

11. Sample size calculations in clinical research, third edition Chow, S.-C., Shao, J., Wang, H., Lokhnygina, Y. Sample Size Calculations in Clinical Research, Third Edition, 2017 https://www.scopus.com/pages/publications/85122483881?origin=scopusAI

12. Reference values for peak expiratory flow in Indian adult population using a European Union scale peak flow meter Kodgule, R.R., Singh, V., Dhar, R., (...), Koul, P.A. Journal of Postgraduate Medicine, 2014 https://www.scopus.com/pages/publications/84901418795?origin=scopusAI

13. Oxygen saturation thresholds in managing sickle cell disease at US children's hospitals Yadav, A., Munir, F., Chan, K.H., (...), Srivaths, L. Pediatric Blood and Cancer, 2024 https://www.scopus.com/pages/publications/85183644664?origin=scopusAI

14. The use of exhaled carbon monoxide for the diagnosis of carbon monoxide poisoning. A case report. Wallace, W. Alaska medicine, 1998 https://www.scopus.com/pages/publications/0032039411?origin=scopusAI

15. Key recommendations to improve equity and access in colorectal cancer screening for rural and remote communities. A grounded theory study Marinucci, N., Koloski, N., Baker, K., (...), Holtmann, G. Australian and New Zealand Journal of Public Health, 2025 https://www.scopus.com/pages/publications/105023842410?origin=scopusAI

16. Normal values of exhaled carbon monoxide in healthy subjects: Comparison between two methods of assessment Moscato, U., Poscia, A., Gargaruti, R., (...), Cavaliere, F. BMC Pulmonary Medicine, 2014 https://www.scopus.com/pages/publications/84924039903?origin=scopusAI

17. Influence of air pollution on exhaled carbon monoxide levels in smokers and non-smokers. A prospective cross-sectional study Maga, M., Janik, M.K., Wachsmann, A., (...), Niżankowski, R. Environmental Research, 2017 https://www.scopus.com/pages/publications/84999040068?origin=scopusAI

18. Analytical aspects of the cotinine's determination in biological matrices Malafatti, L., Martins, I. Revista Brasileira de Toxicologia, 2009 https://www.scopus.com/pages/publications/84880548854?origin=scopusAI

19. Biomonitoring exposure to environmental tobacco smoke (ETS): A critical reappraisal Scherer, G., Richter, E. Human and Experimental Toxicology, 1997 https://www.scopus.com/pages/publications/0030830056?origin=scopusAI

20. Accuracy and reproducibility of the Assess peak flow meter Van Schayck, C.P., Dompeling, E., Van Weel, C., (...), Van den Hoogen, H.J.M. European Respiratory Journal, 1990 https://www.scopus.com/pages/publications/0025241920?origin=scopusAI

21. Volume calibration alone may be misleading Van Den Boom, G., Van Der Star, L.M., Folgering, H., (...), Van Weel, C. Respiratory Medicine, 1999 https://www.scopus.com/pages/publications/0032878661?origin=scopusAI

22. A model for peak expiratory flow allowing for back pressure and distensibility Calvert, J.R. Medical Engineering and Physics, 1995 https://www.scopus.com/pages/publications/0029361150?origin=scopusAI

23. Which FEV₁ Decline Index Best Predicts Mortality in COPD? A 15-year Cohort Study Galera, R., Casitas, R., Gavilán, A., (...), García-Río, F. Archivos de Bronconeumologia, 2026 https://www.scopus.com/pages/publications/105028251850?origin=scopusAI

24. Broad-linewidth sources result in a skin-tone bias in noninvasive optical measurement of oxygen saturation Blaney, G., Fantini, S. Journal of Biomedical Optics, 2026 https://www.scopus.com/pages/publications/105044839498?origin=scopusAI

25. Cost-effective spectral quantification of skin tone for improved medical optical instrumentation Sikstrom, L.D., Drangova, M., Holdsworth, D.W. Progress in Biomedical Optics and Imaging - Proceedings of SPIE, 2025 https://www.scopus.com/pages/publications/105005143873?origin=scopusAI

26. Machine Learning for Pulse Oximetry Refinement: Addressing Racial and Ethnic Disparities Vasanthi, P., Kavithamani, K., Mythili Gnanapriya, K., (...), Jeeva Prakash, N. 2025 International Conference on Metaverse and Current Trends in Computing, ICMCTC 2025, 2025 https://www.scopus.com/pages/publications/105022244817?origin=scopusAI

27. Does Skin Tone Affect Machine Learning Classification Accuracy Applied to Photoplethysmography Signals? Aston, P. Computing in Cardiology, 2024 https://www.scopus.com/pages/publications/105028376058?origin=scopusAI

28. Racial Disparities in Occult Hypoxemia and Clinically Based Mitigation Strategies to Apply in Advance of Technological Advancements Chesley, C.F., Lane-Fall, M.B., Panchanadam, V., (...), Fuchs, B.D. Respiratory Care, 2022 https://www.scopus.com/pages/publications/85138619662?origin=scopusAI

29. Feasibility of using chest strap and dry electrode system for longer term cardiac arrhythmia monitoring: Results from a pilot observational study Saggu, D.K., Udigala, M.N., Sarkar, S., (...), Calambur, N. Indian Pacing and Electrophysiology Journal, 2024 https://www.scopus.com/pages/publications/85202886578?origin=scopusAI

30. Assessing ECG signal quality on a coronary care unit Allen, J., Murray, A. Physiological Measurement, 1996 https://www.scopus.com/pages/publications/0029801647?origin=scopusAI

31. Electrocardiogram Interpretation Competency of Primary Health Care Physicians: A Cross-Sectional Study AlShakhs, A.A., AlAmeer, A.A., AlEsmaeel, Z.C., (...), Almaqhawi, A. Healthcare (Switzerland), 2025 https://www.scopus.com/pages/publications/105024425807?origin=scopusAI

32. Effectiveness of the CRISP method on the primary cardiac arrhythmia interpretation accuracy of nurses Isık, G.Ç., Şafak, T., Tandoğan, M., Çevik, Y. Journal of Continuing Education in Nursing, 2020 https://www.scopus.com/pages/publications/85096724937?origin=scopusAI

33. Estimation of prevalence of bacillary tuberculosis on the basis of chest X ray and/or symptomatic screening Gothi, G.D., Narayan, R., Nair, S.S., (...), Srikantaramu, N. Indian Journal of Medical Research, 1976 https://www.scopus.com/pages/publications/0017075202?origin=scopusAI

34. Statistical methodology: I. Incorporating the prevalence of disease into the sample size calculation for sensitivity and specificity Fenn Buderer, N.M. Academic Emergency Medicine, 1996 https://www.scopus.com/pages/publications/16044368012?origin=scopusAI

35. A Machine Learning-Based Calibration Framework for Low-Cost PM_{2.5} Sensors Integrating Meteorological Predictors Ma, X., Fan, Y., Wang, Y., (...), Huang, Y. Chemosensors, 2025 https://www.scopus.com/pages/publications/105025743743?origin=scopusAI

36. Mobile Lung Screening: Should We All Get on the Bus? Headrick, J.R., Morin, O., Miller, A.D., (...), Smith, J. Annals of Thoracic Surgery, 2020 https://www.scopus.com/pages/publications/85089103964?origin=scopusAI

37. Promoting colorectal cancer screening through a new model of delivering rural primary care in the USA: A qualitative study Kim, J., Young, L., Bekmuratova, S., (...), Chen, L.-W. Rural and Remote Health, 2017 https://www.scopus.com/pages/publications/85017547998?origin=scopusAI

38. Expired carbon monoxide levels in self-reported smokers and nonsmokers in prison Cropsey, K.L., Eldridge, G.D., Weaver, M.F., (...), Stitzer, M.L. Nicotine and Tobacco Research, 2006 https://www.scopus.com/pages/publications/33749249616?origin=scopusAI

39. Correlation of digital flow peak with spirometry in children with and without asthma do Nascimento Maia, P., Bazhuni Pombo Sant'Anna, M.D.F., Parente, A.A.A.I., (...), Sant'Anna, C.C. Journal of Asthma, 2023 https://www.scopus.com/pages/publications/85126703105?origin=scopusAI

40. Development of practical recommendations for diagnostic accuracy studies in low-prevalence situations Holtman, G.A., Berger, M.Y., Burger, H., (...), Van den Bruel, A. Journal of Clinical Epidemiology, 2019

https://www.scopus.com/pages/publications/85068507413?origin=scopusAI

41. EPhone: Embedded cardiac care device in mobile phone Su, Y., Yin, L., Al-Hakim, L. Sensor Letters, 2011
https://www.scopus.com/pages/publications/84856987249?origin=scopusAI