

Contents

1	Abstract	3
2	Introduction	5
3	Theory	7
	Leica software	7
	Image Processing	7
4	Method	9
	Microscope	9
	Automated scanning	10
	Overview images	10
	SHG images	17
	DAPI images	17
	Correlating images with patient data	17
	Collection of SHG images	18
	Technical details	18
	Hardware aspects	18
	Leica software details	19
	Software development	20
5	Result	21
	Segmentation	21
6	Discussion	23

7	Conclusion	25
8	Appendix	27
	Slide map errors	27

Chapter 1

Abstract

What to communicate: goal, overview of experiences made, results

This thesis documents work on automatic microscope imaging of breast tumor tissue micro arrays and how the images can be analyzed for a supplement in cancer diagnosis. The overall research goal has been to classify tumor grade (I, II or III) based on the fiber structure in the tissue samples. Supervised machine learning is the method of analysis, where St. Olavs hospital has supplied a dataset of tissue samples at the tumor peripheral from 924 (TODO update exact number) patients.

Automated microscope scanning is in principle straight forward, but the implementation will be dependent on many aspects of the experimental setup. In general, some of the aspects discussed in this thesis are:

- correcting for systematic errors like intensity variations
- create image analysis algorithms that are robust to experimental variations
- verify that metrics reported by the system are the real physical ones
- finding good compromises between time, signal quality and ease of measurement
- scanning mirror- versus stage coordinate system and reliable stitching
- writing cross-platform software

The general aspects listed above are not unique to the experiments and experimental setup, and could potentially be useful for others. But this thesis will also address issues directly associated to tissue micro arrays and the Leica SP8 microscope:

- adjusting z-plane for large area samples with micrometer precision
- working around Leica software limitations

Results on the dataset was positive/negative. Details on the result.

A proposal for further research with the same dataset is extracting more features from the images and use equivalent methods to find relationships to the clinic data. IF POSITIVE RESULT: Collecting

tissue sample is a part of the standard procedure in breast cancer diagnosis at St. Olavs hospital, and one can do the analysis described in this thesis to further confirm or falsify the result.

Chapter 2

Introduction

What to communicate: motivation, brief summary of chapters

With a population just above 5 million [1](#), three thousand women are diagnosed with breast cancer each year [2](#) in Norway. This makes breast cancer the most common kind of cancer, affecting one of every eleventh woman. Luckily the cancer form is often treatable, and in 2012 there was 649 fatalities caused by breast cancer [3](#). The diagnosis is an act of several steps, and currently contains the following at St. Olavs hospital:

- x-ray mammography
- ultra sound screening
- tissue sample(s)

In particular, pathologists suggest that aggressiveness of a tumor is related to how fiber is aligned at the tumor peripheral. In example straight aligned fibers can be a sign that tumor cells have modified the stroma to promote spreading of cells. The alignment of fibers is a feature which can be extracted by image processing. Since several techniques to extract features is imaginable, supervised machine learning is practical for finding novel approaches.

From report - may use some of this: Over three million published articles on pubmed with keyword cancer shows the huge research effort for understanding, diagnosing and treating cancer diseases. The research focus is mainly on tumor cells, but a segment of interest which is increasing is research on tumor stroma as seen in figure 1.1.

Figure 1.1: Amount of published articles by year on different search terms. The search term tumor cells outnumber the others by two orders of magnitude. Note the logarithmic scale on y-axis.

Tumor stroma is the environment of cells, and it can be suppressing or supporting the function of the tumor cells. It is suggested that in the development of a tumor, the stroma is changing from being suppressive to supportive of the tumor cells [TODO REF].

In particular, collagen fiber is known to be altered in the surroundings of tumor cells under the development towards metastasis. One bio-marker for collagen fibers, is their alignment at the vicinity

of the tumor, which may predict if a tumor is malignant. The fiber alignment can be used as a diagnosis tool for malignant tumor, and an article written at NTNU have studied collagen fiber alignment in a manual qualitative manner.

St. Olav hospital have breast tissue samples from 900 pasients along with clinical data. In total three samples per pasient, one sample inside, one sample at the boundary and one sample outside the tumor. The samples is laid in a matrix on a glass slide, each glass slide having about 130 samples. As microscope scanning and analysis of such a large data set is not straightforward, this project have explored possibilities for automating the process.

To be specific, this thesis will describe method and results for - parameters for obtaining quality SHG images - effective way to scan whole glass slides of 126 samples - machine learning and correlation to clinical data

ML: En hoveddel i arbeidet har vært automatiseringen av TMA. Skrive noe om TMA og hvorfor automatisert analyse er nødvendig... skal lede opp til en beskrivelse av de tekniske utfordringene som er løst.

Chapter 3

Theory

What to communicate: theory and details that are not obvious for understanding the rest of the text

ML: I denne delen bør man primært ha med teori som er nødvendig for å forstå det som kommer i metodedelen. Altså ikke skriv for mye her før strukturen og innholdet er mer klart.)

Leica software

- socket
- CAM
- XML and scanning template (overview ST vs job ST)
- wells and fields

Image Processing

- scikit-image, utils.ipynb, defaults in code blocks
- OCR

ORB and Ransac 1: <https://peerj.com/articles/453/#p-1> ## Scanning microscope - Epi setup - scanning, descanned detectors

focal volume ## Nonlinear light interaction

Chapter 4

Method

What to communicate: experimental setup to reproduce results, description of automatic process, limitations/obstacles specific to experimental setup, brief description of software modules in use

Microscope

The images has been taken with a Leica SP8 microscope using LAS X software version 1.1.0.12420 from Leica Microsystems CMS GmbH. Two lasers was in use, a pulsing Coherent laser and a continious LASOS argon laser. Full specifications of lasers are in table 4.1.

Table 4.1: Lasers

Brand	Model	Specifications
Coherent	Chameleon Vision-S	Modelocked Ti:Sapphire, wavelengths 690-1050 nm, 2500 mW, 80 MHz pulsing, ≈ 75 ps pulse width
LASOS	LGK 7872 ML05	Argon Continious wave, wavelengths 458, 476, 488, 496 and 514 nm, 65mW

The SP8 microscope has an inverted epi-setup, with four descanned detectors and four non descanned detectors. The descanned detectors use a prism along with adjustable mirrors so that specific wavelengths can be picked out in the signal, ranging from TODO. The descanned detectors was used with band pass filters of 525/50 nm and 445/20 nm. Two of the descanned detectors are behind the objective and two on opposite side of the objective behind a collector, which makes it possible to measure both backward and forward light.

Automated scanning

Communicate: the procedure of automatic scanning

The automated scanning aims to lift the burden of manually labor and prevent errors in the imaging process by finding regions with the samples in an overview image. The process consists roughly of the steps:

- Take an overview image with low magnification
- Segment the overview image
- Allow user to confirm or adjust the segmentation
- Scan each region

Overview images was taken with a 10x air objective, equalized and stitched. The equalization step corrects uneven illumination and increases contrast for viewing purposes. To improve robustness of segmentation, a local bilateral population filter was applied to the stitched image before it is thresholded. Each separate region in the segmentation are sorted by their area size, small regions are excluded and the user can exclude or add regions if some of the samples are not detected. Row and column position of the regions are calculated by sorting them by their position in the image. A more detailed description follows.

Overview images

Overview images was taken with an technique similar to bright-field microscopy except that the light source is a scanning laser. The laser in use was the argon laser in table 4.1 with 514 nm emission line, output power set to 2.48% and intensity to 0.10. Forward light was imaged using a 0.55 NA air collector with the non descanned detector having the 525/50 nm bandpass filter. Aperture and detector gain was adjusted so that the histogram of intensities was in the center of the total range without getting peaks at minimum and maximum values.

Zoom 0.75 and 512x512 pixels was chosen, which gives images of $\approx 1500 \mu\text{m}$ (read more about resolution and image size in the discussion). After images is scanned, they are rotated 270 degrees, as Leica LAS store .tif-images with axes swapped in regards to the stage axes.

Uneven illumination

The uneven illumination in the experimental setup is illustrated in figure 4.1(a). By assuming the intensity variation in all pixels are following the slope of the background, equalization was done by dividing each row in the image by the normalized intensity profile of the background.

Listing 4.1: Equalizing an image

```
equalized = img.astype(np.float)      # assure datatype have real division ability
equalized -= images_minimum           # normalize
equalized /= images_maximum - images_minimum
equalized /= intensity_profile        # equalize
equalized[equalized > 1] = 1          # clip values
```




Figure 4.1: (a) Image of glass slide only and no tissue for illustrating the uneven illumination. Dots are impurities in the sample. (b) Original image of sample. The white line is the row with least variance used for equalization. (c) Equalized version of (b). Note that (a), (b) and (c) are displaying values from 130 to 230 to highlight the intensity variation, colorbar is shown to the right.

As seen in code listing 4.1 the image is first normalized. `images_minimum` and `images_maximum` is found by selecting the median of respectively minimum and maximum intensity of all images. By taking the median of all images one avoids outliers and gets the same normalization for all images. Similar technique could be used for normalizing the images after equalization, but clipping gave acceptable results. `intensity_profile` is a curve fit for one of the background rows. The background row was found by selecting the row with least variance (given that the image does have a row with background only). In figure 4.1(b) the row with least variance is indicated with a white line. The same intensity profile is used on all images, and it's fitted to a second degree polynomial to steer clear from noise as illustrated in 4.2(a).

The effect on pixel values can be seen in figure 4.1_intensities (b) and (c), where each dot represents a pixel value with increasing image x-position on the x-axis.




Figure 4.2: (a) Intensities for the line with least variance of figure 4.1(b). The curve is fitted to a second degree polynomial to suppress noise. (b) Intensities for image in figure 4.1(b). Each dot represents a pixel. (c) Intensities for the equalized image in figure 4.1(c). Each dot represents a pixel. Note that the intensities are both spread across the whole intensity range (0-255) and the skewness is fairly straightened out.

Stitching




Figure 4.3: (a) Automatic stitching with Fiji is unreliable, as the image translation calculated by phase correlation is chosen without displacement constraints. (b) Using same overlap for all images gives negligible errors, here using the python package *microscopestitching*.

Due to little signal in areas between samples, automatic stitching with correlation methods are prone to fail. To remedy this, the same overlap was chosen when stitching the overview image. Using the same overlap in this context gives reliable stitching with negligible errors. The overlap is chosen by calculating all overlaps with phase correlation and taking the median. The stitching was put in a python package and can be used as shown in code listing 4.2.

Listing 4.2: Stitching images with the python package **microscopestitching**.

```
from microscopestitching import stitch
from glob import glob

files = glob('path/to/images/*')
images = []
for i, file in enumerate(files):
    # rectangle of 4 rows and len(files)//4 columns
    row = i % 4
    column = i // 4
    images.append((file, row, column))

stitched_image = stitch(images)
```

Segmentation

As seen in figure 4.3(b), the samples at the edge are darker than the samples in the center. To improve this intensity variation, the overview image is filtered with a local bilateral population filter. The filter counts number of neighbour pixels that are outside a specified range. The effect of the filter is less computational demanding and somewhat similar to an entropy filter. Areas with low signal variation




Figure 4.4: Otsu thresholding of figure 4.3(b). (a) Otsu thresholding applied without any filters. Picks out dark areas, but disjointed, especially for brighter sample spots in bottom left. (b) Thresholding after a local bilateral population filter. Quite noisy in the background. (c) Thresholding after local bilateral population and local mean filter. Background noise is gone and sample spots are coherent.

(the background) give low values and areas with high signal variation (the samples) give high values. To reduce noise after the bilateral population filter, a mean filter was applied. The size of structure elements was 9x9 pixels for both filters. Figure 4.4(a), (b) and (c) show how the segmentation is affected by the filters. Code for reproducing the steps are in code listing 4.3.

Listing 4.3: Filter and segment an image with local bilateral population and Otsu thresholding.

```
from skimage.morphology import square
from skimage.filters import threshold_otsu
from leicaautomator.filters import mean, pop_bilateral

selem = square(9)
filtered = pop_bilateral(image, selem)
filtered = mean(filtered, selem)

threshold = threshold_otsu(filtered)
segmented = filtered >= threshold # high values indicate signal
```

After segmentation, regions was sorted by their area size and only the largest regions are kept. Row and column was calculated by sorting regions by position, measuring the distance between them and increment row or column number when there is a peak in the distance to previous region. The code can be seen in code listing 4.4 and figure 4.5 illustrate typical area size (a), position (b) and position derivative (c).

Listing 4.4: ""

```
from skimage.measure import label, regionprops

labels = label(segmented, background=0) # background=0: exclude background
regions = regionprops(labels)          # measure region properties
regions.sort(key=lambda r: -r.area)    # sort by area size, largest first
```




Figure 4.5: (a) Sorted region areas. Area size drops dramatically around region 125 according to number of samples on slide. (b) Regions sorted by position. There is a gap between the positions when row and columns are increasing. (c) X distance to previous region when regions are sorted by x-position. 14 peaks indicate that the image contain 15 columns. Note that x-axes in (a), (b) and (c) doesn't correspond, as the graphs are not sorted by the same attribute.

```

max_regions = 126
if len(regions) > max_regions:
    regions = regions[:max_regions]    # only keep max_regions

for r in regions:
    r.y, r.x, r.y_end, r.x_end = r.bbox # for convenience

for direction in 'yx':
    # same algorithm for row and columns
    regions.sort(key=lambda r: getattr(r, direction))

    previous = regions[0]
    for region in regions:
        # calc distance to previous region
        dx = getattr(region, direction) - getattr(previous, direction)
        setattr(region, 'd' + direction, dx)
        previous = region

```

The whole process of segmentation was done interactive as part of the python package *leicaautomator*, where settings can be adjusted to improve segmentation and regions can be moved, deleted or added with mouse clicks. The interface is shown in figure 4.6.

Calculate stage position from pixel position

After regions was localized, pixel-size in meters was calculated by

$$x_{resolution} = \frac{\Delta x}{\Delta X}. \quad (4.1)$$

Here Δx is displacement in pixels and ΔX is stage displacement in meters read from the overview scanning template in the experiment `AdditionalData/{ScanningTemplate}overview.xml` at XPath `./ScanningTemplate/Properties/ScanFieldStageDistanceX`. Left most left pixel was calculated by




Figure 4.6: The process of segmentation in a graphical user interface. Regions 4,2, 11,7 and 14,1 might be adjusted by the user, all other regions are detected fairly well.

$$X_{start} = X_{center} - \frac{S_x \cdot x_{resolution}}{2}. \quad (4.2)$$

In equation 4.2 X_{center} and S_x is respectively the stage position and number of pixels in the top left image of the overview scan. X_{center} was read from the overview scanning template at `XPath ./ScanFieldArray/ScanFieldData[@WellX="1"][@WellY="1"][@FieldX="1"][@FieldY="1"]/FieldXCoordinate`. The stage x-coordinate for any pixel was then calculated by

$$X = X_{start} + x \cdot x_{resolution}. \quad (4.3)$$

To be able to scan regions of different shape and size, a bounding box for the region was used to calculate the scanning area. Moving the stage to the boundary position will center the boundary in the image, and therefor start position of first image is calculated by

$$X_{start} = X + \frac{\Delta X_{job}}{2}. \quad (4.4)$$

Here, ΔX_{job} is stage displacement between images in the job scanning template. X_{start} will have an error of

$$\epsilon = \frac{1}{2}(\Delta X_{job} - \Delta X_{img}), \quad (4.5)$$

where ΔX_{img} is the total size of the scanned image. This was considered neglectible as $\Delta X_{job} \approx \Delta X_{img}$ and number of columns scanned was calculated by

$$f_x = \lceil \frac{\Delta X}{\Delta X_{field}} \rceil. \quad (4.6)$$

Scanning each region

To avoid unnecessary long stage movements between rows or columns, regions was looped through in a zick-zack pattern, given by their row and column position. For each region the scanning template was edited, the template was loaded and the scan was started through CAM. Single templates was used due to a Leica LAS software limitation; scanning templates with irregular spaced wells can not be loaded. Code listing 4.5 illustrates the process.

Listing 4.5: Scanning

```
from leiscanningtemplate import ScanningTemplate
from leicaautomator import zick_zack_sort
from leicacam import CAM

cam = CAM() # instantiate connection to microscope

# regions sorted as [r(1,1), r(1,2), r(2,2), r(2,1), r(3,1), r(3,2), ...]
# here r(2,1) is region(col=2, row=1)
regions = zick_zack_sort(regions, ('well_x', 'well_y'))

tmpl_path = r"C:\Users\TCS-User\AppData\Roaming\Leica\Microsystems\LAS_X" + \
    r"\MatrixScreener\ScanningTemplates" + "\\\"
tmpl_name = tmpl_path + '{ScanningTemplate}leicaautomator'
for n, region in enumerate(regions):
    # alternate between tmpl_name0/1.xml, due to a
    # bug LAS cannot load the same name twice
    tmpl = ScanningTemplate(tmpl_name + str(n%2) + '.xml')

    tmpl.move_well(1, 1, region.real_x, region.real_y)
    tmpl.write()

    cam.load_template(tmpl.filename)

    # do an autofocus
    cam.autofocus_scan()
    cam.wait_for('inf', 'scanfinished')

    # run the scan job
    cam.start_scan()
    # record output filename
    region.experiment_name = cam.wait_for('relpath')['relpath']

    # continue with next region when scan is done
    cam.wait_for('inf', 'scanfinished')
```


SHG images

SHG images was taken with a 25x/0.95 NA water objective. The pulsed infrared laser was set to 890 nm, intensity 20%, gain 40%, offset 80% and electro-optic modulator (EOM) on. Forward light was measured with non descanned PMT sensor behind a 0.9 NA air collector. Band pass filter in front of the detector was 445/20 nm and gain of detector was adjusted so that signal spanned the whole intensity range. Aperture was set to 24 (maximum).

A resolution of 1024x1024 pixels with 8 bit image depth was used. Frequency of scanning mirror was set to 600 lines/second.

DAPI images

TODO

Correlating images with patient data




Figure 4.7: Top of slide map TP-1. Ids are not incrementing systematically and need to be registered to correlate samples to respective patients. Ids are inside circles and hard to read with OCR. First part of id is same as `ID_deltaker` in patient database, second number is sample number. There should be three samples for each patient.

Slide maps, seen in figure 4.7, and patient database was given by St. Olavs. As the slide maps contained circles, slide maps were filtered to remove all but text before it was read with OCR. The OCR text output was checked for errors programatically (id should be of correct format, id should increment, patients should be registered with correct slide in database column `TP_nr`, each patient should have

three samples). OCR errors was fixed manually and other errors was recorded (see section [Slide map errors](#) in the appendix).

Every pasient id from the slide map was then saved to a stata database along with its slide number, row and column. Code listing 4.6 show how the clinical data was correlated with samples.

Listing 4.6: Get patient outcome of sample on TP-1 row 3 column 5.

```
import pandas as pd

# read databases
locations = pd.read_stata('data/ids/locations.dta')
clinical_data = pd.read_stata('data/clinic_data.dta')

# position query
condition = (locations.TP_nr == 1) & \
            (locations.TP_rad == 3) & \
            (locations.TP_kolonne == 5)

# get patient id
patient_id = locations[condition]['ID_deltaker']

# check exactly 1 patient registered at given row/col
assert len(patient_id) == 1

# clinical data query
condition = clinical_data.ID_deltaker == patient_id.iloc[0]

# get outcome
outcome = clinical_data[condition]['GRAD']
```

Collection of SHG images

- alignment of z-plane
- correlation with patient data (sample map and clinic data)

Technical details

Hardware aspects

- z-plane off by several hundreds of micrometer
- piezo-holder tilted
- slides not necessarily straight, coverslip placement
- too much tilt: out of focus in one image
- tolerated tilt and software autofocus: stitching when edge not from same physical area (especially thick samples)
- signal variations and chosen optimum

- collector 0.55 vs 0.9 when overview vs SHG
- aperture not adjustable from software, resets when using occular
- hard to get same conditions every time (might move to discussion: suggest using test sample routine along with image analysis)
- rotation scanning mirror
- stitch
- finding angle with image registration / phase correlation
- edge of image, intensity variation
- zoom
- correction for overview vs SHG
- HyD shutdown too much light
- HyD behind mirror might get less light, but still good signal
- pinhole adjustment for HyD behind mirror to avoid bright spots?
- reported resolution from LAS not same as stage movement
- use image registration to calculate px-resolution
- calibration of measurement-equipment
- what measurement to trust
- outage and service
- logging, feedback and communication between researchers
- service contracts

Leica software details

The microscope software in use was Leica LAS X version TODO.

- loading template with variable positioned wells not working
- offset first well will offset all wells
- Properties/XStartPosition not used
- no “template-type” property
- must be loaded in GUI first time
 - through CAM opens GUI dialog “Import?”
- CAM only available after manually loading a template in GUI
- GUI automation
- loading modified template with same name
- loading templates automatic goes to position and changes objective
- crashes possible
- trouble if using imersion objective
- switching between AF / job in GUI will automatically switch objective without warning
- trouble if using imersion objective
- mix of 0-indexed and 1-index variables
- files 0 indexed

- cam 1 indexed
- xml 1 indexed (TODO: verify)
- GUI hangs if socket is not read
- loading template should omit .xml from filename
- saving template should not
- not noted in documentation
- “templ.xml.xml not found”
- save template does not update with latest changes in GUI
- XML does not read when missing return char “`\n`”
- not in XML specification
- z-position in template not read
- z-position from CAM sometimes gives “0” instead of real position
- adjusting x/y-coordinate on USB-control panel moves stage to zero or max position

Software development

- Separate of concerns
- modules and code reuse
- publication of software packages and python ecosystem
- leicacam: talking with microscope
- leiscanningtemplate: modify templates
- leicaexperiment: read, stitch, ome.tif experiments
- microscopestitching: reliable stitching with phase corralation (remove outliers vs median)
- leicaautomator: find regions to scan, unifies all of the above
- python cross platform and compilation
- heavy c/c++ dependency
- miniconda
- wheel packages

Utilities (not specific thesis): - fijibin: automate fiji/imagej from python - ipynbcompress: compress images in ipython notebooks

ML: Kan også skrive om spesifikke aspekter ved mikroskopsystemet som har muliggjort/begrenset/forhindret løsningene. All programvare som er utviklet bør omtales her, eventuelt med mer detaljer i et appendiks)

Chapter 5

Result

What to communicate: achievements and show-stopper/hard limitations

Segmentation

ML: Resultat så langt: Kontroll via Python, segmentering, z-correction




Figure 5.1: Comparison of thresholding

Chapter 6

Discussion

What to communicate: discuss results, limitations, possibilities for improvement

ML: Hvilke valg har blitt tatt, hva er viktig for neste bruker, hva er begrensninger, utviklingsmuligheter, pros/cons, hvor bra fungerer det....)

Chapter 7

Conclusion

What to communicate: brief summary of the result and discussion, advice for further work

ML: Automatic imaging and segmentation of TMA has been demonstrated)...and...

Chapter 8

Appendix

Leica LAS design: - user should be mainly in LAS - automating on the side as a supplement - load before CAM can be used - does not load all settings from XML

Slide map errors

```
TP2, row 3, col 6 - pasient id missing in db: 66
TP6, row 1, col 9 - pasient id missing in db: 222
TP3, row 1, col 3 - id 68, wrong TP_nr in db: 3.0 != 2.0
TP6, row 1, col 3 - id 209, wrong TP_nr in db: 6.0 != 4.0
TP6, row 1, col 6 - id 221, wrong TP_nr in db: 6.0 != 5.0
TP22, row 2, col 6 - id 130, wrong TP_nr in db: 22.0 != 3.0
TP22, row 2, col 9 - id 244, wrong TP_nr in db: 22.0 != 5.0
TP22, row 3, col 3 - id 281, wrong TP_nr in db: 22.0 != 6.0
TP22, row 3, col 6 - id 296, wrong TP_nr in db: 22.0 != 6.0
TP22, row 3, col 9 - id 309, wrong TP_nr in db: 22.0 != 6.0
TP22, row 4, col 3 - id 318, wrong TP_nr in db: 22.0 != 6.0
TP22, row 4, col 6 - id 376, wrong TP_nr in db: 22.0 != 7.0
TP22, row 4, col 9 - id 396, wrong TP_nr in db: 22.0 != 8.0
TP22, row 5, col 3 - id 413, wrong TP_nr in db: 22.0 != 8.0
TP22, row 5, col 6 - id 449, wrong TP_nr in db: 22.0 != 9.0
TP22, row 5, col 9 - id 453, wrong TP_nr in db: 22.0 != 9.0
TP22, row 6, col 3 - id 487, wrong TP_nr in db: 22.0 != 10.0
TP22, row 6, col 6 - id 493, wrong TP_nr in db: 22.0 != 10.0
TP22, row 6, col 9 - id 525, wrong TP_nr in db: 22.0 != 10.0
TP22, row 7, col 3 - id 728, wrong TP_nr in db: 22.0 != 15.0
TP3, row 9, col 6 - TP_nr not registered in db for ID_deltaker 140
TP5, row 9, col 9 - TP_nr not registered in db for ID_deltaker 251
TP9, row 10, col 9 - there should be 3 samples: ['467a-1']
TP9, row 11, col 3 - there should be 3 samples: ['467b-1', '467b-2']
TP9, row 12, col 6 - there should be 3 samples: ['471a-1', '471a-2']
TP9, row 12, col 9 - there should be 3 samples: ['471b-1']
TP10, row 8, col 6 - there should be 3 samples: ['507-1', '507-2']
TP10, row 12, col 6 - there should be 3 samples: ['525-2', '525-3']
TP11, row 11, col 6 - there should be 3 samples: ['566-1', '566-2']
```

```

TP3, row 1, col 3 - pasient id did not increment: ['68-1', '68-2', '68-3']
< ['102b-1', '102b-2', '102b-3']
TP4, row 1, col 3 - pasient id did not increment: ['162a-1', '162a-2', '162a-3']
< ['163-1', '163-2', '163-3']
TP6, row 1, col 3 - pasient id did not increment: ['209-1', '209-2', '209-3']
< ['268-1', '268-2', '268-3']
TP11, row 6, col 3 - pasient id did not increment: ['549-1', '549-2', '549-3']
< ['552-1', '552-2', '552-3']
TP22, row 2, col 6 - pasient id did not increment: ['130-1', '130-2', '130-3']
< ['3067-1', '3067-2', '3067-3']

```