

Sleep Need, Debt, Recovery and Stages: Your Body Does Not Keep a Neat Ledger

Sleep need changes across a person's lifespan, people differ in how they respond to sleep loss, and recovery does not happen minute for minute. Sleep stages then make the whole thing look more precise than it really is. A watch can tell us we had 80 minutes of REM and 60 minutes of deep sleep, although that does not mean the body has exactly 15 minutes of REM debt waiting to be paid tomorrow.

There are 4 related ideas that need separating. Sleep need is the amount of sleep a person is likely to require across a 24-hour period. Sleep achieved is the sleep we actually observed, including a main sleep and any valid naps. Recent sleep deficit describes the shortfall accumulated across recent days. Recovery describes the process of returning towards normal function after that shortfall. Stages describe the architecture inside the sleep that happened. They inform us of the picture, but none of them directly tell us the others.

The best place to start sleep need is age. Consensus recommendations are broad because there is genuine variation inside every group. The National Sleep Foundation recommends 8-10 hours for teenagers, 7-9 hours for adults, and 7-8 hours for older adults. [1] The American Academy of Sleep Medicine's conclusion is simpler: healthy adults should regularly get at least 7 hours, with longer sleep remaining appropriate for some young adults, people recovering from sleep debt and people who are ill. [3]

For a system that needs one starting value, the midpoint of the recommended range is a reasonable default:

- 13 to 17 years: 9 hours across 24 hours.
- 18 to 64 years: 8 hours across 24 hours.
- 65 years and older: 7 hours 30 minutes across 24 hours.

These defaults are starting points rather than a complete diagnosis of individual need. Controlled sleep-loss work has found large and repeatable differences in how badly people are affected by the same restriction. Some people maintain attention much better than others, and their subjective sleepiness does not always reveal the difference. [4] [5] This gives us a clear reason to personalise, but habitual sleep alone is a dangerous teacher. Someone who regularly gets 6 hours may have adapted their schedule and their expectations while still carrying measurable impairment.

A personal target should therefore move slowly and stay anchored to the age range. The useful evidence would come from nights with enough sleep opportunity, low alarm pressure, stable timing and reasonable daytime function. A 28-day observation period gives a starting view, with 60 days preferred before making a larger adjustment. The sleep target can move in 15-minute steps no more often than every 14 days. It should not be pushed below the lower age

recommendation simply because short sleep is common. A user-set target can sit alongside this estimate.

Sleep debt starts with a real effect: repeated restriction accumulates. In a controlled study, healthy adults assigned 4 or 6 hours in bed for 14 nights developed cumulative deficits in attention and cognitive performance. People became less reliable at noticing how impaired they were as the sleep restriction continued. [4] Another dose-response study found performance remained stable with 9 hours in bed, while groups receiving 3, 5 or 7 hours showed different levels of degradation and incomplete recovery. [6] The body clearly remembers recent sleep history, but it does not keep a spreadsheet of every minute borrowed.

One long night also does not erase several short ones. Following 5 nights restricted to 4 hours in bed, more recovery sleep produced more improvement, but even a 10-hour recovery opportunity did not return every neurobehavioural measure back to baseline. [7] Reviews of recovery sleep make the same point: sleepiness, mood, attention and metabolic outcomes recover at different speeds, while the amount and number of recovery opportunities both matter. [8] Weekend recovery sleep has also failed to prevent metabolic disruption when insufficient sleep resumes afterwards. [9]

This makes a straight 7-day sum too literal. It treats an hour lost 6 nights ago the same as an hour lost last night, and it lets one long sleep cancel debt even though recovery may still be incomplete. A better practical measure is a recent sleep deficit burden.

The daily shortfall should be calculated from total valid sleep across the relevant 24-hour sleep day:

$$\text{shortfall} = \max(0, \text{core sleep need} - \text{main sleep} - \text{valid naps} - 30 \text{ minutes})$$

For adults, a 30-minute deadband is sensible before adding any shortfall to the persistent debt estimate. Wearables have measurement error and sleep does not need to hit the target to the exact minute. For example, an 8-hour target and 7 hours 45 minutes of observed sleep would produce no meaningful deficit burden. Below that point, only the shortfall beyond the 30-minute allowance is added to debt. This allowance belongs only to the slow-moving debt model. The nightly duration component in Sleep Scores uses the complete shortfall against core need so that a 15 or 30-minute miss can still distinguish an exceptional night from a merely very good one.

Recent burden can then be updated as:

$$\text{burden}_{\text{today}} = \max(0, 0.90 \text{ burden}_{\text{yesterday}} + \text{shortfall} - 0.50 \text{ extra sleep})$$

Extra sleep means valid sleep that exceeds the core need, capped at 120 minutes in one day. The 0.90 carry gives the prior deficit a roughly one-week half-life, while the 0.50 recovery credit prevents one long night from wiping the value away. These constants are design values informed by slow and incomplete recovery, they are not constants discovered in one paper. They should be validated against daytime function and revised if long-term data supports something better. More sophisticated performance models also give recent sleep-wake history greater influence and recover slowly after chronic restriction, which supports this general shape. [10]

This burden should never become negative. Sleep extension before restriction can improve resilience and recovery, so getting extra sleep can still be useful, but a large weekend lie-in shouldn't create a bank that makes future short sleep harmless. [17] Recovery opportunity is a better place to represent this. The core need remains stable; the suggested opportunity can rise when the recent deficit is high.

For an initial implementation, recovery sleep can be spread across several nights:

- Suggested extra opportunity: one third of the current burden, rounded to 15-minute steps and capped at 90 minutes per night.
- High recent physical load: add 0 to 30 minutes of opportunity, without adding the uncompleted amount to sleep debt.
- One severely short night, meaning at least 2 hours below core need, should keep recovery status active for at least 2 adequate nights.
- Three or more severely short nights, or a full night missed, should keep recovery status active for at least 4 adequate nights.
- Recovery can end when the burden falls below 30 minutes and the minimum run of adequate nights has been completed.

The minimum recovery runs are deliberately cautious. Four recovery nights can still leave some outcomes below baseline after severe restriction, so “recovered” cannot be treated as a guarantee that every system is back to normal. [11] The status means recent sleep history no longer gives strong evidence of an active deficit.

Missing sleep must not quietly improve the result. If one night is missing, the previous burden should carry forward without its normal decay and confidence should fall. If 2 or more of the last 7 nights are missing, a precise burden should be withheld or shown as “at least” the burden supported by observed nights. A missing night could contain excellent sleep, terrible sleep or no sleep at all so the safest outcome is to show confidence is low and the data we actually have.

A detected nap can reduce the day’s shortfall and contribute to recovery. This does not make naps identical to a well-timed main sleep. Long or late naps can alter sleep pressure and the following night, and waking from a nap can temporarily create sleep inertia. The cleanest structure is to count valid nap minutes towards 24-hour sleep achieved while keeping nap duration and timing visible as their own context. Short daytime naps have shown cognitive benefits, and sleep intervention reviews in athletes find both naps and sleep extension useful, although the evidence is not strong enough to create an exact universal nap conversion. [12]

Training load deserves similar restraint. Athletes often sleep poorly around travel, competition and early training, and extending sleep can improve some performance outcomes. The intervention literature is still limited and uses quite different protocols. [12] [13] A fixed rule saying every hard workout creates exactly 30 minutes of additional biological need would look more certain than the research is. Recent strain can increase recommended sleep opportunities by up to 30 minutes in a first model, while core need and debt remain anchored to the age and personal target. This stops an uncertain strain modifier from manufacturing debt. This could be modified in future if more research is found.

Sleep stages then give us a different problem. Normal sleep moves through N1 (light sleep), N2 (stable light sleep), N3 (deep sleep) and REM in repeated cycles. N3 sleep is concentrated earlier in the night, with more REM appearing later. Shortening the night can therefore remove a disproportionate amount of later REM even when the percentage from the remaining sleep looks acceptable. Recovery sleep may also increase total sleep, REM, N3 sleep and slow-wave energy (the amount of slow brain-wave activity during deep sleep) as the body responds to prior restriction. [7] A high or low stage amount can be part of a regular response to recent sleep history.

Population values help us understand the broad architecture. A conventional healthy adult night is often described as roughly 5% N1, 50% N2, 20% N3 and 25% REM. These are loose reference values from polysomnography (gold standard lab test), not nightly targets. Large meta-analyses show that total sleep time, efficiency, wake after sleep onset and stage composition

change with age. Slow-wave sleep generally falls from childhood into adulthood, while lighter sleep and wakefulness become more prominent with ageing. [14] More recent PSG work also finds that total sleep and efficiency decline with age while wake after sleep onset rises, with large variation between healthy people. [15]

There is no accepted minimum number of REM or deep minutes that a consumer device can use to declare a good night. REM and N3 have important relationships with memory, learning and the body's recovery processes, although whole-night function depends on the sequence and interaction of stages as well as their duration. The research does not support treating light sleep as wasted sleep or asking users to chase an exact deep-sleep percentage.

Wearables add enough error that population stage ranges become even less suitable as score boundaries. Polysomnography stages sleep from brain activity, eye movements and muscle tone. Consumer devices infer stages from movement and cardiorespiratory signals through algorithms. Reviews repeatedly warn that firmware, device model and algorithm changes matter. [16] In a direct comparison of 6 wearables, sleep-wake agreement was around 86% to 89%, while exact multistage agreement was only 50% to 65% depending on the device. [18] A 2025 meta-analysis also found meaningful differences between wrist devices and polysomnography for total sleep time, efficiency, latency and wake after sleep onset. [19] A systematic review of Fitbit, Garmin and WHOOP found different devices were better at different stages, with no provider consistently perfect. [20]

Stage data should therefore pass quality checks before it influences anything:

- The stage timeline should cover at least 90% of the recorded sleep period.
- Stage minutes should agree with total sleep within 10% or 30 minutes, whichever allowance is larger.
- Intervals must not overlap, run backwards or sit outside the main sleep session.
- The provider, device and algorithm generation should remain consistent with the personal baseline.
- At least 14 valid nights are needed for a provisional stage baseline, with 28 preferred.
- Missing REM or deep sleep remains missing rather than becoming zero.

Once valid, stages should be compared mainly with the person's recent provider-specific pattern. A 7-night median can sit beside a 28-day reference, and a persistent change across at least 3 nights should matter more than one strange breakdown. Age-adjusted polysomnography values can act as broad plausibility checks, but a consumer REM percentage should not be scored as though it came from a laboratory.

This keeps stages as important but conservatively measured evidence. In a sleep score, valid stages should enter directly at 5%. Their limited weight reflects wearable uncertainty rather than limited biological importance. Missing stages should cause no penalty, should never be represented as a score of 100 and should not transfer their full weight to one other weak component. When stage data is missing, the complete behavioural core carries the score subject to the high-score evidence cap defined in Sleep Scores. Short duration, major fragmentation or severe timing disruption still cannot be rescued by an attractive REM chart. "REM debt" and "deep sleep debt" should not be calculated.

Taken together, the clearest implementation for our app is:

- Core sleep need starts from the age midpoint: 8 hours for adults aged 18 to 64 and 7 hours 30 minutes from age 65.
- The target personalises only after at least 28 valid days, moves by no more than 15 minutes every 14 days, and stays inside the recommended age range unless the user explicitly changes it.
- Sleep achieved uses the main sleep plus valid naps in the same 24-hour sleep day.
- A 30-minute allowance protects only the debt estimate from tiny differences and wearable error. The nightly duration score uses the complete shortfall.
- Recent deficit uses the 0.90 carry and 0.50 recovery-credit model above. It remains an estimated burden, not a claim about exact biological sleep owed.
- Extra recovery opportunity is one third of burden, rounded to 15 minutes and capped at 90 minutes per night. Recent high strain may add up to 30 minutes to opportunity without creating debt.
- Missing nights lower confidence and never count as zero shortfall.
- Recovery status requires repeated adequate nights after severe restriction rather than ending after one long sleep.
- Stages pass the 90% coverage and consistency checks, use a 14 to 28-night provider-specific baseline, and account directly for 5% of the score when valid. Missing stages remain missing and invoke the high-score evidence rule in Sleep Scores.
- No REM debt, deep-sleep debt or population stage quota is shown.

The age ranges come from consensus guidance, the slow accumulation and recovery come from controlled restriction studies, and the conservative handling of stages comes from wearable validation work. The exact decay and recovery coefficients still need testing against our own outcomes. That is a much better uncertainty to have than a precise-looking 7-day sum which assumes the body settles sleep debt like a credit card bill.

Bibliography

- [1] Hirshkowitz, M., Whiton, K., Albert, S. M., et al. (2015). National Sleep Foundation's updated sleep duration recommendations: Final report. *Sleep Health*, 1(4), 233-243. <https://doi.org/10.1016/j.sleh.2015.10.004>
- [2] Paruthi, S., Brooks, L. J., D'Ambrosio, C., et al. (2016). Recommended amount of sleep for pediatric populations: A consensus statement of the American Academy of Sleep Medicine. *Journal of Clinical Sleep Medicine*, 12(6), 785-786. <https://doi.org/10.5664/jcsm.5866>
- [3] Watson, N. F., Badr, M. S., Belenky, G., et al. (2015). Recommended amount of sleep for a healthy adult: A joint consensus statement of the American Academy of Sleep Medicine and Sleep Research Society. *Sleep*, 38(6), 843-844. <https://doi.org/10.5665/sleep.4716>
- [4] Van Dongen, H. P. A., Maislin, G., Mullington, J. M., & Dinges, D. F. (2003). The cumulative cost of additional wakefulness: Dose-response effects on neurobehavioral functions and sleep physiology from chronic sleep restriction and total sleep deprivation. *Sleep*, 26(2), 117-126. <https://doi.org/10.1093/sleep/26.2.117>

- [5] Van Dongen, H. P. A., Baynard, M. D., Maislin, G., & Dinges, D. F. (2004). Systematic interindividual differences in neurobehavioral impairment from sleep loss: Evidence of trait-like differential vulnerability. *Sleep*, 27(3), 423-433. <https://pubmed.ncbi.nlm.nih.gov/15164894/>
- [6] Belenky, G., Wesensten, N. J., Thorne, D. R., et al. (2003). Patterns of performance degradation and restoration during sleep restriction and subsequent recovery: A sleep dose-response study. *Journal of Sleep Research*, 12(1), 1-12. <https://doi.org/10.1046/j.1365-2869.2003.00337.x>
- [7] Banks, S., Van Dongen, H. P. A., Maislin, G., & Dinges, D. F. (2010). Neurobehavioral dynamics following chronic sleep restriction: Dose-response effects of one night for recovery. *Sleep*, 33(8), 1013-1026. <https://doi.org/10.1093/sleep/33.8.1013>
- [8] Guzzetti, J. R., & Banks, S. (2022). Dynamics of recovery sleep from chronic sleep restriction. *Sleep Advances*, 4(1), zpac044. <https://doi.org/10.1093/sleepadvances/zpac044>
- [9] Depner, C. M., Melanson, E. L., Eckel, R. H., et al. (2019). Ad libitum weekend recovery sleep fails to prevent metabolic dysregulation during a repeating pattern of insufficient sleep and weekend recovery sleep. *Current Biology*, 29(6), 957-967.e4. <https://doi.org/10.1016/j.cub.2019.01.069>
- [10] Rajdev, P., Thorsley, D., Rajaraman, S., et al. (2013). A unified mathematical model to quantify performance impairment for both chronic sleep restriction and total sleep deprivation. *Journal of Theoretical Biology*, 331, 66-77. <https://doi.org/10.1016/j.jtbi.2013.04.013>
- [11] Yamazaki, E. M., Antler, C. A., Lasek, C. R., & Goel, N. (2021). Residual, differential neurobehavioral deficits linger after multiple recovery nights following chronic sleep restriction or acute total sleep deprivation. *Sleep*, 44(4), zsaa224. <https://doi.org/10.1093/sleep/zsaa224>
- [12] Cunha, L. A., Costa, J. A., Marques, E. A., Brito, J., Lastella, M., & Figueiredo, P. (2023). The impact of sleep interventions on athletic performance: A systematic review. *Sports Medicine - Open*, 9, 58. <https://doi.org/10.1186/s40798-023-00599-z>
- [13] Silva, A. C., Silva, A., Edwards, B. J., et al. (2021). Sleep extension in athletes: What we know so far - A systematic review. *Sleep Medicine*, 77, 128-135. <https://doi.org/10.1016/j.sleep.2020.11.028>
- [14] Ohayon, M. M., Carskadon, M. A., Guilleminault, C., & Vitiello, M. V. (2004). Meta-analysis of quantitative sleep parameters from childhood to old age in healthy individuals: Developing normative sleep values across the human lifespan. *Sleep*, 27(7), 1255-1273. <https://doi.org/10.1093/sleep/27.7.1255>
- [15] Boullos, M. I., Jairam, T., Kendzerska, T., Im, J., Mekhael, A., & Murray, B. J. (2019). Normal polysomnography parameters in healthy adults: A systematic review and meta-analysis. *The Lancet Respiratory Medicine*, 7(6), 533-543. [https://doi.org/10.1016/S2213-2600\(19\)30057-8](https://doi.org/10.1016/S2213-2600(19)30057-8)
- [16] de Zambotti, M., Cellini, N., Goldstone, A., Colrain, I. M., & Baker, F. C. (2019). Wearable sleep technology in clinical and research settings. *Medicine & Science in Sports & Exercise*, 51(7), 1538-1557. <https://doi.org/10.1249/MSS.0000000000001947>
- [17] Rupp, T. L., Wesensten, N. J., Bliese, P. D., & Balkin, T. J. (2009). Banking sleep: Realization of benefits during subsequent sleep restriction and recovery. *Sleep*, 32(3), 311-321. <https://doi.org/10.1093/sleep/32.3.311>
- [18] Miller, D. J., Sargent, C., & Roach, G. D. (2022). A validation of six wearable devices for estimating sleep, heart rate and heart rate variability in healthy adults. *Sensors*, 22(16), 6317.

<https://doi.org/10.3390/s22166317>

[19] Lee, Y. J., Lee, J. Y., Cho, J. H., Kang, Y. J., & Choi, J. H. (2025). Performance of consumer wrist-worn sleep tracking devices compared to polysomnography: A meta-analysis. *Journal of Clinical Sleep Medicine*, 21(3), 573-582. <https://doi.org/10.5664/jcsm.11460>

[20] Schyvens, A.-M., Van Oost, N. C., Aerts, J.-M., et al. (2024). Accuracy of Fitbit Charge 4, Garmin Vivosmart 4, and WHOOP versus polysomnography: Systematic review. *JMIR mHealth and uHealth*, 12, e52192. <https://doi.org/10.2196/52192>