DEMO - Demo Medical Devices Pvt Ltd (pilot, not for use)

SOP-QC-001 Incoming inspection of silicone

Revision 00 - Draft for review

1. Purpose

To inspect each lot of medical-grade silicone compound before it is released from quarantine for use in Foley catheter production.

2. Scope

All receipts of item RM-SILICONE into Quarantine - DMD.

3. Responsibilities

Stores: receive the lot into quarantine. QC Inspector: sample and test. QA Manager: release or reject the lot.

4. Procedure

4.1 Check the supplier certificate of analysis against the specification.

4.2 Take samples as per the sampling plan.

4.3 Test appearance, hardness (Shore A) and cure characteristics.

4.4 Record the results in a Quality Inspection against the lot.

4.5 QA releases the lot to Stores - DMD only when every result conforms.

5. Records

Quality Inspection, lot record and audit trail; kept for 10 years.