



MUNSON HOME MEDICAL TRAVERSE CITY  
3816 W. FRONT STREET  
TRAVERSE CITY  
Michigan, 49684

Phone: 231-935-2100  
Email: MHH-MHMESleep@mhc.net

06/08/2023 - 07/07/2023

## Compliance Report

### Initial compliance period

06/08/2023 - 07/07/2023

Compliance met Yes  
Compliance percentage 70%  
Payor Standard

### Usage

06/08/2023 - 07/07/2023

Usage days 22/30 days (73%)  
    >= 4 hours 21 days (70%)  
    < 4 hours 1 day (3%)  
Usage hours 149 hours 48 minutes  
    Average usage (total days) 5 hours 0 minutes  
    Average usage (days used) 6 hours 49 minutes  
    Median usage (days used) 6 hours 55 minutes  
Total used hours (value since last reset - 07/07/2023) 276 hours

### AirSense 11 AutoSet

Serial number 23231350306  
Mode CPAP  
Set pressure 15 cmH2O  
EPR Fulltime  
EPR level 3

### Therapy

Leaks - L/min Median: 34.5 95th percentile: 62.6 Maximum: 72.1  
Events per hour AI: 4.0 HI: 2.1 AHI: 6.1  
Apnea Index Central: 1.8 Obstructive: 0.4 Unknown: 1.7  
RERA Index 1.3  
Cheyne-Stokes respiration (average duration per night) 7 minutes (2%)

### Usage - hours

