



07/19/2023 - 08/17/2023



Age: 45 years
Gender: Male

Compliance Report

Initial compliance period 07/19/2023 - 08/17/2023

Compliance met	Yes
Compliance percentage	100%
Payor	Medicare

Usage 07/19/2023 - 08/17/2023

Usage days	30/30 days (100%)
>= 4 hours	30 days (100%)
< 4 hours	0 days (0%)
Usage hours	323 hours 21 minutes
Average usage (total days)	10 hours 47 minutes
Average usage (days used)	10 hours 47 minutes
Median usage (days used)	11 hours 14 minutes
Total used hours (value since last reset - 08/17/2023)	931 hours

AirSense 11 AutoSet

Serial number	22231038074
Mode	CPAP
Set pressure	15 cmH2O
EPR	Fulltime
EPR level	3

Therapy



Leaks - L/min	Median: 120.0	95th percentile: 120.0	Maximum: 120.0
Events per hour	AI: 8.7	HI: 0.4	AHI: 9.1
Apnea Index	Central: 0.0	Obstructive: 0.0	Unknown: 8.6
RERA Index			1.0
Cheyne-Stokes respiration (average duration per night)			0 minutes (0%)

Usage - hours





AirView™



Age: 45 years
Gender: Male

Therapy Report

AirSense 11 AutoSet

SN: 22231038074

Usage (hours)

Usage days30/30 (100%)

>= 4 hour days30 (100%)

< 4 hour days0 (0%)

Days not used0 (0%)

Days no data0 (0%)

Used/day (avg.)10.8 hrs.



Leak (L/min)

Set threshold24.0 L/min

Maximum (avg)120.0

95th % (avg)120.0

Median (avg)120.0



Pressure (cmH2O)

ModeCPAP

Set EPRFulltime, 3.0

Set pressure15.0



AHI (events/hour)

AHI9.1

HI0.4

AI8.7

CAI0.0

OAI0.0

UAI8.6

RERA1.0

CSR% (avg)0.0





AirView™



Therapy Report

AirSense 11 AutoSet SN: 22231038074

Tidal Volume(ml)

Maximum (avg) 1621
95th % (avg) 745
Median (avg) 275



Respiratory Rate(breaths/min)

Maximum (avg) 38
95th % (avg) 30
Median (avg) 19



Minute Ventilation(L/min)

Maximum (avg) 23.1
95th % (avg) 12.8
Median (avg) 5.2





AirView™

Therapy Report

AirSense 11 AutoSet

SN: 22231038074

Care Check-In

No available data because the user declined to answer the therapy questions, this feature was turned off on the user's machine or the user did not consent to share their data with myAir.

About Care Check-In

Short patient surveys are conducted on the AirSense 11 and myAir app to better understand how the patient is doing with therapy. If the patient consents, we can share Care Check-In information through AirView for clinical users.

Questions and answers

At the beginning of therapy, the patient is asked a baseline question. A few additional questions are asked, about once per week, during the early stages of therapy. For example:

How sleepy did you feel this week? Extremely, Very, Moderately, Slightly, Not at all.

How is your therapy? Great, OK, Challenging.

What bothers you most? Getting used to therapy, Fitting the mask, Using the machine.

Please rate your usual sleep quality. Very poor, Poor, OK, Good, Excellent

How was your sleep quality in the last week? Very poor, Poor, OK, Good, Excellent