



CHARLESTON
2090 EXECUTIVE HALL RD STE 100
CHARLESTON
South Carolina, 29407

Phone: 843-554-4000
Fax: 843-769-6849
Email: tj@atlanticrespiratory.com



Compliance Report

Compliance

Payor Standard

Usage 04/01/2023 - 10/29/2023

Usage days	207/212 days (98%)
>= 4 hours	200 days (94%)
< 4 hours	7 days (3%)
Usage hours	1,458 hours 37 minutes
Average usage (total days)	6 hours 53 minutes
Average usage (days used)	7 hours 3 minutes
Median usage (days used)	7 hours 9 minutes
Total used hours (value since last reset - 10/29/2023)	1,449 hours

AirSense 10 AutoSet

Serial number	23223420062
Mode	AutoSet
Min Pressure	5 cmH2O
Max Pressure	20 cmH2O
EPR	Fulltime
EPR level	3
Response	Standard

Therapy

Pressure - cmH2O	Median: 6.4	95th percentile: 9.6	Maximum: 11.1
Leaks - L/min	Median: 7.5	95th percentile: 33.2	Maximum: 54.8
Events per hour	AI: 1.3	HI: 0.8	AHI: 2.1
Apnea Index	Central: 0.5	Obstructive: 0.6	Unknown: 0.2
RERA Index			0.1
Cheyne-Stokes respiration (average duration per night)			1 minutes (0%)


Usage - hours





CHARLESTON
2090 EXECUTIVE HALL RD STE 100
CHARLESTON
South Carolina, 29407

Phone: 843-554-4000
Fax: 843-769-6849
Email: tj@atlanticrespiratory.com



Therapy Report

AirSense 10 AutoSet

SN: 23223420062

Usage (hours)

Usage days	207/212
>= 4 hour days	200 (94%)
< 4 hour days	7 (3%)
Days not used	1 (0%)
Days no data	4 (2%)
Used/day (avg.)	7.0 hrs.




Leak (L/min)

Set threshold	24.0 L/min
Maximum (avg)	54.8
95th % (avg)	33.2
Median (avg)	7.5



Pressure (cmH2O)

Mode	AutoSet
Set EPR	Fulltime, 3.0
Set Max Pressure	20.0
Set Min Pressure	5.0
Maximum (avg)	11.1
95th % (avg)	9.6
Median (avg)	6.4



AHI (events/hour)

AHI	2.1
HI	0.8
AI	1.3
CAI	0.5
OAI	0.6
UAI	0.2
RERA	0.1
CSR% (avg)	0.3

