

Felt taken off the ceiling: room error below 3 kHz 2.028 → 2.107 dB (+0.079, worse)

2026-09-24/shield-removed → 2026-09-24/ceiling-felt-removed · below 3 kHz, 479 of 946 cells moved >0.05 dB toward flat and 403 away from it · 95 tones · 11 calibrated capsules

