

Absorbing wedges removed from the floor (flat foam base left): room error below 3 kHz 1.400 → 2.054 dB (+0.654, worse)

2026-09-24/bare-day2 → 2026-09-24/floor-no-wedges · below 3 kHz, 287 of 946 cells moved >0.05 dB toward flat and 612 away from it · 95 tones · 11 calibrated capsules

