

Band	Limit	Tones	Bare floor	
Hz	± dB	per band	worst mic, dB	cells in limit
250	2.5	3	5.71	52 %
315	2.5	8	2.60	52 %
400	2.5	8	2.05	70 %
500	2.5	8	2.25	68 %
630	2.5	8	1.64	82 %
800	2.0	8	1.42	72 %
1000	2.0	9	0.89	82 %
1250	2.0	8	0.72	90 %
1600	2.0	8	1.39	93 %
2000	2.0	8	0.98	94 %
2500	2.0	8	0.81	86 %
5000*	2.0	4	1.77	95 %
6300*	3.0	5	1.36	100 %

BARE FLOOR

Against the hemi-anechoic ISO values

Limits: ±2.5 dB up to 630 Hz, ±2.0 dB from 800 to 5000 Hz, ±3.0 dB at 6300 Hz. Colour: worst microphone as a share of the limit (yellow at the limit, red at twice).

2 of 13 bands over the limit · 80 % of cells in limit

Room error below 3 kHz: 2.05 dB

How the floor should look in a hemi-anechoic room

A hemi-anechoic room has a hard, reflecting floor, for example a concrete slab, with absorbing walls and ceiling. The floor is part of the intended sound field and is not treated. Ours is different: the foam base is still on the floor, so the floor still absorbs and this test is not a true hemi-anechoic one.