

One-layer folding screen placed in the chamber: room error below 3 kHz 1.445 → 1.626 dB (+0.180, worse)

2026-09-25/corner-stuffed → 2026-09-25/screen-1-layer · below 3 kHz, 374 of 946 cells moved >0.05 dB toward flat and 513 away from it · 95 tones · 11 calibrated capsules

