

All felt removed from the chamber: room error below 3 kHz 2.753 → 2.159 dB (-0.594, better)

2026-09-24/felt-foil-in → 2026-09-24/felt-all-removed · below 3 kHz, 584 of 946 cells moved >0.05 dB toward flat and 313 away from it · 95 tones · 11 calibrated capsules

