
Chamber top-5 vs top performer (2026-09-23-25)

Room-error total <3kHz, dB, lower=flatter. Repeat 0.01dB, handling 0.04dB, overnight drift 0.1-0.2dB. Grid 95 tones 257Hz-6.35kHz, 3-5kHz excluded (sphere mode). Top = floor-carpet 1.277.


#	setup	total	bands 250-400/400-630/630-1k/1-1.6k/1.6-3k	vs top
1	floor-carpet	1.277	1.74/1.51/1.30/0.81/0.95	0
2	cleanup-1/backed-5cm	1.301/1.303	1.56/1.62/1.43/0.81/1.03	+0.02
3	closer	1.311	1.99/1.28/1.29/0.86/0.97	+0.03
4	cleanup-2	1.314	1.61/1.72/1.40/0.79/0.98	+0.04
5	chaotic-carpet	1.317	1.73/1.45/1.48/0.90/0.97	+0.04

AVOID 250-400Hz (modes+floor) and 3-5kHz (sphere m=1, 11dB@4362Hz). TRUST 1-3kHz. 400Hz-1kHz use with per-position correction (du Plessis <0.55dB precedent; quarter-wave: cm porous = 3kHz treatment; gate 1.98ms -> valid >500Hz).


floor-carpet.png




carpet-floor-vs-ceiling.png




backed-5cm-day3.png




backed-5cm-vs-best-day3.png



closer-day3.png




closer-vs-best-day3.png



cleanup-2-day3.png

Clean-up 2: more junk removed: room error below 3 kHz 1.301 → 1.314 dB (+0.014, worse)


2026-09-25/cleanup-1 → 2026-09-25/cleanup-2 · below 3 kHz, 398 of 946 cells moved >0.05 dB toward flat and 392 away from it · 95 tones · 11 calibrated capsules




chaotic-carpet-vs-wedges-day2.png

Chaotic carpet on flat foam vs wedges only: room error below 3 kHz 1.400 → 1.317 dB (-0.083, better)


2026-09-24/bare-day2 → 2026-09-24/chaotic-carpet · below 3 kHz, 462 of 946 cells moved >0.05 dB toward flat and 413 away from it · 95 tones · 11 calibrated capsules




chaotic-1-2-3-day2.png



bare-day2.png



floor-no-wedges-day2.png



floor-states-compare.png

Floor states compared: an evenly laid carpet is the worst of all

