Cleaning Kit for BactoSense User Manual

English

Document ID: 40108-02

Compiled on 12 November 2026

https://manual.bnovate.net/other-manuals/cleaning-kit

Cleaning Kit for BactoSense

General information

The Cleaning Kit deep-cleans the fluidic system and manual sampling device of BactoSense. Its reagents clear biological and non-biological contamination arising from the instrument's intended use. Effectiveness against contamination from other applications cannot be guaranteed. Attempts to clean with other chemicals or materials are the user's sole responsibility.

Store the kit at **2–25 °C**, in the dark. Under these conditions, the shelf-life is **one year**. The kit provides **five biological and five non-biological decontaminations**.

Safety information

WEAR PROTECTIVE GLOVES

In case of skin contact, rinse immediately and abundantly with fresh water. In case of pain or irritation, consult a doctor.

WEAR PROTECTIVE SAFETY GLASSES

In case of eye contact, flush eyes immediately with flowing water for 10–15 minutes, holding eyelids apart, and consult a doctor.

TOXIC, IRRITATING AND CORROSIVE REAGENTS

Some reagents are toxic or irritating. Wear protective equipment; do not ingest or inhale. Some reagents are corrosive to metals.

FOLLOW INSTRUCTIONS

Carefully follow this manual and the BactoSense user manual.

Contents

Item	Quantity	SDS document number	Use
Alcohol wipes	10	—	Single use
Cleaning solution	5	50102	Single use
Decontamination solution	5	50103	Single use
Rinsing solution	10	50104	Single use
Tube holder	2	—	Reusable

Discard single-use items according to local waste disposal regulations. Safety data sheets are available from bNovate or your local distribution partner upon request.

When to use the Cleaning Kit

Indications	Protocol
The manual sampling device may have been contaminated through improper storage; the instrument stopped in an error state with a loaded sample; or persistent biological contamination cannot be cleared with the Wash Station.	Biological contamination
Visible deposits on the sampling needle; frequent **High baseline in SSC** or **Air bubble detection** warnings; or electronic noise in the dotplots much higher than usual, exceeding 4 on the FL1 axis.	Non-biological contamination

Biological contamination

This protocol clears biological contamination from the instrument and manual sampling device.

DURATION AND MATERIALS

Approximately **20 minutes**. Requires one alcohol wipe, one Decontamination solution, one Rinsing solution and one Tube holder.


PROTECT YOURSELF WHEN HANDLING REAGENTS

Carefully follow the instructions and wear protective gloves and safety glasses.

1. Make sure the manual sampling device is tightly installed on BactoSense.
2. Unpack an alcohol wipe. Wipe the needle from top to bottom to remove contaminants. Let it dry for **30 seconds**.
3. Take a Decontamination solution tube and Tube holder. Remove their caps and insert the opened tube into the holder. Attach the holder containing the tube to the manual sampling device.

   USE THE WHOLE REAGENT TUBE

   Insert the entire tube, rather than pouring its reagents into the holder.



4. Open **Home** → **Manual mode** → **Start**.
5. Select **Prime** in **Protocol** and configure **two replications** in **Replications**.
6. Press **Next** and **Start**.
7. After both Prime processes finish, remove the holder and discard the Decontamination solution tube.
8. Uncap a Rinsing solution tube, insert the entire tube into the holder and attach the holder to the manual sampling device.
9. Press **New**, or open **Home** → **Manual mode** → **Start** again.
10. Select **Prime**, with **two replications**.
11. Press **Next** and **Start**.
12. After both Prime processes finish, remove the holder and discard the Rinsing solution tube.
13. Close the holder with its cap and return it to the Cleaning Kit box.
14. The decontamination sequence is complete. Follow After cleaning.

Non-biological contamination

This protocol clears non-biological contamination from the instrument and manual sampling device.

> **DURATION AND MATERIALS**
>
> Approximately **3.5 hours**. Requires one alcohol wipe, one Cleaning solution, one Rinsing solution and one Tube holder.

> **PROTECT YOURSELF WHEN HANDLING REAGENTS**
>
> Carefully follow the instructions and wear protective gloves and safety glasses.

1. Make sure the manual sampling device is tightly installed on BactoSense.
2. Unpack an alcohol wipe. Wipe the needle from top to bottom to remove contaminants. Let it dry for **30 seconds**.
3. Take a Cleaning solution tube and Tube holder. Remove their caps, insert the entire opened tube into the holder, as illustrated above, and attach the holder to the manual sampling device.
4. Open **Home** → **Manual mode** → **Start**.
5. Select **Prime**, with **two replications**.
6. Press **Next** and **Start**.
7. After the protocols finish, **wait 3 hours**.
8. Remove the holder and discard the Cleaning solution tube.
9. Uncap a Rinsing solution tube, insert the entire tube into the holder and attach it to the manual sampling device.
10. Press **New**, or open **Home** → **Manual mode** → **Start** again.
11. Select **Prime**, with **three replications**.

12. Press **Next** and **Start**.
13. After all three Prime processes finish, remove the holder and discard the Rinsing solution tube.
14. Close the holder with its cap and return it to the Cleaning Kit box.
15. The decontamination sequence is complete.

After cleaning

PRIME BEFORE THE NEXT SAMPLE

To measure a sample now, attach it to the sampling device and first perform **two Prime processes**. This ensures the highest accuracy of your results.

If you do not plan to measure samples, attach a clean, empty tube or the Wash Station. Perform **two Prime processes** before measuring the next sample.