

Supply a laboratory ultrapure-water skid with reverse-osmosis membranes, UV oxidation, resistivity monitor, TOC analyzer and automatic sanitization. A small feed pump is included inside the skid as a replaceable accessory. Provide calibration certificates, validation protocol and one-year consumables.