

Purchase a benchtop gas chromatograph mass spectrometer with autosampler, capillary column oven, electron-ionization source, spectral library and nitrogen generator. Provide IQ/OQ documents, traceable calibration, training and two-year service. The laboratory schedule references IS 88888:2024 for instrument performance.