Tuesday, 31 October 2023 10:37 AM

Elemental Analysis

+ Atomic Absorption Spectroscopy (Bulk technique)

+ ICP

Analyzer: AES, OES, MS → Ar used fuel source

MS > AES, OES } MS has 3 order of magnitude resolution

Single wavelength → monochromator

Atomic Absorption Spectroscopy (AAS): vs. ICP → plasma ionizes & excites every e⁻

+ Solid samples are digested → & then atomized in chamber

↓

Lamp generates only specific wavelengths (instead of ICP)