

PRODUCT QUALITY COMPLAINT (PQC) FORM

PRODUCT: HepaGuard 5000 IU Pre-Filled Syringes

BATCH/LOT NUMBER: PF-4401

PROBLEM: 4 cracked glass syringe barrels in box. Leakage observed prior to use.

PHOTO MENTIONED: Yes (photos attached to incident log)

REPORTER: Elena Rostova, PharmD (Lead Pharmacist)