

Nozzle Clogging and Delivery Uniformity of Metered Inhalers

ABSTRACT: Investigation of propellant crystallization causing valve obstruction in PulmoMist.

1. Problem Definition

Quality complaints regarding clogged nozzles.

Batch PM-5049 evaluated.

2. Remediation

Clean nozzle orifice with alcohol swab.

Redesign actuator cap orifice.