

PRODUCT QUALITY COMPLAINT - ACTUATOR DEFECT

PRODUCT: PulmoMist Inhaler 100mcg

BATCH/LOT NUMBER: PM-5049

PROBLEM: Actuator nozzle completely clogged, zero dose discharge on depression.

PHOTO MENTIONED: Yes (photos included in QA dossier)

REPORTER: Pharmacist Mark Vance