

ABSTRACT

Title of dissertation:

CALIBRATION OF THE
LARGE UNDERGROUND XENON DETECTOR:
AND EVENT LEVEL FLUCTUATIONS
IN LIQUID XENON MEASURED WITH TRITIUM

Attila Dobi, Doctor of Philosophy, 2014

Dissertation directed by:

Professor Carter Hall
Department of Physics

... Abstract here...

CALIBRATION OF THE LARGE UNDERGROUND XENON
DETECTOR AND EVENT LEVEL FLUCTUATIONS IN LIQUID
XENON MEASURED WITH TRITIUM

by

Attila Dobi

Dissertation submitted to the Faculty of the Graduate School of the
University of Maryland, College Park in partial fulfillment
of the requirements for the degree of
Doctor of Philosophy
2014

Advisory Committee:

Professor Carter Hall, Chair/Advisor
Professor Elizabeth Beise
Professor Massimo Ricotti
Professor Thomas Cohen
Professor Zackaria Chacko

Acknowledgments

Table of Contents

List of Figures	vi
1 Introduction	1
1.0.1 Galactic Rotation Curves	2
1.0.2 Big Bang Nucleosynthesis	4
1.0.3 Cosmic Microwave Background	5
1.0.4 Baryonic Acoustic Oscillations	6
1.0.5 Gravitational Lensing	9
1.1 Dark Matter Particles	10
1.1.1 Weakly Interacting Massive Particles	12
1.2 WIMP Dark Matter Searches	13
1.2.1 Direct Detection of WIMPs	15
1.2.2 WIMP Detection Experiments	20
1.3 Outline of Thesis	21
List of Abbreviations	1
2 The LUX Detector	24
2.1 Introduction	24
2.2 The LUX TPC	25
2.2.1 Target Xenon	27
2.2.2 Background Rejection	28
2.2.3 The Drift Field inside the LUX TPC	31
2.3 Light and Charge Signals in Liquid Xenon	33
2.3.1 Electronic Recoils (ER)	34
2.3.2 Nuclear Recoils (NR)	36
2.3.3 Energy and Position Reconstruction	37
2.4 Identifying S ₁ , S ₂	38
2.5 LUX Science Result (WIMP limit)	42

3	Spacial Dependent Correction of the S1 and S2 Signals	49
3.1	^{83m}Kr Calibration	51
3.2	^{83m}Kr Mixing in Liquid Xenon	55
3.3	S2 Electron Lifetime and x,y Correction	55
3.4	S1 Spacial Dependant Correction	59
3.5	Application to x,y,z Corrections in Data Processing	65
4	Calibrating the Combined Energy Scale	69
4.1	Anti-Correlation Space	72
4.2	Refitting in Combined Energy Space	73
4.3	Finding Errors with MCMC	79
4.4	Combined Energy Space	80
4.5	Light Collection and Electron extraction	82
4.6	Tritium Beta Spectrum	84
5	Event Level Fluctuations	90
5.1	An Introduction to Fluctuations	90
5.2	Modeling Intrinsic Detector Resolution	95
5.3	Measuring Recombination Fluctuations with Mono-Energetic Sources	101
5.4	Measuring Recombination Fluctuations in Discrete Energy Bins . .	108
5.4.1	Application to ^{83}Kr	119
5.4.2	Application to Simulated Tritium Data	121
5.5	Extracting Recombination Fluctuations from Tritium Calibration Data	126
5.5.1	Extracting Recombination fraction From Tritium Data	130
5.5.2	Modeling the ER Band	131
5.6	Measuring Alpha From the Tritium Data	133
5.7	Extracting Recombination Fluctuations from ^{137}Cs Calibration . . .	135
5.8	Recombination Fluctuations, The Bigger Picture	137
5.8.1	Encounter Recombination Probability	139
5.8.2	Clusters of Encounter Recombination	147
5.9	Conclusion	155
6	Ionization and Scintillation Yield from Tritium Calibration	157
6.1	Correcting for Spectral Shape for Finite Resolution	158
6.1.1	Calculating the Observed Energy	159
6.1.2	Smearing a Toy Spectrum	161
6.2	Light Yield, Charge Yield and Comparison to NEST Modeling	163
6.2.1	Tritium S1 and S2 vs. NEST	164
6.2.2	Tritium Energy Spectrum	167
6.2.3	Results for Light Yield, Charge Yield and Recombination . . .	171
6.3	Measuring Light Yield, Charge Yield and Recombination, Corrected for Spectral Shape	174
6.3.1	Tritium S1 and S2 Correction	174
6.4	Ionization and Scintillation Yield After Correction	178

6.5	Comparison of Light Yield Measured With Tritium to Other Measurements With Compton Scatters	183
6.5.1	The Standard Candle: Light Yield from ^{83m}Kr	183
6.5.2	Comparing Tritium Scintillation Yield with Compton Scatters	185
6.6	Summery	188
7	Tritiated-Methane Calibration Source and the ER Band	190
7.1	The need for an Internal Calibration Source	190
7.2	Tritiated-Methane as a Calibration Source	192
7.3	Removal of CH_4 from LUX	194
7.4	Light Yield Quenching from CH_4 in LUX	196
7.4.1	Tritiated Methane injection into the LUX detector	198
7.4.2	Mixing of Tritiated Methane in Liquid Xenon	199
7.4.3	Definition of Electronic Recoil Band. ER and NR Discrimination	201
7.5	ER Band Gaussianity	205
7.6	S1 Threshold For Golden Events, Using Tritium	206
7.7	Conclusion	208
	Bibliography	210

List of Figures

1.1	Measured rotational velocities vs. radius in the Milky Way galaxy. The velocity distribution is expected to fall off as $\sim \frac{1}{r}$ beyond the radius of the luminous disk, shown as the dashed magenta line. The Navarro-Frank-White model in solid black, along with other models listed in the caption. The velocity distribution appears consistent with that expected due to a halo of mass surrounding the galaxy, well beyond the observed luminous disk [1].	3
1.2	Maru the cat explains the cosmic microwave background. Consider that Maru is a photon and the box size represents local energy densities of the universe. The scale of the the box size is inversely proportional to the local energy density and temperature. At the time of last scatter all boxes containing Marus cease to exist. The Marus are now left to propagate freely through the universe, with their configurations unchanged. Figures a-d show Maru the cat contained within increasing box sizes corresponding to decreasing energy densities.	7
1.3	Improvement of resolution from COBE, first to discover anisotropies in the CMB, to WMAP and Planck which have set stringent limits on cosmological parameters by mapping variations in temperature of 1 part in 100,000. Image credit: NASA/JPL-Caltech/ESA.	8
1.4	BAO peak obtained from BOSS [2]. The correlation function has a peak at of $100h^{-1}$ Mpc, expected with a dark matter density of $\Omega_bh^2 \sim 0.2$	9
1.5	The concentration of mass in the bullet-cluster as observed from X-rays, emitted by baryonic matter, in pink, and the concentration of mass from gravitational lensing in blue. [Composite image credit: X-ray: NASA/CXC/CfA/M.Markevitch et al.; Optical: NASA/STScI; Magellan/U.Arizona/D.Clowe et al.; Lensing Map: NASA/STScI; ESO WFI; Magellan/U.Arizona/D.Clowe et al.]	11

1.6	Methods for detecting WIMPs from interactions of (q) standard model particles and (χ) dark matter particles χ . WIMPs can be detected through the annihilation of dark matter particles into standard model particles, $\chi\chi \rightarrow qq$. WIMPs can be produced by colliding standard model particles, $qq \rightarrow \chi\chi$. Or one can look for the scatter of a dark matter particle off a standard model particle, $\chi q \rightarrow \chi q$	14
1.7	Dark matter density vs. distance from the galactic center, calculated from galactic rotation curves of the Milky Way galaxy. The Earth is located at 8.3 kpc. Figure from [1].	16
1.8	Plot of WIMP event rate per kg/day/keV vs. Nuclear recoil energy (keV) for several target nuclei, using parameters from [3]. Figure taken from [32]. At low detection threshold xenon is the most attractive target nuclei. The fall in the xenon curve is due to decoherence described by the Helm factor [4].	19
2.1	Photo of the outer vessel of the LUX detector from inside the water tank.	25
2.2	Illustration of the LUX detector's internals. The detector contains two arrays of PMTs on the top and bottom housing 61 PMTs each. Teflon panels on the edges of the active region are used to reflect scintillation signals. The vertical distance between the two PMT arrays is 60 cm, and the diameter to the outer edge of the teflon panels is 50 cm.	26
2.3	Simulation of expected gamma background events inside the LUX detector, using the single scatter cut requirement[5]. There is an additional factor of 1000 background rejection within the fiducial volume where the WIMP search is conducted (inside the black dashed lines) .	29
2.4	The ER type and NR type bands from a) AmBe and ^{252}Cf (Red) and b) tritium calibration (Blue). The band means are solid lines and the 10-90% CL are shown as dashed lines. The ER to NR discrimination by using the charge to light ratio was measured to 99.6 ± 0.1 % at 50% NR acceptance.	31
2.5	Field grids in the LUX detector during the 2013 science run. T,A,G,C,B are biassed to -1,+3.5,-1.5,-10,-2 (all in kV), respectively. The PMTs are biased to -1.2 kV on average. The figure on the right shows the electric field model in the drift region between the cathode and gate for drift distance z vs. detector radius r. On average the drift field is 170 V/cm with variation from 140 V/cm to 200 V/cm from cathode to gate. Electric field model from [6].	32

2.6	An electronic recoil (ER) event in xenon. The energy deposited is converted primarily to ionization and roughly one tenth for excitation. Only several percent is lost to heat. Xenon excitons and recombining electron ion pairs for xenon dimers which de-excite producing very ultra violet (VUV) scintillation light at 175nm producing the primary scintillation signal (S1). Electrons that do not recombine are drifted by an electric field into the gas phase where they are accelerated producing the secondary scintillation (S2) signal.	35
2.7	A nuclear recoil (NR) event in xenon. The energy deposited goes mainly towards heat (phonons), the remaining energy is split evenly between ionization and excitation. Xenon excitons and recombining electron ion pairs for xenon dimers which de-excite producing very ultra violet (VUV) scintillation light at 175nm producing the primary scintillation signal (S1). Electrons that do not recombine are drifted by an electric field into the gas phase where they are accelerated producing the secondary scintillation (S2) signal.	36
2.8	Event diagram.	38
2.9	2keV ER event as seen by each PMT channel of the LUX detector. The S1 signal summed across all channels is overlaid on the top left, and the S2 signal summed across all channels is overlaid on the top right.	39
2.10	Density plot of prompt fraction vs. Pulse Area. Top: ^{83m}Kr data set. Bottom: Tritium data set. Populations of single electrons, single photons and the S1 S2 pairs associated with γ , β and α are highlighted as rectangles.	41
2.11	All single scatter events seen in the active region of the LUX detector over the course of the first science run passing all cuts listed in table 2.3 excluding the fiducial cut. The dashed cyan box indicates the fiducial volume.	43
2.12	The remaining events passing the quality cuts listed in table 2.3. The charge to light ratio (S2/S1) is plotted vs. S1 (proportional to energy) to show the separation of ER and NR type events. The 10-90% CF limits of the ER and NR band are plotted as the dashed blue and red curves, respectively. The band means are solid. S2 _b stands for the S2 signal on the bottom PMT array.	45
2.13	LUX detector is consistent with a P value of 0.35 of the background only hypothesis. The 90% upper C.L. cross section for various spin independent masses are shown in figure 2.13 in blue. Also shows are limits from Xenon100 (red), CDMS II (green), ZEPLIN III (magenta), and one sigma signal claimed for DAMA/LIBRA (shaded grey), CDMS II Silicone (shaded green), CRESST II (shaded yellow).	47
3.1	A simplified decay diagram of ^{83}Rb and ^{83m}Kr , from [7].	51

3.2	The S1 and S2 of two ^{83m}Kr events. Top Figures (a): Show a ^{83m}Kr event in which the two x-rays have overlapped within 60 ns, the S2 arrives about 93 μs later. Bottom Figures (b): Show a ^{83m}Kr event in which the two x-rays have overlapped within 220 ns, the S2 arrives about 205 μs later. The LUX pulse finder classifies events within a $1\mu\text{s}$ window as a single S1. The S2 pulses are insensitive to the timing separation of the dual decay as electron diffusion smears the pulses two together as the electrons drift. The PMT hit map for these events are shown below in figure 3.3.	53
3.3	The S1 and S2 of two ^{83m}Kr events. Top Figures (a): Show a ^{83m}Kr event in which the two x-rays have overlapped within 60 ns. Bottom Figures (b): Show a ^{83m}Kr event in which the two x-rays have overlapped within 220 ns. The black, open circle represents the location of the event. The S1 hit pattern is diffuse with more light collected on the bottom arrays due to total internal reflection at the liquid surface. The S2 is localized in the top PMT arrays in x,y at the location where the electrons are extracted, and diffuse on the bottom due to scattering. The summed waveforms for these events are shown above in figure 3.2.	54
3.4	Distribution of ^{83m}Kr events 10 minutes after the injection. The source mixes uniformly throughout the liquid xenon illuminating all regions of the active volume. The solid black lines represent the fiducial volume used for the WIMP search.	56
3.5	Left: Fits to the mean of $S2_b$ of the ^{83m}Kr data in several z slices. Right, the exponential fit to the means of $S2_b$ vs. drift time used to extract electron lifetime τ . The electron lifetime is found to be $\tau = 753.1 \pm 3.2 \mu\text{s}$. The exponential fit to the means deviates near the top and bottom of the active region since the charge yield from the ^{83m}Kr decay is sensitive to the varying electric field. The data shown was taken on May 10, 2013 (lux10_20130510_T1250) and contains 700,000 ^{83m}Kr events.	57
3.6	Left: Response of $S2_b$ vs. x,y normalized to the response at the center ($x=y=0$). The region of larger response around $x=0, y=-25$ is likely from an enhanced extraction field between the anode and gate wires. Right: Response of $S2_b$ vs. x,y after correcting the data using $\mathcal{N}\mathcal{F}_{S2_b}$	58
3.7	Improvement of resolution in $S2_b$ after applying the z and x,y,z correction. Black: The uncorrected data. Red: The data with only z dependent correction, the electron lifetime correction. Blue: The data with full x,y,z dependent correction. The data shown was taken on May 10, 2013 (lux10_20130510_T1250) and contains 700,000 ^{83m}Kr events.	59
3.8	S1 x,y,z response normalized to the center of the detector. The interpolated map represents the inverse of the normalization factor $\mathcal{N}\mathcal{F}$	61

3.9	S1 x,y,z response normalized to the center of each z slice. There are greater variations near the very top and bottom 4 z slices as the solid angle for light hitting the PMT arrays increases. For the central slices the x,y response is uniform due to increased scattering off the teflon panels.	62
3.10	S1 x,y,z response normalized to the center of the detector after the data has been corrected and normalized to the detector center. The remaining variations in the fiducial volume ($r < 18\text{cm}$) is less than 1%. Near the top and bottom edges the deviation increases to as much as 3% due to the interpolation of the correction becoming poorly constrained.	64
3.11	Improvement of resolution in S1 after applying the z and x,y,z correction. Black: The uncorrected data. Red: The data with only z dependent correction. Blue: The data with full x,y,z dependent correction. The data shown was taken data on May 10, 2013 (lux10_20130510_T1250) and contains 700,000 $^{83\text{m}}\text{Kr}$ events.	65
3.12	Electron lifetime measured using $^{83\text{m}}\text{Kr}$ calibrations during the LUX science run in 2013. The blue dashed lines show the boundaries of the WIMP search from April 21 to Aug 8, 2013. The red dashed lines indicate circulation loss events.	66
3.13	Measured response of to light from $^{83\text{m}}\text{Kr}$ calibrations at the center of the LUX detector during of the LUX science run in 2013. The blue dashed lines show the boundaries of the WIMP search from April 21 to Aug 8, 2013. The red dashed lines indicate circulation loss events.	67
4.1	LUX data in anti correlation space (S1 vs. S2), the black lines indicate the initial cuts by eye used to isolate populations of constant energy. In both figures diagonals represent lines of constant energy with a slope depending on the local recombination probability. The centroids found by an unbind maximum likelihood analysis are shows as a black X, for sources show in table 4.1.	73
4.2	Doke plot showing the best fit for the energy calibration parameters g1 and g2 using S1 and S2 means extracted from anti-correlation space. The first three blue points, from the left, are from ^{83}Kr calibrations at 50, 100 and 170 V/cm respectively. The open circle was from the K-shell xenon X-ray and was not used for the fit as it's absolute energy and origin from the skin of the detector is uncertain.	74
4.3	Doke plot showing the best fit for the energy calibration parameters g1 and g2 using S1 and S2 extracted from a combined energy space. The first three blue points, from the left, are from ^{83}Kr calibrations at 50, 100 and 170 V/cm respectively. The open circle was from the K-shell xenon X-ray and was not used for the fit as it's absolute energy and origin from the skin of the detector is uncertain.	75

4.4	S1 fits to sources at nominal field of 170 [V/cm] unless otherwise noted. Source and energy in keV from top left to bottom right: a) ^{131}Xe : 163, b) ^{127}Xe : 207, c) ^{127}Xe & $^{129\text{m}}\text{Xe}$: 236.8, d) ^{127}Xe : 410, e) ^{214}Bi : 609, f) ^{137}Cs : 661.6, g) $^{83\text{m}}\text{Kr}$: 41.5 - at 50 [V/cm], h) $^{83\text{m}}\text{Kr}$ 41.5 - at 100 [V/cm], i) $^{83\text{m}}\text{Kr}$ 41.5	76
4.5	S2 fits to sources at nominal field of 170 [V/cm] unless otherwise noted. Source and energy in keV from top left to bottom right: a) ^{131}Xe : 163, b) ^{127}Xe : 207, c) ^{127}Xe & $^{129\text{m}}\text{Xe}$: 236.8, d) ^{127}Xe : 410, e) ^{214}Bi : 609, f) ^{137}Cs : 661.6, g) $^{83\text{m}}\text{Kr}$: 41.5 - at 50 [V/cm], h) $^{83\text{m}}\text{Kr}$ 41.5 - at 100 [V/cm], i) $^{83\text{m}}\text{Kr}$ 41.5	77
4.6	Doke plot of the data showing the light yield vs. charge yield. Only solid circles are used for the fit, the open circle is the Xe K shell X-ray from the detector edge.	78
4.7	MCMC for the linear fit to the Doke plot. There is a strong negative correlation between the slope m and intercept b which results from the degeneracy between gains g1 and g2.	80
4.8	Combined energy scale. a) The xenon activation lines from early in the run. b) ^{137}Cs calibration data. c) All calibration data including the $^{83\text{m}}\text{Kr}$ calibration. d) Xenon X-ray.	81
4.9	Single electron distribution as seen by the bottom PMT array fitted with a skew Gaussian model to account for the underling Poisson statistics of observing between 9 and 10 photo electrons (PE) for each electron leaving the liquid and entering to the gas phase at a higher field causing electro-luminescence. The μ of the fit represents the true mean of the skew Gaussian distribution.	84
4.10	The tritium energy spectrum reconstructed from the data (black). Along with LUX SIM (blue), and the true tritium beta spectrum (blue) and a tritium spectrum smeared with detector resolution (green).	86
4.11	Threshold calculated from difference of simulated Tritium S1, S2 and energy spectra at 170 V/cm. a) S1 b) S2 _b , c) Combined Energy . . .	88
4.12	Threshold calculated from difference of simulated Tritium S1, S2 and energy spectra at 100 V/cm. a) S1 b) S2 _b , c) Combined Energy . . .	89
5.1	Fluctuations of the 164 keV line from ^{131}Xe in the LUX detector, with the number of electrons plotted vs. the number of photons. The blue and magenta arrows labeled as $\sigma_{n_{\gamma\text{Det}}}$ and $\sigma_{n_{e\text{Det}}}$ are the size of fluctuations in the light and charge channels due to the resolution of the LUX detector. The black arrow represents the size of recombination fluctuations σ_r . The value of $\langle r \rangle$ is the average observed recombination fraction or the average recombination probability r_p . The red lines represent constant $\langle r \rangle \pm \sigma_r$ assuming the expected binomial variance of equation 5.1. The black lines represent constant $\langle r \rangle \pm \sigma_r$ measured from the data. The black dashed line is represent constant energy or constant number of quanta.	92

5.2	The projection of the population in figure 5.1 onto the combined quanta axis ($n_\gamma + n_e$) = E/W. Along this projection the dominant recombination fluctuations cancel out, leaving on the components from detector resolution $\sigma_{n_{\gamma\text{Det}}}$ and $\sigma_{n_{e\text{Det}}}$	93
5.3	The fluctuations in the S1-x,y,z corrected signal for the 9.4 keV from $^{83\text{m}}\text{Kr}$. At 9.4 keV the statistical variance is dominate over recombination fluctuations, $\sigma_{S1} = 8.15$. The resolution is consistent with that expected from statistical fluctuations of 8.3 ± 0.1 PE, from equation 5.8.	98
5.4	Populations of calibration sources in discrimination space $\log\left(\frac{n_e}{n_\gamma}\right)$ vs. combined energy keV _{ee} . The ovals represent the combination of σ_R , $\sigma_{n_{\gamma\text{Det}}}$, $\sigma_{n_{e\text{Det}}}$ in black, white, red respectively.	106
5.5	Measured values of σ_R , $\sigma_{n_{\gamma\text{Det}}}$, $\sigma_{n_{e\text{Det}}}$ vs. Energy on the left and vs. quanta in photons, electrons and ions, respectively, on the right. Measured using sources listed in Table 5.3.	107
5.6	A simulated $^{83\text{m}}\text{Kr}$ source with infinite detector resolution and recombination fluctuations set to 90 quanta. The recombination fluctuation move along the line of constant energy about the center, shown in black with bin width of $\Delta_E = 1$ keV. The events between the black lines and projected onto the n_γ and n_e axes. The variance in n_γ and n_e are $\sigma_{n_\gamma}^2 = \chi_{n_\gamma}^2 = \sigma_R^2$ and $\sigma_{n_e}^2 = \chi_{n_e}^2 = \sigma_R^2$	110
5.7	A simulated $^{83\text{m}}\text{Kr}$ source with no recombination fluctuations and with detector resolution, $\sigma_{n_{\gamma\text{Det}}} = 171$ and $\sigma_{n_{e\text{Det}}} = 51$. The solid black lines represents constant energy about the center having width of $\Delta_E = 1$. The events that fall between the black dashed lines are projected onto the n_γ and n_e axes. The square root of the observed variance in light and charge are show on the labels with and without the energy cut, as χ and σ respectively. The energy cut has swept out a component of variance χ_{Det} that is shared between the light and charge channels.	112
5.8	A simulated $^{83\text{m}}\text{Kr}$ source with no recombination fluctuations and with detector resolution, $\sigma_{n_{e\text{Det}}} = \sigma_{n_{\gamma\text{Det}}} = 100$. The solid black line represents constant energy about the center having a width of $\Delta_E = 1$. The events that fall between the black dashed lines are projected onto the n_γ and n_e axes. The square root of the observed variance in light and charge are show on the labels with and without the energy cut, as χ and sigma respectively.	116

5.9 Data from a ^{83m}Kr data set, plotting the number of photons against the number of electrons. The red-dashed oval represents where the population would lie if recombination fluctuations were the dominant fluctuation. The solid black line represents constant energy about the center, and the region between the black dashed lines represents the cut in energy with bin width $\Delta_E = 1$. The events that fall between the black dashed lines are projected onto the n_γ and n_e axes. The observed standard deviation in n_γ and n_e is shown on the plot labels as χ and σ .	119
5.10 Number of photons plotted against the number of electrons from a simulated tritium data set having only recombination fluctuations. Left: Tritium events propagated from 1 keV bins from 1-18 keV. Right, the tritium events propagated from 0.1 keV bins from 1-18 keV. The recombination fluctuations move events along the lines of constant energy.	121
5.11 The number of photons plotted against the number of electrons from a simulated tritium data set having recombination fluctuations and detector resolution. The tritium events propagated from 0.1 keV bins from 1-18.6 keV. Events move along the lines of constant energy due to recombination fluctuations and χ_{Det} .	122
5.12 Light yield (left) and Charge yield (right) of a simulated tritium spectrum, with the fit to the centroid show in black. The variance in number of photon and electrons per energy bin is used to extract recombination fluctuations. Recombination fluctuations and detector resolution χ_{Det} move events up and down along lines of constant energy.	123
5.13 Simulated tritium spectrum with the detector resolution of the LUX detector and an initial guess of recombination fluctuations. The detector resolution, labeled as $\sigma_{S1_{\text{Det}}}$ and $\sigma_{S2_{\text{Det}}}$ is shown in red and green, respectively. The value of recombination fluctuation σ_R input into the simulation is shown in magenta. The methodology described in this previous section is applied to extract σ_R from the electron and photon spectrum, black and blue respectively. The statistical component of the standard deviation in each bin is χ_{Det} is shown in green, equation 5.47. The plots show cases for various bin widths in keV: a) $\Delta E = 0.1$ b) $\Delta E = 0.25$ c) $\Delta E = 0.5$ d) $\Delta E = 1$.	125
5.14 a: Density plot of number of photons vs. energy in keV using the tritium calibration data at 170 V/cm. b: Number of electrons vs. energy in keV using the tritium calibration data at 170 V/cm. c: Number of photons vs. energy in keV using the tritium calibration data at 100 V/cm. d: Number of electrons vs. energy in keV using the tritium calibration data at 100 V/cm. The data has been corrected for spectral shape. The black line indicates the quadratic fit to the centroid of the population.	127

- 5.20 Recombination fluctuations measurements are labeled on the plot and include data from tritium at 170 V/cm, tritium at 100 V/cm, ^{137}Cs calibration, the line sources used for the energy scale calibration listed in table 4.1 and a ^{57}Co calibration at a variety of electric fields ranging from 60 to 5000 V/cm from [8]. Left: the observed recombination fluctuation σ_R measurements vs. the standard deviation expected from a purely binomial process, equation 5.51. The red curves represent power law fits to the data. 138
- 5.21 Illustration of recombination for the case of a single ion, blue box highlighted in black at the center. The total recombination probability for one ion is the combination of the dominant self recombination probability (r_s) plus the sum of the encounter recombination probabilities (r_e) from the remaining N-1 electrons. This process is repeated for N number of ion electron pairs leading to a binomial variance given in equation 5.54. 141
- 5.22 Encounter recombination probability r_e extracted from the tritium and ^{137}Cs calibration data at 170 V/cm, derived from equation 5.58. Left plotted vs. energy in keV. Right: plotted vs. the number of ions. The solid red line represents the overall average of r_e from the calibration data is $r_e = 0.0042$ 144
- 5.23 The red crosses indicate the total recombination probability r for the calibration sources labeled on the plot. The black x-s indicate the encounter recombination probability r_e . The blue points indicate the tritium data at 100 V/cm. The data includes tritium at 170 V/cm (black), tritium at 100 V/cm (blue), ^{137}Cs and data from Dahl for ^{57}Co ranging from 60 to 5000 V/cm [8]. 145
- 5.24 Fit to the observed ER recombinations modeled with binomial fluctuations where the recombination probability includes an encounter recombination component. Case one (red) is with assuming a constant $r_e = 0.0042$, extracted from the tritium and ^{137}Cs data. Case two (green) is using $r_e = 0.0011 + 0.006r$. Case two is in significantly better agreement with the field dependent data from Dahl, as it accounts for field effects [8]. The fit for case two (green) deviates less than 30% from the data. 146
- 5.25 Illustration of recombination for the case of a single ion occurring in N_C clusters with number of encounters N_e . The total recombination probability for all clusters is equal to the average of each individual cluster. In each cluster the recombination probability is equal to the number of interaction N_e times the encounter recombination probability r_e . The process is repeated for N_C clusters, resulting in an additional factor N_i in the observed variance while maintaining the correct average recombination probability r 149

5.26 The data shown are labeled on the plot and include calibration data of tritium data at 170 V/cm, tritium at 100 V/cm (blue), ^{137}Cs at 170 V/cm, and data over a range of fields using ^{57}Co from Dahl. [8]. a) The encounter recombination probability r_ϵ plotted vs. energy. b) The number of encounters per cluster N_ϵ vs. energy. c) The number of clusters N_C vs. energy.	151
5.27 The measured recombination fluctuations from calibration data vs. energy. The red line indicates the expected recombination fluctuation of the clustering model given in equation 5.65, assuming a cluster size of $N_\epsilon = 50$. $N_\epsilon = 50$ corresponds to an average encounter recombination probability of $r_\epsilon = 0.01$. The deviation of the fit from the data is less than 30%. The $^{83\text{m}}\text{Kr}$ data point is the result of two decays and is expected to fall below the curve by 40% considering the sum of the individual variances.	153
6.1 Smearing of a linear spectrum using equation 6.1 for an arbitrary energy scale E . The counts in each shaded bin are redistributed into normalized Gaussians (blue) with the resolution σ_E growing like \sqrt{E} . The spectrum smeared with detector resolution is the sum of Gaussians shown in red.	159
6.2 Top Left: A linearly decaying spectrum, in blue. The black curve represents the sum of the Gaussians assuming a constant resolution. Top Right: A linearly decaying spectrum, in blue. The black curve represents the sum of the Gaussians with a $\sigma = \sqrt{E}$ dependent resolution. Bottom: The observed mean, with finite resolution, compared to the real mean with infinite resolution. The black points are for the case with linear resolution and the red points represent the case with $\sigma = \sqrt{E}$ dependent resolution.	163
6.3 a): In Black S1 tritium S1 spectrum extracted from the data. In blue, The NEST light yield curve. In red, the NEST light yield curve with recombination fluctuations. Dashed magenta is NEST light yield with smearing from equations 6.7. b): The ratio of the real S1 mean to the S1 observed mean vs. the observed mean after smearing. Note, the S1 threshold at about 3 PE in S1. c): In Black S2 tritium spectrum extracted from the data. In blue, The NEST light yield curve. In red, the NEST light yield curve with recombination fluctuations. Dashed magenta is NEST light yield with smearing from equations 6.8. d): The ratio of the real S2 mean to the S2 observed mean vs. the observed mean after smearing.	167
6.4 The tritium energy spectrum reconstructed from the data (black). Along with LUX SIM (red), the true tritium beta spectrum (dashed blue) and a tritium spectrum smeared with detector resolution of equation 6.11 (green).	169

- 6.5 Left, mapping from real Monte Carlo energy to observed energy plotted vs the observed energy after applying a finite resolution using LUXSIM. Right, comparing the correction determined from the Monte Carlo (Red) to the detector smearing model (black) given in equation 6.11. The dashed lines represent the uncertainty in the measured values. The agreement is within errors from 1 to 18 keV_{ee}. The Energy threshold is 50% at 1.0 keV_{ee}. 170
- 6.6 Top: Extracted recombination fluctuation from the tritium data from fluctuations in photons and electrons (Black and Red respectively). Bottom right: mean number of quanta in photons, electrons, ions, excitons vs. energy keV for the tritium calibrations. Bottom left: Recombination fraction and the one sigma (shaded) vs. energy keV. . 172
- 6.7 Light yield and charge yield from tritium data without spectral shape correction at 170 V/cm in black. The shaded region represents the one sigma systematic uncertainty on g1 and g2. The NEST yield prediction and its corresponding 1 sigma is shaded in blue. NEST interpolation is shown in magenta to energies where the model is not vetted. Note that the statistical errors are about the size of the data points and the dominant uncertainty illustrated by the shaded region is 100% correlated bin-to-bin. A one sigma shift up in light yield corresponds to a one sigma shift down in charge yield and visa versa. 173
- 6.8 a): In Black S1 tritium S1 spectrum extracted from the data. In red, the S1 spectrum based upon the LY and QY measured from the tritium data after applying recombination fluctuations. In dashed blue, the expected S1 spectrum after applying finite detector resolutions of equation 6.7. b): The ratio of the real mean to the observed mean vs. the observed mean after smearing the tritium photon spectrum with detector resolution. Note the S1 threshold at about 3 PE in S1. c): In black, tritium S2 spectrum from the data. In red, the S2 spectrum based upon the LY and QY measured from the tritium data after applying recombination fluctuations. In dashed blue, the expected S2 spectrum after applying finite detector resolutions of equation 6.8. d): The ratio of the real mean to the observed mean vs. the observed mean after smearing the tritium electron spectrum from NEST with detector resolution. Note the S2 threshold at about 400 PE in S2. . 176
- 6.9 Means of the light yield and charge yield from tritium data corrected for spectral shape along with the 1 sigma statistical errors. 178
- 6.10 Light yield and charge yield from tritium data corrected for spectral shape along with the 1 sigma systematic constraint on g1 and g2. The blue and magenta curve are NEST extrapolation and interpolation, respectively. 180
- 6.11 In black, the light yield and charge yield from tritium data corrected for spectral shape. In red, the light yield and charge yield from tritium data uncorrected for spectral. The shaded region represent the one sigma systematic error from the constraint on g1 and g2 shaded. . . 182

6.12 Left: The light yields of the 9.4 and 32.1 keV decay of ^{83m}Kr plotted vs. timing separation, for events separated by more than 1200 ns. The point at 600 ns is calculated assuming the 32.1 keV yield remains flat to 200 ns as observed in [9] and [10]. Right: the histogram of ^{83m}Kr events vs timing separation finding a best fit to the half life of 159.7 ± 7.5 consistent with the measured value of 154 ns [11] [12].	184
6.13 Left: S1 ^{83m}Kr peaks at zero field. Right: shows the histogram of the separated 32.1 and 9.4 keV decays plotted above vs. time. The half life fit to the population is in good agreement with the measured half-life of 154 ns [11] [12].	185
6.14 LY relative to the light yield of the 32.1 keV decay of ^{83m}Kr at zero field. The black and blue bands represent the results from tritium at 170 and 100 V/cm respectively. The shaded region represents the systematic error due to the one sigma constraints on g1 and g2. Magenta points are Compton scattering measurements from [13]. The grey and red represent zero field and 450 V/cm Compton Scattering measurements from [10].	187
7.1 True tritium beta spectrum from [14].	193
7.2 Removal of natural methane observed by the integrated xenon sampling system prior to the tritiated-methane injections. The red points indicate measurements at the getter outlet, we find a 97% one pass removal efficiency at a flow rate of 25 SLPM. The blue curve shows the improved upper limit on the effect of outgassing from the plastics. The black dashed lines shows the exponential fit to the natural methane removal from the xenon with a time constant of 5.9 ± 0.07 hours. 5×10^{-3} ppb (g/g) is the limit of detection for methane.	195
7.3 In black, the response to scintillation from ^{83m}Kr at the center of the detector normalized to the first data point before the natural methane (CH_4) injection. The dashed magenta lines represent the time window from the beginning of the natural methane injection to the time the background of 5 ppt is reached. The blue points represent that methane concentration in the gas returning from the bulk liquid of the detector. The concentration in the liquid xenon is roughly 1/6 of the concentration measured in the gas phase due to solubility.	197
7.4 The rate of single scatter events with S1 below 100 PE in the fiducial volume. 100 PE in S1 is about 18.6 keV _{ee} , the endpoint to the tritium beta spectrum. The magenta and red curves are fits to the first and second tritium injection's removal rate. Note, the removal of tritiated-methane is consistent with the natural methane removal rate measured independently.	199

7.5	The distribution of tritium events vs. detector radius squared. The solid black line represents the fiducial volume. Right: The distribution of tritium events vs. XY in the region between the gate and the cathode. The solid black line represents the fiducial volume and the black circles represent the locations of PMTs (photo multiplier tubes).	200
7.6	Charge to light ratio $\log_{10}(S2_b/S1)$ plotted vs. S1, the energy contours are also shown. Top, the tritium data uncorrected for spectral shape. Bottom, with the spectral shape correction discussed in chapter 6. The calibration consists of over 112,000 tritium beta decays between 1 and 50 PE in S1 ($1 - 10 \text{ keV}_{ee}$), in the fiducial volume of the detector. The black circles represent the ER band mean and 10% to 90% bounds, $\pm 1.28\sigma$. The red solid line is the NR band mean determined using a DD neutron generator calibration. The dashed red lines indicate the 10% to 90% bounds of the NR band.	202
7.7	Discrimination vs. S1 using over 112,000 tritium beta decays between 1 and 50 PE in S1 ($1 - 10 \text{ keV}_{ee}$), using tritium data at 170 V/cm. Top, the tritium data uncorrected for spectral shape. Bottom, with the spectral shape correction discussed in chapter 6. The dashed horizontal lines indicate the mean leakage fraction from 1 to 50 PE.	204
7.8	ER band Gaussianity from 0-50 PE with the centroid of the ER band subtracted off, using tritium data at 170 V/cm. Top, the tritium data uncorrected for spectral shape. Bottom, with the spectral shape correction discussed in chapter 6. The dashed vertical red line indicates the mean of the NR band from 1 to 50 PE.	206
7.9	The S1 detection efficiency measured from tritium data at 170 V/cm (black square) and 100 V/cm (blue circle) is plotted along with the best fit to an error function to the higher statistics 170 V/cm data. The red triangles are detection efficiency determined from LED pulsing.	207

Chapter 1: Introduction

Astronomical observations hinting at the existence of dark matter were first observed in 1932 by Oort [15] and more precisely in 1937 by Zwicky [16]. Both noted discrepancies in galactic mass measurements when comparing the luminous mass to that required to support galactic rotational velocities measured by red shifts. Oort had noted up to a factor of ten more mass than luminous mass in the Sombrero Galaxy and Zwicky found a factor of 500 for the Coma cluster. Both observations were far to large to be accounted for by light absorption, indicating the existence of dark matter to account for the missing mass. Since then more evidence for the existence of dark matter has been compiled, including big bang nucleosynthesis (BBN), anisotropies in the cosmic microwave background (CMB), baryonic acoustic oscillations (BAO), formation of large structures, galactic rotation curves, and gravitational lensing. All independent techniques leading to a unified conclusion for the existence of non-baryonic and non-luminous matter. Individually some pieces of evidence, such as galactic rotation curves, can be explained by modifications to general relativity (GR), but not all simultaneously. The existence of non-relativistic, dark matter particles are required in order to unify the current observations. This dark matter does not couple to the electromagnetic force and is thus able to avoid our

standard optical detection techniques, making its presence felt on large scales via gravity.

1.0.1 Galactic Rotation Curves

There are two common methods for measuring the mass of a galaxy or cluster of galaxies. First, one can use the total luminosity and the known distance to the galaxy to determine the luminous mass, (the mass corresponding to the visible light). Second, the rotational velocities of stars orbiting the galactic center can be mapped and used to determine the mass distribution as a function of galactic radius. Rotational velocities of stars around galactic centers at large distances can be measured with Doppler shift, with more recent measurement relying on the 21 cm H1 line from hydrogen as the standard candle. The rotational velocities of objects orbiting galaxies are highly non-relativistic, moving at speeds on the order 100 km/s. At the outer edges of the luminous galactic centers, typically past 5 kpc, the velocity distribution is expected to fall off as predicted by Newtonian mechanics ($\sim \frac{1}{r}$). Yet observations from as early as 1932 indicate that velocity distributions tend to remain constant with radius suggesting that the objects are rotating inside a solid body of dark matter [15] [16] [17] [1]. The velocity distributions measured for the Milky Way galaxy are show in 1.1.




Figure 1.1: Measured rotational velocities vs. radius in the Milky Way galaxy. The velocity distribution is expected to fall off as $\sim \frac{1}{r}$ beyond the radius of the luminous disk, shown as the dashed magenta line. The Navarro-Frank-White model in solid black, along with other models listed in the caption. The velocity distribution appears consistent with that expected due to a halo of mass surrounding the galaxy, well beyond the observed luminous disk [1].

1.0.2 Big Bang Nucleosynthesis

Big band nucleosynthesis (BBN) accounts for the relative abundances of light elements in the universe today, including H, D, He³, He⁴ and Li⁷ [18]. BBN took place in a relatively short time window beginning several seconds after the big bang when the universe cooled below 10¹¹K (10 MeV) and ceased as the temperature cooled below 10⁹K (100 keV). Under the temperature conditions of BBN it was energetically favorable for free protons and neutrons to undergo nuclear fusion. This is the only mechanism to produce the light elements we see today. The heavier elements were later fused together in stars and ejected upon the star's death into the cosmos. Nuclear cross sections of protons, neutrons and light elements have been measured to high precision and can be combined with the expansion rate of the universe to precisely predict the relic abundances of baryonic matter. Observations constrain the abundances of the light elements to be H~ 75%, D~ 25%, He⁴ ~ 0.01%, Li⁷ ~ 10⁻¹⁰ %. The ratio of D/H has been used to constrain the relic density of baryonic matter to be

$$\Omega_b h^2 = \frac{p_b}{p_c} = 0.02202 \pm 0.00046 \quad [19] \quad (1.1)$$

where h is the Hubble constant (H) dividend by 100 (H₀/100), p_b is the baryonic density and p_c is the critical density required for a flat universe (verified by the CMB). We can write the ith density component as:

$$\Omega_i \equiv \frac{p_b}{p_c} = \frac{8\pi G \rho_i}{3H^2} \quad (1.2)$$

where G is the gravitational constant. The baryon density measured using BBN is constrained to within 1% and is in agreement with the latest constraints from Plank's CMB data, $\Omega_b h^2 = 0.02205 \pm 0.00028$ [20].

1.0.3 Cosmic Microwave Background

The early universe consisted of a plasma opaque to photons as they scattered off free electrons. As the temperature fell below the binding energy of hydrogen 13.6 eV, electrons could bind with free protons forming neutral atoms making the universe transparent to photons. The mean temperature of decoupling was actually at 0.25 eV (\sim 4000 K) as photons still scatter frequently near the binding energy of hydrogen [18]. After making a final scatter, photons decoupled from electrons effectively attaining a mean free path on the scale of the universe. The photons from the CMB can be observed today at the red shift temperature of 2.72548 ± 0.00057 K [21] from the time of last scatter, 379,000 years after the big bang.

The CMB has encoded within it a wealth of information about the universe as it was at the time of decoupling. The the information encoding is illustrated in figure 1.2 using Maru the cat. Consider that Maru is a photon and the box size represents local energy densities of the universe. The smaller boxes represent areas of higher energy density and temperature. At the time of last scatter all boxes containing Marus cease to exist. The cats now propagate freely through the universe with their configuration unchanged. As the universe expands, so will the cats. When the Marus finally reach our telescopes, 13 billion years later, the shape and squeezing

of each Maru informs us of the box size (temperature) from which each Maru has emanated. Using this information from multiple Marus the distribution of box sizes at the time of last scatter can be mapped, revealing areas of slightly larger boxes and areas of slightly smaller boxes. This is roughly the idea behind measuring anisotropies in the cosmic microwave background using microwave telescopes.

Ever more precise measurements from COBE, WMAP and Planck have been able to probe slight temperature variations to 1 part in 100,000 as seen in figure 1.3. Table 1.1 shows the constraints on cosmological parameters set by Planck. The results are in good agreement with baryonic density derived from BBN and predict a dark matter component of 25.8%.

Parameter	Value	Definition
$\Omega_b h^2$	0.2214 ± 0.00024	Baryon energy density
$\Omega_c h^2$	0.1187 ± 0.0017	Cold dark matter energy density
$\Omega_m h^2$	0.1423 ± 0.0029 *	Total matter energy density
Ω_Λ	0.692 ± 0.010	Dark energy density
Ω_K	-0.0005 ± 0.0065 (95%)	Curvature
Σ_{m_v}	< 0.230	Sum of neutrino masses [eV]
H_0	67.77 ± 0.77	Hubble Constant [$\text{km s}^{-1} \text{Mpc}^{-1}$]

Table 1.1: Cosmological parameters from Planck+WP+highL+BAO [20]. * Only Planck.

1.0.4 Baryonic Acoustic Oscillations

The universe 379,000 years after the big bang was uniformly distributed, with only small variations observed in the CMB temperature of 1/100,000 [22]. Before decoupling took place, gravity pulled baryons and dark matter into high density regions resulting in an opposing outward force from photon pressure. The outward force




Figure 1.2: Maru the cat explains the cosmic microwave background. Consider that Maru is a photon and the box size represents local energy densities of the universe. The scale of the the box size is inversely proportional to the local energy density and temperature. At the time of last scatter all boxes containing Marus cease to exist. The Marus are now left to propagate freely through the universe, with their configurations unchanged. Figures a-d show Maru the cat contained within increasing box sizes corresponding to decreasing energy densities.




Figure 1.3: Improvement of resolution from COBE, first to discover anisotropies in the CMB, to WMAP and Planck which have set stringent limits on cosmological parameters by mapping variations in temperature of 1 part in 100,000. Image credit: NASA/JPL-Caltech/ESA.

from the photon pressure was only felt by the baryons whereas the dark matter component does not couple to photons. The competing attractive and repulsive forces gave rise to baryonic acoustic oscillations (BAO), with density regions propagating as spherical sound waves. The peaks of the waves are separated by a characteristic radius called the sound horizon (r), which is sensitive to the initial dark matter and baryon densities [23]. Anisotropies in the CMB power spectrum probe these oscillations as discussed previously in section 1.0.3. At the time of decoupling the photon pressure ceased providing the opposing force allowing the gravitational restoring force to dampen the oscillations. If this picture is propagated forward in time, we expect that areas of the CMB that were denser would cluster, thus statistically

the universe is expected to have large scale structures on the order of the sound horizon r . Measurements of galaxy clusters by the Sloan Digital Sky Survey [24] and BOSS [2] are consistent with the sound horizon expected from anisotropies in the CMB, with a preferred scale of $100h^{-1}$ Mpc (~ 150 Mpc) between large scale structures. Figure 1.4 shows the result from BOSS using Lyman- α absorption in the quasar emission spectrum due to the presence of neutral hydrogen in the intergalactic medium [2].




Figure 1.4: BAO peak obtained from BOSS [2]. The correlation function has a peak at of $100h^{-1}$ Mpc, expected with a dark matter density of $\Omega_b h^2 \sim 0.2$.

1.0.5 Gravitational Lensing

The Bullet cluster is composed of two galaxy clusters which have recently collided and passed through each other. The collision has caused the ordinary, baryonic

matter to heat and emit X-rays that are observed and used to map the luminous mass distribution [25]. However, the observed concentration of mass is not consistent with the center of mass observed using gravitational lensing via GR [26]. The lensing indicates the presence of a dark-matter shell which, unlike the ordinary matter, has passed through undisturbed due to its lack of interactions. Figure 1.5 shows the concentration of mass in the bullet-cluster as observed from X-rays, emitted by baryonic matter, in pink and the concentration of mass from gravitational lensing in blue. The X-ray mapping from Chandra, when compared with gravitational lensing studies of the Bullet cluster, clearly demonstrate a decoupling of the dark matter center of mass from the baryonic center of mass induced by the cluster's recent collision.

1.1 Dark Matter Particles

The evidence for the existence of missing mass outlined in the previous section motivates us to examine solutions to account for those cosmological phenomena. It is natural to first try to solve the anomaly with standard model particles, considering is the existence of Massive Compact Halo Objects (MACHOs). A MACHO is a Jupiter size object that could add additional mass at the outer edges of galaxies. However, since MACHOs are baryonic the hypothesis is disfavored by precision measurements (BBN, CMB) that limit baryonic mass to only 2.2% of the total [27] [20] (see table 1.1). A non-baryonic candidate drawn from the standard model are neutrinos, which are known to be massive and only interact weakly [28]. However, large




Figure 1.5: The concentration of mass in the bullet-cluster as observed from X-rays, emitted by baryonic matter, in pink, and the concentration of mass from gravitational lensing in blue. [Composite image credit: X-ray: NASA/CXC/CfA/M.Markevitch et al.; Optical: NASA/STScI; Magellan/U.Arizona/D.Clowe et al.; Lensing Map: NASA/STScI; ESO WFI; Magellan/U.Arizona/D.Clowe et al.]

scale structure formation require that the universe have a ‘cold’ (non-relativistic) dark matter component in order to become gravitationally bound to galaxies. Neutrinos are highly relativistic and would fail to reproduce structuring of the universe observed today [29].

1.1.1 Weakly Interacting Massive Particles

A model with Weakly Interacting Massive Particles (WIMP) can be used to account for the dark matter component of the universe. The new particle would be massive, non-baryonic, and couple via the weak force. The WIMP is theorized to have a mass and cross section on the order of the weak interaction. In the early universe the number density of WIMPs and photons would have been roughly equal as there was sufficient thermal energy keep the creation and annihilation in equilibrium.



where χ represents WIMPs and q are standard model particles. The reaction can go in either direction as long as the thermal temperature of the universe is greater than WIMP mass, holding the relative abundances constant .

As the universe expanded and cooled, production of WIMPs from standard model particles ceased as the temperature of the universe dropped below the WIMP mass, leaving only WIMP annihilation into standard model particles. The annihilation would continue indefinitely, leaving only a small number density at the tail of an exponentially falling Boltzmann distribution remaining today. However, if the universe's expansion is fast compared to the annihilation then the WIMPs would fail to find each other and their number density could ‘freeze out’. The freeze out condition requires that the rate of the Universe’s expansion H is greater than the number density of WIMPs times the cross section.

$$H > \Gamma_A \equiv n_\chi \langle \sigma_A v \rangle \quad (1.4)$$

H is the Hubble constant, n_χ is the number density of WIMPs and $\langle \sigma_A \nu \rangle$ is the thermally averaged annihilation cross-section. The annihilation process can be described by the Boltzmann equation

$$\frac{dn_\chi}{dt} = -3Hn_\chi - \langle \sigma_A \nu \rangle (n_\chi^2 - n_{\chi_{eq}}^2) \quad (1.5)$$

where the first term represents the dilutions of WIMP number density with three degrees of freedom, n_χ^2 is from the annihilation process, $\chi\chi \rightarrow qq$ and $n_{\chi_{eq}}^2$ are from the reverse process $qq \rightarrow \chi\chi$. Equation 1.5 does not have an analytic solution but has been solved numerically[30]. The relic density is estimated within 10% to be,

$$\Omega_\chi h^2 = \frac{3 \times 10^{-27} \text{cm}^3 \text{s}^{-1}}{\langle \sigma_A \nu \rangle} \sim \frac{10^{-10} \text{GeV}^{-2}}{\langle \sigma_A \nu \rangle} \quad (1.6)$$

Using a typical weak scale cross-section in equation 1.6,

$$\langle \sigma_A \nu \rangle \sim \frac{\alpha^2}{m_{weak}^2} \sim 10^{-9} \text{GeV}^{-2} \quad (1.7)$$

We find that the relic WIMP density reduces to $\Omega_\chi h^2 = 0.1$ for a particle with a weak scale interaction. This is in good agreement with the expected cold dark matter component of the universe $\Omega_c h^2 = 0.12029$ [20].

1.2 WIMP Dark Matter Searches

There are three methods for detecting WIMPs other than looking for its gravitational effects. First, we can look for the annihilation of dark matter particles into standard model particles. Second, we can try to produce dark matter particles by colliding standard model particles in accelerators. Third, we can search for the rare

collisions of cosmic dark matter particles with ordinary matter. The third method is rather attractive as it involves directly observing a collision with a dark matter particle.




Figure 1.6: Methods for detecting WIMPs from interactions of (q) standard model particles and (χ) dark matter particles χ . WIMPs can be detected through the annihilation of dark matter particles into standard model particles, $\chi\chi \rightarrow qq$. WIMPs can be produced by colliding standard model particles, $qq \rightarrow \chi\chi$. Or one can look for the scatter of a dark matter particle off a standard model particle, $\chi q \rightarrow \chi q$.

1.2.1 Direct Detection of WIMPs

WIMPs could have masses in the GeV to TeV range and would comprise a quarter of the total mass of the universe. The local density of dark matter around the earth, at 8 kpc from the galactic center, is about 0.3 GeV/cm³, estimated from the galactic rotation curve of the Milky-Way with the assumption of a halo-like distribution (figure 1.7 and [1]). Assuming that the WIMP mass is on the order of the weak scale, 100 GeV, there are roughly three WIMPs per liter of space locally. The velocity of WIMPs near the Earth is similar to the orbital velocity of objects about the galactic center (240km/s at 8.3 kpc), figure 1.1. WIMPs, being highly non-relativistic, would scatter coherently on nuclei with a cross-section corresponding to $\sim A^2$.

WIMP scattering on nuclei can be expressed as a classical inelastic collision.

$$E_{\max} = r \cdot E_0 = \frac{r}{2} M_\chi v^2 \quad (1.8)$$

where E_{\max} is the most frequent energy deposit, E_0 is the average WIMP energy, M_χ is the WIMP mass and v is the WIMP velocity. The kinematic factor r for isotropic scattering off a target mass M_T in the laboratory frame is given by (using $\frac{1}{2}$ for the average of $1-\cos\theta$.):

$$r = \frac{4M_\chi M_T}{(M_\chi + M_T)^2} \quad (1.9)$$

Assuming classic billiard ball scattering off a xenon nucleus we calculate the expected energy deposits for various WIMP masses ranging from 1 to 1000 GeV/c², the




Figure 1.7: Dark matter density vs. distance from the galactic center, calculated from galactic rotation curves of the Milky Way galaxy. The Earth is located at 8.3 kpc. Figure from [1].

results are listed in table 1.2.

M_χ [GeV/c ²]	r	E_{\max} [keV]
1	0.032	0.01
10	0.28	0.90
100	0.99	31.7
10000	0.40	124

Table 1.2: The kinematic factor r and most common energy deposit E_{\max} for a WIMP of mass M_χ scattering off a xenon nucleus. Calculated using xenon as the target mass ($M_{Xe} = 122 \text{ GeV}/c^2$), and a WIMP velocity of 240 km/s ($8 \times 10^{-4} c$).

To calculate the WIMP-target scattering event rate we follow the derivation given by Lewin and Smith [3]. The differential rate of the WIMP nuclear recoils will be an exponentially decaying spectrum.

$$\frac{dR}{dE_R} = \frac{R_0}{E_0 r} e^{-E_R/E_0 r} \quad (1.10)$$

E_R is the recoil energy, E_0 is the most probable WIMP kinetic energy, r is the kinematic factor, R is the event rate per unit mass and R_0 is the total rate per unit mass. The event rate dR scattering off a target size A can be written as

$$dR = \frac{\mathcal{N}_0}{A} \sigma \nu dn \quad (1.11)$$

where \mathcal{N}_0 is Avagadro's number, A is the atomic mass, σ is the atomic cross-section, ν is the WIMP velocity and dn is the differential number density of WIMPs given by:

$$dn = \frac{n_o}{k} \mathcal{F}(\nu, \nu_E) d^3\nu \quad (1.12)$$

where $k = (\pi \nu_0^2)^{3/2}$ as $\nu_{esc} \rightarrow \infty$, an approximation good to within 0.5% for the Milky Way. n_o is the particle number density ($n_o = p_\chi/m_\chi$). The WIMP velocity distribution $\mathcal{F}(\nu, \nu_E)$ is assumed to be ideal gas described by a Maxwellian distribution:

$$\mathcal{F}(\nu, \nu_E) = e^{-(\nu + \nu_E)^2/\nu_0^2} \quad (1.13)$$

where ν is the WIMP velocity, ν_E is the earth velocity, ν_0 is the average velocity (about 230 km/s). We now rewrite equation 1.10 in terms of an integral over all possible velocities rather than energy and plug in the result for dR (1.11) and dn (1.12) leading to :

$$\frac{dR}{dE_R} = \frac{R_0}{E_0 r k} \frac{1}{2\pi\nu_0^2} \int_{\nu_{min}}^{\nu_{max}} \frac{1}{\nu} \mathcal{F}(\nu, \nu_E) d^3\nu \quad (1.14)$$

R_0 absorbs the constants $R_0 = \frac{2}{\pi^{1/2}} \frac{N}{A} \frac{\rho_\chi}{M_\chi} \sigma_T \nu_o$.

Having solved for the differential rate we now calculate the spin independent cross section for WIMPs scattering off nucleons of an atom (σ_T). We write the cross section as a sum off scattering off protons and neutrons in the nucleus. We use the fact that nucleon coupling for protons and neutrons is approximately equal [31].

$$\sigma_T = \frac{4\mu^2 A}{\pi} [Z \cdot f_p + (A - Z)f_n] \approx \frac{4\mu^2 A^2}{\pi} \sigma_n \quad (1.15)$$

where μ is the reduced mass of the WIMP nucleon system given by,

$$\mu = \frac{M_\chi M_n}{M_\chi + M_n} \quad (1.16)$$

Finally we must add the specific nuclear form factor for the specific target atom to account for decoherence, described by the Helm factor [4] $F(q)$. The cross section for spin independent scattering off the target nucleus can be written as a product of the idealized cross section and Helm factor:

$$\sigma_T(q) = \sigma_T F^2(q) = \frac{4A^2}{\pi} \left(\frac{M_\chi M_n}{M_\chi + M_n} \right)^2 \sigma_n F^2(q) \quad (1.17)$$

The spin-independent cross section for WIMPs is found to be proportional to the atomic number squared (A^2). The event rate per nuclear recoil energy is plotted for several potential WIMP search target nuclei in figure 1.8.




Figure 1.8: Plot of WIMP event rate per kg/day/keV vs. Nuclear recoil energy (keV) for several target nuclei, using parameters from [3]. Figure taken from [32].

At low detection threshold xenon is the most attractive target nuclei. The fall in the xenon curve is due to decoherence described by the Helm factor [4].

Xenon, being a relatively heavy element, ($A=131$), is an ideal candidate for a WIMP dark-matter search at low energy thresholds. Xenon detectors today have achieved thresholds as low as $3 \text{ keV}_{\text{nr}}$ [33]. Other common detection mediums are germanium, which is rather expensive on the ton scale, and argon which is inexpensive but contains a troublesome radioactive isotope (^{39}Ar). To probe dark matter cross sections the next generation experiments must be bigger and contain

less radioactive background contamination. With current limits on the WIMP cross section a ton-scale xenon experiment may only detect a handful of events per year.

1.2.2 WIMP Detection Experiments

Several experiments are currently conducting WIMP dark matter searches using a variety of targets, including xenon [33, 34, 35, 36], argon [37, 38], germanium [39, 40], fluorine [41, 42]. In the event that a WIMP strikes a target in the detector, it will primarily interact with the nucleus, deposit energy, and traverse throughout the detector without a second interaction. Neutrons could also interact with atomic nuclei and fake a WIMP signal, however after the initial energy deposit they are likely to interact again. Thus, neutrons can be rejected by removing multiple scatter events in the detectors. The most common source of backgrounds are electromagnetic in nature: gammas and betas from the rock surrounding the experiment, detector components, and internal to the xenon. As for the case of neutrons, the likelihood of a single scatter within the detector is small for a β or γ particle. Naked beta decays are the most troublesome, appearing as a single energy deposit in the detector medium. Fortunately, electronic recoil events can be discriminated from WIMP-like nuclear recoil events by more than 1/100 using the ratio of charge-to-light or charge-to-phonons produced in the interaction. Xenon based experiments currently having the best limits on WIMP nucleon cross sections [33].

1.3 Outline of Thesis

In this section we have reviewed the cosmological evidence for the existence of dark matter and have overviewed the WIMP model. The WIMP theory can be tested by observing their scattering off target nuclei.

In Chap. 2, the LUX detector, a liquid xenon time projection chamber (TPC), is described. We discuss how the light (S1) and charge (S2) signals from energy deposits are selected. The S1 and S2 signals are then used to reconstruct the energy, reject backgrounds and define the x,y,z coordinate of each event. We conclude the chapter with the most recent LUX science results which holds the worlds leading limit for spin independent WIMP nucleon scattering cross section.

In Chap. 3, the spacial dependent correction of the S1 and S2 signals are discussed. A $^{83\text{m}}\text{Kr}$ line source is injected periodically producing hundreds of thousands of events in the LUX detector. The calibration data is used to create the position dependent corrections which are then applied to the physics data.

In Chap. 4, the energy scale calibration of the LUX detector using line sources is discussed. The energy scale calibration is validated in the WIMP search region of interest (1-5 keV) by a reconstructing the beta spectrum of a tritium calibration source (1-18 keV).

In Chap. 5 recombination fluctuations inherent to liquid xenon are overviewed. We model detector resolution based on the PMT response to single photons and electrons along with light collection and charge extraction efficiency. We then proceed to extract recombination and statistical fluctuations are then extracted from

line source calibrations. The method is then adapted to the case of continuous energy spectra allowing for the extraction of recombination fluctuations from the tritium data. Having extracted the mean recombination fraction and its variance the electronic recoil band inherent to the liquid xenon at our applied electric field is known. The data extracted from the tritium calibration source can be used to model background discrimination in liquid xenon. We end with a discussion of what the recombination fluctuations imply about the underlying physical process that produces them, currently unknown for liquid xenon.

In Chap. 6 the results for ionization and scintillation yield measured using the tritium calibration source is presented. We discuss the spectral shape corrections used to undo the smearing of the true tritium spectrum caused by detector resolution. We compare our scintillation yield measurements (from 1-16 keV) with recent Compton scattering results.

In Chap. 7 we overview the development and deployment of the tritiated-methane calibration source. The source was successfully injected, used to calibrate and removed from the LUX detector producing high purity calibration data in the fiducial volume (140,000 tritium and 3 ± 2 background events). The tritium calibration data was used to determine the electronic recoil band in terms of the observables S1 and S2 and define the background rejection for the WIMP search.

In Chap. 8 future work is discussed in terms of detector calibrations, understanding event level fundamental fluctuations, and tighter constraints on the WIMP limit. Using the method for implementing the tritium source we can now consider calibration with other radioactive isotopes.

Chapter 2: The LUX Detector

2.1 Introduction

The LUX experiment is located 4850 ft underground (4300 m w.e.) at the Sanford Underground Research Facility in Lead, South Dakota. After running for 85.3 live days in 2013, LUX has set the most sensitive limit for a spin-independent WIMP scattering cross section [33] and is expected to achieve five times the sensitivity after a 300 day run ending in 2015.

Nobel elements are promising candidates for WIMP detection. They are easy to purify and are transparent to their own scintillation light. Xenon is especially favorable due to its large atomic mass (131.3 amu) and high liquid phase density (~ 2.9 kg/l) which provides both an excellent target for coherent WIMP scattering while simultaneously providing excellent stopping power from external radioactivity. Xenon is also free of any long lived isotopes which contribute backgrounds for the WIMP search. There are well established techniques to remove and monitor the residual, troublesome radio isotopes of ^{39}Ar and ^{85}Kr found in the atmosphere from which the xenon is distilled [43] [44] [45] [46].

WIMPs, being electrically neutral, primarily interact with the xenon target nuclei producing nuclear recoil (NR) events whereas typical backgrounds in the

detector, gammas and betas, interact with the atomic electrons producing electronic recoils (ER). In liquid xenon ER events can be further discriminated from NR events a factor of 100 or more by measuring the charge to light ratio of the interaction, as explained in section 2.2.2.

2.2 The LUX TPC



Figure 2.1: Photo of the outer vessel of the LUX detector from inside the water tank.

Figures 2.1 shows the LUX detector held in place by a stainless steel frame and the water tank that surrounds it. The water tank provides shielding from gammas and neutrons emanating from the surrounding rock, the water tank PMTs used as an active veto are also pictured along the sides of the water tank. Figure 2.2 show a cross sector of the LUX detector's inner and outer vessel. The LUX detector is a two phase xenon time projection chamber (TPC) [33]. The detector contains two PMT




Figure 2.2: Illustration of the LUX detector’s internals. The detector contains two arrays of PMTs on the top and bottom housing 61 PMTs each. Teflon panels on the edges of the active region are used to reflect scintillation signals. The vertical distance between the two PMT arrays is 60 cm, and the diameter to the outer edge of the teflon panels is 50 cm.

arrays on the top and bottom with 61 PMTs each for a total of 122 PMTs, with quantum efficiencies ranging from 30-40%. The active region consists of a 49 cm length between the cathode and gate grid with a 47 cm diameter of the dodecagonal geometry. The drift field between the cathode and gate is 170 V/cm resulting in

an electron drift velocity of 1.51 mm/ μ s. The liquid level terminates 5-6mm above the gate grid. The liquid level is precisely maintained by a weir reservoir into which xenon between the anode and gate spills into. The anode grid is 1.0 cm above the gate grid and creates an extraction field of 6 [kV/cm] where electrons are removed from the liquid and accelerated causing electroluminescence in the gas phase. LUX contains a gross mass of 370 kg of xenon of which 250 kg are in the active region.

2.2.1 Target Xenon

The target xenon is commercially available natural xenon with standard isotopic abundances (table 2.1), initially distilled to \sim 1 part per million (ppm) residual air contamination (N₂, O₂, Ar). The commercially available xenon also contained \sim 100 parts per billion (ppb) of krypton which is far greater than the background allowance of 5 parts per trillion (ppt). Krypton contains trace amounts of a beta emitter ⁸⁵Kr and is a troublesome internal background dissolved uniformly directly in the detection medium (xenon). The unwanted krypton was removed from the bulk xenon using a gas chromatography technique [45], with the removal independently verified before the science run with a xenon gas analysis technique developed for EXO-200 and LUX [44]. The xenon purity was monitored daily throughout the 2013 science run by an in situ gas analysis system that will be described in detail in section [Sampling]. Just one standard liter of air contains enough krypton to raise the concentration in the LUX xenon above the background goals. Daily krypton monitoring ensured that the krypton content in the xenon remained constant at 4

parts per trillion over the 2013 science run[5].

Electronegative impurities such as N₂, O₂ and H₂O attenuate electrons drifting through the xenon and must be removed in order to properly reconstruct events originating deep in the detector. These impurities continuously emanate from detector components degrading the free electron attenuation length. The accumulation of electronegative impurities is countered by circulating the xenon at 26.5 SLPM through a heated zirconium getter. The gross mass of 370 kg has a turn over time of 1.65 days. The monitoring of several impurities using the in situ ahas analysis system is described in detail in section [Sampling]. Throughout the 2013 science the electron attenuation length was measured to be 75 cm to 150 cm, corresponding to 70% to 50% charge loss for events originating from the bottom of the active region.

Isotope	Natural Abundance (%)
¹²⁴ Xe	0.09
¹²⁶ Xe	0.09
¹²⁸ Xe	1.92
¹²⁹ Xe	26.44
¹³⁰ Xe	4.08
¹³¹ Xe	21.18
¹³² Xe	26.86
¹³⁴ Xe	10.44
¹³⁶ Xe	8.87

Table 2.1: Xenon isotopic abundances, from [47]

2.2.2 Background Rejection

Xenon has a density of ~ 2.9 g/cm giving it excellent stopping power for shielding against external radioactivity. External radioactivity with energies in the WIMP

search region of interest (below 50 keV) only penetrate millimeters in xenon, being completely absorbed in the outer edges of the detector. Gammas in the MeV range have mean free paths on the order of 3 cm and are likely to be cut by the single scattering requirement . The most troublesome of the gamma backgrounds are from PMT materials containing trace amounts of uranium and thorium located inside the TPC. Figure figure 2.3 shows a simulation of expected gamma background events inside the LUX detector, using the single scatter cut requirement. By cutting out the edge events we can gain at least an additional factor of 1000 background rejection within the fiducial volume (inside the black dashed lines) [5].




Figure 2.3: Simulation of expected gamma background events inside the LUX detector, using the single scatter cut requirement[5]. There is an additional factor of 1000 background rejection within the fiducial volume where the WIMP search is conducted (inside the black dashed lines) .

Another means for discriminating background events is through measuring the charge to light ratio of each event. WIMPs will produce nuclear recoils whereas gammas and betas interact primarily with atomic electrons, resulting in different charge to light ratios of a given energy deposit. Using AmBe and ^{252}Cf neutron sources along with a tritium calibration source the NR to ER discrimination factor was measured to be 99.6 ± 1 % at 50% NR acceptance. Meaning that only one in 250 of the residual gamma and beta background events is expected to fake a WIMP signal when cutting out half of the potential nuclear recoil candidates. The ER type and NR type bands are shown in figure 2.4 with the band means as solid lines (Blue and Red, respectively) and the 10-90% CL as dashed lines. The internal tritium calibration source will be discussed in further details in section (Tritium Section).




Figure 2.4: The ER type and NR type bands from a) AmBe and ^{252}Cf (Red) and b) tritium calibration (Blue). The band means are solid lines and the 10-90% CL are shown as dashed lines. The ER to NR discrimination by using the charge to light ratio was measured to $99.6 \pm 0.1\%$ at 50% NR acceptance.

2.2.3 The Drift Field inside the LUX TPC

The LUX TPC contains five wire grids used to control the electric fields inside the detector. The grids and their spacings are shown in figure 2.5, labeled from top to bottom as Top (T), Anode (A), Gate (G), Cathode (C), Bottom (B). The grids T,A,G,C,B are biassed at -1,+3.5,-1.5,-10,-2 (all in kV), respectively. The PMTs are biased to -1.2 kV on average. The field created in the active region between the cathode and gate is also shown in figure 2.5. On average the drift field is 170 V/cm with variation from 140 V/cm to 200 V/cm from cathode to gate. The extraction

region between the anode and gate has a 6 kV/cm field. This extraction field is used to create the secondary scintillation (S2) signal via electroluminescence as the electrons are extracted and accelerated. The top and bottom grids serve to shield the PMTs from the anode and cathode biases.




Figure 2.5: Field grids in the LUX detector during the 2013 science run. T,A,G,C,B are biassed to -1,+3.5,-1.5,-10,-2 (all in kV), respectively. The PMTs are biassed to -1.2 kV on average. The figure on the right shows the electric field model in the drift region between the cathode and gate for drift distance z vs. detector radius r . On average the drift field is 170 V/cm with variation from 140 V/cm to 200 V/cm from cathode to gate. Electric field model from [6].

2.3 Light and Charge Signals in Liquid Xenon

When energy is deposited in the active region of the xenon TPC it is converted to ionization, excitation and heat.

$$\begin{aligned} E &= W(n_i + n_{ex}) + \text{Heat} \\ E &= W(n_\gamma + n_e) + \text{Heat} \end{aligned} \quad (2.1)$$

where E is the energy of the deposition in keV, n_q is number of quanta (photons + electrons), n_i , n_{ex} , n_γ and n_e are the number of ions, excitons, photons and electrons respectively. The work function (W) for xenon has been measured to be 13.7 ± 0.2 [eV/quanta] [8]. The number of photons observed for a given energy deposit arise from the excitons that de-excite and from ions which recombine with freed electrons. The number of electrons corresponding to a given energy deposit is equal to the number of ions that did not recombine with a freed electron.

$$\begin{aligned} n_\gamma &= n_{ex} + n_i r = n_i(r + \alpha) \\ n_e &= n_i(1 - r) \end{aligned} \quad (2.2)$$

Where r represents the electron-ion recombination probability and α represents the ratio n_{ex}/n_{ion} . The model given in equation 2.2 states that for each additional photon produced from recombination a corresponding electron is lost, and visa versa. The value of α for an ER event is approximately 0.06 and is expected to be independent

of energy [48] [49] [8]. For nuclear recoils α is approximately 1 [8]. The light and charge production in liquid xenon will be discussed in further detail below. Some useful properties of xenon are listed in table listed in table 2.2.

Parameter	Value	Ref.
Scintillation wavelength	174-178 nm	[50]
W (work function)	13.7 ± 0.2 [eV/quanta]	[8]
Xe ₂ [*] singlet lifetime	3.1 ± 0.7 ns	[51] [52] [53]
Xe ₂ [*] triplet lifetime	24 ± 1 ns	[51] [52] [53]
Recombination time	7.5 ns **	[54] [53]
Liquid density at boiling point	2.95 g/l	[55]

Table 2.2: Properties of xenon. ** The expected recombination time in the LUX detector. Recombination time ranges from 0 to 46 ns depending on electric field, energy deposit, and interaction type [53] [54].

2.3.1 Electronic Recoils (ER)

For an electronic recoil event the energy lost to heat is only about 5% [56] thus, equation 2.1 is valid for use with electronic recoils, we simply drop the small loss to heat. A schematic of an ER event is shown in figures 2.6 which we will overview. When an incoming beta or gamma interacts with the electron of the xenon atom the energy deposited is converted primarily to ionization, with roughly 6% excitation and $\sim 5\%$ is lost to heat [49] [56]. Excitons arise from ionized xenon atoms that bond together forming diatomic molecules (Xe₂^{*}). Xenon excitons will de-excite with characteristic time constants of 2.2 and 27 ns for the singlet and triplet state, respectively, producing ~ 175 nm scintillation light. Ion-electron pairs produced via ionization can also recombine, with probability r , producing additional excitons resulting in the production of additional ~ 175 nm scintillation light. The characteristic




Figure 2.6: An electronic recoil (ER) event in xenon. The energy deposited is converted primarily to ionization and roughly one tenth for excitation. Only several percent is lost to heat. Xenon excitons and recombining electron ion pairs for xenon dimers which de-excite producing very ultra violet (VUV) scintillation light at 175nm producing the primary scintillation signal (S1). Electrons that do not recombine are drifted by an electric field into the gas phase where they are accelerated producing the secondary scintillation (S2) signal.

recombination time constant is 45 ns [52]. Each initial exciton or recombining ion produces one scintillation photon, as written in equation 2.2. The two paths for photon production overlap in time and sum to produce the primary scintillation signal (S1). The S1 signal is collected by the PMT arrays within 500 ns. Electrons that escape recombination , with probability $1-r$, begin to drift upwards under the influence of the electric field between the cathode and gate (shown in 2.5). The electrons eventually reach the liquid-gas interface where they are extracted into the gas. As they accelerate, the extracted electrons produce a larger secondary scintillation signal (S2) that is proportional the the number of electrons extracted. The drift times for the electrons in the 49 cm long active region range from 1 to 324 μ s with an average drift velocity of 1.51 mm/ μ s. Thus, the S2 signal is well separated from the S1.

2.3.2 Nuclear Recoils (NR)

For nuclear recoils the energy lost to heat is more than half the total energy deposition [56]. This energy is lost through elastic collisions with other xenon atoms that fall below the ionization threshold. The energy lost to heat is characterized by an energy dependent Lindhard factor (\mathcal{L}) [57], written as:

$$E = \mathcal{L}^{-1} W(n_\gamma + n_e) \quad (2.3)$$

A schematic of a NR event is shown in figure 2.7. The signal production follows the same process as described above for an ER event but with greater amount of energy going towards heat and exciton production.




Figure 2.7: A nuclear recoil (NR) event in xenon. The energy deposited goes mainly towards heat (phonons), the remaining energy is split evenly between ionization and excitation. Xenon excitons and recombining electron ion pairs for xenon dimers which de-excite producing very ultra violet (VUV) scintillation light at 175nm producing the primary scintillation signal (S1). Electrons that do not recombine are drifted by an electric field into the gas phase where they are accelerated producing the secondary scintillation (S2) signal.

The additional energy lost to heat leaves less energy available for excitation and

ionization for an NR event. Further, NR events produce roughly equal amounts of ionization and excitation whereas ER events produce mostly ionization [56] [8]. Relative to an ER event, a NR event will have less electron ion pairs leading to a reduction of the S2 signal and enhancement of the S1. Thus, the ratio of S2 to S1 for a NR event is quenched compared to an ER event with an equivalent energy deposition. The quenching of the charge to light ratio is what leads to additional discrimination between nuclear and electronic recoil demonstrated in figure 2.4.

2.3.3 Energy and Position Reconstruction

In order to reconstruct the true energy of an event we will need to know its nature, ER or NR. For ER events we work in units of electronic equivalent energy, (keV_{ee}), using equation 2.1 neglecting the small heat loss. ER calibrations will be discussed in greater detail in section [ER calibration]. For NR events the energy is reconstructed in terms of nuclear recoil equivalent energy (keV_{nr}), using equation 2.3 with the Lindhard factor measured from calibration data given in [58] [59]. An illustration of an energy deposition in the LUX detector is shown in figure 2.8. The time difference between the S1 and S2 pulse defines the drift time, the drift time gives the measure of the z coordinate (depth). The hit pattern of the S2 signal on the top PMT array measures the x,y coordinates of the event. From the S1 and S2 signals the full x,y,z position and energy deposit of the event can be reconstructed.




Figure 2.8: Event diagram.

2.4 Identifying S1, S2

The primary and secondary scintillation signals can be identified by their unique properties. The S1 signal has a fast rise time and decays on the order of 10s of nanoseconds as the dimers of xenon produced through excitation and recombination de-excite (time constants listed in table 2.2). The S2 signal arrives several μs later with the electron population spread out spatially about its centroid due to diffusion, transverse and radial [60]. The characteristic S2 signal is thus one with a slow rise and corresponding slow fall. It resembles a bell curve, as the diffused electron

population arrives, peaking at the centroid of the distribution. A 2 keV event as seen by all 122 PMT channels is showing in figure 2.9. The S1 pulse is fast and the S2 pulse is much larger with a slower rise-time. The S2 pulse is larger since a single electron creates hundred of photons as it is accelerated in the extraction region.




Figure 2.9: 2keV ER event as seen by each PMT channel of the LUX detector. The S1 signal summed across all channels is overlaid on the top left, and the S2 signal summed across all channels is overlaid on the top right.

To identify S1 and S2 populations we define the variable Prompt Fraction as the area covered in the first 10% of the pulse normalized to the total area. The calculation is performed on the summed waveform after a first pass which defines the pulse's start and end timestamp. The separation of population density when plotting the total Pulse Area (measured in detected photo electrons [PE]) vs. Prompt Fraction is shown in figure 2.10, for the case of a ^{83m}Kr data set (41.5 keV)

and a tritium calibration data set (1-18.5 keV). The population of single electrons, single photons and the S1 S2 pairs associated with γ , β and α interactions are well separated and are highlighted as rectangles. The upper left corner is the single photon population, their areas consist of approximately 1 PE, the definition of the PMTs response to a detected photon. The single electron population is labeled SE and peaks at roughly 20 PE with a prompt fraction of -1. For all S1 pulses the prompt fraction is found to be between 1 and 0.3 (\log_{10} of 0 and -0.5) for the entire range of Pulse Area. The pulse area is a proxy for energy deposit, spanning from 1 keV tritium events to 7 MeV alphas. The populations of S2 from ^{83m}Kr , tritium, γ , β and α are found to have Prompt Fractions more than an order of magnitude smaller than their corresponding S1.

The S1 and S2 signal corresponding to an event are identified using a prompt fraction selection that had been tuned to calibration data. Valid S1 and S2 signals that spill into the single electron and single photon region at low energies can be identified by requiring the pulses be paired. For the WIMP search we define ‘golden’ events consisting of single scatters with a single S1 paired with a subsequent S2 pulse with a timing separation not exceeding the maximum drift time inside the TPC. With the golden requirement each event has a well defined x,y coordinate and z making it possible to correct the signals for geometry and electron attenuation.




Figure 2.10: Density plot of prompt fraction vs. Pulse Area. Top: $^{83\text{m}}\text{Kr}$ data set.

Bottom: Tritium data set. Populations of single electrons, single photons and the

S1 S2 pairs associated with γ , β and α are highlighted as rectangles.

2.5 LUX Science Result (WIMP limit)

The first science run of the LUX detector consisted of 85.3 live days from April 21, 2013 to Aug 8, 2013. A total of 83,673,413 triggers were recorded with 160 remaining as golden after applying quality cuts, listed in table 2.3.

Cut	Events Remaining
all triggers	83, 673, 413
detector stability	82, 918, 902
single scatter	6, 585, 686
S1 energy (2 - 30 phe)	26, 824
S2 energy (200 - 3300 phe)	20, 989
single electron background	19, 796
fiducial volume	160

Table 2.3: Data quality cuts used for the WIMP search results presented in [33].

Detector stability cuts remove the live time in which liquid level, gas pressure or grid voltages were out of normal ranges. The single scatter cut requires a single S1 with a subsequent S2 within a time window of $324 \mu\text{s}$, the maximum time required for electrons to traverse the active region. An area cut was also placed on both the S1 and S2 in order to narrow the energy region of interest. The minimum S2 requirement of 200 PE ensures the quality of the x,y position reconstruction, with ~ 8 extracted electrons. An additional cut was placed around time windows with anomalously high single electron rates. All single scatter WIMP search events before applying the fiducial cut are shown in figure 2.11, the vast majority of events occurring at the edges of the detector. The fiducial cut reduces residual radioactivity from the detector surface and PMTs by another two orders of magnitude. The fiducial cut consists of a radial cut at radius less than 18 cm from the detector

center. The z coordinate in drift time is defined to be 0 μs at the liquid surface and 324 μs at the cathode. The fiducial cut in z required that event drift times be between 38 and 305 μs , corresponding to 6 to 46 cm below the liquid surface (drift velocity = 1.51 mm/ μs). The fiducial cut is shown as the dashed cyan line in figure 2.11.




Figure 2.11: All single scatter events seen in the active region of the LUX detector over the course of the first science run passing all cuts listed in table 2.3 excluding the fiducial cut. The dashed cyan box indicates the fiducial volume.

Within the fiducial volume 160 events remain which meet our WIMP search energy requirement. The energy cut is placed in terms of S1 from 2-30 PE. As explained in chapter [Energy Calibrations] this corresponds to roughly 1.0 to 6 keV_{ee} or 3 to 25 keV_{nr}. We choose to select events based on S1 because it is directly observed, whereas true energy depends on the nature of the event (ER or NR) and must be inferred. The ER and NR discrimination band was measured using

calibration data and shown in figure 2.12. The blue and red bands represent the 10% to 90% confidence bounds of events being ER and NR type, respectively. The ER band was measured using a tritium calibration source (β^-) and the NR band was measured with neutrons from AmBe and ^{252}Cf along with NEST simulations [59]. The ER/NR discrimination at 50% NR acceptance was measured to be 99.6 ± 0.1 %. This value serves as a proxy for background events rejection, which is ultimately treated with a profile likelihood method on an event by event basis. Both the S1 and S2 signal have been corrected for spacial dependance which is discussed in further detail in Chapter . Due to cross talk or shorts, two PMTs on the top array and one on the bottom were left unbiassed. In order to avoid misreconstructing events extracted around the unbiased top PMTs only the bottom PMT array was used for the S2 signal ($S2_b$).




Figure 2.12: The remaining events passing the quality cuts listed in table 2.3. The charge to light ratio (S_2/S_1) is plotted vs. S_1 (proportional to energy) to show the separation of ER and NR type events. The 10-90% CF limits of the ER and NR band are plotted as the dashed blue and red curves, respectively. The band means are solid. S_2 _b stands for the S_2 signal on the bottom PMT array.

All 160 remaining events in the fiducial volume are consistent with being ER type events. The main source of BG events include residual ^{85}Kr , activated ^{127}Xe and ^{214}Pb from ^{222}Rn , described in further detail in [5]. The residual ER background rate in the WIMP region of interest as found to be 3.6 ± 0.3 mDRU ($10^{?3}$ cnts/keVee/kg/day) with an expectation of $2.6 \pm 0.2_{\text{stat}} \pm 0.4_{\text{sys}}$ mDRU.

A profile likelihood test is conducted on all WIMP search candidates remaining after the cuts listed in table 2.3. Testing the charge to light ratio (S_2 _b/ S_1) vs. energy (S_1) for likelihood of being an expected ER background. The signal model for the

test is derived from AmBe and ^{252}Cf neutron calibrations. The background rates input into the profile likelihood were independently measured and modeled with LUXSIM using NEST, described in further detail in [5] [33] [59]. The WIMP signal model was generated using an isothermal halo with a Maxwellian distribution, with a local WIMP density of 0.3 GeV/cm³ (as discussed in section 1.2). The galactic escape velocity input into the model is 544 km/s (cutting off the high end of WIMP velocity distribution), with an average WIMP velocity of 220 km/s. The earth's seasonal velocity being 245 km/s with respect to the galactic center. The result from the 2013 science run with the LUX detector is consistent with a P value of 0.35 for the background only hypothesis. The 90% upper C.L. cross section for various spin independent masses are shown in figure 2.13. The minimum cross section reported occurs at 7.6×10^{-46} cm² for a WIMP mass of 33 GeV/c² [33]. The LUX result is a factor of two improvement in WIMP cross section sensitivity over the Xenon100 limit reported in 2012 [34] and is in tension with reported WIMP signal claims from CoGent [40], CDMSLite (Silicone) [61], CRESST II [62] and DAMA/LIBRA [63].




Figure 2.13: LUX detector is consistent with a P value of 0.35 of the background only hypothesis. The 90% upper C.L. cross section for various spin independent masses are shown in figure 2.13 in blue. Also shows are limits from Xenon100 (red), CDMS II (green), ZEPLIN III (magenta), and one sigma signal claimed for DAMA/LIBRA (shaded grey), CDMS II Silicone (shaded green), CRESST II (shaded yellow).

As mentioned above the ER band is measured with a novel tritium calibration source and will be discussed in further detail in Chapter [TritiumSource]. Since the initial science run we have gathered twenty times the tritium statistics to further study the ER band, with 150,000 events in the fiducial volume. We also spent several months deploying a DD neutron generated source to further study the NR band mean. The results from the improved ER and NR calibrations will be reported

in this thesis along with the implications for an improved WIMP limit. An updated profile likelihood reanalysis of the LUX WIMP search result will be submitted for publication in late Fall of 2014.

Chapter 3: Spacial Dependent Correction of the S1 and S2 Signals

In this chapter we address the spacial dependent corrections applied to the S1 and S2 signal. During the 2013 science run $^{83\text{m}}\text{Kr}$ injections were performed periodically and used to calculate the position dependent corrections that were applied to the WIMP search data. Knowing the x,y,z position of each event provides a powerful tool for reducing variations in the signals due to detector geometry. The better we can correct for position dependance of the S1 and S2 signals the smaller the observed variations will be, leading to better ER, NR discrimination and ultimately better signal to background for the WIMP search.

After the energy deposit occurs in the active region of the LUX detector the freed electrons are drifted via electric field towards the liquid surface. The S2 light is emitted as the electrons are extracted at a given x,y position from the liquid surface and accelerated by a 6 kV potential, traversing 5mm from the liquid surface to the anode. The dominant effect to correct the data for is the free electron lifetime. As charge is drifted from the event site to the extraction region (0-47 cm) electronegative impurities in the liquid latch onto them. S2s of equal sizes are exponentially attenuated with increasing depth in the detector by impurities such as O₂, H₂O, N₂ in the xenon. (7 ppb O₂ corresponds to roughly 100 μs lifetime [64]). The S2 signal

has x,y dependent variations due to non uniformities in the extraction field, tilt in the liquid level, and non uniformities in the anode-gate wire separation (potential wire grids sagging).

The S1 light propagates isotropically from the interaction site and has about a 30% variation in light collection efficiency between events near the top and bottom of the detector due to geometric effects. About 2/3 of the S1 light is collected on the bottom PMT arrays due to total internal reflection at the liquid gas interface. The closer the event to the bottom PMTs the larger the solid angle is into the bottom PMT, increasing the probability of the detecting a photon and producing a photo electron (PE). Other position-dependent effects include the photon absorption length which is negligible at the purities achieved in LUX, and teflon reflectivity which is > 90% in liquid xenon [65] [66]. The S1 position dependent correction used in the LUX analysis normalizes the photon detection probability of all events to the center of the active region, an arbitrary choice corresponding roughly to the average light response.

It should also be noted that variation in light yield and charge yield due to the non uniformity of the electric field are also folded into x,y,z deponent correction. The effect of light yield and charge yield for ^{83m}Kr due to dependance of recombination on the local electric field is on the order of 10% [59]. The field increases from about 140 to 200 V/cm from the cathode to gate in the LUX detector, the electric field model is shown in 2. Also, all S1 and S2 signals are measured as photo electron (PE) pulse areas that have been calibrated by pulsing LEDs (450 ns) located inside the TPC. Any quantum efficiency (QE) or gain variations in the 122 PMTs which

are not properly normalized by the individual gain corrections are also folded into the position dependent corrections measured from the ^{83m}Kr calibrations.

3.1 ^{83m}Kr Calibration

Throughout the science run periodic ^{83m}Kr injections were performed to calculate position dependent corrections. ^{83m}Kr is produced from the decay ^{83}Rb with a half life of 86.2 days. The ^{83}Rb source used is housed in charcoal and plumbed directly into the LUX circulation system. A decay scheme of ^{83}Rb and ^{83m}Kr is shown in figure 3.1.






Figure 3.1: A simplified decay diagram of ^{83}Rb and ^{83m}Kr , from [7].

The daughter ^{83m}Kr is continually produced in the charcoal housing, having a half-life 1.8 hours. ^{83m}Kr decays via an electron capture first emitting a 32.1 keV x-ray followed by a 9.4 keV x-ray with a half life of 154 ns between the two [67] [9]. For the vast majority of the decays the combined S1 pulse corresponding to 41.55 keV is observed, since the minimum S1 pulse separation in the LUX reconstruction is

1000 ns. Figure 3.2 shows the waveforms (S1 and S2) for two ^{83m}Kr events with the the decay of the 32.1 and 9.4 keV x-ray split by 60 and 220 ns, both S1s shown are classified as a single S1 event. The timing separation between the S1 and S2 is used to infer the drift distance to be 14.0 and 40.0 cm below the liquid surface, respectively. The reduction of the S2 signal that originated from deeper in the detector is apparent by comparing the amplitudes of the S2, with roughly 50% charge loss. The S2 pulses are insensitive to the timing separation of the dual decay as electron diffusion smears the pulses two together as the electrons drift through the active region before extraction [60]. The PMT hit map for the events is shown in figure 3.3.




(a) $^{83\text{m}}\text{Kr}$ event with 60 ns timing separation.




(b) $^{83\text{m}}\text{Kr}$ event with 220 ns timing separation.

Figure 3.2: The S1 and S2 of two $^{83\text{m}}\text{Kr}$ events. Top Figures (a): Show a $^{83\text{m}}\text{Kr}$ event in which the two x-rays have overlapped within 60 ns, the S2 arrives about 93 μs later. Bottom Figures (b): Show a $^{83\text{m}}\text{Kr}$ event in which the two x-rays have overlapped within 220 ns, the S2 arrives about 205 μs later. The LUX pulse finder classifies events within a 1 μs window as a single S1. The S2 pulses are insensitive to the timing separation of the dual decay as electron diffusion smears the pulses two together as the electrons drift. The PMT hit map for these events are shown below in figure 3.3.



(a) $^{83\text{m}}\text{Kr}$ event with 60 ns timing separation.



(b) $^{83\text{m}}\text{Kr}$ event with 220 ns timing separation.

Figure 3.3: The S1 and S2 of two $^{83\text{m}}\text{Kr}$ events. Top Figures (a): Show a $^{83\text{m}}\text{Kr}$ event in which the two x-rays have overlapped within 60 ns. Bottom Figures (b): Show a $^{83\text{m}}\text{Kr}$ event in which the two x-rays have overlapped within 220 ns. The black, open circle represents the location of the event. The S1 hit pattern is diffuse with more light collected on the bottom arrays due to total internal reflection at the liquid surface. The S2 is localized in the top PMT arrays in x,y at the location where the electrons are extracted, and diffuse on the bottom due to scattering. The summed waveforms for these events are shown above in figure 3.2.

3.2 ^{83m}Kr Mixing in Liquid Xenon

The ^{83m}Kr source is introduced when needed into the LUX detector by flushing the charcoal housing with xenon and diverting the flow inline with the main circulation path. The ^{83m}Kr source and delivery into the xenon detector is described in more detail in [9]. The relatively short half-life of 1.8 hours allows for several injections per week without interrupting WIMP search data taking. Once injected the source is uniformly mixed into the liquid xenon within a matter of minutes and can be used to calculate corrections for the XYZ the response of the detector. ^{83m}Kr has proved useful for calibrating liquid xenon detectors [9] [10] [13]. The solubility of krypton in liquid xenon will be discussed further in Chapter[Sampling]. Figure 3.4 shows the uniform distribution of ^{83m}Kr events in the LUX detector thirty minutes after the injection. Once uniformly mixed, the decay of ^{83m}Kr produces a well defined mono energetic peak in the detector (the S1, S2, and energy spectra are shown in section 4). Measuring the spectral peak vs. x,y,z allows for the detector's spacial dependent response to be gauged over the course of the science run.

3.3 S2 Electron Lifetime and x,y Correction

As mentioned previously, the S1 and S2 signals collected on the top and bottom PMT arrays have well defined x,y,z positions computed from the S2 hit pattern (x,y) and the signal separation in time (z). Throughout the 2013 science run the electron lifetime was measured between 500 to 1000 μs , or an attenuation length of




Figure 3.4: Distribution of $^{83\text{m}}\text{Kr}$ events 10 minutes after the injection. The source mixes uniformly throughout the liquid xenon illuminating all regions of the active volume. The solid black lines represent the fiducial volume used for the WIMP search.

75 to 150 cm with a drift velocity of $1.51\text{mm}/\mu\text{s}$. The measured electron lifetimes (or attenuation lengths) correspond to between 70% and 50% reduction in S2 signal from the cathode grid, the bottom of the active region.

The electron lifetime is calculated by binning the detector in drift time into 60 slices. In each bin a Gaussian is fit to extract the mean. Once the mean in each slice is calculated, an exponential is fit to the mean S2 response vs. z. The exponential time constant the measure of the characteristic drift time $\tau [\mu\text{s}]$, shown in figure 3.5. The characteristic attenuation length is $\lambda = \tau v_{drift}$, where v_{drift} is the electron drift velocity. For this analysis we use the S2 response of the bottom PMT array ($S2_b$) as it was used for the 2013 WIMP search analysis discussed in 2.

The z corrected $S2_b$ from each signal is calculated as follows:

$$S2_{b-z} = S2_b \cdot \exp\left(\frac{\text{drift time}[\mu\text{s}]}{\tau[\mu\text{s}]}\right) \quad (3.1)$$




Figure 3.5: Left: Fits to the mean of S2_b of the ^{83m}Kr data in several z slices. Right, the exponential fit to the means of S2_b vs. drift time used to extract electron lifetime τ . The electron lifetime is found to be $\tau = 753.1 \pm 3.2 \mu\text{s}$. The exponential fit to the means deviates near the top and bottom of the active region since the charge yield from the ^{83m}Kr decay is sensitive to the varying electric field. The data shown was taken on May 10, 2013 (lux10_20130510_T1250) and contains 700,000 ^{83m}Kr events.

Where $S2_{b-z}$ is the z corrected S2_b signal and τ is the free electron lifetime. After correcting the dominant z dependent electron attenuation, corrected to 0 drift time, we calculate the normalization factor ($\mathcal{N}\mathcal{F}$) that will be used to correct for the x,y dependent variations in the S2 signal.

The normalization is calculated by creating a 25 x 25 grid on the x-y plane, corresponding to 2 cm x 2 cm x,y bins. For each bin the average S2_b light response is determined by fitting a Gaussian. Figure 3.6 (left) shows the measured S2_b response to 700,000 ^{83m}Kr decays normalized to the response at the center, x=y=0. This map represents the inverse of the normalization factor that we call $\mathcal{N}\mathcal{F}(x, y)$. $\mathcal{N}\mathcal{F}$ is then applied to the S2_b data by using a spline interpolation of the x,y coordinate of each event relative to the bin centers $\mathcal{N}\mathcal{F}(x, y)$. Figure 3.6 (right) shows the S2_b response after correcting the data relative to the center x=y=0 using $\mathcal{N}\mathcal{F}(x, y)$.

After applying the x,y correction the variation decreases from 10% to 1% in the inner 18 cm radius of the detector (the fiducial volume).




Figure 3.6: Left: Response of $S2_b$ vs. x, y normalized to the response at the center ($x=y=0$). The region of larger response around $x=0, y=-25$ is likely from an enhanced extraction field between the anode and gate wires. Right: Response of $S2_b$ vs. x, y after correcting the data using $\mathcal{N}\mathcal{F}_{S2_b}$.

After correcting for z and x, y we can define the position-dependent (x, y, z) corrected $S2_b$ signal, which we will call $S2_{bc}$ calculated as follows:

$$S2_{bc} = S2_{b-z} \cdot \mathcal{N}\mathcal{F}_{S2_b}(x, y) \quad (3.2)$$

where $S2_{bc}$ is the x, y, z corrected $S2_b$ signal and $\mathcal{N}\mathcal{F}_{S2_b}(x, y)$ is the Normalization Factor of the bottom PMT array for $S2_b$ s and is a function of x, y . The interpolation of the inverse of $\mathcal{N}\mathcal{F}_{S2_b}(x, y)$ along the x, y grid is plotted in figure 3.6.

Figure 3.7 shows the improvement in the $S2_b$ signal after applying the z and x, y correction. After applying the z correction to $S2_b$ there is a fractional improvement in resolution of 18%, for the case of an 750 μs electron lifetime. The correcting

in the x,y plane provides an additional 4.9% improvement in resolution to the S2_b signal.




Figure 3.7: Improvement of resolution in S2_b after applying the z and x,y,z correction. Black: The uncorrected data. Red: The data with only z dependent correction, the electron lifetime correction. Blue: The data with full x,y,z dependent correction. The data shown was taken data on May 10, 2013 (lux10_20130510_T1250) and contains 700,000 ^{83m}Kr events.

3.4 S1 Spacial Dependant Correction

To measure the x,y,z dependent Normalization Factor for S1 ($\mathcal{N}\mathcal{F}_{S1}$), we divide the detector into a 25 x 25 x 16 x,y,z mesh with each voxel having dimensions of 2 cm x 2 cm x 20 μ s. To achieve sufficient statistics for the correction we require at least 400,000 ^{83m}Kr events, about 40 events per voxel to define the mean. Monthly high stats calibrations are performed that yield about 1 million counts to providing precise $\mathcal{N}\mathcal{F}$ correction maps. Unlike the S2 correction, which is highly dependent on purity, the S1 has been found to be invariant to within a percent over the course

of the science run, thus the monthly calibrations with high statistics are sufficient to provide the position dependent correction.

Figure 3.10 shows the response of the detector to ^{83m}Kr normalized to the center of the detector in 16 slices of z , each with a $2\text{ cm} \times 2\text{ cm}$ x,y grid. The plotted maps are normalized to the center of the detector and represent the inverse of the normalization factor ($\mathcal{N}\mathcal{F}_{S1}$). We choose to normalize the center of the detector as it represents the average light collection efficiency of the detector. Though the dominant correction is the z -dependance, there is also substructure in x,y to each z slice which is illustrated in figure 3.9, where we have normalized each slice to its own center. It is evident that near the top and bottom there are additional geometric effects around the radial edges, whereas in the central z slices the uniformity in x,y is much better due to the diffusion of the light scattering on the teflon.




Figure 3.8: S1 x,y,z response normalized to the center of the detector. The interpolated map represents the inverse of the normalization factor $\mathcal{N}\mathcal{F}$.




Figure 3.9: S1 x,y,z response normalized to the center of each z slice. There are greater variations near the very top and bottom 4 z slices as the solid angle for light hitting the PMT arrays increases. For the central slices the x,y response is uniform due to increased scattering off the teflon panels.

We define the position dependent (x,y,z) corrected S1 signal as $S1_c$, normalized to the center of the detector ($x=y=0$ and $z=160\mu s$) calculated as follows:

$$S1_c = S1_{(x,z,y)} = S1 \cdot \mathcal{N}\mathcal{F}_{S1}(x, y, z) \quad (3.3)$$

where $S1_c$ is the x,y,z corrected S1 signal and $\mathcal{N}\mathcal{F}_{S1}(x, y, z)$ is the Normalization Factor of the sum of all PMTs for S1s and is a function of x,y,z . The interpolation of the inverse of $\mathcal{N}\mathcal{F}_{S1}(x, y, z)$ along the x,y grid in z slices is plotted in figure 3.10. The normalization factor is applied to the S1 data by using a spline interpolation of the x,y,z coordinate of each event relative to the bin centers $\mathcal{N}\mathcal{F}_{S1}(x, z, y)$. Figure 3.10 shows the S1 response after correcting the data relative to the center of the detector. After applying the x,y,z correction the position dependent variations decrease to less than 1% in the inner radial 18 cm of the detector (the fiducial volume), with as much as 3% variations near the top and bottom edges where the interpolation fails.

Figure 3.11 shows the improvement in the S1 signal after applying the z and x,y,z correction. With z -only correction to S1 there is a fractional improvement in resolution of 31.5%. The combined correction in z and the x,y plane provides an additional 2.0% improvement in resolution over the z only correction.




Figure 3.10: S1 x,y,z response normalized to the center of the detector after the data has been corrected and normalized to the detector center. The remaining variations in the fiducial volume ($r < 18\text{cm}$) is less than 1%. Near the top and bottom edges the deviation increases to as much as 3% due to the interpolation of the correction becoming poorly constrained.




Figure 3.11: Improvement of resolution in S1 after applying the z and x,y,z correction. Black: The uncorrected data. Red: The data with only z dependent correction. Blue: The data with full x,y,z dependent correction. The data shown was taken data on May 10, 2013 (lux10_20130510_T1250) and contains 700,000 ^{83m}Kr events.

3.5 Application to x,y,z Corrections in Data Processing

As mentioned earlier, the purpose of the periodic ^{83m}Kr calibrations is to measure the position-dependent S1 and S2 corrections over the course of the 2013 science run. Before processing the WIMP search data the calibration sets were processed and a MYSQL table of electron lifetimes and corrections maps were populated for each calibration date. The electron lifetime applied to each WIMP search data set was a linear interpolation between calibration dates. For the S2 x,y correction the nearest $\mathcal{N}\mathcal{F}_{S2}(x, y)$ entry in time was used. Combined these produced the corrected $S2_c$ quantity to be used for the WIMP analysis. For the S1 x,y,z corrections, the

nearest $\mathcal{NF}_{S1}(x, y, z)$ entry is used to produce the corrected $S1_c$ quantity via a spline interpolation in x, y, z . For both the $S2-x, y$ and $S1-x, y, z$ correction the time dependence is assumed to be negligible as these are geometric effects, whereas the electron lifetime varies with liquid purity each day. The electron lifetime and the stability of the $S1$ correction over the course of the 2013 science run is shown in figure 3.12 and 3.13. While the electron lifetime needs frequent monitoring, the $S1$ x, y, z response is fixed over several months of running. For the results discussed in the subsequent sections of the thesis we will only work with the x, y, z corrected $S1$ and $S2_b$ pulses ($S1_c$ and $S2_{bc}$ respectively) .




Figure 3.12: Electron lifetime measured using ^{83m}Kr calibrations during the LUX science run in 2013. The blue dashed lines show the boundaries of the WIMP search from April 21 to Aug 8, 2013. The red dashed lines indicate circulation loss events.




Figure 3.13: Measured response of to light from $^{83\text{m}}\text{Kr}$ calibrations at the center of the LUX detector during of the LUX science run in 2013. The blue dashed lines show the boundaries of the WIMP search from April 21 to Aug 8, 2013. The red dashed lines indicate circulation loss events.

^{83m}Kr Calibration Set	Number of ^{83m}Kr Events
lux10_YYYYMMDDThhmm	
lux10_20130320T1430	136,877
lux10_20130328T1437	444,622
lux10_20130405T1417	187,059
lux10_20130411T1524	99,724
lux10_20130425T1047	104,231
lux10_20130426T1019	92,024
lux10_20130429T1447	133,652
lux10_20130501T1508	91,465
lux10_20130503T1457	108,898
lux10_20130506T1328	45,678
lux10_20130510T1250	670,895
lux10_20130510T1607	499,743
lux10_20130510T2008	113,347
lux10_20130511T0014	44,372
lux10_20130517T1542	138,432
lux10_20130520T1504	216,709
lux10_20130524T1503	28,975
lux10_20130528T1546	11,569
lux10_20130531T1421	125,921
lux10_20130604T1421	110,219
lux10_20130607T1512	106,315
lux10_20130610T1518	116,349
lux10_20130617T1457	61,737
lux10_20130621T1533	78,707
lux10_20130626T1517	25,124
lux10_20130628T1444	44,134
lux10_20130701T1646	71,410
lux10_20130709T1009	106,230
lux10_20130712T1427	104,150
lux10_20130717T1424	66,801
lux10_20130720T1045	88,945
lux10_20130723T1452	88,626
lux10_20130726T1431	35,056
lux10_20130729T1004	59,906
lux10_20130807T1403	27,015
lux10_20130812T1546	113,560
lux10_20130823T0953	107,820
lux10_20130829T1005	479,676

Table 3.1: ^{83m}Kr sets used to calculating corrections for the 2013 LUX WIMP search [33].

Chapter 4: Calibrating the Combined Energy Scale

This section outlines the method used to calibrate the energy scale of the LUX detector for electronic recoils. The idea behind the method is to take calibration data with multiple sources and/or electric fields then combine the measured scintillation signals, primary (S1) and secondary (S2) in order to reconstruct energy. For a given energy deposit in liquid xenon an amount of quanta released is proportional to a work function W , for nuclear recoils we must also consider heat loss. The quanta created at the interaction site are the results of electron-ion pairs and excitons produced by the recoiling xenon nucleus that become photons and electrons, equation 4.1.

$$\begin{aligned} E &= W(n_i + n_{ex}) \\ E &= W(n_\gamma + n_e) \end{aligned} \tag{4.1}$$

where E is the energy of the deposition in keV, n_i , n_{ex} , n_γ and n_e are the number of ions, excitons, photons and electrons respectively. The work function (W) for xenon has been measured to be 13.7 ± 0.2 eV/quanta [8]. Excitons quickly de-excite and contribute to the primary scintillation signal (S1). Ions that recombine with their electron pairs produce scintillation light (S1), while those electrons that do not recombine are collected several μ s later in the extraction region as the larger

secondary scintillation signal S2. The interactions of NR and ER type events were discussed previously in 2.3.

There are two knobs to turn that tune the recombination fraction and probe combined energy space over a variety of S1 and S2, we can either change the energy of the source or adjust the drift field. The larger the spread in S1 and S2 the more constrained the combined energy scale will be. Measuring both light and charge allows for a vastly improved resolution compared with only using a single S1 or S2 only space, since recombination fluctuations cancel out if energy is reconstructed correctly. Using equation 4.1 and assuming that the heat loss is negligible for electronic recoils (ER) we can reconstruct energy by knowing the work function and the conversion from measured S1(light) and S2(charge) signals to the number of quanta ($n_\gamma + n_e$) liberated by the interaction. We define gain-1 (g_1) and gain-2 (g_2) as the conversion from the average number of photons and electrons propagated from the interaction site to the observed signal by the PMT arrays as a photo electron (PE), given in equation 5.2. Note for the value of S2 in this section we only use the signal on the bottom PMT array S_{2b} .

$$\begin{aligned}\langle n_\gamma \rangle &= \frac{\langle S1 \rangle}{g_1} \\ \langle n_e \rangle &= \frac{\langle S2 \rangle}{g_2}\end{aligned}\tag{4.2}$$

the combined energy of equation 4.1 can be written in terms of the observable S1 and S2 as

$$E = W \left(\frac{n_\gamma}{g_1} + \frac{n_e}{g_2} \right)\tag{4.3}$$

By using multiple line-sources with known energies we can extract a best fit

for the value of the gains (g_1, g_2) by making a Doke plot [68] [49]. The mono energetic lines used for the purposes of the calibration are listed in table 4.1. For each calibration point we calculate the mean light yield and charge yield and fit a line, S_1/E and S_2/E respectively, (Equation 4.4).

$$\begin{aligned} S_1/E &= \frac{n_\gamma}{(n_\gamma + n_e)} \times \frac{g_1}{W} \\ S_2/E &= \frac{n_e}{(n_\gamma + n_e)} \times \frac{g_2}{W} \end{aligned} \quad (4.4)$$

Fitting equation 4.4 to a line yields

$$\begin{aligned} \left(\frac{S_1}{E}\right) &= \left(\frac{g_1}{W}\right) - \left(\frac{S_2}{E}\right) \left(\frac{g_1}{g_2}\right) \\ y &= \frac{S_1}{E}, x = \frac{S_2}{E} \end{aligned} \quad (4.5)$$

$$y = m \cdot x + b$$

The x and y intercepts from Equation 4.5 can be used to solve for g_1 and g_2 .

$$\begin{aligned} g_1 &= b \cdot W \\ g_2 &= \frac{g_1}{m} = \frac{b \cdot W}{m} \end{aligned} \quad (4.6)$$

The values of g_1, g_2 are degenerate and highly correlated such that the ratio of $g_1:g_2$ is always a constant, a reduction in g_1 can be compensated by an increase in g_2 and still yield the same number of initial quanta and visa versa. Breaking the degeneracy requires data over a wide range of S_1 and S_2 values near the intercept of the Doke plot. Due to the strong correlation in the fit parameters the data is fit by minimizing the likelihood and the errors in intercept and slope are determined using MCMC (Markov Chain Monte Carlo).

Source	Energy [keV]	Decay Type
Xe K shell	29.7, 34	X-ray
^{83m}Kr	41.55**	Internal Conversion
^{131}Xe	163.9	Internal Conversion
^{127}Xe	203 or 375	^{127}I daughter γ -emission
	33.8	Kb shell X-ray
	5.3	L shell X-ray
^{129m}Xe	236.1	Internal Conversion
^{214}Bi	609	γ -emission
^{137}Cs	661.6	Photo-absorption

Table 4.1: Mono energetic peaks used for g1 g2 calibration. ** ^{83m}Kr data was taken at 50 and 100 [V/cm] along with the standard field of 170 [V/cm].

4.1 Anti-Correlation Space

The first step in calibrating the energy scale is to plot the observables S1 vs. S2, by doing this the anti correlation between light and charge at a given energy become apparent, figure 4.1. For the data presented here a fiducial cut was placed at a radius of less than 18 [cm] and drift distance between 6 and 46 [cm] which greatly reduces the background event rate. To extract g1 g2 we first determine the average values of S1 and S2 at each known energy. Initially loose diagonal cuts are placed by eye on the populations, figure 4.1. Next, using a un-binned maximum likelihood fit the mean and sigma are estimated and then refit using $\pm 2\sigma$ of the initial distribution to remove tails from backgrounds. With the initial estimate for the mean S1 and S2

response to a given energy the gains g_1, g_2 are determined. The resulting value of g_1 and g_2 is found to be 0.096 ± 0.009 and 5.94 ± 1.68 respectively, the fit is shown in figure 4.2. The values of g_1 and g_2 represent a best fit to the underlying recombination theory where for each additional photon there is a corresponding reduction of one electron and visa versa. The method for extracting the uncertainties using MCMC will be discussed later in section 4.3.




Figure 4.1: LUX data in anti correlation space (S_1 vs. S_2), the black lines indicate the initial cuts by eye used to isolate populations of constant energy. In both figures diagonals represent lines of constant energy with a slope depending on the local recombination probability. The centroids found by an unbind maximum likelihood analysis are shown as a black X, for sources shown in table 4.1.

4.2 Refitting in Combined Energy Space

From the first attempt to find g_1, g_2 in figure 4.2 there are discrepancies between the data and the fit, however this first result is only a crude estimate derived from anti correlation space. Once we have an initial estimate of gains g_1, g_2 a combined energy




Figure 4.2: Doke plot showing the best fit for the energy calibration parameters g_1 and g_2 using S_1 and S_2 means extracted from anti-correlation space. The first three blue points, from the left, are from ^{83}Kr calibrations at 50, 100 and 170 V/cm respectively. The open circle was from the K-shell xenon X-ray and was not used for the fit as it's absolute energy and origin from the skin of the detector is uncertain.

scale can be constructed with significantly improved resolution over the initial guess, due to the fact that recombination fluctuations are canceled. With the improved resolution the data are fit around the combined energy peaks using an unbinned maximum likelihood fit to a normal distribution, and then the data refitted around 1.5σ of the initial fit. The fits used to extract the means and sigmas of the S_1 and S_2 signals at a given energy are show in figures 4.4 and 4.5. We iterate this technique twice as the convergence is rapid, in this case the initial value of g_1 and g_2 derived from anti-correlation space are already a close approximation to the true value. The

resulting value of g_1 and g_2 is found to be 0.097 ± 0.008 and 5.75 ± 1.4 respectively, the fit is shown in figure 4.3. After refitting there is a significant improvement over figure 4.2, especially the xenon activation lines in the center, which is due to better peak finding in combined energy space over anti-correlation space.

$$\begin{aligned} g1 &= 0.097 \pm 0.008 \\ g2 &= 5.75 \pm 1.4 \end{aligned} \tag{4.7}$$




Figure 4.3: Doke plot showing the best fit for the energy calibration parameters g_1 and g_2 using S_1 and S_2 extracted from a combined energy space. The first three blue points, from the left, are from ^{83}Kr calibrations at 50, 100 and 170 V/cm respectively. The open circle was from the K-shell xenon X-ray and was not used for the fit as it's absolute energy and origin from the skin of the detector is uncertain.




Figure 4.4: S1 fits to sources at nominal field of 170 [V/cm] unless otherwise noted.

Source and energy in keV from top left to bottom right: a) ^{131}Xe : 163, b) ^{127}Xe : 207, c) $^{127}\text{Xe} \& ^{129\text{m}}\text{Xe}$: 236.8, d) ^{127}Xe : 410, e) ^{214}Bi : 609, f) ^{137}Cs : 661.6, g) $^{83\text{m}}\text{Kr}$: 41.5 - at 50 [V/cm], h) $^{83\text{m}}\text{Kr}$ 41.5 - at 100 [V/cm], i) $^{83\text{m}}\text{Kr}$ 41.5 .

Figure 4.6 is the final Doke plot for multiple peaks the theory describes the data well using the optimal fit for g1 and g2, for each increase in number of photons there is a corresponding decrease in the number of electrons and visa versa. The




Figure 4.5: S2 fits to sources at nominal field of 170 [V/cm] unless otherwise noted.

Source and energy in keV from top left to bottom right: a) ^{131}Xe : 163, b) ^{127}Xe : 207, c) ^{127}Xe & $^{129\text{m}}\text{Xe}$: 236.8, d) ^{127}Xe : 410, e) ^{214}Bi : 609, f) ^{137}Cs : 661.6, g) $^{83\text{m}}\text{Kr}$: 41.5 - at 50 [V/cm], h) $^{83\text{m}}\text{Kr}$ 41.5 - at 100 [V/cm], i) $^{83\text{m}}\text{Kr}$ 41.5 .

relatively large error on g2 is due to the distance of the data points from the x-intercept. As stated before the values of g1 and g2 can be locally degenerate as long as their ratio remains a constant. Thus for future studies it will be important to

probe more of the parameter space in order to place a tighter constraint on gains g1 and g2.




Figure 4.6: Doke plot of the data showing the light yield vs. charge yield. Only solid circles are used for the fit, the open circle is the Xe K shell X-ray from the detector edge.

4.3 Finding Errors with MCMC

The error bars reported in this section on g1 and g2 are from the error in the slope and intercept of the linear fit in the Doke plot derived using MCMC. For calculating the error in slope and intercept three random walkers were used at each data point and allowed to take 500 steps. The MCMC takes into account the covariance of the parameters, shown in figure 4.7 as a two dimensional Gaussian. There is a strong negative correlation between the slope m and intercept b which is the result of the degeneracy between gains g1 and g2 used to reconstruct energy by combining the light and charge signal. Thus, the error on g1 and g2 is such that for the positive maxima deviation in g1 we reach the negative maxima of the error on g2, and visa versa. Using standard reduced χ^2 for fitting and calculating errors in the slope and intercept would be underestimated the true error by a factor of five as it does not account for the degeneracy of the anti-correlated gains.




Figure 4.7: MCMC for the linear fit to the Doke plot. There is a strong negative correlation between the slope m and intercept b which results from the degeneracy between gains $g1$ and $g2$.

4.4 Combined Energy Space

With the values of $g1$ and $g2$ known the combined energy of events can be reconstructed with a significant improvement over using only the light or charge channel. In combined energy space recombination fluctuations are removed by the anti correlation of light and charge production and any residual smearing is due to intrinsic detector resolution (discussed later in section :). Figure 4.8 shows the energy histograms of the data used for the fits to gains $g1$ and $g2$ including the xenon activation lines and the ^{137}Cs calibration, along with a zoom in of the xenon K shell Xray at

34 keV. With the energy scale calibrated we can now reconstruct the energy of the events and convert the measured S1 and S2 signals to fundamental quanta using the gains g1,g2 allowing us to untangle instrumental and recombination fluctuations and measure light and charge yields.




Figure 4.8: Combined energy scale. a) The xenon activation lines from early in the run. b) ^{137}Cs calibration data. c) All calibration data including the $^{83\text{m}}\text{Kr}$ calibration. d) Xenon X-ray.

4.5 Light Collection and Electron extraction

The value of g1 represents the mean efficiency for collecting photons at the center of the LUX detector, the response to S1 light is flat fielded and normalized to the detector center (section Kr calibrations). The measured value of $g1 = 0.097 \pm 0.008$ implies a 9.7% probability of a photon propagated from the center of the detector striking a PMT and being converting into a photo electron (PE) signal. The value of g2 represents the average number of PE collected for each electron that escaped recombination at the initial interaction site and then drifted, by the electric field, towards liquid-gas interface where it is extracted with some probability ϵ . The value of g2 can be thought of as the average single electron size in PE times the extraction probability ϵ .

$$g2 = \epsilon \times SE \quad (4.8)$$

where S2 is the average pulse area of a single electron. The S2 signals are corrected for depth as impurities exponentially attenuate electrons drifting through the xenon, section 3. The LUX detector is low enough in threshold to observe single electrons being extracted from the liquid. Comparing the value of g2 derived from the Doke method with the single electron size is a good sanity check on the energy scale calibration. As the electrons are extracted from the liquid they are accelerated by larger field between the gate and anode where they electro-luminesce, a single extracted electron creates tens to hundreds of photons which are collected by both PMT arrays. We can cut on the single electron population (small S2 pulses without an associated S1) and measure the single electron size along with the extraction

efficiency efficiency. The extraction efficiency is defined as the probability that an electron will be extracted from the liquid into the gas in a region, held at 3.5 kV in the liquid between the anode and the gate. For a given event the extraction of electrons is a binomial processes with a rate approaching unity for fields above 5 kV [54] [43]. Figure 4.9, shows the single electron size as measured by the bottom PMT array (used for S2 pulses in the LUX detector to avoid saturation). The population is modeled by a skew Gaussian due to the Poisson nature of measuring only a handful of photo electrons (PE) per extracted electron. The mean of the distribution is found to be 9.7 PE/e^- with a width of $\sigma_{\text{SE}} = 3.6 \text{ Phe/e}^-$. Thus, the extraction efficiency is g^2 over the single electron size and is found to be $(59.3 \pm 14)\%$, given the extraction field the value is in good agreement with previous measurements in other xenon detectors [54] [43].




Figure 4.9: Single electron distribution as seen by the bottom PMT array fitted with a skew Gaussian model to account for the underlying Poisson statistics of observing between 9 and 10 photo electrons (PE) for each electron leaving the liquid and entering to the gas phase at a higher field causing electro-luminescence. The μ of the fit represents the true mean of the skew Gaussian distribution.

4.6 Tritium Beta Spectrum

The energy calibration in the WIMP search region can be tested by using the tritium calibration source described in Chapter[]]. Tritium has a Q value of 18.6 keV [69], a mean beta energy of 5.6 keV [70] and a mode of 3.4 keV [14] making it ideal for calibrating the LUX detector at the lowest energies. The tritium beta spectrum produces events at energies well below the detector threshold. Therefore, by comparing the reconstructed energy to the true tritium beta spectrum we can extract the energy threshold. We account for the detector resolution, smearing, by applying

the empirically determined resolution measured in Chapter []. Smearing the spectrum with the empirically determined detector resolution is in good agreement with that expected from the simulation package of LUXSIM.

Figure 6.4 (a,b) shows the reconstructed energy from a tritium calibration at the default field setting of 170 V/cm. The calibration contains 140,000 tritium events with an expected 2 ± 1 background events in black. In red, a simulated tritium beta spectrum from the LUXSIM package with modeled detector resolution and having gone through data processing framework. In blue and green are the theoretical tritium beta spectrum with infinite detector resolution and with the added resolution of the LUX detector, respectively. Figure 6.4 (c,d) shows the same calibration but at a lower drift field setting of 100 V/cm with only 4,500 tritium events and an expected $1\pm$ background events.




Figure 4.10: The tritium energy spectrum reconstructed from the data (black).

Along with LUX SIM (blue), and the true tritium beta spectrum (blue) and a tritium spectrum smeared with detector resolution (green).

The reconstructed energy spectrum is in good agreement with the expected tritium beta spectrum with detector resolution, using both LUXSIM and the empirically determined resolution. The detector threshold reaches 100% at about 1.5 keV making the tritium beta peak clearly visible providing crucial cross check of

the reconstructed energy around the WIMP search region of interest (1-5 keV). The 18.6 keV endpoint is another good low energy calibration point. We find that the end point of the reconstructed energy spectrum is consistent with that expected when convolving the true tritium beta spectrum with detector resolution, visible in log space. Though the energy scale for ER events was events was calibrated using mono energetic sources well above the tritium Q value the reconstructed tritium beta spectrum lines up with the expectation all the way down to the 1.5 keV threshold.

Figure 4.11 and 4.12 show the S1,S2 and Energy threshold attained by comparing the data to the expected photon, electron and energy spectrum, at 170 and 100 V/cm. The energy threshold is set by the light collection of the much smaller S1 signal. For the energy threshold we find roughly 50% efficiency at 1 keV_{ee} approaching 100% at 1.5 keV_{ee} regardless of the applied field. The S1 threshold is 50% at 2 PE and surpassed 90% above 3 PE. This translates to an S2_b threshold for golden events of 50% at 300 PE approaching 100% a 400 PE. Extracting the threshold will be discussed in further detail in Section




Figure 4.11: Threshold calculated from difference of simulated Tritium S1, S2 and energy spectra at 170 V/cm. a) S1 b) S2_b, c) Combined Energy .




Figure 4.12: Threshold calculated from difference of simulated Tritium S1, S2 and energy spectra at 100 V/cm. a) S1 b) S2_b, c) Combined Energy .

Chapter 5: Event Level Fluctuations

In this chapter we discuss extracting recombination fluctuations from line sources and continuous spectrum such as tritium. The method outlined will be used to measure recombination fluctuations from the tritium calibration data down to 1 keV. We begin by modeling the intrinsic resolution of the LUX detector, based on counting statistics. We then separate the fluctuations in light and charge collection from recombination fluctuations using line source calibrations. Once the variances from light and charge collection are modeled the recombination fluctuations from continuous spectra can be extracted, specifically for the tritium beta spectrum. We conclude with the results for recombination and recombination fluctuations as measured from tritium beta decay in the LUX detector along with a measure of the exciton-to-ion ratio α for ER events. Since we use photons and electrons (S1 and S2)to discriminate background events in LUX understanding recombination fluctuations are of great importance.

5.1 An Introduction to Fluctuations

When Xenon TPCs where first developed it was expected that the resolution in the ionization and scintillation channels would be dominated by detector resolution.

The only fluctuations fundamental to liquid xenon is theorized to be from the Fano factor along with an additional binomial variance in electron ion pair recombination. Recombination models from [71] and [72], used in [59], assume that the total observed recombination is the result each electron-ion pairs interacting with its-self, geminate recombination. The variance for such a process with recombination probability r_p acting on n_i number of ions is $(1 - r_p)r_p n_i$. Thus, a liquid xenon detector with infinite resolution should observe fluctuations governed by,

$$\begin{aligned}\sigma_{n_\gamma}^2 &= r_p^2 F n_i + (1 - r_p)r_p n_i \\ \sigma_{n_e}^2 &= (1 - r_p)^2 F n_i + (1 - r_p)r_p n_i\end{aligned}\tag{5.1}$$

where F is the Fano factor, $F = 0.05$ in liquid xenon [56], and r_p is the recombination probability. Note, equation 5.1 will be derived later in this section, given in equations 5.25 and 5.26. The recombination probability r_p is equal to the average observed recombination $\langle r \rangle$ as the number of events gets large, given in table 5.1.

The variance in equation 5.1 is in fact orders of magnitude off, and has been an unsolved mystery since the first xenon TPCs was built. Fortunately, these large fluctuations in scintillation and ionization are 100% anti-correlated and cancel when both light and charge is combined to measure energy. We refer to these fluctuations as recombination fluctuations σ_R . The 100% anti-correlation implies, that for each additional electron-ion recombination a single photon is produced at the cost of a single electron, and visa versa.

Let us consider the 164 keV line from ^{131}Xe used to produce the Doke plots in section 4. An illustration with all the fluctuations in units of quanta is shown in figure 5.1, with the number of photons (S_1/g_1) plotted vs the number of electrons

$(S2/g_2)$. The values of g_1 and g_2 were determined from calibrations in section 4.




Figure 5.1: Fluctuations of the 164 keV line from ^{131}Xe in the LUX detector, with the number of electrons plotted vs. the number of photons. The blue and magenta arrows labeled as $\sigma_{n\gamma_{\text{Det}}}$ and $\sigma_{n_e_{\text{Det}}}$ are the size of fluctuations in the light and charge channels due to the resolution of the LUX detector. The black arrow represents the size of recombination fluctuations σ_R . The value of $\langle r \rangle$ is the average observed recombination fraction or the average recombination probability r_p . The red lines represent constant $\langle r \rangle \pm \sigma_r$ assuming the expected binomial variance of equation 5.1. The black lines represent constant $\langle r \rangle \pm \sigma_r$ measured from the data. The black dashed line is represent constant energy or constant number of quanta.

the blue and magenta arrows labeled as $\sigma_{n\gamma_{\text{Det}}}$ and $\sigma_{n_e_{\text{Det}}}$ are the size of fluctuations in the light and charge measurement due to the resolution of the LUX detector. The value of $\langle r \rangle$ is the average observed recombination fraction and can be thought of as the average recombination probability r_p . The red lines represent constant $\langle r \rangle \pm \sigma_r$ assuming the expected binomial variance of equation 5.1. The expected

recombination (in red) is small compared to fluctuations from detector resolutions and is more than a factor of ten less than that observed. The black lines represent constant $\langle r \rangle \pm \sigma_r$ measured from the data. The size of recombination fluctuations σ_R are dominant over detector resolution stretching the island size along lines of constant energy (black dashed line). The island is not stretched exactly along the black dashed line as the non negligible component from detector resolution also warp the population. The projection of the population onto the combined quanta axis ($n_\gamma + n_e$), or energy E/W , is shown in figure 5.2.




Figure 5.2: The projection of the population in figure 5.1 onto the combined quanta axis ($n_\gamma + n_e$) = E/W . Along this projection the dominant recombination fluctuations cancel out, leaving on the components from detector resolution $\sigma_{n_\gamma \text{Det}}$ and $\sigma_{n_e \text{Det}}$.

In the combined light and charge axis we find that the recombination fluctuations have vanished. Leaving only the statistical variance from detector resolution. Some

useful definitions of parameters which will be discussed in this section are given in table 5.1.

Parameter	Description	Definition
n_γ	Number of photons	$\langle S_1 \rangle / g_1$
n_e	Number of electrons	$\langle S_2 \rangle / g_2$
n_i	Number of ions	$(n_\gamma + n_e) / (1 + \alpha)$
n_{ex}	Number of excitons	αn_i
α	Exciton to ion ratio	n_{ex} / n_i
$\sigma_{n_{\gamma_{\text{Det}}}}$	n_γ detector fluctuations	equation 5.17
$\sigma_{n_{e_{\text{Det}}}}$	n_e detector fluctuations	equation 5.17
r	Recombination fraction	$\left(\frac{n_\gamma}{n_e} - \alpha \right) / \left(\frac{n_\gamma}{n_e} + 1 \right)$
r_p	Recombination probability	$\langle r \rangle$
R	Recombined ions	$\langle r \rangle n_i$
σ_R	Recombination fluctuations	$\sigma_{\langle r \rangle} n_i$

Table 5.1: LUX detector parameters, used to measure statistical fluctuations in the light (S_1) and charge (S_2) channels. Where the S_1 and S_2 signals have been corrected for position dependance outlined in section 3. The values of g_1 and g_2 where measured in section 4.

For the rest of this section we will build on the example from figure 5.1 to better understand recombination fluctuations. We will explore recombination fluctuations with line-source calibrations then expand the picture to deal with continuous sources.

5.2 Modeling Intrinsic Detector Resolution

Intrinsic statistical fluctuations in light and charge (S1 and S2) collection in the LUX detector lead to a spread in collected quanta. To measure effects from recombination fluctuations and the Fano factor, we must first decouple the detector component of resolution. We use the model described in [73] [8] in which the measured scintillation and ionization signals (S1 and S2 measured in PE) are related to the number of photons and electrons by gains g_1 and g_2 , equation 5.2 and 5.3. Specifically, the average number of photons and electrons produced for a given energy deposit are proportional to the average S1 and S2 signals.

$$\langle n_\gamma \rangle = \frac{\langle S1 \rangle}{g_1} \quad (5.2)$$

$$\langle n_e \rangle = \frac{\langle S2 \rangle}{g_2} \quad (5.3)$$

where the gain g_1 represents photon detection efficiency, the probability of a photon from an energy deposit striking a PMT and producing a photo electron signal (PE). Gain g_2 represents the average S2 signal of a single electron multiplied by the electron extraction efficiency ϵ . Here, we are only using the S2 of the bottom PMT array and is corrected for electron-lifetime. The fluctuations in photons and electrons are related to the observables S1 and S2 by equations 5.2 and 5.3

$$\sigma_{n_{\gamma_{\text{stat}}}}^2 = \frac{\sigma_{S1_{\text{stat}}}^2}{g_1^2} \quad (5.4)$$

$$\sigma_{n_{e_{\text{stat}}}}^2 = \frac{\sigma_{S2_{\text{stat}}}^2}{g_2^2} \quad (5.5)$$

where $\sigma_{n_{\gamma_{\text{stat}}}}^2$ and $\sigma_{n_{e_{\text{stat}}}}^2$ represent the variance in the average number of measured photons and electrons as measured through S1 and S2, respectively. The variances in S1 and S2, $\sigma_{S1_{\text{stat}}}^2$ and $\sigma_{S2_{\text{stat}}}^2$, are the observable quantities with the detector. The variance in terms of quanta (n_{γ} and n_e) must be considered in terms of the PE being counted by the PMTs. Note, here we are working with the x,y,z corrected signals outlined in section 3.

Before proceeding to derive the statistical variance in light and charge, we overview a list of terms and their values in the LUX detector, given in table 5.2.

The statistical variance of the x,y,z corrected S1 signal, in equation ??, can be broken into two linearly independent parts. First we consider the binomial variance, for each event there are n_{γ} number of PE to be collected by the PMTs with probability g_1 . Here, n_{γ} should be thought of as the number of trials. The binomial variance is,

$$\sigma_{S1_{\text{Bin}}}^2 = (1 - g_1)g_1 n_{\gamma} \quad (5.6)$$

where $\sigma_{S1_{\text{Bin}}}^2$ is the binomial variance of the S1 light collection process with probability g_1 and n_{γ} number of trials. Each PE that was collected by a PMT then undergoes a second fluctuation due to the resolution of the PMTs. The variance from PMT resolution can be written as

$$\sigma_{S1_{\text{PMT}}}^2 = g_1 n_{\gamma} \sigma_{\text{PE}}^2 \quad (5.7)$$

Parameter	Definition	Value
g_1	photon detection probability	0.097 ± 0.008 [PE/n _{γ}]
σ_{PE}	single PE resolution, all PMTs	0.50 [PE/n _{γ}]
$g_2 = \epsilon \text{SE}_b$	average electron signal	5.75 ± 1.4 [PE/n _e]
SE_b	single electron size, bottom PMT array	9.70 ± 0.1 [PE/n _e]
ϵ	electron extraction probability	0.593 ± 0.144
σ_{SE_b}	single electron resolution, bottom PMT array	3.6 [PE/n _e]
κ	fraction of non-attenuated electrons	0.85

Table 5.2: LUX detector parameters, used to measure statistical fluctuations in the light (S1) and charge (S2) channels. Where the S1 and S2 signals have been corrected for position dependance outlined in section 3. The values of g_1 and g_2 where measured in section 4.

where $\sigma_{\text{S1}_{\text{PMT}}}^2$ is the average variance due to PMT resolution. The average single PE resolution of the PMTs is σ_{PE}^2 and is multiplied by the number of PE collected, $g_1 n_\gamma$.

Combining the two linearly independent processes of equation 5.6 and 5.7 leads to the the result in equation 5.8.

$$\sigma_{\text{S1}_{\text{stat}}}^2 = (1 - g_1 + \sigma_{\text{PE}}^2)(g_1 n_\gamma) = (1 - g_1 + \sigma_{\text{PE}}^2)\text{S1} \quad (5.8)$$

where $\sigma_{\text{S1}_{\text{stat}}}^2$ is the statistical variance of the S1 signal. The variance in the measured

number of photons is then equation 5.8 divided by g_1^2 , using equation 5.4.

$$\sigma_{n_{\gamma}^{stat}}^2 = \frac{(1 - g_1 + \sigma_{PE}^2)}{g_1} n_{\gamma} \quad (5.9)$$

We test equation 5.8 and 5.9 by using the 9.4 keV S1 signal from ^{83m}Kr . At that energy the dominant fluctuation is due the statistical variance of light collection, we can ignore contributions from recombination fluctuations and instrumental fluctuations (will be discussed later in this section). The S1 signal is shown in figure 5.3.




Figure 5.3: The fluctuations in the S1-x,y,z corrected signal for the 9.4 keV from ^{83m}Kr . At 9.4 keV the statistical variance is dominate over recombination fluctuations, $\sigma_{S1} = 8.15$. The resolution is consistent with that expected from statistical fluctuations of 8.3 ± 0.1 PE, from equation 5.8.

We find that the observed $\sigma_{S1} = 8.15$ PE is in good agreement with the expectation from equation 5.8 of 8.3 ± 0.1 PE. Or in terms of photons, $\sigma_{n_{\gamma_{\text{stat}}}} = 84.1$ with the expectation from equation 5.9 of 85.7 ± 1 . Having used the values listed in table 5.2.

The variance of the x,y,z corrected S2 signal, given in 5.5, is also comprised of several independent processes. Each electron that reaches the liquid-gas interface without being attenuated will either be extracted into the gas producing SE_b number of PE or not. The binomial variance of such a process is,

$$\sigma_{S2_{\text{bino}}}^2 = (1 - \epsilon)\epsilon(\kappa n_e)SE_b^2 \quad (5.10)$$

where ϵ is the electron extraction probability, and κn_e is the average number of electrons that reach the liquid surface from the interaction site, also the number of trials. Recalling that κ is the average electron probability of an electron to not be attenuated. For the case of LUX with an electron lifetime of $\sim 1000 \mu s$ and an average drift time of $160 \mu s$ the value of $\kappa = 0.85$ and is listed in table 5.2. Each electron that gets extracted then multiplies producing SE_b PE in the bottom PMT array, this multiplicative factor must be squared in the variance.

Next, we consider the spread of the single electron size as measured by the bottom PMT array, σ_{SE_b} . The variance from the PMT resolution for the $\epsilon(\kappa n_e)$ number of electrons extracted is,

$$\sigma_{S2_{\text{PMT}}}^2 = \epsilon(\kappa n_e)\sigma_{SE}^2 \quad (5.11)$$

where $\epsilon \kappa n_e$ is the number of extracted electrons each with PMT resolution σ_{SE}^2 .

For the S2 signal we must also consider the additional variance from electron attenuation as the electrons drift, we model the process with a Poisson probability of electron capture in each Z slice of the detector. The variance from each Z slice depends of the average number of electrons that will be attenuated. The probability of attenuation at each slice in drift-time T is $P(T) = 1 - e^{-T/\tau}$, where τ for the data sets to be considered is $1000 \mu\text{s}$. The drift region considered in the fiducial volume is from 38 to $304.5 \mu\text{s}$. The average variance from events in the fiducial can be given by equation 5.12.

$$\sigma_{n_{e_{att}}}^2 = n_e \frac{\int_{T_{min}}^{T_{max}} (1 - e^{-T/\tau}) dT}{\int_{T_{min}}^{T_{max}} dT} = 0.155 \times n_e \quad (5.12)$$

Combining the variances from equations 5.13 5.11 and 5.12 leads to the the result for the statistical variance in the observed S2 signal,

$$\begin{aligned} \sigma_{S2_{stat}}^2 &= (1 - \epsilon)\epsilon(\kappa n_e)SE_b^2 + \epsilon(\kappa n_e)\sigma_{SE}^2 + g_2^2\sigma_{n_{e_{att}}}^2 \\ \sigma_{S2_{stat}}^2 &= ((1 - \epsilon)SE_b\kappa + \sigma_{SE}^2\kappa/g_2 + 0.155g_2) S2 \end{aligned} \quad (5.13)$$

$$\sigma_{n_{e_{stat}}}^2 = \frac{(1 - \epsilon)\epsilon SE_b^2 + \epsilon\sigma_{SE}^2}{g_2^2} \kappa n_e + \sigma_{n_{e_{att}}}^2 \quad (5.14)$$

Using equations 5.9, 5.14 and table 5.2 we calculate the intrinsic detector resolution for S1 and S2 signals in the LUX detector, given equation 5.15. Note, the intrinsic resolution in S2 is subdominant to that of S1, since on average one

electron multiplies to about ten photons detected by the bottom PMT array. Also listed in 5.16, are the instrumental fluctuations with a linear dependance on quanta measured using calibrations discussed in section 5.3. The total variance that the detector observes in the light and charge is the linear combination of the statistical and instrumental variance given in equation eq:SigDet.

$$\begin{aligned}\sigma_{n_{\gamma_{\text{stat}}}} &= 3.45 \pm^{0.17}_{0.15} \sqrt{n_{\gamma}} \\ \sigma_{n_{e_{\text{stat}}}} &= 1.04 \pm^{0.26}_{0.20} \sqrt{n_e}\end{aligned}\quad (5.15)$$

where $\sigma_{n_{\gamma_{\text{stat}}}}$ and $\sigma_{n_{e_{\text{stat}}}}$ are the statistical fluctuations outlined in this section.

$$\begin{aligned}\sigma_{n_{\gamma_{\text{inst}}}} &= \frac{6.4 \pm 1.7}{100} \times n_{\gamma} \\ \sigma_{n_{e_{\text{inst}}}} &= \frac{6.6 \pm 0.9}{100} \times n_e\end{aligned}\quad (5.16)$$

where $\sigma_{n_{\gamma_{\text{inst}}}}$ and $\sigma_{n_{e_{\text{inst}}}}$ are instrumental fluctuations extracted in section 5.3 that grow like number of quanta n but only appear to turn on above 200 keV.

$$\begin{aligned}\sigma_{n_{\gamma_{\text{Det}}}}^2 &= \sigma_{n_{\gamma_{\text{stat}}}}^2 + \sigma_{n_{\gamma_{\text{inst}}}}^2 \\ \sigma_{n_{e_{\text{Det}}}}^2 &= \sigma_{n_{e_{\text{stat}}}}^2 + \sigma_{n_{e_{\text{inst}}}}^2\end{aligned}\quad (5.17)$$

where $\sigma_{n_{\gamma_{\text{Det}}}}^2$ and $\sigma_{n_{e_{\text{Det}}}}^2$ are the fluctuations in counting photons and electrons, respectively, due to detector resolution.

5.3 Measuring Recombination Fluctuations with Mono-Energetic Sources

To model recombination we start with the assumption that for a given energy deposit in liquid xenon the number of quanta produced is equal to the number of excitons

and the number of ions [73].

$$\begin{aligned}\frac{E}{W} &= n_q = n_i + n_{ex} = n_i(1 + \alpha) \\ \frac{E}{W} &= n_\gamma + n_e = \frac{n_\gamma}{g_1} + \frac{n_e}{g_2}\end{aligned}\quad (5.18)$$

where E is energy in keV, W is the work function in keV/quanta, n_q is the number of quanta, n_i is the number of ions, n_{ex} is the number of excitons and α is the exciton-to-ion ratio. The theoretical value of the number of excitons produced to ions is $\frac{n_{ex}}{n_i} = \alpha = 0.20$ [68] and is not expected to change as a function of energy [48] [49] [8]. For the subsequent equations in this section we will simplify equations 5.18 to,

$$\begin{aligned}\alpha &= 0.20 \\ n_i &= \frac{E}{W} \frac{1}{(1 + \alpha)} = \frac{n_\gamma + n_e}{(1 + \alpha)}\end{aligned}\quad (5.19)$$

Equation 5.19 gives us a simple model for the number of ions produced for a given interaction. We work in number of ions for convince as the recombination fluctuations act only on ions. The only variation in quanta thus far is due to a Fano factor governing the variation in initial quanta produced.

$$\sigma_{n_i}^2 = F \times n_i \quad (5.20)$$

where F is the Fano factor. The value of F for liquid xenon is small and has a theoretical value of 0.05 [56].

We now describe the observed scintillation and ionization signals that are measured in the LUX detector, S1 and S2 respectively, as a function of n_i . The number of photons observed for a given energy deposit arise from the excitons that de-excite and from ions which recombine with freed electrons.

$$n_\gamma = n_{ex} + n_i \times r = n_i \times (r + \alpha) \quad (5.21)$$

The number of electrons corresponding to a given energy deposit is equal to the number of ions that did not recombine with a freed electron.

$$n_e = n_i \times (1 - r) \quad (5.22)$$

The recombination fraction of each event can be solved for in terms of number of photons and electrons produced,

$$r = \frac{\frac{n_\gamma}{n_e} - \alpha}{\frac{n_\gamma}{n_e} + 1} \quad (5.23)$$

where r represents the electron-ion recombination probability of each event.

Two key measurable quantities from the scintillation and ionization signals are the average recombination fraction $\langle r \rangle$ and the spread in recombination probability $\sigma_{\langle r \rangle}$. The average recombination fraction $\langle r \rangle$ can be interpreted as the electron-ion pair recombination probability r_p . The recombination fluctuation in units of quanta is

$$\sigma_R = \sigma_{\langle r \rangle} \times n_i \quad (5.24)$$

where σ_R is the recombination fluctuation. As mentioned earlier, the recombination

fluctuations are much larger than those expected from the binomial variance of a binomial process with probability r_p or the Fano factor, illustrated in figure 5.1.

We now combine the variance from the Fano factor, recombination and detector resolution (equation 5.17) and solve for the total observed variance in photons and electrons given in 5.25 and 5.26

$$\sigma_{n_\gamma}^2 = \sigma_{n_{ex}}^2 + \sigma_{n_i}^2 r_p^2 + \sigma_{\langle r \rangle}^2 n_i^2 + \sigma_{n_{\gamma_{Det}}}^2 = [\sigma_{n_{ex}}^2 + n_i F r_p^2] + \sigma_{\langle r \rangle}^2 n_i^2 + \sigma_{n_{\gamma_{Det}}}^2 \quad (5.25)$$

$$\sigma_{n_e}^2 = \sigma_{n_i}^2 (1 - r_p)^2 + \sigma_{\langle r \rangle}^2 n_i^2 + \sigma_{n_{e_{Det}}}^2 = [n_i F (1 - r)^2] + \sigma_{\langle r \rangle}^2 n_i^2 + \sigma_{n_{e_{Det}}}^2 \quad (5.26)$$

where $\sigma_{n_{ex}}^2$ is the variance in exciton production, F is the Fano factor, the term $\sigma_{\langle r \rangle}^2 n_i^2$ is the recombination fluctuation σ_R and the rest of the variables are described in table 5.1.

Using a line source, we measure combined energy (equation 5.18) and $\sigma_{n_\gamma}^2$, $\sigma_{n_e}^2$ and σ_E^2 . Dropping the contribution form the Fano factor and the the number of excitons in equations 5.25 and 5.26 the variance of $\sigma_{n_\gamma}^2$ and $\sigma_{n_e}^2$ is a linear combination of detector resolution and recombination fluctuations.

$$\sigma_{n_\gamma}^2 = \sigma_{n_{\gamma_{Det}}}^2 + \sigma R^2 \quad (5.27)$$

$$\sigma_{n_e}^2 = \sigma_{n_{e_{Det}}}^2 + \sigma R^2 \quad (5.28)$$

this concept is illustrated in figure 5.1. The value of recombination fluctuation σ_R can be determined by rearranging equations 5.27 and 5.28. As shown in figure 5.2, when the light and charge signals are combined the rustling variance σ_E^2 contains no

recombination fluctuations, as recombination fluctuations are 100% anti-correlated in light and charge.

$$\sigma_R^2 = \frac{1}{2} \left(\sigma_{n_\gamma}^2 + \sigma_{n_e}^2 - \frac{\sigma_E^2}{W^2} \right) \quad (5.29)$$

where the variance in photons $\sigma_{n_\gamma}^2$, electrons $\sigma_{n_e}^2$ and quanta $\frac{\sigma_E^2}{W^2}$ are all observable quantities with a line source. And can be rewrite in terms of the observable quantities S1 and S2.

$$\sigma_R^2 = \frac{1}{2} \left(\frac{\sigma_{S1}^2}{g_1^2} + \frac{\sigma_{S2}^2}{g_2^2} - \frac{\sigma_E^2}{W^2} \right) \quad (5.30)$$

We now have a method to extract the recombination fluctuation σ_R using a line calibration source. To complete our treatment of line sources, the variance in the light and charge due to detector resolution can also be measured.

$$\sigma_{n_{\gamma_{\text{Det}}}}^2 = \frac{\sigma_{S1}^2}{g_1^2} - \sigma R^2 \quad (5.31)$$

$$\sigma_{n_{e_{\text{Det}}}}^2 = \frac{\sigma_{S2}^2}{g_2^2} - \sigma R^2 \quad (5.32)$$

Using equation 5.30, 5.31 and 5.32 along with the measurements of g_1 g_2 , we construct a combined energy and deconvolve the recombination fluctuations from variances in the light and charge observed by the detector. The result is shown in figure 5.4, The black white and red lines represent σ_R , $\sigma_{n_{\gamma_{\text{Det}}}}$, $\sigma_{n_{e_{\text{Det}}}}$, respectively. With the values and sources listed in table 5.3.

The values of σ_R , $\sigma_{n_{\gamma_{\text{Det}}}}$, $\sigma_{n_{e_{\text{Det}}}}$ from table 5.3 are plotted in figure 5.5.

A variance with a linear and root term is fit to the data with the results given in 5.33. The linear term corresponds to instrumental fluctuations and the root term




Figure 5.4: Populations of calibration sources in discrimination space $\log\left(\frac{n_e}{n_\gamma}\right)$ vs. combined energy keV_{ee}. The ovals represent the combination of σ_R , $\sigma_{n_\gamma\text{Det}}$, $\sigma_{n_e\text{Det}}$ in black, white, red respectively.

Source	Energy [keV]	σ_R	$\sigma_{n_\gamma\text{Det}}$	$\sigma_{n_e\text{Det}}$	σ_E/W
K-shell X-ray	~ 32	52.6 ± 23	244 ± 5	116 ± 3	269 ± 5
^{83m} Kr	41.55	89.1 ± 0.6	171 ± 0.3	51.2 ± 0.3	173 ± 0.2
¹³¹ Xe	163.9	637 ± 20	388 ± 31	83 ± 77	375 ± 8
¹²⁷ Xe	208.3	916 ± 59	506 ± 96	260 ± 133	568 ± 26
¹²⁷ Xe, ^{129m} Xe	236.8	1001 ± 23	441 ± 50	253 ± 71	491 ± 9
¹²⁷ Xe	408.8	1294 ± 44	1166 ± 39	949 ± 40	1562 ± 22
²¹⁴ Bi	609	2488 ± 109	2298 ± 99	1627 ± 107	3291 ± 60
¹³⁷ Cs	661.6	2686 ± 34	2059 ± 38	1368 ± 46	2564 ± 17

Table 5.3: Extracted fluctuations from the line source calibration data in units of quanta. The method of extracting the quantities is given in 5.30, 5.31 and 5.32 and illustrated for the case of ¹³¹Xe in figures 5.1 and 5.2. Note, the K-shell X-ray may include a fairly large systematic as no radial cut was made in order to observe the signal near the detector edge.




Figure 5.5: Measured values of σ_R , $\sigma n_{\gamma_{\text{Det}}}$, $\sigma n_{e_{\text{Det}}}$ vs. Energy on the left and vs. quanta in photons, electrons and ions, respectively, on the right. Measured using sources listed in Table 5.3.

corresponds to statistical fluctuations. Instrumental fluctuations are proportional to the signal size and cause the fluctuations in $\sigma n_{\gamma_{\text{Det}}}$, $\sigma n_{e_{\text{Det}}}$ at high energies to deviate from statistical fluctuations alone.

$$\begin{aligned}\sigma_{n_{\gamma_{\text{Det}}}}^2 &= \sigma_{n_{\gamma_{\text{Stat}}}}^2 + \sigma_{n_{\gamma_{\text{Inst}}}}^2 = (0 \pm 10 \cdot \sqrt{n_{\gamma}})^2 + ((6.4 \pm 1.8)/100 \cdot n_{\gamma})^2 \\ \sigma_{n_{e_{\text{Det}}}}^2 &= \sigma_{n_{e_{\text{Stat}}}}^2 + \sigma_{n_{e_{\text{Inst}}}}^2 = (1 \pm 4 \cdot \sqrt{n_e})^2 + ((6.6 \pm 0.6)/100 \cdot n_e)^2 \\ \sigma_R^2 &= ((5.5 \pm 0.5)/100 \cdot n_q)^2\end{aligned}\quad (5.33)$$

With the limited data we can not tightly constrain the root-n term. The statistical fluctuation expected from equation 5.15 is consistent with the observed values without the instrumental component for calibration energies at and below 236.8 keV. The instrumental fluctuations appear to turn on above ~ 200 keV and may be due to and may be due to ripples in the liquid surface caused by xenon bubbles or other systematics.

5.4 Measuring Recombination Fluctuations in Discrete Energy Bins

The previous section outlined a method to extract recombination fluctuations from line source calibrations, equation 5.30. In this section the formalism to measure recombination fluctuations when considering events binned in energy will be outlined. The consideration of discrete binning is crucial when dealing with a continual energy spectrum. Take the tritium beta spectrum as an example, we lose the ability to independently measure $\sigma_{n_\gamma}^2$, $\sigma_{n_e}^2$, σ_E^2 and are only left with a smear of n_γ , n_e , E . However, there are two key pieces of information still left at our disposal. First, the combined energy can be reconstructed from constraints on g_1 and g_2 , and even corrected for spectral shape and detector resolution (discussed later in section 6.1). Second, know the dependence of detector resolution as a function of quanta, given in 5.15 and 5.16. It will be shown, that knowing g_1 , g_2 and the intrinsic detector resolution is sufficient to measure recombination fluctuations for a continual energy spectrum.

To simplify the picture we introduce new variables when dealing with events having been cut in energy of bin of width Δ_E . After cutting in energy, we define the remaining variance projected onto light and charge as $\chi_{n_\gamma}^2$ and $\chi_{n_e}^2$, which is analogous to $\sigma_{n_\gamma}^2$ and $\sigma_{n_e}^2$ of equations 5.27 5.28, respectively. We denote the component of variance from detector resolution contained in the slice of energy as χ_{Det} . These concepts will be clarified in the subsequent examples and are summarized in table 5.4.

Parameter	Definition	Analogy Without Binning
$\chi^2_{n_\gamma}$	Total variance with E cut projected onto n_γ	$\sigma_{n_\gamma}^2$
$\chi^2_{n_e}$	Total variance with E cut projected onto n_e	$\sigma_{n_e}^2$
χ^2_{Det}	Variance from detector resolution shared between n_γ and n_e , with E cut	$\sigma_{n_{\gamma\text{Det}}}^2 \sigma_{n_{e\text{Det}}}^2$

Table 5.4: Useful definitions when considering cutting events in a bin of energy and projecting onto the n_γ and n_e axis.

We will overview the treatment of binning by considering two cases using a simulated $^{83\text{m}}\text{Kr}$ line-source. For each case, the energy is defined using equation 5.18 and a cut is energy of width $\Delta_E = 1$ is made about the energy center. First, variance from detector resolution is turned off leaving only recombination fluctuations. The value of recombination fluctuations in terms of quanta is set to $\sigma_R = 90$ in the simulation, similar to the value extracted from the $^{83\text{m}}\text{Kr}$ data in table 5.3. Figure 5.6 shows the result of such a system.




Figure 5.6: A simulated $^{83\text{m}}\text{Kr}$ source with infinite detector resolution and recombination fluctuations set to 90 quanta. The recombination fluctuation move along the line of constant energy about the center, shown in black with bin width of $\Delta_E = 1 \text{ keV}$. The events between the black lines and projected onto the n_γ and n_e axes. The variance in n_γ and n_e are $\sigma_{n_\gamma}^2 = \chi_{n_\gamma}^2 = \sigma_R^2$ and $\sigma_{n_e}^2 = \chi_{n_e}^2 = \sigma_R^2$.

In figure 5.6 we see that recombination fluctuations move events along the diagonal of constant energy, as expected. The standard deviation in light and charge is found by the fit to be 90.3 ± 0.3 , consistent to what was input into the simulation, 90. For the case of only recombination fluctuations and perfect detector resolution we find,

$$\begin{aligned} \sigma_{n_\gamma}^2 &= \chi_{n_\gamma}^2 = \sigma_R^2 \\ \sigma_{n_e}^2 &= \chi_{n_e}^2 = \sigma_R^2 \end{aligned} \tag{5.34}$$

The key point demonstrated in figure 5.6 is that all contribution from recombination fluctuations are included when slicing along a contour of energy about the center, no matter how thin we make the cut. Note, with infinite detector resolution the population becomes a delta function in energy about the central value.

Next, we visualize how the intrinsic detector resolution, $\sigma_{n_{\gamma_{\text{Det}}}}^2$ and $\sigma_{n_{e_{\text{Det}}}}^2$, appears along a contour of constant energy for a simulated $^{83\text{m}}\text{Kr}$ source. The recombination fluctuations are set to zero and the fluctuations from detector resolution are set to $\sigma_{n_{\gamma_{\text{Det}}}} = 171$ and $\sigma_{n_{\gamma_{\text{Det}}}} = 51$, close to the true values of the LUX detector 5.3. The slice in combined energy is illustrated in figure 5.6, using a bin width of $\Delta_E = 1 \text{ keV}$ (73 quanta).




Figure 5.7: A simulated $^{83\text{m}}\text{Kr}$ source with no recombination fluctuations and with detector resolution, $\sigma_{n_\gamma\text{Det}} = 171$ and $\sigma_{n_e\text{Det}} = 51$. The solid black lines represents constant energy about the center having width of $\Delta_E = 1$. The events that fall between the black dashed lines are projected onto the n_γ and n_e axes. The square root of the observed variance in light and charge are show on the labels with and without the energy cut, as χ and σ respectively. The energy cut has swept out a component of variance χ_{Det} that is shared between the light and charge channels.

In the example pictured in figure 5.6 we find that the energy cut leaves a reduced statistical component to be projected onto n_γ and n_e . The reduction in variance projected onto n_γ is an order of magnitude. However, this is expected as only a small portion of n_γ is included with the energy cut shown in figure 5.6.

With the examples shown in figure 5.6 and 5.9 we can now visualize the effect of an energy cut on fluctuations in the light (n_γ) and charge signals (n_e). The

slice, by definition, contains only events which fluctuated about the contour of constant energy thus, automatically including all recombination fluctuations and only a reduced component from detector resolution (χ_{Det}).

To solve for the value of χ_{Det} we write the detector-resolution fluctuations $\sigma_{n_{\gamma\text{Det}}}$ and $\sigma_{n_{e\text{Det}}}$ as a function of combined energy in number quanta.

$$n_q = E/W = n_\gamma + n_e \quad (5.35)$$

where n_q is the number of quanta (photons + electrons), E is energy in keV, and $W=0.0137\pm0.001$ quanta/keV. Next, consider the the slope induced by the statistical variance of $\sigma_{n_{\gamma\text{Det}}}$ and $\sigma_{n_{e\text{Det}}}$ as a function quanta n_q . The slope between n_γ with respect to total quanta n_q (energy) is the variance in n_γ over the variance in n_q , and the same for n_e . The result for the slope M considering a detector with resolution $\sigma_{n_{\gamma\text{Det}}}$ and $\sigma_{n_{e\text{Det}}}$ is,

$$\begin{aligned} M &= \tan(\theta_{n_{\gamma\text{Det}}}) = \frac{\sigma_{n_{\gamma\text{Det}}}^2}{\sigma_{n_{\gamma\text{Det}}}^2 + \sigma_{n_{e\text{Det}}}^2} \\ 1 - M &= \tan(\theta_{n_{e\text{Det}}}) = \frac{\sigma_{n_{e\text{Det}}}^2}{\sigma_{n_{\gamma\text{Det}}}^2 + \sigma_{n_{e\text{Det}}}^2} \end{aligned} \quad (5.36)$$

where M is the slope of n_γ with respect to n_q . Once M is defined, the angle between n_e and n_q is the complementary slope ($1-M$). Recalling that $(n_\gamma + n_e) = n_q$, in equation 5.35. The slope M can also be thought of as $\tan(\theta)$ where θ is the angle between n_γ and n_q , ranging from 0 to $\pi/4$. The slop of the population induced by the statistical variance of n_γ and n_e with respect to the quanta (or E/W) axis can be expressed interns of the slope M .

$$\begin{aligned} n_\gamma &= Mn_q \\ n_e &= (1 - M)n_q \end{aligned} \tag{5.37}$$

Using equation 5.37 and 5.35 we solve for the for a small variation in n_q induced by the fluctuations in n_γ and n_e along a contour of constant n_q . First holding the number of photons n_γ constant and allowing for small variations in n_e and then holding the number of electrons n_e constant and allowing for small variations in n_γ .

$$\begin{aligned} \delta n_q &= \frac{\partial n_\gamma}{\partial n_q} \delta n_q \Big|_{n_\gamma} + \frac{\partial n_e}{\partial n_q} \delta n_q \Big|_{n_e} \\ \delta n_q &= \frac{\partial(n_q - n_e)}{\partial n_q} \delta n_e + \frac{\partial(n_q - n_\gamma)}{\partial n_q} \delta n_\gamma \\ \delta n_q &= \left[\frac{\partial n_q}{\partial n_q} - \frac{\partial n_e}{\partial n_q} \right]^{(1-M)} \delta n_e + \left[\frac{\partial n_q}{\partial n_q} - \frac{\partial n_\gamma}{\partial n_q} \right]^M \delta n_\gamma \\ \delta n_q &= (M)\delta n_e + (1 - M)\delta n_\gamma \end{aligned} \tag{5.38}$$

squaring the result from equation 5.38 the variance along an infinitely thin line of constant energy induced by statistical fluctuations is,

$$\chi_{\text{Det}}^2 = \text{Var}(\delta n_q) = M^2 \sigma_{n_{e\text{Det}}}^2 + (1 - M)^2 \sigma_{n_{\gamma\text{Det}}}^2 \tag{5.39}$$

where M is given in equation 5.36 and the cross term $\delta n_\gamma \delta n_e$ has been dropped as the statistical variance from detector resolution ($\sigma_{n_{\gamma\text{Det}}}^2$ and $\sigma_{n_{e\text{Det}}}^2$) are uncorrelated.

Combining equation 5.39 and the equation for slope M in equation 5.36 simplifies the expression for χ_{Det}^2 to,

$$\chi_{\text{Det}}^2 = \frac{\sigma_{n_{\gamma_{\text{Det}}}}^2 \sigma_{n_{e_{\text{Det}}}}^2}{\sigma_{n_{\gamma_{\text{Det}}}}^2 + \sigma_{n_{e_{\text{Det}}}}^2} \quad (5.40)$$

the expression for χ_{Det}^2 describes the amount of statistical variance that will be included with the recombination fluctuations when events are cut along contours of constant energy. The only information needed to determine are χ_{Det}^2 is $\sigma_{n_{\gamma_{\text{Det}}}}^2$ and $\sigma_{n_{e_{\text{Det}}}}^2$, which are given in equation 5.17 as a function of quanta.

The example shown in figure 5.9 can now be understood in terms of χ_{Det}^2 . The simulation had used values of $\sigma_{n_{\gamma_{\text{Det}}}} = 171$ and $\sigma_{n_{e_{\text{Det}}}} = 51$. Solving for M using equation 5.36 we find, $M = 0.92$, meaning that the angle θ of the bulk population between the n_γ axis and the line of constant quanta (E/W) is 42.6° . Plugging the values into equation 5.40 the value of $\chi_{\text{Det}} = 49$ quanta. This is within 10% of the value of χ_{n_γ} and χ_{n_e} observed in the simulated data set, not bad considering that we used an infinitely thin line approximation. Later we will expand upon equation 5.38 to account for the finite bin width.

Let us briefly consider the implication of equation 5.39 and 5.40. In the case of equal detector resolution in light and charge, $\sigma_{n_{e_{\text{Det}}}}^2 = \sigma_{n_{\gamma_{\text{Det}}}}^2$ the value of M is 0.5 leading to, $\chi_{\text{Det}}^2 = \frac{1}{2}\sigma_{n_{e_{\text{Det}}}}^2$. Curiously, regardless of the resolution in light and charge, $\sigma_{n_{e_{\text{Det}}}}^2 \neq \sigma_{n_{\gamma_{\text{Det}}}}^2$, the statistical variance in a cut of energy is always less than the variance of the best channel. Specifically for the LUX detector, the implication of equation 5.39 is that the statistical variance measured in a cut on energy will

be less than that of the S2 statistical uncertainty, this will aid in our extraction of recombination fluctuations. The illustration for the case of $\sigma_{n_e \text{Det}} = \sigma_{n_\gamma \text{Det}} = 100$ is shown in figure 5.8, as expected the square root of the variance along the line of constant energy is about $100/\sqrt{2} = 70.7$.




Figure 5.8: A simulated $^{83\text{m}}\text{Kr}$ source with no recombination fluctuations and with detector resolution, $\sigma_{n_e \text{Det}} = \sigma_{n_\gamma \text{Det}} = 100$. The solid black line represents constant energy about the center having a width of $\Delta_E = 1$. The events that fall between the black dashed lines are projected onto the n_γ and n_e axes. The square root of the observed variance in light and charge are show on the labels with and without the energy cut, as χ and sigma respectively.

To complete the treatment of χ_{Det}^2 we add the contribution to variance from the finite bin width Δ_E . The residual variance arises from projecting the rotated population onto n_γ or n_e , the rotation having a slope of M and $(1-M)$ respectively, given in equation 5.36. Note, the residual variance from the rotation can also be removed by working with the centroid subtracted projections of number of photons and electrons as a function of energy, as discussed later in section 5.4.2. Considering the bin width, the total variance from detector resolution projected in n_γ and n_e is

$$\begin{aligned}\chi_{n_{\gamma\text{Det}}}^2 &= \chi_{\text{Det}}^2 + \frac{(MW\Delta_E)^2}{12} \\ \chi_{n_{e\text{Det}}}^2 &= \chi_{\text{Det}}^2 + \frac{((1-M)W\Delta_E)^2}{12}\end{aligned}\quad (5.41)$$

where χ_{Det}^2 is defined in 5.40, M is given in equation 5.36, W is the work function in quanta/keV, Δ_E is the bin width in energy keV, the normalization of 1/12 arrises from rotating a uniform population about its center. The value of $\chi_{n_{\gamma\text{Det}}}^2$ is the variance from statistical fluctuations after cutting along a line of constant energy having width Δ_E and then projecting onto the n_γ axis. Similarly, $\chi_{n_{e\text{Det}}}^2$ is the variance from statistical fluctuations after cutting along a line of constant energy and projecting onto the n_e axis.

The correction given in equation 5.41 is applied to the example in figure 5.9, which used a bin width of 1 keV (73 quanta). The value of χ_{Det} was previously calculated to be 49 quanta. The additional contribution from bin width results in $\chi_{n_{\gamma\text{Det}}} = 52.5$ quanta and $\chi_{n_{e\text{Det}}} = 49$ quanta. These values are consistent with those observed in the simulation shown in figure with 5.9, $\chi_{n_\gamma} = 52.6 \pm 0.6$ and $\chi_e = 50.0$

± 0.7 .

We now combine the two cases illustrated in figures 5.6 and 5.9, having recombination only and then detector resolution only. The combined variance projected in n_γ and n_e can be determined, given a cut in energy. The contribution to variance from the two can be added as they are independent processes.

$$\chi_{n_\gamma}^2 = \sigma_R^2 + \chi_{n_{\gamma\text{Det}}}^2 \quad (5.42)$$

$$\chi_{n_e}^2 = \sigma_R^2 + \chi_{n_{e\text{Det}}}^2 \quad (5.43)$$

$$(5.44)$$

where σ_R^2 are the recombination fluctuations, and the values $\chi_{n_{\gamma\text{Det}}}^2$ and $\chi_{n_{e\text{Det}}}^2$ are given in equation 5.41. The recombination fluctuation can be measured by rearranging equation 5.42 and 5.43

$$\sigma_{R_\gamma}^2 = \chi_{n_\gamma}^2 - \chi_{n_{\gamma\text{Det}}}^2 \quad (5.45)$$

$$\sigma_{R_e}^2 = \chi_{n_e}^2 - \chi_{n_{e\text{Det}}}^2 \quad (5.46)$$

Using equations 5.45 and 5.46 we can extract recombination fluctuations directly from the n_γ (S1) and n_e (S2) signals. The quantities $\chi_{n_\gamma}^2$ and $\chi_{n_e}^2$ are the observables and χ_{Det}^2 is determined from calibrations.

5.4.1 Application to ^{83}Kr

We now have a method to extract recombination fluctuations from data binned in energy in equation 5.45 and 5.46. Before applying the method to the tritium beta spectrum we test the method using a $^{83\text{m}}\text{Kr}$ calibration data set. The result is shown in figure 5.9 for a bin width of 1 keV, and listed for several bin width in table 5.5.




Figure 5.9: Data from a $^{83\text{m}}\text{Kr}$ data set, plotting the number of photons against the number of electrons. The red-dashed oval represents where the population would lie if recombination fluctuations were the dominant fluctuation. The solid black line represents constant energy about the center, and the region between the black dashed lines represents the cut in energy with bin width $\Delta_E = 1$. The events that fall between the black dashed lines are projected onto the n_γ and n_e axes. The observed standard deviation in n_γ and n_e is shown on the plot labels as χ and σ .

σ_R 5.30 (Quanta)	Δ_E (keV)	Count	$\sigma_{R_\gamma} = \sqrt{\chi^2_{n_\gamma} - \chi^2_{n_{\gamma_{\text{Det}}}}}$ (Quanta)	$\sigma_{R_e} = \sqrt{\chi^2_{n_e} - \chi^2_{n_{e_{\text{Det}}}}}$ (Quanta)
89.1 \pm 0.6	0.025	1518	87.2 \pm 2.9	87.1 \pm 2.9
	0.05	3124	85.0 \pm 2.0	84.9 \pm 2.0
	0.1	6269	87.8 \pm 1.3	87.6 \pm 1.3
	0.2	12508	90.0 \pm 1.0	89.7 \pm 1.0
	0.25	15557	88.5 \pm 0.8	88.3 \pm 0.8
	0.5	30826	92.5 \pm 0.6	91.2 \pm 0.6

Table 5.5: The extraction of recombination fluctuations from a ^{83m}Kr data set using various bin widths ΔE about the central value of energy. The standard deviation of recombination fluctuations σ_R are calculated from the fluctuations observed in n_γ and n_e using equations 5.45 and 5.46. The value of χ_{Det} is 49 quanta.

The ^{83m}Kr data set contained 400,000 events in the fiducial volume of the detector. Using equations 5.45 and 5.46 we calculate recombination fluctuations, finding good agreement with the method described for in equation 5.29. The agreement between the two methods is within 5% and implies that the detector resolution has been modeled well. Having demonstrated the method for a line source calibration source the next step is to apply the method outlined in this section to simulated tritium data and see how well the method can extract test values of recombination fluctuations.

5.4.2 Application to Simulated Tritium Data

In this section the method outlined in section 5.4 for line sources will be applied to the tritium beta spectrum. To first order, the treatment of a continuous spectrum is identical to the method outlined for ^{83m}Kr if we consider that the tritium beta spectrum is comprised of multiple line sources. It will be shown that the formalism developed can be applied to continuous spectra by testing the concepts using simulations.

As was done for the ^{83m}Kr source, we start by simulating a tritium beta spectrum with only recombination fluctuations. Figure ?? on the left, shows a tritium beta spectrum propagated at every 1 keV energy bin and on the right from every 0.1 keV.




Figure 5.10: Number of photons plotted against the number of electrons from a simulated tritium data set having only recombination fluctuations. Left: Tritium events propagated from 1 keV bins from 1-18 keV. Right, the tritium events propagated from 0.1 keV bins from 1-18 keV. The recombination fluctuations move events along the lines of constant energy.

As expected, we find that the recombination fluctuations events along the lines of constant energy. Figure 5.10 is analogous to the illustration shown for the case of

^{83m}Kr in figure 5.6. We now add in detector resolution (S1 and S2 fluctuations) and propagate events from every 0.1 keV bin along the tritium spectrum. The detector resolution included is that modeled for the LUX detector. The resulting spectrum is shown in figure 5.11.




Figure 5.11: The number of photons plotted against the number of electrons from a simulated tritium data set having recombination fluctuations and detector resolution. The tritium events propagated from 0.1 keV bins from 1-18.6 keV. Events move along the lines of constant energy due to recombination fluctuations and χ_{Det} .

The recombination fluctuations can now be extracted by plotting the photon and electron populations as a function of energy. The results from the simulation are shown in figure 5.12. By plotting the photon and electron spectra as a function of energy we immediately see the effect of recombination fluctuations and detector resolution χ_{Det} , as these cause the fluctuations up and down in constant energy.




Figure 5.12: Light yield (left) and Charge yield (right) of a simulated tritium spectrum, with the fit to the centroid show in black. The variance in number of photon and electrons per energy bin is used to extract recombination fluctuations. Recombination fluctuations and detector resolution χ_{Det} move events up and down along lines of constant energy.

Before extracting the recombination fluctuation from the simulated data shown in figure 5.12 we must consider the slopes of the population M . The slope of the n_γ was defined earlier in equation 5.36 purely as a function of the detector resolution. However, the slope of a continuous spectrum is effected by the functional form of light yield and charge yield as a function of energy. This issue is easily solved by fitting a line to the population. For this analysis we fit a quadratic to n_γ as a function of energy, the slope of the fit is by definition the value of M . Note, for the case of LUX the value of M is dominated by statistical fluctuations and using 5.12 is a good approximation.

The quadratic fit to both n_γ and n_e as a function of energy is used to subtract the centroid of the population calculating the variance in each energy bin. Making

the centroid subtraction removes the additional term due to slope and bin width given in 5.41, allowing us to only dealing with χ_{Det} without having to keep track of the energy bin size. Recombination fluctuations are calculated by binning the data in energy, extracting the raw variance in n_γ and n_e in each bin (identical to fitting a Gaussian) and subtracting off χ_{Det}^2 , following equation 5.46. The value of χ_{Det}^2 is explicitly,

$$\chi_{\text{Det}}^2 = M^2 \left((1.04 \pm 0.26 \sqrt{n_e})^2 + \left(\frac{6.6 \pm 0.9}{100} n_e \right)^2 \right) + (M - 1)^2 \left((3.45 \pm 0.17 \sqrt{n_\gamma})^2 + \left(\frac{6.4 \pm 1.7}{100} n_\gamma \right)^2 \right) \quad (5.47)$$

where the values of $\sigma_{n_{\gamma_{\text{Det}}}}$ and $\sigma_{n_{e_{\text{Det}}}}$ are taken from equation 5.17 and have been input to equation 5.36. Note, the value of M for the case of a continuous source should be taken from a fit to the slope of n_γ as a function of quanta (E/W). With equation 5.47 the value of χ_{Det} in every energy bin is known, since we measure the mean value of n_γ and n_e for each. The result of extracting recombination fluctuations is shown in figure 5.13 for various energy bin widths.

We find good agreement between the recombination fluctuations extracted from the simulation to the true value. Importantly, we have demonstrated that the concept of χ_{Det} that was introduced to deal with fluctuations in energy bins is in fact applicable to a continuous spectrum. We see in figure 5.13 that χ_{Det} reduces to the detector resolution of the best quanta collection channel (S2 for LUX). This is exactly what we found from the examples with the $^{83\text{m}}\text{Kr}$ line source and it is non-trivial that the result would hold for a continuous spectrum. Thus, as long as the recombination fluctuation are greater than χ_{Det} the value of recombination can




Figure 5.13: Simulated tritium spectrum with the detector resolution of the LUX detector and an initial guess of recombination fluctuations. The detector resolution, labeled as $\sigma_{S1_{\text{Det}}}$ and $\sigma_{S2_{\text{Det}}}$ is shown in red and green, respectively. The value of recombination fluctuation σ_R input into the simulation is shown in magenta. The methodology described in this previous section is applied to extract σ_R from the electron and photon spectrum, black and blue respectively. The statistical component of the standard deviation in each bin is χ_{Det} is shown in green, equation 5.47. The plots show cases for various bin widths in keV: a) $\Delta E = 0.1$ b) $\Delta E = 0.25$ c) $\Delta E = 0.5$ d) $\Delta E = 1$.

be determined to good precision event with a considerable error on χ_{Det} , as the two fluctuations add in quadrature. The analytic solution for extracting recombination outlined in this section is sufficient to first order. We are now ready to apply this method to the tritium calibration data.

5.5 Extracting Recombination Fluctuations from Tritium Calibration Data

In this section we apply the methods outlined in this chapter and use them to extract the recombination fluctuations from the tritium data. The first step in this process was calibrate the energy scale solving for g_1 and g_2 as outlined in 4. Second, the S1 and S2 signals of the tritium calibration data have been corrected for spectral shape, will be discussed in 6.1. Finally, having modeled and measured the statistical and instrumented variances for light collection of the LUX detector 5.17, 5.15, 5.16 the amount of fluctuations from detector resolution on each energy contour is known, 5.47.

The events from the tritium calibration data are binned in energy and the mean value of S1 and S2_b in calculated for each. The S1 and S2 signal is converted into mean number of photons and electrons using g_1 and g_2 . Using the mean number of photons and electro in each bin the value of χ^2_{Det} is calculated using 5.47. Next the total variance in the number of photons and electrons in each energy bin is calculated, defined as $\chi^2_{n_\gamma}$ and $\chi^2_{n_e}$ respectively. We then solve for $\sigma_{R_\gamma}^2$ and $\sigma_{R_e}^2$ given in equation 5.48,

$$\sigma_R^2 = \sigma_{R_\gamma}^2 = \sigma_{R_e}^2 = \chi^2 - \chi^2_{\text{Det}} \quad (5.48)$$

where the subscripts γ and e denote counting photon and electron respectively. The density plot of n_γ and n_e is shown in figure 5.14 and the value of recombination

fluctuation is shown in 5.15, for both the 170 V/cm and 100 V/cm data. The 170 V/cm data contained 140,000 tritium beta decays in the fiducial volume and the 100 V/cm data contains 4,500 events.




Figure 5.14: a: Density plot of number of photons vs. energy in keV using the tritium calibration data at 170 V/cm. b: Number of electrons vs. energy in keV using the tritium calibration data at 170 V/cm. c: Number of photons vs. energy in keV using the tritium calibration data at 100 V/cm. d: Number of electrons vs. energy in keV using the tritium calibration data at 100 V/cm. The data has been corrected for spectral shape. The black line indicates the quadratic fit to the centroid of the population.




Figure 5.15: The figures on the left in 5.15 (a: 170 V/cm, c: 100 V/cm) show the extracted recombination fluctuation σ_R from the light (black) and charge (red) channel denoted with subscript γ and e respectively, note they are identical. Also shown are the fluctuations in light collection $\sigma_{\gamma\text{Det}}$ (blue), charge collection $\sigma_{e\text{Det}}$ (cyan), and their manifestation in a combined energy bin, χ_{Det} (magenta). The figures on the right in 5.15 (b: 170 V/cm, d: 100 V/cm) show the mean and one sigma standard deviation of the measured number of photons (blue) and electrons (red). Also shown is the total quanta (in black) which is the sum of photons and electrons and the expected number of ions (magenta) and excitons (cyan) using $\alpha = 0.20$.

The figures on the left in 5.15 (a: 170 V/cm, c: 100 V/cm) show the extracted recombination fluctuation σ_R from the light (black) and charge (red) channel denoted with subscript γ and e respectively, note they are identical. Also shown is the

detector resolution of light collection $\chi_{\sigma_{\text{Det}}}$ (blue), charge collection $\sigma_{e_{\text{Det}}}$ (cyan), and their manifestation in an energy bin χ_{Det} (magenta). In regions where the measured recombinations are larger than the fluctuations from χ_{Det} any error from the constraint on g_1 and g_2 is negligible. At the higher energy bins the uncertainty grows the measurements become statistics limited, the size of the bin to bin fluctuations represent the statistical errors. The figures on the right in 5.15 (b: 170 V/cm, d: 100 V/cm) show the total quanta (black) which is the sum of the photons (blue) and electrons (red) and the expected number of ions (magenta) and excitons (cyan), using the exciton-to-ion ratio $\alpha = 0.20$. Since χ_{Det} is solved for in terms of photons and electrons the means of the of number of photons and electrons in each energy bin must be measured first.

5.5.1 Extracting Recombination fraction From Tritium Data

Having measured the mean number of photons and electrons in each bin the value of recombination probability r_p can be determined by taking the average value of recombination for a late number of events, $\langle r \rangle$ using equation 5.23. Figure 5.16 shows the measurement of the recombination fraction r for the 170 V/cm and 100 V/cm tritium calibration data. The shaded region represents the one sigma of the recombination fraction which can be thought of in terms of the recombination fluctuation, $\sigma_r = \sigma_R/n_{\text{ions}}$. We find that the bands converge below 4 keV_{ee} meaning that the light yields and charge yields also converge (discussed in section 6). Having extracted the apparent recombination probability r_p we can begin to think about recombination fluctuations in terms of a binomial process where an electron ion pair will either recombine or not, with a probability $r_p = \langle r \rangle$.




Figure 5.16: Recombination Fraction at 170 V/cm (black) and 100 V/cm (blue). The shaded regions represent the one sigma of the observed fluctuations in recombination fractions σ_r . The dashed line at 1.0 keV_{ee} represents the 50% detection threshold.

5.5.2 Modeling the ER Band

The ultimate goal of this section, and the tritium calibrations, are to be able to make predictions about WIMP sensitivity at various electric fields in the WIMP search energies of interest, 1-5 keV_{ee}. Having extracted recombination fluctuations from the data, we have the ability to reconstruct the electronic recoil band as it would appear with infinite detector resolution. Starting with the recombination fluctuations, which are fundamental to the xenon, the base ER band can be modeled by adding the resolution of any detector. Using the mean and width of the ER band

along with the mean of the NR band we can make predictions for WIMP background rejection.

The mean of the ER band in the commonly used discrimination variable $\log_{10}(S2_b/S1)$ can be written as a function of number of ions,

$$\log_{10}(S2_b/S1) = \log_{10} \left(\frac{(1-r)N_i}{(r+\alpha)N_i} \right) + \log_{10} \left(\frac{g_2}{g_1} \right) \quad (5.49)$$

where the observed charge and light signals $S2_b$ and $S1$ have been converted to recombination probability r , number of ions N_i and the exciton to ion ratio α using equations ??, ?? and 5.19. The variance of the band can be written as,

$$\text{Var}_{\log_{10}(S2_b/S1)} = \frac{1}{(\log(10))^2} \times \sigma_R^2 \left(\frac{-(\alpha+1)}{(1-r)(r+\alpha)N_i} \right)^2 \quad (5.50)$$

Which has been written in terms of the number of ions N_i , the recombination fraction r , and the measured recombination fluctuation σ_R , defined to be $\sigma_r \times N_i$. The result of the ER band's mean population and its corresponding 1 sigma fluctuation are shown in figure 5.17 for the case of 100 V/cm (blue) and 170 V/cm (black). This result shows the ER band with recombination fluctuations only. One can add light and charge collection fluctuations in quadrature to complete the modeling specific to any detector.




Figure 5.17: The result of the ER band's mean population and its corresponding 1 sigma fluctuation for the case of 100 V/cm (blue) and 170 V/cm (black). We find an overlap below 4 keV_{ee} where the additional strength of the drift field is higher improving threshold of discrimination. Above 4 keV_{ee} the bands separate as the higher drift field increased the charge extraction leading to better discrimination.

5.6 Measuring Alpha From the Tritium Data

Having measured r_p and σ_r from the tritium calibration data the exciton to ion ratio α determined by requiring that as the number of ions tends to one the recombination fluctuations tend to that of a binomial process. This is justified, as at low energies the handful of ion-electron pairs will either recombine or not with recombination

probability r_p . Such a process has a binomial variance written as,

$$\text{Var}_{\text{Bino}} = (1 - r_p)r_p N_i \quad (5.51)$$

where r_p is the recombination probability and N_i is the number of ions which can be thought of as the number of trials for the binomial process.

The value of alpha can be extracted by requiring that the recombination fluctuations σ_R^2 tend to that of equation 5.51. In figure 5.18 the y axis shows the ratio of the measured recombination standard deviation (σ_R) to the standard deviation of a purely binomial process. The plot on the right shows the expected binomial standard deviation on the x axis. The best alpha is one in which the observed standard deviation converges with that of a binomial process as the binomial variance tends to 1. The figure on the left has the number of ions available for recombination on the x axis. As the number of ions approaches one the standard deviation of recombination should become that of a binomial process. A single ion will either recombine or not with probability r . The extrapolation is made by fitting the lowest energy bins above 90% threshold 1.3 to 3 keV. Going below the value of one on the y axis implies that recombining electron-ion pairs have a variance better than binomial, which is unphysical if it is a random process. The best intercept converging to a purely binomial process is with $\alpha = 0.18 \pm 0.02$, consistent with the measurement in [68] and not 0.06 as used in [8]. Note, for this analysis we continue to use $\alpha = 0.20$ as measured by [68].




Figure 5.18: Determining the best α using the tritium calibration data, with $\alpha = 0.3$ (blue), 0.2(black), 0.15(magenta) and 0.06(red). Left: The y axis is the ratio of the measured standard deviation of recombination to that of a binomial processes and is plotted vs. the expected binomial standard deviation on the x axis. The best α is one for which the observed standard deviation converges with that of a binomial process as the binomial variance tends to 1. Right, the same y axis as on the left but plotted vs. the number of ions available for recombination. As the number of ions approaches one the standard deviation of recombination should become that of a binomial process. A single ion will either recombine or not with probability r . The best intercept converging to a purely binomial process (black star) is with $\alpha = 0.20$. Falling below the value of one on the y axis implies that recombining electron-ion pairs have a variance better than binomial, which is unphysical if it is a random process. Note, the fits use only data above 90% threshold at 1.3 keV, starting from the third data point from the left. The higher end cut off at 3 keV corresponds to the end of the fitted lines.

5.7 Extracting Recombination Fluctuations from ^{137}Cs Calibration

To expand the picture of recombination fluctuation to higher energies the same method used for the tritium calibration was applied to Compton scatters from an external ^{137}Cs calibration source. The ^{137}Cs source provides ER calibration data from the backscatter peak around 150 keV to the photo peak at 662 keV. Figure 5.19 on the left shows the measured mean number of photons, electron in each energy bin along with their one sigma fluctuation (shaded). The number of excitons and

ions are also shown assuming an $\alpha = 0.20$. Once the mean number of photons and electrons are measured the recombination probability is determined and plotted on the right in figure 5.19. The inflection around 662 keV is due to the sharp rise and fall of the photo peak skewing the measurement of number of photons and electrons.




Figure 5.19: Left: The mean and one sigma standard deviation of the measured number of photons (blue) and electrons (red). Also shown is the total quanta (in black) which is the sum of photons and electrons and the expected number of ions (magenta) and excitons (cyan) using $\alpha = 0.20$. Right: The recombination probability r (solid black) and the one sigma fluctuation σ_r (shaded).

5.8 Recombination Fluctuations, The Bigger Picture

We have now measured the recombination probability and fluctuation over a wide range of energies and at two fields for tritium (100 and 170 V/cm). The calibrations range from the 1.0 keV 50% threshold with tritium to about 700 keV with the ^{137}Cs calibration, and include the line sources used for the energy calibration in 4 and table 4.1. Also shown is a ^{57}Co calibration at a variety of electric fields ranging from 60 to 5000 V/cm from [8]. Figure 5.20 on the left shows the observed recombination fluctuation σ_R measurements vs. the standard deviation expected from a purely binomial process (equation 5.51). At our field of 100 and 170 V/cm we find good agreement with a simple power law fit which can be thought of as the fluctuation receiving an amplification over the underlying binomial process.

Figure 5.20 on the right shows the measured recombination fluctuation σ_R vs. the number of ions available for recombination N_i . The x axis is chosen to be number of ions as recombination fluctuations only act on ions and not excitons, the conversion to energy on the x axis is simply $E = W \times n_i(1 + \alpha)$. It is found that the measured recombination fluctuation can be well described by a generic power law fit.

There is no physical justification for the power law fit in figure 5.20 to the apparent binomial variance amplification. This is a fatal flaw in our recombination model. The observed variance is off by roughly a factor of N_{ions} , surpassing 1000 in the cesium data. The recombination model we are using assumes that every electron ions pair is a separate entity that either recombines or not, thus the variance should




Figure 5.20: Recombination fluctuations measurements are labeled on the plot and include data from tritium at 170 V/cm, tritium at 100 V/cm, ¹³⁷Cs calibration, the line sources used for the energy scale calibration listed in table 4.1 and a ⁵⁷Co calibration at a variety of electric fields ranging from 60 to 5000 V/cm from [8]. Left: the observed recombination fluctuation σ_R measurements vs. the standard deviation expected from a purely binomial process, equation 5.51. The red curves represent power law fits to the data.

be binomial. There is another issue with the recombination model. The decline of recombination probability as the energy (or N_i) tends to zero is unphysical yet is observed in the data, shown in figure 5.16. The probability of the electron ion-pair to recombine should be insensitive to the number of electron-ion pairs created by the energy deposit. We take this as a first clue to solving the problem and outline the consequences of allowing for electron-ion pair mixing with some encounter probability in the next two sub-sections. Culminating with a model which will simultaneously reproduce the observed recombination fraction, explain its decline at low energy, and yield the correct recombination fluctuations.

5.8.1 Encounter Recombination Probability

The apparent amplification of the observed ER fluctuations over that of a binomial process is troubling. What causes the ER events to be so erratic over their nuclear recoil counterparts which can be well described by binomial fluctuation [8]. The key difference that lends a clue to solving the puzzle is that for a given energy nuclear recoils will produce significantly less electron-ion pairs. It has been observed for ER events that as the number of ions goes to one the recombination fluctuations do indeed become more binomial like, shown in figure 5.18. Before we move on, it should be noted that the method outlined in this section will succeed in explaining recombination fluctuation but will fail to reproduce the recombination fraction probability. This subsection is meant as a discussion on encounter recombination probability leading to the model in the subsection to follow that will bridge measuring recombination probability and its variance.

In order to tackle the issue of recombination fluctuations we introduce encounter recombination probably r_ϵ . The term r_ϵ will couple freed electrons to ions other than its own. There are well motivated arguments to be made to include a r_ϵ term. First, in liquid xenon the decay of the calibration source ^{83m}Kr has been observed to receive an enhancement of several percent in the light yield of the second 9.4 keV following the first decay of 32.1 keV [67]. This can be attributed to the second decay occurring surrounded by a ball of charge from the first decay resulting in an enhanced encounter recombination probability and increased light yield. The two ^{83m}Kr decays are separated by a half life to 154 ns [11]. The shorter the

timing separation between the two decays the greater the light yield enhancement, with light yield enhancement observed past 1000 ns [9], [10] [LUX data shows enhancement out to 2000 ns could add this plot to thesis...]. This lends evidence that freed electrons can be attracted to ions while diffusing from the interaction site on the time scales of hundreds of nano seconds. Further, the idea of encounter probability was worked out by Mozumder noting the need for encounter recombination probabilities of 0.01 in order to explain the ion production rate in liquid xenon [74].

In this subsection we set out to model the variations resulting from the recombination porbability containing a component from encounter recombination probability. Once the model for variance is worked out we can fit to the tritium and ^{137}Cs calibration data in order to extract the encounter recombination probability.

First we start with expression for the variance of a binomial process with some recombination probability r .

$$\text{Var}_r = (1 - r)rN_i \quad (5.52)$$

where N_i is the number of ions, and can be thought of as the number of trials. Next we split the total observed recombination probability into two components.

$$r = r_s + r_e \quad (5.53)$$

where r_s is the self recombination probability and r_e is the encounter recombination probability. The expectation is that $r_e \ll r_s$, [74]. The variance resulting from the two terms can be considered as independent processes occurring subsequently thus,

the total variance of the process is the sum of the individual variances for each ion. In our modeling every free electron has an average encounter probability r_ϵ with each ion. The variance of N_i ions is given in equation 5.52 and is illustrated in figure 5.21 for the case of one ion. At this point we pause to point out the fatal fall in this theory. As the number of ion-electron pairs grow all ions will recombine as each ion is unable to avoid recombination from the bombardment of N_i electrons. The observed recombination probability for this process becomes $r = r_s + N_i \cdot r_\epsilon$. With that note, we proceed to learn more about encounter recombination probability.




Figure 5.21: Illustration of recombination for the case of a single ion, blue box highlighted in black at the center. The total recombination probability for one ion is the combination of the dominant self recombination probability (r_s) plus the sum of the encounter recombination probabilities (r_ϵ) from the remaining $N-1$ electrons. This process is repeated for N number of ion-electron pairs leading to a binomial variance given in equation 5.54.

$$\text{Var}_{(r_s+r_e)} = \sum_{N_i} r_s(1-r_s) + \sum_{N_i} r_e(1-r_e)(N_i - 1) \quad (5.54)$$

the first term of equation 5.54 is the binomial variance of a single electron-ion pair with some probability r_s to recombine. The second term is the binomial variance of the average encounter probability r_e with all other electrons excluding its own escaped electron, a total of $(N_i - 1)$. Both terms are summed over all possible ions, accounting for all N_i ions with $N_i - 1$ electrons available to be encountered. Due to the relatively slow mobility of ions vs electrons, the ions are treated as fixed with the freed electrons having some probability of encounter an ion. This is a simplistic model that treats the encounter probability as an overall average for all electron-ion pair combinations. Assuming N_i is large equation 5.54 can be simplified to

$$\text{Var}_{(r_s+r_e)} = r_s(1-r_s)N_i + r_e(1-r_e)N_i^2 \quad (5.55)$$

The result of splitting the recombination probability into self and encounter recombination is subtle, yet has huge implications. Comparing equation 5.52 to 5.54 we find that the binomial variance of the process with encounter recombination probability will grow like N_i^2 as opposed to the binomial variance of a self recombination process that grows like N_i . To better understand the amplification of the binomial fluctuation observed in the data, figure 5.20, we define an amplification term as the ratio of the binomial variance with encounter recombination probability to that of a binomial process with self recombination probability r .

$$\mathcal{A} = \frac{\text{Var}_{(r_s+r_\epsilon)}}{\text{Var}_r} \quad (5.56)$$

$\text{Var}_{(r_s+r_\epsilon)}$ and Var_r are given in equations 5.55 and 5.52, respectively.

We will treat two cases. First, we will assume that $r_\epsilon \ll r_s$ and $r_s \simeq r$. Second, we will hold the ratio of r_ϵ/r to be a constant. The second case is motivated by the idea that electric field and energy dependance that governs self recombination probability also applies to encounter recombination probability.

The amplification of the binomial variance from equation 5.56 is

$$\mathcal{A} = \frac{N_i}{N_i} \left(\frac{r_s(1-r_s) + r_\epsilon(1-r_\epsilon)N_i}{r(1-r)} \right) \quad (5.57)$$

assuming that $r_\epsilon \ll r_s$ and $r_s \simeq r$ equation 5.57 can be simplified to,

$$\mathcal{A} = \left(1 + \frac{r_\epsilon(1-r_\epsilon)N_i}{r(1-r)} \right) \quad (5.58)$$

Using equation 5.58 the value of encounter recombination probability r_ϵ can be extracted from the tritium and ^{137}Cs data using N_i , r , and \mathcal{A} . Where the value of binomial amplification \mathcal{A} is the extracted from the data defined as the recombination fluctuation σ_R over σ_R -binomial, shown in figure 5.20. The result of extracting encounter recombination probability r_ϵ is shown in figure 5.22. The overall average of r_ϵ from the calibration data is $r_\epsilon = 0.0042$ varying from 0.002 to 0.007, in good agreement with Mozumder [74].




Figure 5.22: Encounter recombination probability r_e extracted from the tritium and ^{137}Cs calibration data at 170 V/cm, derived from equation 5.58. Left plotted vs. energy in keV. Right: plotted vs. the number of ions. The solid red line represents the overall average of r_e from the calibration data is $r_e = 0.0042$.

Taking the value of r_e as a constant of 0.0042 we find that the observed ER recombination fluctuations are infact consistent with binomial fluctuation at our field of 170 and 100 V/cm, shown later in figure 5.24. This is a step in the right direction for understanding recombination fluctuations. Our data is limited to only two electric fields at 100 and 170 V/cm making it difficult to model field dependance. To expand the model we include data from Dahl, using a ^{57}Co source with fields ranging from 60 to 5000 V/cm [8]. We then extract r_e , shown as the black points in figure 5.23.

The data from Dahl, shown in 5.23, provides good motivation to proceed with our second assumption, modeling the ratio of r_e/r to be a constant. There appears to be correlation between the recombination probability and the encounter recombination probability. This correlation is sensible, considering that as the electric field is increased the freed electrons can escape the ions more readily. Thus, both




Figure 5.23: The red crosses indicate the total recombination probability r for the calibration sources labeled on the plot. The black x-s indicate the encounter recombination probability r_ϵ . The blue points indicate the tritium data at 100 V/cm. The data includes tritium at 170 V/cm (black), tritium at 100 V/cm (blue), ^{137}Cs and data from Dahl for ^{57}Co ranging from 60 to 5000 V/cm [8].

the self and encounter recombination decline as a function of applied electron field. However, the assumption that r_ϵ and r are always correlated is to be taken with a grain of salt, and is not supported by the tritium data. Between 2.5 and 10 keV the recombination probability r and r_ϵ become anti-correlated. With that caveat mentioned, we proceed with the second case.

$$r_\epsilon = r_{\epsilon_0} + \mathcal{C}r \quad (5.59)$$

where \mathcal{C} is a constant linking the observed recombination probability r to r_ϵ . The best fit for both cases is show in figure 5.24. Case one, is with a global average of $r_\epsilon = 0.0042$, extracted from the tritium and ^{137}Cs data. Case two, is using $r_\epsilon =$

$0.0011 + 0.006r$. The fit for case one is within 30% when excluding the data from Dahl, and the fit for the second case deviates less than 30% from all of the data.




Figure 5.24: Fit to the observed ER recombinations modeled with binomial fluctuations where the recombination probability includes an encounter recombination component. Case one (red) is with a constant $r_{\epsilon} = 0.0042$, extracted from the tritium and ^{137}Cs data. Case two (green) is using $r_{\epsilon} = 0.0011 + 0.006r$. Case two is in significantly better agreement with the field dependent data from Dahl, as it accounts for field effects [8]. The fit for case two (green) deviates less than 30% from the data.

The simplistic model outlined in this subsection demonstrates that recombination fluctuation of electronic recoil can indeed be the result of binomial statistics. Remarkably, a sub 1% component of recombination probability r from encounter probability leads to an apparent binomial amplification factor of 1 to 1000, shown in the right plot in figure 5.20. The tiny value of r_{ϵ} makes a significant impact when considering the combinatorics of the recombination process. This is good progress, we have found a simplistic way to gain an additional factor of N_i for recombination

fluctuations vs. energy, instead of applying a power law fit to binomial amplification with no physical basis. However, we still need to clean up the problem of total ion-electron pair recombination resulting from such a model. We have found the correct variance but have failed to produce the correct recombination probability. The fix the the issue is described in the next subsection.

5.8.2 Clusters of Encounter Recombination

In the previous subsection a model was introduced which could predict the observed recombination fluctuations of ER events. However, the model failed to produce the correct recombination probability. We found that as the energy (or N_i) rises the recombination probability tends to 1, and beyond as $r = r_s + N_i \cdot r_\epsilon$. The zeroth order correction to this problem is to remove the self recombination term r_s and treat self recombination as just another encounter recombination probability, $r_s = r_\epsilon$. Next, we will not allow the full N_i electrons to interact with each ion as this is unphysical. Instead, only a small number of the total ions N_ϵ are allowed to have encounters for each energy deposit, illustrated in 5.21. The value of N_ϵ is set by the requirement to reproduce the correct recombination probability. The observed recombination probability r is then written as,

$$r = r_s + (N_\epsilon - 1)r_\epsilon = N_\epsilon r_\epsilon \quad (5.60)$$

where r is the observed recombination probability at a given energy, r_s is the self recombination probability taken to be equal to the encounter recombination prob-

ability r_ϵ , and N_ϵ is the average number of electron encounters for each ion. Note that the value of N_ϵ used is an average and can also vary event to event depending on the ER track geometry. The equation for recombination probability in 5.60 inadvertently solves the other issue of the original recombination model. By taking self recombination r_s to be equal to r_ϵ the observed recombination probability r naturally vanishes as the energy (N_i) tends to zero, as observed in the data shown in figure 5.16. The binomial variance of such a process is,

$$\text{Var}_{N_\epsilon r_\epsilon} = (1 - N_\epsilon r_\epsilon)N_\epsilon r_\epsilon N_i \quad (5.61)$$

where $\text{Var}_{N_\epsilon r_\epsilon}$ is the recombination variance in number of quanta squared for a given interaction with N_i ions, encounter recombination probability r_ϵ with an average of N_ϵ encounters. The variance derived in equation 5.61 grown like N_i and suffers the same problem as our original model. The variance needs to grow like N_i^2 to explain the observed recombination fluctuations in the data, as seen in the previous subsection.

The solution is surprisingly straight forward considering the two requirements. First, in order to maintain the correct recombination probability r the value of encounter interactions N_ϵ must be fixed. Second, to get the additional factor of N_i in variance the process of size N_ϵ must be repeated, with the repetition scaling like N_i . Remember, we are only allowing a small number, N_ϵ , of the total ions available ions to have encounters. Consider the picture in figure 5.21 but for N_ϵ encounters repeated N_i/N_ϵ times in N_C clusters, such a process is illustrated in figure 5.25. The

number of clusters N_C for a given in equation 5.62.

$$N_C = N_i/N_e \quad (5.62)$$




Figure 5.25: Illustration of recombination for the case of a single ion occurring in N_C clusters with number of encounters N_e . The total recombination probability for all clusters is equal to the average of each individual cluster. In each cluster the recombination probability is equal to the number of interaction N_e times the encounter recombination probability r_e . The process is repeated for N_C clusters, resulting in an additional factor N_i in the observed variance while maintaining the correct average recombination probability r .

The value N_e can be considered as the number of encounters per cluster, or generalized to the number of ions per cluster. The electron-ion pairs have N_e encounters in each cluster with an average recombination probability r . The total recombination probability of the system of clusters remains equal to $r = N_e r_e$ without blowing up to unity. Since the process is repeated for N_C clusters the variance from equation 5.61 becomes

$$\text{Var}_{N_\epsilon r_\epsilon} = \sum_{N_C} (1 - N_\epsilon r_\epsilon) N_\epsilon r_\epsilon N_i \quad (5.63)$$

where $\text{Var}_{N_\epsilon r_\epsilon}$ is the recombination variance of equation 5.61 for N_C clusters. Plugging in the value of N_C from equation 5.62 we find

$$\text{Var}_{N_\epsilon r_\epsilon} = (1 - N_\epsilon r_\epsilon) N_\epsilon r_\epsilon N_i (N_i/N_\epsilon) \quad (5.64)$$

where the multiplication by N_C is written as N_i/N_ϵ . Further simplifying equation 5.64 and recalling that $r = N_\epsilon r_\epsilon$,

$$\begin{aligned} \text{Var}_{N_\epsilon r_\epsilon} &= (1 - N_\epsilon r_\epsilon) r_\epsilon N_i^2 \\ \text{Var}_{N_\epsilon r_\epsilon} &= (1 - r) r N_i^2 \\ \text{Var}_{N_\epsilon r_\epsilon} &= (1 - r) r \frac{N_i^2}{N_\epsilon} \end{aligned} \quad (5.65)$$

Taking the result for variance given in equation 5.65 we calculate the apparent amplification factor \mathcal{A} over that of a binomial process, our original recombination theory.

$$\begin{aligned} \mathcal{A} &= \frac{\text{Var}_{N_\epsilon r_\epsilon}}{\text{Var}_r} = \frac{(1 - r)r N_i^2 / N_\epsilon}{(1 - r)r N_i} \\ \mathcal{A} &= \frac{N_i}{N_\epsilon} = N_C \end{aligned} \quad (5.66)$$

the amplification factor in equation 5.66 remarkably simple. The amplification over that of a binomial process reduces to the number of clusters N_C . Note, the value of N_ϵ is the ratio of r/r_ϵ . This is the key to the cluster recombination model. We have picked up an additional factor N_i in the variance while maintaining the

correct average recombination probability r . Using equation 5.66 we can extract the encounter probability r_e , the cluster size N_e and the number of clusters N_C from our calibration data.





Figure 5.26: The data shown are labeled on the plot and include calibration data of tritium data at 170 V/cm, tritium at 100 V/cm (blue), ^{137}Cs at 170 V/cm, and data over a range of fields using ^{57}Co from Dahl. [8]. a) The encounter recombination probability r_e plotted vs. energy. b) The number of encounters per cluster N_e vs. energy. c) The number of clusters N_C vs. energy.

The average value of r_ϵ over a wide range of energies is found to be 0.01, shown in figure 5.26. The value of r_ϵ determined by using the cluster-encounter model is consistent with that noted by Mozumder in 1995 to explain ion fluctuations in liquid xenon [74]. The value of $r_\epsilon = 0.01$ corresponds to an average value of encounters $N_\epsilon = 50$, which can be thought of as the cluster size. Figure 5.26 (c) also shows the number of clusters vs. energy, which grows linearly with N_i . Note, the number of clusters is the factor by which the binomial fluctuations of the initial self-recombination model are amplified at a given energy.

The number of interaction per cluster N_ϵ , or cluster size, sets the encounter probability r_ϵ in this model. The value of N_ϵ appears to level off around 25 as energy tends to zero and varies between 25 to 100 over a wide range of energies and electric fields. Both r_ϵ and N_ϵ appear to have dependencies on electric field and energy. Which is not surprising as both the track geometry and interaction probability depend on electric field and energy. The data from Dahl over the range of 60 to 5000 V/cm especially illuminates the dependancies vs. field. We proceed with a simplified model and take the value of N_ϵ as a constant. The average value of $N_\epsilon = 50$ is used to calculate the expected variance at each energy using equation 5.66, the result is shown in figure 5.27.




Figure 5.27: The measured recombination fluctuations from calibration data vs. energy. The red line indicates the expected recombination fluctuation of the clustering model given in equation 5.65, assuming a cluster size of $N_\epsilon = 50$. $N_\epsilon = 50$ corresponds to an average encounter recombination probability of $r_\epsilon = 0.01$. The deviation of the fit from the data is less than 30%. The ^{83m}Kr data point is the result of two decays and is expected to fall below the curve by 40% considering the sum of the individual variances.

Figure 5.27 shows the measured recombination fluctuations vs. energy along with that expected for the clustering model given in equation 5.65, for $N_\epsilon = 50$, corresponding to an average encounter recombination probability of $r_\epsilon = 0.01$. The deviation of the fit and the data is less than 30%, including Dahl's data taken at electric fields ranging from 60 to 5000 V/cm.

With the clustered encounter model we can now account for the recombination

fluctuations and understand why recombination tends to zero as energy tends to zero. Having explained these two PEnomena, the clustered-encounter model is a significant improvement over the initial recombination model. The flaw of the initial recombination model was to assume that the recombination process only consists of ion-electron pair self recombination. This assumption is what lead to the observed variance being off by a factor of N_i and the inability to explain the dependance of recombination fraction as energy tends to zero. The picture of ER band mean and width is now complete, only requiring recombination fraction r as an input. The recombination fraction is dependent on the energy of the interaction and electric field, modeled by Thomas & Imel [71] and NEST [58], [59].

5.9 Conclusion

There have been numerous steps in this section culminating in expanding our knowledge by extracting as much information as possible from the calibration sources. We have measured that the best exciton to ion ratio α measured to be 0.20 in the WIMP search energies, which is consistent with the measurement from [68]. The value of alpha was constrained by extrapolating the recombination fluctuations from the tritium data from 3 to 1.2 keV and requiring that for a single ion-electron pair the fluctuation be purely binomial, shown in figure 5.18.

The recombination model presented in this section can be used to predict the ER band for any xenon detector, as shown in figure 5.17. The most critical results are those specific to our WIMP search, 10-100 GeV WIMPs, which are focused in the range of 1 to 5 keV_{ee} and well covered by the tritium calibration data. Having extracted the values of r and σ_r for ER events the generic mean and band widths can be determined. Thus, the ER band shape can be determined for any xenon detector with the application of the additional variance from the specific detector resolution. The knowledge of this band shape can be used to make predictions about the background rejection power of a given experiment.

It is surprising to find that changing the drift field from 100 V/cm to 170 V/cm had only an epsilon impact on the mean of ER band below 4 keV_{ee}, figure 5.17. Further, there was no impact on the energy threshold since the light and charge yields merge at the threshold of 1 keV. A more dramatic field dependance was expected from [8] and [59]. However, the low energy region never been probed

to such high precision as with the tritium calibration using the LUX detector. To expand upon the modeling at low energies it will be useful for the next science run using the LUX detector to take tritium calibration data at a verity of fields. This will allow us to predict exactly how much additional NR and ER discrimination can be achieved by increasing the field.

Chapter 6: Ionization and Scintillation Yield from Tritium Calibration

In this section we measure the scintillation yield and ionization yield from the tritium calibration data. Using the measurement of gains g_1 and g_2 in chapter 4, the average number of photons, electrons and the corresponding combined energy can be determined. With that information the light yield (n_γ/keV) and charge yield (n_e/keV) are extracted from tritium data down to 1 keV_{ee}. Before the yields can be measured, the effect of finite detector resolution convolved with the tritium spectral shape must be accounted for. Detector resolution was characterized in chapter 5 and will be used to model the smeared tritium spectra as observed by the LUX detector. Once the spectral shape has been corrected we report the values of light yield and charge yield measured at 170 and 100 V/cm. The results are compared to two recent measurements for light yield in the keV range using Compton scatters. This provides a crucial cross check that the ER band calibration using the tritium beta source is valid for use with the more generic backgrounds found in WIMP search data consisting of Compton scatter from high energy gammas. At low energies the light yields and charge yields from betas and gammas are expected to be identical [58] [59].

6.1 Correcting for Spectral Shape for Finite Resolution

The distribution of tritium events convolved with the detector's finite resolution for S1 (scintillation) and S2 (ionization) causes the observed mean of the spectra to shift from the actual mean. The shift is non-trivial and depends on the spectral shape and the functional form of the resolution over a range of energies. A large negative derivative of the spectral shape will tend to pull the observed mean to lower values, and a large positive slope will pull the observed mean to higher values. Figure 6.1 and equations 6.2 and 6.4 demonstrate a simple model to solve for the relation between observed mean and actual mean. Consider, for example, a linearly declining distribution. Starting with infinite detector resolution we set up bins of width Δx . To account for finite energy resolution we distribute the counts in each rectangular bin into Gaussians centered at μ_i , with a spread of σ_i , and normalized to the area of the bin $N_i \times \Delta x$ with amplitude c_i . Each rectangular bin(i) can be written as a Gaussian $G(i)$:

$$c_i = \frac{N_i \times \Delta x}{\sigma_i \sqrt{2\pi}}$$

$$G_i(x) = c_i \times \exp\left(\frac{-(x - \mu_i)^2}{2\sigma_i^2}\right)$$
(6.1)

where N_i is the count in the i^{th} bin, Δx is the bin width, μ_i is the bin center and σ_i is the resolution at the i^{th} bin. Figure 6.1 show the application of equation 6.1 to a

linear energy distribution with a \sqrt{E} dependent σ . The observed distribution is the sum of the Gaussians, shown in red.




Figure 6.1: Smearing of a linear spectrum using equation 6.1 for an arbitrary energy scale E . The counts in each shaded bin are redistributed into normalized Gaussians (blue) with the resolution σ_E growing like \sqrt{E} . The spectrum smeared with detector resolution is the sum of Gaussians shown in red.

6.1.1 Calculating the Observed Energy

After modeling the finite resolution with Gaussians the mean observed at each bin can be calculated from the overlap of all bins weighted by the corresponding means. We can write the observed mean in the i^{th} bin, ν_i , in terms of the bin centers μ and

overlapping areas of all bins using the normalizations c_i from equation 6.1:

$$\nu_i = \frac{\sum_{j=1}^n \mu_j \int_{\mu_i - \frac{\Delta x}{2}}^{\mu_i + \frac{\Delta x}{2}} G_j(x) dx}{\sum_{j=1}^n \int_{\mu_i - \frac{\Delta x}{2}}^{\mu_i + \frac{\Delta x}{2}} G_j(x) dx} \quad (6.2)$$

Equation 6.2 can be solved in terms of error function and complimentary error function. First we will generalize a formula to solve for the overlapping area from the j^{th} bin into the i^{th} bin.

$$A_{i,j} = \int_{\mu_i - \frac{\Delta x}{2}}^{\mu_i + \frac{\Delta x}{2}} G_j(x) dx = \begin{cases} c_i \operatorname{erf}\left(\frac{\Delta x}{\sigma_i \sqrt{2}}\right), & j = i \\ \frac{c_j}{2} \operatorname{erfc}\left(\frac{|\mu_j - \mu_i| - \frac{\Delta x}{2}}{\sigma_j \sqrt{2}}\right) - \frac{c_i}{2} \operatorname{erfc}\left(\frac{|\mu_j - \mu_i| + \frac{\Delta x}{2}}{\sigma_j \sqrt{2}}\right), & j \neq i \end{cases} \quad (6.3)$$

The error function and complementary error function are defined in equation 6.4 and the coefficient c_i is defined in equation 6.1.

$$\begin{aligned} \operatorname{erf}(x) &= \frac{2}{\sqrt{\pi}} \times \int_0^x \exp(-t^2) \\ \operatorname{erfc}(x) &= \frac{2}{\sqrt{\pi}} \times \int_x^\infty \exp(-t^2) = 1 - \operatorname{erf}(x) \end{aligned} \quad (6.4)$$

As μ approaches zero the Gaussian distribution of equation 6.1 begins to spill over into negative values, which in some cases may be unphysical. For instance, the Gaussian assumption leads to negative photons. We can chose to ignore this area or make the distribution more Poisson-like by bouncing the Gaussian back at $\mu = 0$.

The formula for accounting for the area of the reflected Gaussian is described in 6.5. Ultimately this assumption has little impact on the S1 and S2 analysis because the threshold cut off well before the zero interface is reached, but it does make the distributions more Poisson like near the zeroth bins. Equation 6.5 is the same as 6.3 with the bin center μ_i mapped to $-\mu_i$.

$$B_{i,j} = \frac{c_j}{2} \operatorname{erfc} \left(\frac{|\mu_j + \mu_i| - \frac{\Delta x}{2}}{\sigma_j \sqrt{2}} \right) - \frac{c_j}{2} \operatorname{erfc} \left(\frac{|\mu_j + \mu_i| + \frac{\Delta x}{2}}{\sigma_j \sqrt{2}} \right) \quad (6.5)$$

Finally, we solve for the observed mean in the i^{th} bin by summing all the Gaussian overlaps $A_{i,j} + B_{i,j}$ (equations 6.3,6.5), weighting the overlapping area from each bin by the corresponding bin center μ_j . The result is shown in equation 6.6 and is equivalent to equation 6.2 when the area from the reflected Gaussian is not considered, $B_{i,j}=0$.

$$\nu_i = \frac{\sum_{j=1}^n \mu_j \cdot (A_{i,j} + B_{i,j})}{\sum_{j=1}^n (A_{i,j} + B_{i,j})} \quad (6.6)$$

6.1.2 Smearing a Toy Spectrum

To demonstrate the application of equation 6.6 we use it to smear a toy linearly decaying spectrum. By modifying the dependence of σ_i on μ_i we can better understand the effects of the spectral shape and the functional form of the resolution.

Figure 6.2 shows the effect of the finite resolution on a linearly decaying spectral shape. Using a constant resolution σ the observed mean, when accounting for finite resolution, shifts down due to the spectral shape. In the case with $\sigma_i \sim \sqrt{\mu_i}$

the observed mean at first shifts higher as the increasing width at higher value bin centers, even with lower counts, out weighs the lower bin centers with higher counts and narrower widths. In both cases as the bin centers approach zero the observed mean shifts higher due to an imposed threshold at zero, where Poisson statistics take over and the Gaussian characterization leads to a loss of events below zero. Thus, for the sake of the toy model in figure 6.2 we only characterize the relation between the real mean and the observed mean from the second bin center. It is also worth mentioning that for the case of having a varying resolution in figure 6.2 the shift in spectral shape seems minor, yet there is a significant 20% deviation in the observed mean of the last bin.




Figure 6.2: Top Left: A linearly decaying spectrum, in blue. The black curve represents the sum of the Gaussians assuming a constant resolution. Top Right: A linearly decaying spectrum, in blue. The black curve represents the sum of the Gaussians with a $\sigma = \sqrt{E}$ dependent resolution. Bottom: The observed mean, with finite resolution, compared to the real mean with infinite resolution. The black points are for the case with linear resolution and the red points represent the case with $\sigma = \sqrt{E}$ dependent resolution.

6.2 Light Yield, Charge Yield and Comparison to NEST Modeling

The first attempt to remove the effect of the tritium spectral shape and finite detector resolution is to use the NEST model from [59]. NEST stands for Nobel Element

Simulation Technique. NEST fits data from previous xenon and argon detectors to recombination models, producing predictions for light yield n_γ/keV and charge yield (n_e/keV) at a verity of electric fields, energy deposits, and particle types. We take the light yield (LY) and charge yields (QY) from NEST and convolve it with the known tritium energy spectrum to produce the S1 and S2 spectra. The S1 and S2 spectra are then smeared with detector resolution and recombination fluctuations, determined in chapter 5. It is found that the S1 and S2 spectra from the data deviate from the NEST model making it difficult to reverse-engineer the effect of smearing. However, taking the NEST model to be correct within 20%, the spectral shape correction is calculated and found to be small. We proceed to extract LY, QY and recombination fluctuation (σ_R) from the tritium data without any correction producing a model that is more accurate than NEST. It should be noted that the NEST model which has not been confirmed at our electric field and energy.

6.2.1 Tritium S1 and S2 vs. NEST

The spectral shape correction for the mean of the observed S1 and S2 signal from tritium beta decay can be found using equation 6.6. We start with NEST to get the expected S1 and S2 tritium spectrum. The variance of S1 and S2 arise from recombination fluctuations and detector resolution (statistical and instrumental fluctuations), given in equation 5.15 and 5.16. We use equations 6.7 and 6.8 to smear the photon and electron yields, essentially putting in detector resolution by hand. Then, by applying equations 6.1-6.6 the S1 and S2 bin centers after smearing can

be mapped back to the true bin centers before smearing.

$$\begin{aligned}\sigma_{S1_R}^2 &= g_1^2(\sigma_R^2) \\ \sigma_{S1_{\text{Det}}}^2 &= g_1^2(\sigma_{n_{\gamma_{\text{stat}}}}^2 + \sigma_{n_{\gamma_{\text{inst}}}}^2) \\ \sigma_{S1}^2 &= \sigma_{S1_R}^2 + \sigma_{S1_{\text{Det}}}^2\end{aligned}\tag{6.7}$$

$$\begin{aligned}\sigma_{S2_R}^2 &= g_2^2(\sigma_R^2) \\ \sigma_{S2_{\text{Det}}}^2 &= g_2^2(\sigma_{n_{e_{\text{stat}}}}^2 + \sigma_{n_{e_{\text{inst}}}}^2) \\ \sigma_{S2}^2 &= \sigma_{S2_R}^2 + \sigma_{S2_{\text{Det}}}^2\end{aligned}\tag{6.8}$$

where g_1 and g_2 are the gains to convert S1 and S2 to number of photons and electrons, respectively. The values $\sigma_{S1_R}^2$ and $\sigma_{S2_R}^2$ are the variances in S1 and S2 given only recombination fluctuations, this is what a detector with infinite resolution would observe. Detector resolution is comprised of the statistical and instrumental fluctuations in light and charge collection written as $\sigma_{S1_{\text{Det}}}^2$ and $\sigma_{S2_{\text{Det}}}^2$, this would be the resolution given no recombination fluctuations. The total variance in S1 and S2 observed by a detector with finite resolution is the sum of recombination fluctuations and detector resolution $\sigma^2 = \sigma_R^2 + \sigma_{\text{Det}}^2$ (where the subscripts S1 and S2 have been removed). The use of Gaussian sigma down to low S1 is an acceptable approximation since the underlying distribution actually consists of the number of photons, $n_\gamma = \frac{S1}{g1}$. With $g1=0.097$ there are still 30 photons near the S1 threshold of 3 PE. The S2 threshold for golden events is around 400 PE, $n_e = \frac{S2}{g2}$. With $g2=5.75$ there are still 70 electrons near the lower end of the tritium spectrum.

Figure 6.3 (a,c) shows the application of smearing from equation 6.7 and 6.8

to the expected S1 and S2 tritium spectrum, respectively, overlaid with the data. The mapping of the observed S1 and S2 mean values to the real S1 and S2 mean values is shown in 6.3 (b,d). The mapping from observed mean to real mean is the result of starting with infinite resolution containing recombination fluctuations only (σ_R) and applying the model as outlined in 6.1 with detector resolution (σ_{Det} of equations 6.7 and 6.8).

We find that that peak location of the S2 spectra from NEST deviates from the data by up to 20%. These discrepancies maybe arising from the error in g_1 and g_2 which could systematically shift light yield and charge yield by the appropriate amount. However, modifying g_1 and g_2 only induces a horizontal shift left or right, and the data indicates the need to modify the derivative of LY and QY from NEST.

Since the means of the model do not line up with the data the calculated corrections in figure 6.3 can't be applied. In order to account for detector resolution as outlined in section 6.1 we must have a reasonably accurate initial guess of the spectrum with infinite resolution. In this case we do not. However, even though the means from NEST light and charge yields are off we find that the effect of recombination fluctuations and detector resolution is relatively small. The correction for the S1 ranges from +30% to -20% and the S2 correction from +2% to -10%. We takes these as small enough to proceed with extracting light yield, charge yield and recombination without any correction in order to construct a more accurate model than our initial NEST prediction. Having outlined a method for mapping the observed S1 and S2 means to their real values we will now gauge the effect of finite resolution on the observed total energy.




Figure 6.3: a): In Black S1 tritium S1 spectrum extracted from the data. In blue, The NEST light yield curve. In red, the NEST light yield curve with recombination fluctuations. Dashed magenta is NEST light yield with smearing from equations 6.7. b): The ratio of the real S1 mean to the S1 observed mean vs. the observed mean after smearing. Note, the S1 threshold at about 3 PE in S1. c): In Black S2 tritium spectrum extracted from the data. In blue, The NEST light yield curve. In red, the NEST light yield curve with recombination fluctuations. Dashed magenta is NEST light yield with smearing from equations 6.8. d): The ratio of the real S2 mean to the S2 observed mean vs. the observed mean after smearing.

6.2.2 Tritium Energy Spectrum

Unlike the S1 and S2 spectra which are dependent on light and charge yield, the tritium energy spectrum is well known [14]. The tritium spectrum is perhaps the most studied beta spectra, so there is no need to rely on modeling. Also, recombination

fluctuations cancel out in combined energy space leaving only detector resolution to be applied for smearing, given equation 6.11. The accuracy of the smearing model described in equations 6.1-6.6 can be tested by comparing it against the energy observed after a full simulation, which accounts for detector geometry and having been processed by the full offline framework. Using equation ??, ?? and 6.9 we solve for the the variance of E (σ_E^2),

$$E = W(n_\gamma + n_{e^-}) \quad (6.9)$$

$$\sigma_E^2 = W^2(\sigma_{n_\gamma}^2 + \sigma_{n_{e^-}}^2) \quad (6.10)$$

$$\sigma_E^2 = W^2(a_\gamma^2 n_\gamma + a_e^2 n_{e^-}) \quad (6.11)$$

W is the work function $0.0137 \pm 0.002 \frac{\text{keV}}{\text{N}_{\text{quanta}}}$, n_γ and n_e and number of photons and electrons respectively. The constants a_γ and a_e represent the coefficients of the \sqrt{n} term for the statistical uncertainty given in equation 5.15. There is also an instrumental component which is proportional to n_γ and n_e . However, the instrumental term is subdominant at the low tritium energies with coefficients given in equation 5.16.

Starting with the tritium energy spectrum with infinite resolution we apply the empirically determined resolution in equation 6.11. Figure 6.4 shows the comparisons of the true tritium spectrum, the spectrum with smearing from equation 6.11, the expectation from LUXSIM and the data. The smearing from the model described in equations 6.1-6.6 is found to be almost identical to the output of LUXSIM.




Figure 6.4: The tritium energy spectrum reconstructed from the data (black). Along with LUX SIM (red), the true tritium beta spectrum (dashed blue) and a tritium spectrum smeared with detector resolution of equation 6.11 (green).

Comparing the true tritium spectrum, the data, LUXSIM and the spectrum with 6.11, we find that for the cases with finite resolution the endpoint flare out above 16 keV. With the endpoint reaching out past 20 keV instead of terminating at 18.6 keV. This effect is precisely what the modeling in section 6.1 attempts to undo. Clearly events observed at 20 keV must have fluctuated up from bins below the tritium endpoint at 18.6 keV [69]. Most importantly, besides the additional fluctuation around the endpoint the difference in spectral shape after accounting for finite resolution is hardly noticeable.

The mapping for observed energy to true energy is calculated for both LUXSIM and the simpler smearing model of equation 6.11. From the LUXSIM data the initial Monte Carlo values of each event are compared to the final reconstructed energies. Each energy deposit is simulated with photon and electron propagation along with light collection in the LUX detector. Figure 6.5 shows the results for mapping

observed energy to real energy using both smearing methods.




Figure 6.5: Left, mapping from real Monte Carlo energy to observed energy plotted vs the observed energy after applying a finite resolution using LUXSIM. Right, comparing the correction determined from the Monte Carlo (Red) to the detector smearing model (black) given in equation 6.11. The dashed lines represent the uncertainty in the measured values. The agreement is within errors from 1 to 18 keV_{ee}. The Energy threshold is 50% at 1.0 keV_{ee}.

The shift in the observed mean from the true mean for both LUXSIM and the model of equation 6.11 are in good agreement down to the threshold of 1.5 keV, the agreement with simulation is always within 1%. Below 2 keV the model predicts the ratio of true energy to observed energy to rise as there are greater number of events at higher energy spilling over to lower energy. The simulation however does not show this behavior leading to a 5% discrepancy in the 1 keV bin. Comparing the modeled detector resolution to the more complex LUXSIM simulations provides a proof of principle of the model.

The effect of detector resolution on skewing the true energy to observed energy

is found to be minor, blowing up only at the tail end of the tritium spectrum. Over 95% of the tritium event occur between 1 and 15 keV were the correction is less than 5%. With this information we proceed with extracting light and charge yields without concern about incorrectly assigning the energy bin.

6.2.3 Results for Light Yield, Charge Yield and Recombination

As shown in figure 6.3 the S1 and S2 spectral shape is not a good match with the light yield model from NEST, thus applying a correction to the observed means using NEST is not prudent. Fortunately, both for the S1 and S2 the spectral shape correction is less than 10% in the region where the vast majority of the tritium events are populated. Further, the reconstructed energy is also valid to within 5%. Knowing this we can move forward with extracting a more accurate light yield and recombination fluctuations.

By the same method outlined in section 5.5, the number of photons, electrons and recombination fluctuation is extracted from the tritium data. This is the result of the raw data uncorrected for the spectral shape. The result is shown in figure 6.6.

Having measured number of photons, electrons in each energy bin we compare the tritium data to NEST yields, shown in figure 6.7. The disagreement between the data and the NEST yields was expected since before the S1 and S2 tritium spectrum did not line up, shown in figure 6.3 . Though the means do not match the measured light yield is within 1 sigma considering the large systematic uncertainty in gains




Figure 6.6: Top: Extracted recombination fluctuation from the tritium data from fluctuations in photons and electrons (Black and Red respectively). Bottom right: mean number of quanta in photons, electrons, ions, excitons vs. energy keV for the tritium calibrations. Bottom left: Recombination fraction and the one sigma (shaded) vs. energy keV.

g1 and g2. The figure also shows the one sigma prediction of the yields from NEST [59] shaded in blue where the model is interpolated and magenta where the model is extrapolated.




Figure 6.7: Light yield and charge yield from tritium data without spectral shape correction at 170 V/cm in black. The shaded region represents the one sigma systematic uncertainty on g1 and g2. The NEST yield prediction and its corresponding 1 sigma is shaded in blue. NEST interpolation is shown in magenta to energies where the model is not vetted. Note that the statistical errors are about the size of the data points and the dominant uncertainty illustrated by the shaded region is 100% correlated bin-to-bin. A one sigma shift up in light yield corresponds to a one sigma shift down in charge yield and visa versa.

The yields and recombination fluctuations measured from the tritium data

can be used to construct an improved model over the initial NEST prediction. As mentioned previously, the effect of detector resolution can be ignored in the region where the vast majority of the tritium events occur. Having built a new model for yields we now proceed to patch up the 20-30% shifts which occur at the edges of the S1 and S2 tritium spectra.

6.3 Measuring Light Yield, Charge Yield and Recombination, Corrected for Spectral Shape

Using the light yield and charge yield measurements extracted from the uncorrected tritium data we improve the light yield and charge yield model over the initial NEST-based model. In this section we will take the information gathered in the previous section to calculate the spectral shape correction for the tritium S1 and S2. With the improved model for LY and QY we can determine the efficiency for detecting S1s, S2s and the energy threshold. Finally, after applying the correction we can extract the true light yield and charge yield information from the tritium data.

6.3.1 Tritium S1 and S2 Correction

We now proceed to calculate the spectral shape correction using the method outlined in section 6.2.1 with NEST yields replaced by those measured from the uncorrected tritium data. Figure 6.8 (a,c) shows the application of smearing from equation 6.7 and 6.8 applied to the light and charge yield extracted from the uncorrected tritium

data. The mapping of the observed S1 and S2 to the real S1 and S2 is also shown in the figure 6.8 (b,d). The correction is determined by using the extracted yields, applying the gains g_1 and g_2 , and convolving it with a true tritium beta spectrum along with the measured recombination fluctuations, equation 6.7 and 6.8. Given infinite detector resolution this is the spectrum the LUX detector would observe for S1 and S2, labeled on the figure as $LY-T + \sigma_R$ and $QY-T + \sigma_R$ respectively. After applying detector resolution from equation 5.17 to the plotted $LY-T + \sigma_R$ and $QY-T + \sigma_R$ spectrum we calculate the mapping from the real S1 and S2 means to the observed means using the model outlined in 6.1. The final spectrum with recombination fluctuations and detector resolution is labeled as $LY-T + \sigma_R + \sigma_{S1}$ and $QY-T + \sigma_R + \sigma_{S2}$.

As expected, we find good agreement between the smeared tritium S1 and S2 spectra with the data. The spectral shape correction found for both S1 and S2 is consistent with those found by using NEST previously, shown in figure 6.8. This gives us confidence that we can apply the mapping of true mean to observed to the data.

The meaning of the spectral shape corrections for S1 and S2 shown in figure 6.8 (b) and (d) can be understood as a mixture of spectral shape and varying resolution. As the measured value of S1 drops to 1 PE (20% detection efficiency) we find the correction factor rises to a factor of 1.3. Even though the count rate is growing in lower S1 bins the narrowing S1 resolution cancels out the spill over from lower S1s as compared to the overlap from larger S1s with a lower count. This effect causes an observed S1 mean of 1 PE to actually be comprised of events with a mean of 1.3




Figure 6.8: a): In Black S1 tritium S1 spectrum extracted from the data. In red, the S1 spectrum based upon the LY and QY measured from the tritium data after applying recombination fluctuations. In dashed blue, the expected S1 spectrum after applying finite detector resolutions of equation 6.7. b): The ratio of the real mean to the observed mean vs. the observed mean after smearing the tritium photon spectrum with detector resolution. Note the S1 threshold at about 3 PE in S1. c): In black, tritium S2 spectrum from the data. In red, the S2 spectrum based upon the LY and QY measured from the tritium data after applying recombination fluctuations. In dashed blue, the expected S2 spectrum after applying finite detector resolutions of equation 6.8. d): The ratio of the real mean to the observed mean vs. the observed mean after smearing the tritium electron spectrum from NEST with detector resolution. Note the S2 threshold at about 400 PE in S2.

PE. On the other end the correction is straight forward. As the beta spectrum is dropping sharply to reach 0 at the Q value of 18.6 keV [69] events from the more populated lower energy bins spill over into higher energy regions. This effect causes

the observed mean at the endpoint of 18.6 keV to actually be comprised of events with an average energy of 15 keV. Note, that the energy spectrum with detector resolution extends to 21-22 keV due to upward statistical fluctuations in S1 and S2. The S2 spectrum exhibits the same behavior at the high end above the peak of the spectrum around 1000 PE. As the S2 approaches threshold of 400 PE there is a < 3% correction to the mean to account for the spill over of the more populated regions to the right. This is different than case of S1 at low PE as the S2 resolution is about a factor of three better at the threshold.

6.4 Ionization and Scintillation Yield After Correction

The light and charge yield which can now be extracted from the tritium data are unique properties of liquid xenon given for ER interaction at low energies. Figure 6.9 shows the data used to calculate the means after having corrected for the S1 and S2 spectral shape. Two tritium calibration data sets are shown, one with high statistics at 170 V/cm containing 140,000 events and the second at 100 V/cm with a modest 2,500 events. (Both numbers correspond to events in the fiducial volume).





Figure 6.9: Means of the light yield and charge yield from tritium data corrected for spectral shape along with the 1 sigma statistical errors.


The means used for LY and QY are the population means in each slice of energy for our best value of g1 and g2. The errors shown are the statistical errors along with a small systematic component from the difference of the population mean from the Gaussian mean. The systematic offset from the constraint of g1 and g2 are treated in figure 6.10. With the potential for tighter constraints on the value of gains g1 and g2, the remaining uncertainty in the measurement of LY and QY would be less than 3% below 10 keV. The tritium calibration source has the potential to be used to determine the light and charge yield in liquid xenon to less than 3% including at the detector threshold. This calibration source hold great promise considering that few yield measurements exist below 5 keV.

Figure 6.10 shows the mean yields with the one sigma bands from the uncertainty in gains g1 and g2. The errors are anti-correlated, thus a shift up in light yields corresponds to a shift down in charge yield preserving the energy. The figure also shows the one sigma prediction of the yields from NEST [59] shaded in blue where the model is interpolated and magenta where the model is extrapolated.


The overlap between the data and the NEST model is within one sigma considering errors in g1 and g2. However, as the errors in g1 and g2 are systematic and 100% correlated bin-to-bin they can only shift the curves up or down. Even under such a shift, the shape of the tritium data would not agree perfectly with the NEST prediction. As the statistical uncertainty alone constrains the means to better than 3% below 10 keV for the 170 V/cm data set.




(a) n_γ , 170 V/cm



(b) n_e , 170 V/cm



(c) n_γ 100 V/cm



(d) n_e , 100 V/cm

Figure 6.10: Light yield and charge yield from tritium data corrected for spectral shape along with the 1 sigma systematic constraint on $g1$ and $g2$. The blue and magenta curve are NEST extrapolation and interpolation, respectively.

The comparison of measured light and charge yield before and after the tritium spectral shape correction is shown in figure 6.11 . The band in red is the result from using the raw data and combined energy, the band shown in blue is after the correction. The yields before and after the correction overlap with the exception of the last 15-18 keV bins which are pulled back as the events reconstructed at those energies on average were the result of events with 20% less energy, having upward fluctuated. In the middle regions however the spectral shape correction of the energy, the photon (S1) and electron (S2) spectrum are canceled. Recalling that yield is defined as n_γ/E , and n_e/E , with $E = W(n_\gamma + n_e)$. Thus, on average the upward or downward fluctuations in collected photons and electron are canceled by the corresponding fluctuations in reconstructed energy E.

One question remaining to be answered is if the yields and ER band defined by tritium beta decays are consistent with the more generic Compton Scatter backgrounds, which are expected to be found in the WIMP search data. It is expected that betas and gammas are indistinguishable below 10 keV [58] [59]. In the following section we will compare the light yield results from the tritium data with recent Compton scattering measurements that have probed light yield in xenon down to 1.5 keV.




Figure 6.11: In black, the light yield and charge yield from tritium data corrected for spectral shape. In red, the light yield and charge yield from tritium data uncorrected for spectral. The shaded region represent the one sigma systematic error from the constraint on g_1 and g_2 shaded.

6.5 Comparison of Light Yield Measured With Tritium to Other Measurements With Compton Scatters

In this section we compare the LY measured with tritium in LUX to that measured with Compton scatters by other researchers. When comparing measurements from different xenon detectors it is prudent to report the result relative to that of a standard calibration source. The light yields reported for low energy Compton scatter measurements from [10] [13] are normalized to the first 32.1 keV decay of ^{83m}Kr . Before comparing the light yields results from the tritium calibration we first need to measure the light yield in LUX at zero field from ^{83m}Kr .

6.5.1 The Standard Candle: Light Yield from ^{83m}Kr

The ^{83m}Kr source was discussed in greater in chapter 3. The decay of ^{83m}Kr consists primarily of the emission of two internal conversion electrons at 32.1 keV and 9.4 keV, with a half life of 154 ns between the two [9] [11] [12]. The combined signal (41.6 keV) is found by the LUX pulse finder in the majority of cases. However, the combined signal is not useful as a standard calibration. The second 9.4 keV decay receives an enhancement in light yield due to increased recombination probability from the presence of ions and electrons from the initial 32.1 keV decay. It has been observed that the light yield enhancement depends upon the decay time separation, out to 1000 ns [9] and [10]. In the LUX detector we have also observe the enhancement of the light yield of the second 9.4 keV out to 2000 ns shown in

figure 6.12. Our ability to split pulses with 100% efficiency starts at 1200 ns.




Figure 6.12: Left: The light yields of the 9.4 and 32.1 keV decay of $^{83\text{m}}\text{Kr}$ plotted vs. timing separation, for events separated by more than 1200 ns. The point at 600 ns is calculated assuming the 32.1 keV yield remains flat to 200 ns as observed in [9] and [10]. Right: the histogram of $^{83\text{m}}\text{Kr}$ events vs timing separation finding a best fit to the half life of 159.7 ± 7.5 consistent with the measured value of 154 ns [11] [12].

Fortunately, the first 32.1 keV appears to have no time dependance as it decays under normal circumstances in the xenon, without the presence of additional free ions or electron [10], [13]. Using a $^{83\text{m}}\text{Kr}$ data set at zero field the yield of the 32.1 keV decay was determined. Since the S2 (charge) signal is unavailable at zero drift field we rely on the top-bottom asymmetry of the PMTs to define the Z coordinate of the event, $\frac{\text{top-bottom}}{\text{top+bottom}}$. We must know at least the Z coordinate in order to apply the position dependent corrections outlined in chapter 3. The XY correction is subdominant to the Z-dependent correction for the S1 signal and can be ignored. Each event, given a top-bottom asymmetry, is mapped to a detector depth

Z allowing for the Z -dependent correction to be applied. The correction normalizes the pulse area PE to the detector center (241.6 mm below the gate grid). The result for the zero field data set is shown in figure 6.13 and is reported in table 6.1.




Figure 6.13: Left: S1 ^{83m}Kr peaks at zero field. Right: shows the histogram of the separated 32.1 and 9.4 keV decays plotted above vs. time. The half life fit to the population is in good agreement with the measured half-life of 154 ns [11] [12].

Table 6.1 shows the measured scintillation of the 32.1 keV internal conversion electron from ^{83m}Kr using the LUX detector at various fields. The table also includes the NEST predictions [59]. Electron mobility and charge separation at the interaction site increases with drift field leading to less recombination, causing scintillation yield to be quenched. The field effect is more dramatic at higher energies than in the low energy regions probed by the tritium data.

6.5.2 Comparing Tritium Scintillation Yield with Compton Scatters

We report the measured light yield of tritium relative to the 32.1 keV decay of ^{83m}Kr at zero field, defined as \mathcal{R}_e . The comparison is done as a proof of principle

Field V/cm	S1 PE	Photons $\langle n_\gamma \rangle$	Yield $\langle n_\gamma \rangle / \text{keV}$	NEST $\langle n_\gamma \rangle / \text{keV}$
0	216.2 ± 5.0	2228.9 ± 50.5	69.4 ± 1.6	64.2 ± 3.2
50	195.0 ± 0.7	2010.3 ± 7.2	62.6 ± 0.2	59.8 ± 3.0
100	178.4 ± 0.7	1839.2 ± 7.2	57.3 ± 0.2	55.8 ± 2.8
170	171.4 ± 0.9	1767.0 ± 9.2	55.0 ± 0.3	51.9 ± 2.6

Table 6.1: Field dependance of the light yield form the 32.1 keV decay of $^{83\text{m}}\text{Kr}$ along with the NEST [59] predictions.

that the light yield from betas and gammas overlap at low energies, at least within the rather large systematic uncertainties. Further, this is a cross check that the residual $< 10 \times 10^{-12} \text{ g/g}$ concentration of methane injected for the tritiated-methane calibration had negligible impact on the light yield, as expected from previous measurements with methane[75]. The result for \mathcal{R}_e is shown in figure 6.14, with the one sigma regions plotted as bands for the tritium data.

We find good agreement between the centroids of the tritium data at 100 and 170 V/cm with the zero field and 450 V/cm Compton scattering measurements from [13] and [10]. The finding are consistent with the expectation that tritium light yield data at 100 and 170 V/cm lie between the zero field light yield measurements the light yield at 450 V/cm. The error bars from the Compton scattering measurement are rather large due to the uncertainty in scattering angle and the need for Monte Carlo to reconstruct the energy deposit in the liquid xenon. In those measurements, the combined energy of the deposit in the liquid xenon is uncertain as both experiments were done in light-only collection mode [13] [10] (even for the 450 V/cm measurement). The errors on the tritium results are systematic and dominated by the constraint of g1 and g2 and are comparable with the errors on the Compton




Figure 6.14: LY relative to the light yield of the 32.1 keV decay of $^{83\text{m}}\text{Kr}$ at zero field.

The black and blue bands represent the results from tritium at 170 and 100 V/cm respectively. The shaded region represents the systematic error due to the one sigma constraints on g_1 and g_2 . Magenta points are Compton scattering measurements from [13]. The grey and red represent zero field and 450 V/cm Compton Scattering measurements from [10].

scattering measurements.

If the constraint on g_1 and g_2 in LUX is improved by future calibrations, then the errors on the tritium data will shrink significantly to below 3% making the tritium calibration a powerful tool for calibrating both light yields and charge

yields all the way to the energy threshold. The tritium light and charge yields reported here extend down to 1 keVee corresponding to the 50% threshold of LUX. The measurements also confirm, within systematic errors, that the ER band as calibrated by the tritium data is applicable to the more generic Compton scatter backgrounds in the WIMP search region of 1 to 5 keV_{ee}. Compton scatters comprises about 2/3 of the expected background in LUX with the remaining 1/3 being from the beta decay of ⁸⁵Kr [5].

6.6 Summery

In this section we have extracted the light and charge yields using the tritium calibration source in energies ranging from 1 to 16 keV_{ee}, shown in figures 6.9 and 6.10. The light and charge yields measured are a fundamental property electronic recoils in liquid xenon, in this work the liquid density was 2.888 g/cm³. We find agreement, within the systematic errors, between the light yields measured with the tritiated-methane source with Compton scattering measurements down to 1.5 keV_{ee}, figure 6.14. The result supports the model that low energy betas leave identical tracks to low energy Compton scatters in liquid xenon [58] [59]. Having measured the yields, the S1 and S2 signals can be modeled for any background source only requiring the energy spectrum as an input. The ratio of charge (S2) to light (S1) characterizes a xenon detector's ability to reject background events from WIMP candidates at a particular energy. We have also found that the light and charge yields measured at 170 and 100 V/cm merge below 5 keV_{ee}, indicating that below this energy recom-

bination is insensitive to field. The results for ER and NR discrimination using the tritium calibration source will be discussed in the next chapter.

Chapter 7: Tritiated-Methane Calibration Source and the ER Band

In this final chapter we overview the development, implementation and the main results of the tritiated-methane calibration source. The source was developed to be a method of calibrating the electronic recoil (ER) band in current and future large scale xenon detectors (+100 kg). As opposed to external ER calibration sources, the tritiated-methane is internal and mixes with the bulk xenon. Internal sources have a great advantage over external sources as they trivially overcome the formidable stopping power of liquid xenon, illuminating the fiducial volume of the detector. The tritiated-methane calibration source was ultimately used to characterize background ER events from potential NR signals of WIMP. Once the ER band is defined the background rejection of WIMP candidates at a particular energy is treated with a profile likelihood analysis, described in [33].

7.1 The need for an Internal Calibration Source

Over the past two decades liquid noble TPCs used for dark-matter experiments have been growing in size to net more kg-days of exposure for WIMP searches. With the additional mass, noble detectors benefit from the self shielding properties of the dense liquid inside the inner fiducial volume of the detector, as the outer

volume is used as an active veto [43]. With current generation liquid xenon detectors containing +100 kg, the detectors are vitally insensitive to external gamma radiation in the WIMP search region of interest of 1-10 keV_{ee} [5] [34] [35] [36]. Being insensitive to external radioactivity is great for improving the signal to background ratio for WIMP searches. Unfortunately, it is also equally as great for shielding against external calibration sources. With plans for even larger xenon detectors already moving forward [76] [77], we must develop a new method for introducing controlled radioactivity for calibration purposes.

The fiducial volume of the LUX detector is surrounded by more than than 6 cm of liquid xenon providing excellent shielding from both external backgrounds and calibration sources [5]. For example, a 100 keV gamma has a mean free path of about 2 mm in liquid xenon and would require thirty mean free paths to penetrate into the fiducial volume. A higher energy source such as ¹³⁷Cs (662 keV) has a longer mean free path of 4 cm however, the probability of a low energy deposit from forward scattering in the fiducial followed by an escape is greatly suppressed. We would expect less than one ER event per day between 1-10 keV_{ee} in the fiducial region if calibrating with an external ¹³⁷Cs or Th source, [5]. Further, calibrating with high energy sources or high rate sources brings along systematic uncertainties from high energy deposits near the detector edge and high rates would overwhelm the DAQ.

7.2 Tritiated-Methane as a Calibration Source

In order to overcome the issues with external calibrations, the source used to calibrate the ER band must satisfy three requirements. First, it should illuminate the WIMP search band with single scatter events (1-10 keV_{ee}). Second, it must be able to mix with the xenon and be delivered as an internal source. Third, it should either have a short half-life or be easily separated from xenon by commercially available bettering technology. ^{83m}Kr has been developed and used as an internal source with liquid xenon detectors [9] [10]. However, in LUX ^{83m}Kr only produces a mono energetic peak at 42.1 keV_{ee} being too high in energy to calibrate the WIMP search region. In order to populate the ER band with single scatter events between 1- 10 keV a beta emitter should be injected. Tritium is an ideal candidate, satisfying all but the removal requirement. Tritium has a Q value of 18.6 keV [69], a mean beta energy of 5.6 keV [70] and a mode of 3 keV [14], see figure 7.1.

Tritium has a half life is 12.3 years [78]. Thus, it is not practical to wait for the tritium activity to decay away. The tritium must be removed after the calibration is complete in order to continue a low background WIMP search run. Hydrogen, being chemically identical to tritium, can be removed from the bulk xenon by standard gettering technology [79]. However, the removal is complicated by tritium's high diffusion rate into plastics and other detector materials [80]. A botched tritium injection would mean the end of a dark matter experiment.

To mitigate the effect of tritium diffusion into plastics we use a tritiated-methane source instead of bare tritium. Tritiated-methane consists of a tritium




Figure 7.1: True tritium beta spectrum from [14].

atom and three hydrogen atoms bound to a carbon molecule, and is chemically identical to methane. Using tritiated-methane reduces the diffusion of tritium into detector internals by an order of magnitude [80]. The weak molecular bond, < 1 eV, to the methane molecule does not impact the nuclear physics of the tritium beta decay. Also, methane is known to be soluble in liquid xenon and is sometimes used to quench scintillation light when injecting relatively large amounts into the liquid [81], [75], [82]. Finally, We have studied the removal of methane by the commonly used SAES heated zirconium getter and found that significant amounts of methane can be removed from xenon at flow rates commonly employed in large scale liquid xenon experiments [83],[84].

7.3 Removal of CH₄ from LUX

The injection and removal of tritiated-methane into a liquid xenon vessel containing plastics was first conducted in an experimental setup at UMD. Even with conservative estimates for diffusion rates into plastics, we could not be certain about the behavior of tritiated-methane in the much larger LUX detector. Prior to injecting the tritiated-methane into LUX a much larger natural methane injection was performed in order to characterize diffusion, being chemically identical tritiated-methane. A xenon gas analysis system, developed at UMD, allowed for in situ purity analysis from several ports plumbed directly into the LUX xenon circulation loop. The analysis system has ppt (part-per-trillion) sensitivity to CH₄ and better than ppt sensitivity to Kr. The compact system allows for hourly sampling and is used for detection of several key impurities (N₂, O₂, He, Ar, Kr, CH₄, H₂). The design is compact, the analysis can be done on site, and is significantly less expensive than the more complex techniques used by [85] and [86]. The purity analysis technique is described in [84], [83], [87], [44], and will be described for the LUX system in a future publication.

Figure 7.2 shows the results from the xenon gas analysis system for a 50 ppb (g/g) methane injection into the bulk xenon. After the first hour, the 50 ppb of methane mixes into the liquid and appears as 300 ppb in the gas returning from the bulk liquid. The enhancement in the gas phase is due to the solubility of methane dissolved in liquid xenon and is characterized by the Henry's constant .

Having in situ sampling provided crucial diagnostics for methane removal and




Figure 7.2: Removal of natural methane observed by the integrated xenon sampling system prior to the tritiated-methane injections. The red points indicate measurements at the getter outlet, we find a 97% one pass removal efficiency at a flow rate of 25 SLPM. The blue curve shows the improved upper limit on the effect of outgassing from the plastics. The black dashed lines shows the exponential fit to the natural methane removal from the xenon with a time constant of 5.9 ± 0.07 hours. 5×10^{-3} ppb (g/g) is the limit of detection for methane.

diffusion into plastic components in the LUX detector. We measured the one pass purification efficiency for methane to be 97% by the SAES PS4-MT15-R-1 MonoTorr getter ([79]) at a flow rate of 27 SLPM. The getter's health for hydrocarbon removal is important to check prior to an injection, as an aged zirconium getter will lose its ability to remove CH_4 before failing for N_2 and O_2 [83]. Crucially, we confirmed that natural methane could be removed by more than five orders of magnitude without residual back diffusion. The plateau seen in the end of figure 7.2 is caused by the 5 ppt limit for methane detection rather than diffusion from plastic components which absorb the impurity when it is initially injected. Taking the background as the limit

for back diffusion we infer a diffusion constant of less than $1.6 \times 10^{-3} \text{ cm/sec}^{1/2}$.

The results from the natural methane injection gave us the confidence to proceed knowing that the goal of reducing the tritium rate to less than 5% of background could be met in the LUX detector. The purification time constant for natural methane removal was measured to be 5.9 ± 0.07 hours as seen in figure 7.2. The removal time constant is 1/6 of that expected based on xenon circulation rates alone, and is exactly the ratio of methane concentration in the gas to the methane concentration in the liquid. The enhanced purification time constant is reasonable as the methane is purified from the gas phase where it is more abundant, with the equilibrium concentration in the gas above the liquid set by the solubility.

7.4 Light Yield Quenching from CH₄ in LUX

It is well known that at high concentrations, of several percent, methane will quench scintillation in liquid xenon [81], [75], [82] [88]. The quenching of scintillation in liquid argon has been observed with methane concentrations as low as 10 ppb (parts-per-billion) [89]. When performing a tritiated-methane calibration for the LUX detector, containing 350 kg of xenon, the amount of natural methane injected is <10 ppt (parts-per-trillion). This is far to low to cause any shift in the observed light yield and impact the ER band measurement. To prove the case, a natural methane injection of 100 ppm was performed along with periodic ^{83m}Kr calibrations that were used to track the light yield of the line source normalized at the center of the detector. Figure 7.3 shows the result of ^{83m}Kr calibration along with the

methane concentration in the gas measured by the analysis system.




Figure 7.3: In black, the response to scintillation from $^{83\text{m}}\text{Kr}$ at the center of the detector normalized to the first data point before the natural methane (CH_4) injection. The dashed magenta lines represent the time window from the beginning of the natural methane injection to the time the background of 5 ppt is reached. The blue points represent that methane concentration in the gas returning from the bulk liquid of the detector. The concentration in the liquid xenon is roughly 1/6 of the concentration measured in the gas phase due to solubility.

In figure 7.3, there are two $^{83\text{m}}\text{Kr}$ data sets that had methane concentrations above background levels. The first measurement is made with greater than 25 ppb in the gas corresponding to greater than 4 ppb in the liquid. The second contained greater than 50 ppt in the gas corresponding to greater than 83 ppt in the liquid. The shifts in yield are purely systematic, as the two light yield measurements containing methane fall between the measured yields of the first (prior to and injection) and last (well below 5ppt) data points. We constrain light yield quenching induced by 4 ppb (g/g) of methane in liquid xenon to < 1%. Note, that a typical tritiated-methane calibration contains roughly three orders of magnitude less methane than

4 ppb (g/g).

7.4.1 Tritiated Methane injection into the LUX detector

Following the natural methane test, the tritiated-methane injection was conducted at the end of the first underground science run, on Aug 8th 2013. An absolute activity of 20 mBq of tritiated-methane was injected at the purifier's outlet while circulating. A removal time constant of 6.0 ± 0.5 hours was measured in the liquid volume, figure 7.4, and is consistent with the natural methane removal measured in the gas by the sampling system 7.2. After a day of circulating through the getter the tritium decay rate had fallen below detectable levels confirming the effective removal of the tritiated-methane with the getter. A second, larger injection of 800 mBq was performed a week later yielding a similar removal time constant of 6.4 ± 0.1 . The second injection produced 20,000 beta decays in the LUX detector before being completely removed, 7000 of which were in the fiducial volume and could be used to calibrate the ER band in the WIMP search region of 1-50 PE (about 1-8 keVee). The injections were performed while circulating and with the getter actively purifying in order to maintain detector purity and stability. Prior to LUX detector upgrades in December 2013, a total of 10 Bq of tritiated-methane was injected into the LUX detector and successfully removed providing over 140,000 beta decays within the fiducial volume.




Figure 7.4: The rate of single scatter events with S1 below 100 PE in the fiducial volume. 100 PE in S1 is about 18.6 keV_{ee}, the endpoint to the tritium beta spectrum. The magenta and red curves are fits to the first and second tritium injection's removal rate. Note, the removal of tritiated-methane is consistent with the natural methane removal rate measured independently.

7.4.2 Mixing of Tritiated Methane in Liquid Xenon

Tritium events appear uniformly distributed in the liquid volume several minutes after injecting the tritiated-methane inline with the xenon gas circulation path.

Figure 7.5 shows the R² vs. Z distribution of tritium events thirty minutes after an injection. The events shown cover the region from the gate to the cathode and radially outward to the edge of the detector.




Figure 7.5: The distribution of tritium events vs. detector radius squared. The solid black line represents the fiducial volume. Right: The distribution of tritium events vs. XY in the region between the gate and the cathode. The solid black line represents the fiducial volume and the black circles represent the locations of PMTs (photo multiplier tubes).

7.4.3 Definition of Electronic Recoil Band. ER and NR Discrimination

The electronic recoil band in the fiducial volume of the LUX detector was calibrated to unprecedented accuracy using the tritium source. The calibration data was acquired in a 40 hour time window in which less than four out of the 140,000 events in the fiducial are expected to be non tritium [5]. Nearly every data point (99.997%) that will be presented in the subsequent figures is the result of a tritium beta decay in the fiducial volume of the LUX detector. Figure 7.6 shows the tritium calibration data with fits to the mean of the ER band along with the 10-90% confidence bounds ($\pm 1.28\sigma$), at a drift field of 170 V/cm. Also shown in red is the nuclear recoil (NR) mean measured using a mono energetic DD neutron generator calibration, described in [future publication].




Figure 7.6: Charge to light ratio $\log_{10}(S2_b/S1)$ plotted vs. $S1$, the energy contours are also shown. Top, the tritium data uncorrected for spectral shape. Bottom, with the spectral shape correction discussed in chapter 6. The calibration consists of over 112,000 tritium beta decays between 1 and 50 PE in $S1$ ($1 - 10$ keV_{ee}), in the fiducial volume of the detector. The black circles represent the ER band mean and 10% to 90% bounds, $\pm 1.28\sigma$. The red solid line is the NR band mean determined using a DD neutron generator calibration. The dashed red lines indicate the 10% to 90% bounds of the NR band.

The ER band defined by the tritium data shown in figure 7.6 has been corrected for spectral shape discussed in chapter 6. The spectral shape correction recovers the true ER band for a flat energy spectrum, left uncorrected the ER band would peak up slightly faster in the lowest S1 bins, below 5 PE. The discrimination between ER and NR events at 50% NR acceptance can be determined by counting the number of tritium events that fall below the mean of the NR band. Note, below 15 PE in S1 the NR acceptance actually becomes greater than 50% due to the asymmetric shape of the NR band at low energies, which works in our favor [58] [59]. The values of the leakage fraction at 50% NR acceptance per each 1 PE bin in S1 are shown in 7.7. Discrimination and leakage fraction is listed listed at several acceptance fractions in table 7.1.

NR %	Leakage Fraction $\times 10^{-3}$	Discrimination %	Leakage Fraction* $\times 10^{-3}$	Discrimination* %
90	52 ± 0.70	94.8 ± 0.07	49 ± 0.60	95.1 ± 0.06
80	18.8 ± 0.40	98.1 ± 0.04	17.4 ± 0.40	98.3 ± 0.04
70	8.6 ± 0.30	99.1 ± 0.03	7.6 ± 0.30	99.2 ± 0.03
60	4.2 ± 0.20	99.58 ± 0.02	3.7 ± 0.20	99.63 ± 0.02
50	2.1 ± 0.14	99.79 ± 0.014	1.8 ± 0.13	99.82 ± 0.013
40	0.97 ± 0.09	99.90 ± 0.009	0.87 ± 0.09	99.91 ± 0.009
30	0.49 ± 0.07	99.95 ± 0.007	0.44 ± 0.06	99.96 ± 0.006
20	0.23 ± 0.05	99.977 ± 0.005	0.22 ± 0.04	99.978 ± 0.004
10	0.08 ± 0.03	99.992 ± 0.003	0.08 ± 0.03	99.992 ± 0.003

Table 7.1: Leakage fraction at various NR acceptance % over the range of 1-50 PE in S1, using tritium data at 170 V/cm. For the case without spectral shape correction, and with spectral shape correction to the left (columns marked with *). Note, these are conservative estimates assuming Gaussian behavior about the NR mean. Below 15 PE in S1 the NR band actually becomes bottom heavy, which works in our favor [58] [59].




Figure 7.7: Discrimination vs. S1 using over 112,000 tritium beta decays between 1 and 50 PE in S1 ($1-10\text{ keV}_{ee}$), using tritium data at 170 V/cm . Top, the tritium data uncorrected for spectral shape. Bottom, with the spectral shape correction discussed in chapter 6. The dashed horizontal lines indicate the mean leakage fraction from 1 to 50 PE.

7.5 ER Band Gaussianity

With the high statistics tritium data set we could test the Gaussianity of the ER band in the WIMP search region. The charge to light ratio used to discriminate ER and NR events (plotted as $\log(S_2/S_1)$) and has been assumed to have Gaussian behavior in past experiments. The largest tritium calibration yielded 112,000 tritium beta decays with only four expected to be non tritium events [5] in LUX detector's fiducial region. Figure 7.8 shows the same ER band of figure 7.6 but with the centroid subtracted off, and is plotted for cases of without and with the spectral shape correction. We find that below 3 sigma of the ER mean the data begins to deviate from Gaussian behavior, potentially due to instrumental fluctuations. For example, the probability of finding an ER event below the NR mean at 4 sigma below the ER mean is nearly 10 times higher than the probability that would be assigned by a Gaussian PDF. If not treated properly by data driven calibrations, this could be falsely interpreted as a WIMP event by a profile likelihood analysis.




Figure 7.8: ER band Gaussianity from 0-50 PE with the centroid of the ER band subtracted off, using tritium data at 170 V/cm. Top, the tritium data uncorrected for spectral shape. Bottom, with the spectral shape correction discussed in chapter 6. The dashed vertical red line indicates the mean of the NR band from 1 to 50 PE.

7.6 S1 Threshold For Golden Events, Using Tritium

The tritiated-methane calibration source provides beta decays with energies well below the energy threshold of the detector 1.5keV_{ee}. The energy threshold was

measured by comparing to the combined energy to the true tritium spectrum in chapter 4. The S1 detection efficiency, or the efficiency for detecting golden events, can be determined in the same manner by taking the ratio of the S1 spectrum observed to expected S1 spectrum. The observed S1 spectrum overlaid with the expected S1 spectrum is shown in figure 6.8 a). The result for detection efficiency of S1 measured using tritium data at 170 and 100 V/cm is shown in figure 7.9, along with the efficiency measured using LED calibrations. The efficiency for detecting S1 signals at 2 PE is found to be 75%, rising to >95% at 3 PE.




Figure 7.9: The S1 detection efficiency measured from tritium data at 170 V/cm (black square) and 100 V/cm (blue circle) is plotted along with the best fit to an error function to the higher statistics 170 V/cm data. The red triangles are detection efficiency determined from LED pulsing.

7.7 Conclusion

In this chapter we have described the development and implementation of a tritiated-methane calibration source. The primary application of the tritiated-methane calibration is to characterize the ER band in the WIMP search regions of 1-50 PE (1-10 keV_{ee}). This is of great importance, as it gauges the detectors ability to reject background (ER) events from potential WIMP signals (NR). Having both high statistics and high data purity from the tritium calibration, the future LUX profile likelihood analysis will not have to rely on the assumption that the ER band can be characterized by a Gaussian PDF. All previous xenon dark-matter searches using liquid xenon have assumed ER band Gaussianity in the WIMP search region, all having been calibration statistics limited. We find that the use of a Guassian PDF, for the LUX detector, fails below 3 sigma of the ER mean which is just below the NR mean, shown in figure 7.8. The deviation maybe arising from instrumental fluctuations or is perhaps a fundamental property of electronic recoils.

Using the calibration data collected, a purely data driven PDF can be constructed to characterize the distribution of ER events in each S1 bin, greatly improving the accuracy of the next WIMP limit. Much more fundamental xenon physics could be probed with the tritium calibration source, discussed in chapters 4, 5,6. The tritium calibration was used to test the energy scale calibration over the range from 1 to 18 KeV_{ee}, and to measure the light yield, charge yield and recombination fluctuation, and the detection thresholds down to 1 keV_{ee}. The data taken with the LUX detector has allowed for improvements to the NEST model down to 1 keV_{ee}.

and at our fields of 100 and 170 V/cm.

Bibliography

- [1] M. Weber and W. de Boer. Determination of the local dark matter density in our Galaxy. *AAP*, 509:A25, jan 2010. doi:10.1051/0004-6361/200913381.
- [2] Anse Slosar, et al. Measurement of Baryon Acoustic Oscillations in the Lyman-alpha Forest Fluctuations in BOSS Data Release 9. *JCAP*, 1304:026, 2013. doi:10.1088/1475-7516/2013/04/026.
- [3] J.D. Lewin and P.F. Smith. Review of mathematics, numerical factors, and corrections for dark matter experiments based on elastic nuclear recoil. *Astropart.Phys.*, 6:87–112, 1996. doi:10.1016/S0927-6505(96)00047-3.
- [4] Richard H. Helm. Inelastic and elastic scattering of 187-mev electrons from selected even-even nuclei. *Phys. Rev.*, 104:1466–1475, Dec 1956. doi:10.1103/PhysRev.104.1466.
- [5] D. S. Akerib, et al. Radiogenic and muon-induced backgrounds in the LUX dark matter detector. *Astroparticle Physics*, 62:33–46, March 2015. doi:10.1016/j.astropartphys.2014.07.009.
- [6] Scott Hertel. private communication, 2014.
- [7] M. Arenz and R. Vianden. 83mkr, a potentially powerful pac probe. *Hyperfine Interactions*, 222(1-3):73–76, 2013. ISSN 0304-3843. doi:10.1007/s10751-012-0723-3.
- [8] C. E. Dahl. *Ph.D. Thesis*. Ph.D. thesis, Princeton University, 2009.
- [9] L.W. Kastens, S. Bedikian, S.B. Cahn, A. Manzur, and D.N. McKinsey. A 83Krm Source for Use in Low-background Liquid Xenon Time Projection Chambers. *JINST*, 5:P05006, 2010. doi:10.1088/1748-0221/5/05/P05006.
- [10] L. Baudis, et al. Response of liquid xenon to compton electrons down to 1.5 kev. *Phys. Rev. D*, 87:115015, Jun 2013. doi:10.1103/PhysRevD.87.115015.

- [11] I. Ahmad, et al. Half-lives of isomeric states in fe57 and kr83. *Phys. Rev. C*, 52:2240–2241, Oct 1995. doi:10.1103/PhysRevC.52.2240.
- [12] S. L. Ruby, Y. Hazoni, and M. Pasternak. Lifetimes of the low-energy m1 transitions in la137 and kr83. *Phys. Rev.*, 129:826–828, Jan 1963. doi:10.1103/PhysRev.129.826.
- [13] E. Aprile, et al. Measurement of the scintillation yield of low-energy electrons in liquid xenon. *Phys. Rev. D*, 86:112004, Dec 2012. doi:10.1103/PhysRevD.86.112004.
- [14] P Venkataramaiah, K Gopala, A Basavaraju, S S Suryanarayana, and H Sanjeeviah. A simple relation for the fermi function. *Journal of Physics G: Nuclear Physics*, 11(3):359, 1985.
- [15] J. H. Oort. The force exerted by the stellar system in the direction perpendicular to the galactic plane and some related problems. *BAIN*, 6:249, August 1932.
- [16] F. Zwicky. On the Masses of Nebulae and of Clusters of Nebulae. *API*, 86:217, October 1937. doi:10.1086/143864.
- [17] K. G. Begeman, A. H. Broeils, and R. H. Sanders. Extended rotation curves of spiral galaxies: dark haloes and modified dynamics. *Monthly Notices of the Royal Astronomical Society*, 249(3):523–537, 1991. doi:10.1093/mnras/249.3.523.
- [18] Robert H. Brandenberger. Formation of structure in the universe. 1995.
- [19] Ryan Cooke, Max Pettini, Regina A. Jorgenson, Michael T. Murphy, and Charles C. Steidel. Precision measures of the primordial abundance of deuterium. 2013.
- [20] P.A.R. Ade et al. Planck 2013 results. XVI. Cosmological parameters. *Astron. Astrophys.*, 2014. doi:10.1051/0004-6361/201321591.
- [21] D. J. Fixsen. The Temperature of the Cosmic Microwave Background. *APJ*, 707:916–920, December 2009. doi:10.1088/0004-637X/707/2/916.
- [22] P.A.R. Ade et al. Planck 2013 results. I. Overview of products and scientific results. 2013. doi:10.1051/0004-6361/201321529.
- [23] Daniel J. Eisenstein, Hee-jong Seo, and 1 White, Martin J. On the Robustness of the Acoustic Scale in the Low-Redshift Clustering of Matter. *Astrophys.J.*, 664:660–674, 2007. doi:10.1086/518755.
- [24] W. J. Percival, et al. Baryon acoustic oscillations in the Sloan Digital Sky Survey Data Release 7 galaxy sample. *mnras*, 401:2148–2168, February 2010. doi:10.1111/j.1365-2966.2009.15812.x.

- [25] Maxim Markevitch. Chandra observation of the most interesting cluster in the universe. 2005.
- [26] Douglas Clowe, et al. A direct empirical proof of the existence of dark matter. *Astrophys.J.*, 648:L109–L113, 2006. doi:10.1086/508162.
- [27] Peter J. Kernan and Lawrence M. Krauss. Refined big bang nucleosynthesis constraints on Omega (baryon) and N (neutrino). *Phys.Rev.Lett.*, 72:3309–3312, 1994. doi:10.1103/PhysRevLett.72.3309.
- [28] Q.R. Ahmad et al. Direct evidence for neutrino flavor transformation from neutral current interactions in the Sudbury Neutrino Observatory. *Phys.Rev.Lett.*, 89:011301, 2002. doi:10.1103/PhysRevLett.89.011301.
- [29] George R. Blumenthal, S.M. Faber, Joel R. Primack, and Martin J. Rees. Formation of Galaxies and Large Scale Structure with Cold Dark Matter. *Nature*, 311:517–525, 1984. doi:10.1038/311517a0.
- [30] Gerard Jungman, Marc Kamionkowski, and Kim Griest. Supersymmetric dark matter. *Phys.Rept.*, 267:195–373, 1996. doi:10.1016/0370-1573(95)00058-5.
- [31] Andriy Kurylov and Marc Kamionkowski. Generalized analysis of weakly interacting massive particle searches. *Phys.Rev.*, D69:063503, 2004. doi:10.1103/PhysRevD.69.063503.
- [32] V. Chepel and H. Araújo. Liquid noble gas detectors for low energy particle physics. *Journal of Instrumentation*, 8:R04001, April 2013. doi:10.1088/1748-0221/8/04/R04001.
- [33] S. Akerib, D. et al. First results from the lux dark matter experiment at the sanford underground research facility. *Phys. Rev. Lett.*, 112:091303, Mar 2014. doi:10.1103/PhysRevLett.112.091303.
- [34] E. Aprile, et al. Dark Matter Results from 225 Live Days of XENON100 Data. *Physical Review Letters*, 109(18):181301, November 2012. doi:10.1103/PhysRevLett.109.181301.
- [35] Mengjiao Xiao et al. First dark matter search results from the PandaX-I experiment. 2014.
- [36] K. Abe, et al. Light WIMP search in XMASS. *Physics Letters B*, 719:78–82, February 2013. doi:10.1016/j.physletb.2013.01.001.
- [37] A. Wright. The DarkSide Program at LNGS. In S. Giani and et al., editors, *Astroparticle, Particle, Space Physics and Detectors For Physics Applications - Proceedings of the 13th ICATPP Conference. Edited by Giani Simone et al. Published by World Scientific Publishing Co. Pte. Ltd., 2012. ISBN #9789814405072, pp. 414-420, pages 414–420*. August 2012. doi:10.1142/9789814405072_0061.

- [38] A. Hime. The MiniCLEAN Dark Matter Experiment. *ArXiv e-prints*, October 2011.
- [39] CDMS II Collaboration, et al. Dark Matter Search Results from the CDMS II Experiment. *Science*, 327:1619–, March 2010. doi:10.1126/science.1186112.
- [40] C. E. Aalseth, et al. CoGeNT: A search for low-mass dark matter using p-type point contact germanium detectors. *prd*, 88(1):012002, July 2013. doi:10.1103/PhysRevD.88.012002.
- [41] S. Archambault et al. Constraints on Low-Mass WIMP Interactions on ^{19}F from PICASSO. *Phys.Lett.*, B711:153–161, 2012. doi:10.1016/j.physletb.2012.03.078.
- [42] E. Behnke, et al. First dark matter search results from a 4-kg CF_3I bubble chamber operated in a deep underground site. *prd*, 86(5):052001, September 2012. doi:10.1103/PhysRevD.86.052001.
- [43] E. Aprile and T. Doke. Liquid xenon detectors for particle physics and astrophysics. *Rev. Mod. Phys.*, 82:2053–2097, Jul 2010. doi:10.1103/RevModPhys.82.2053.
- [44] Attila Dobi, et al. Detection of krypton in xenon for dark matter applications. *Nucl.Instrum.Meth.*, A665:1–6, 2011. doi:10.1016/j.nima.2011.11.043.
- [45] A.I. Bolozdynya, P.P. Brusov, T. Shutt, C.E. Dahl, and J. Kwong. A chromatographic system for removal of radioactive ^{85}Kr from xenon. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 579(1):50 – 53, 2007. ISSN 0168-9002. doi: DOI:10.1016/j.nima.2007.04.011. Proceedings of the 11th Symposium on Radiation Measurements and Applications.
- [46] K. Abe, et al. Distillation of liquid xenon to remove krypton. *Astroparticle Physics*, 31(4):290 – 296, 2009. ISSN 0927-6505. doi:DOI:10.1016/j.astropartphys.2009.02.006.
- [47] John R de Laeter, et al. Atomic weights of the elements. review 2000 (iupac technical report). *Pure and Applied Chemistry*, 75(6):683–800, 2003.
- [48] M. Miyajima, et al. Average energy expended per ion pair in liquid argon. *Phys. Rev. A*, 9:1438–1443, Mar 1974. doi:10.1103/PhysRevA.9.1438.
- [49] T. Takahashi, et al. Average energy expended per ion pair in liquid xenon. *Phys. Rev. A*, 12:1771–1775, Nov 1975. doi:10.1103/PhysRevA.12.1771.
- [50] T. Doke and K. Masuda. Present status of liquid rare gas scintillation detectors and their new application to gamma-ray calorimeters. *Nucl.Instrum.Meth.*, A420:62–80, 1999. doi:10.1016/S0168-9002(98)00933-4.

- [51] S Kubota, M Hishida, and J Raun. Evidence for a triplet state of the self-trapped exciton states in liquid argon, krypton and xenon. *Journal of Physics C: Solid State Physics*, 11(12):2645, 1978.
- [52] Akira Hitachi, et al. Effect of ionization density on the time dependence of luminescence from liquid argon and xenon. *Phys.Rev.*, B27:5279–5285, 1983. doi:10.1103/PhysRevB.27.5279.
- [53] Jeremy Mock, et al. Modeling Pulse Characteristics in Xenon with NEST. *JINST*, 9:T04002, 2014. doi:10.1088/1748-0221/9/04/T04002.
- [54] J. V. Dawson, et al. A study of the scintillation induced by alpha particles and gamma rays in liquid xenon in an electric field. *Nuclear Instruments and Methods in Physics Research A*, 545:690–698, June 2005. doi:10.1016/j.nima.2005.01.343.
- [55] David P. Lide, editor. *CRC Handbook of Chemistry and Physics*. CRC PRESS, 81 edition, 2000.
- [56] T. Doke, A. Hitachi, S. Kubota, A. Nakamoto, and T. Takahashi. Estimation of Fano factors in liquid argon, krypton, xenon and xenon-doped liquid argon. *Nucl.Instrum.Meth.*, 134:353–357, 1976. doi:10.1016/0029-554X(76)90292-5.
- [57] Jens Lindhard, V Nielsen, M Scharff, and PV Thomsen. *Integral equations governing radiation effects*. Munksgaard i komm., 1963.
- [58] M. Szydagis, et al. NEST: A Comprehensive Model for Scintillation Yield in Liquid Xenon. *ArXiv e-prints*, June 2011.
- [59] Matthew Szydagis, Adalyn Fyhrie, Daniel Thorngren, and Mani Tripathi. Enhancement of NEST Capabilities for Simulating Low-Energy Recoils in Liquid Xenon. *JINST*, 8:C10003, 2013. doi:10.1088/1748-0221/8/10/C10003.
- [60] W. . Chen, et al. Measurement of the transverse diffusion coefficient of charge in liquid xenon. *ArXiv e-prints*, September 2011.
- [61] R. Agnese et al. Search for Low-Mass Weakly Interacting Massive Particles Using Voltage-Assisted Calorimetric Ionization Detection in the SuperCDMS Experiment. *Phys.Rev.Lett.*, 112(4):041302, 2014. doi:10.1103/PhysRevLett.112.041302.
- [62] G. Angloher, et al. Results from 730 kgdays of the cresset-ii dark matter search. *The European Physical Journal C*, 72(4):1971, 2012. ISSN 1434-6044. doi:10.1140/epjc/s10052-012-1971-8.
- [63] R. Bernabei, et al. Final model independent result of DAMA/LIBRA-phase1. *European Physical Journal C*, 73:2648, December 2013. doi:10.1140/epjc/s10052-013-2648-7.

- [64] A. Dobi, et al. A xenon gas purity monitor for EXO. *Nuclear Instruments and Methods in Physics Research A*, 659:215–228, December 2011. doi:10.1016/j.nima.2011.09.017.
- [65] C. Silva, et al. A model of the reflection distribution in the vacuum ultra violet region. *Nuclear Instruments and Methods in Physics Research A*, 619:59–62, July 2010. doi:10.1016/j.nima.2009.10.086.
- [66] C. Silva, et al. Reflectance of polytetrafluoroethylene for xenon scintillation light. *Journal of Applied Physics*, 107(6):064902, March 2010. doi:10.1063/1.3318681.
- [67] V. Eckardt, et al. Calibration of the STAR forward time projection chamber with krypton-83m. 2001.
- [68] Tadayoshi Doke, et al. Absolute scintillation yields in liquid argon and xenon for various particles. *Japanese Journal of Applied Physics*, 41(3R):1538, 2002.
- [69] Sz. Nagy, T. Fritioff, M. Bjorkhage, I. Bergström, and R. Schuch. On the q-value of the tritium β -decay. *EPL (Europhysics Letters)*, 74(3):404, 2006.
- [70] D. P. Gregory and D. A. Landsman. Average decay energy of tritium. *Phys. Rev.*, 109:2091–2091, Mar 1958. doi:10.1103/PhysRev.109.2091.
- [71] J. Thomas and D. A. Imel. Recombination of electron-ion pairs in liquid argon and liquid xenon. *Phys. Rev. A*, 36:614–616, Jul 1987. doi:10.1103/PhysRevA.36.614.
- [72] J. B. Birks. *Proc. Phys. Soc.*, A64:874, 1951.
- [73] R. L. Platzman. *Int.J.Appl.Radiat.Isotopes*, page 116, 1961.
- [74] A. Mozumder. Free-ion yield and electron-ion recombination rate in liquid xenon. *Chemical Physics Letters*, 245(45):359 – 363, 1995. ISSN 0009-2614. doi:[http://dx.doi.org/10.1016/0009-2614\(95\)01024-4](http://dx.doi.org/10.1016/0009-2614(95)01024-4).
- [75] K. N. Pushkin, et al. Scintillation light, ionization yield and scintillation decay times in high pressure xenon and xenon methane. *Nuclear Science, IEEE Transactions on*, 54(3):744–750, June 2007. ISSN 0018-9499. doi:10.1109/TNS.2007.894815.
- [76] D. C. Mallings, et al. After LUX: The LZ Program. *ArXiv e-prints*, October 2011.
- [77] E. Aprile and XENON1T collaboration. The XENON1T Dark Matter Search Experiment. *ArXiv e-prints*, June 2012.
- [78] L. L. Lucas and M. P. Unterweger. Comprehensive review and critical evaluation of the half-life of tritium. *Journal of Research of NIST*, 105(4):541 – 551, 2000. ISSN 0883-2889. doi:<http://dx.doi.org/10.6028/jres.105.043>.

- [79] *PS4-MT3/15-R/N SPECIFICATIONS.*
- [80] Hitoshi Miyake, Masao Matsuyama, Kan Ashida, and Kuniaki Watanabe. Permeation, diffusion, and solution of hydrogen isotopes, methane, and inert gases in/through tetrafluoroethylene and polyethylene. *Journal of Vacuum Science Technology A: Vacuum, Surfaces, and Films*, 1(3):1447–1451, Jul 1983. ISSN 0734-2101. doi:10.1116/1.572038.
- [81] A Bondar, et al. Two-phase argon and xenon avalanche detectors based on gas electron multipliers. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 556(1):273–280, 2006.
- [82] Eido Shibamura, et al. Drift velocities of electrons, saturation characteristics of ionization and w-values for conversion electrons in liquid argon, liquid argon-gas mixtures and liquid xenon. *Nuclear Instruments and Methods*, 131(2):249 – 258, 1975. ISSN 0029-554X. doi:[http://dx.doi.org/10.1016/0029-554X\(75\)90327-4](http://dx.doi.org/10.1016/0029-554X(75)90327-4).
- [83] A. Dobi, et al. Study of a zirconium getter for purification of xenon gas. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 620(23):594 – 598, 2010. ISSN 0168-9002. doi:<http://dx.doi.org/10.1016/j.nima.2010.03.151>.
- [84] D.S. Leonard, et al. A simple high-sensitivity technique for purity analysis of xenon gas. *Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment*, 621(1-3):678 – 684, 2010. ISSN 0168-9002. doi:DOI:10.1016/j.nima.2010.04.152.
- [85] E. Aprile, T. Yoon, A. Loose, L. W. Goetzke, and T. Zelevinsky. An atom trap trace analysis system for measuring krypton contamination in xenon dark matter detectors. *Review of Scientific Instruments*, 84(9):093105, September 2013. doi:10.1063/1.4821879.
- [86] Sebastian Lindemann and Hardy Simgen. Krypton assay in xenon at the ppq level using a gas chromatographic system and mass spectrometer. *Eur.Phys.J.*, C74:2746, 2014. doi:10.1140/epjc/s10052-014-2746-1.
- [87] A. Dobi, et al. Xenon purity analysis for EXO-200 via mass spectrometry. *Nucl.Instrum.Meth.*, A675:40–46, 2012. doi:10.1016/j.nima.2012.01.066.
- [88] J. Escada, et al. A Monte Carlo study of the fluctuations in Xe electroluminescence yield: pure Xe vs Xe doped with CH₄ or CF₄ and planar vs cylindrical geometries. *Journal of Instrumentation*, 6:8006, August 2011. doi:10.1088/1748-0221/6/08/P08006.
- [89] B.J.P. Jones, et al. The Effects of Dissolved Methane upon Liquid Argon Scintillation Light. *JINST*, 8:P12015, 2013. doi:10.1088/1748-0221/8/12/P12015.