042 Non-Progr Instruction Manual:Layout 1 8/11/11 1:37 PM Page 1

INSTRUCTION MANUAL

Hearing Healthcare Provider

Made in U.S.A.

Audina HEARING INSTRUMENTS, INC
P. O. Box 180877
Casselberry, FL 32718-0877
audina.net

ver: 1.2 08.01.11

Non-Programmable Analog

Hearing Instruments

- Full Shell
- Half Shell
- Mini Canal
- Canal
- CIC

Audina HEARING INSTRUMENTS, INC

042 Non-Progr Instruction Manual:Layout 1 8/11/11 1:37 PM Page 3

Non-Programmable Analog Hearing Instruments

Welcome

Welcome to a world of better hearing! We understand that hearing loss affects one's quality of life. That is why we are dedicated to building the highest quality hearing instruments. Your new hearing instruments will soon be indispensable to you and will bring new pleasure and enjoyment to your life.

Our company was established in 1990 and is well-known internationally for offering high quality, reliable, state-of-the-art technology. These hearing instruments have been adjusted to your individual hearing loss to precisely fit your hearing and communication requirements. If necessary, your hearing healthcare provider can make further modifications to enhance your listening experience.

A period of adjustment may be needed to become accustomed to your new hearing instruments and the renewed world of sounds you will experience. The adjustment time varies for each individual and may be as little as a couple of weeks to as long as several months. Allow yourself time to become familiar with those sounds you have not heard for a long time.

Before operating your hearing instruments, please read this manual thoroughly.

Table of Contents

Inserting Hearing Instruments 1
Serial Numbers 1
Additional Options 2
Battery Information 3
Telephone Use 5
Cell Phone Compatibility 6
Adjusting Volume 7
Cleaning 8
Care and Storage 9
Do's and Don'ts 10
Warranty Information 12
Returning for Service 13
Troubleshooting 15
Binaural Amplification 17
Children with Hearing Loss 18
Notice to Prospective Users 19
Notice to Providers 20
Owner Information 22

Inserting Your Hearing Instrument

These hearing instruments have been custom made for your ears. The hearing aids should fit comfortably and easily in your ears. Your hearing healthcare provider will provide instruction on proper insertion and removal.

Interpretation of Serial Number

Each hearing instrument has its own unique number located on the exterior of the hearing aid.

Blue writing or blue coloring denotes a left hearing instrument. Red writing or red coloring denotes a right hearing instrument. The first two digits of the serial number represent the year in which the instrument was manufactured. The second two digits of the serial number represent the month in which the hearing instrument was manufactured.



Additional Options

Toggle Switch

This feature can be used to switch back and forth between two distinct listening situations. For example, this feature may be linked with your telecoil (see below) so that you can switch to that feature when you answer the telephone. Your hearing healthcare provider will instruct you on how to best use your toggle switch.

Directional Microphone

When activated, this feature may reduce distracting noise arriving from behind the hearing aid user.

Telecoil Only

This option allows use of the telephone without the whistling, or feedback, that can often occur when the phone is placed next to a hearing aid. This may be desirable for someone using certain assistive listening devices that work in conjunction with hearing aids.

Batteries

To replace your hearing instrument battery, follow these steps:

Step 1:
Lift the notch at the edge of the compartment and swing open the door. Remove the battery.





Step 2:
Remove the colored tab on the battery, and insert the battery with + side up.



Step 3:
Swing the door into the closed position. The compartment should close easily. Do not force it. If resistance is noted, check that the battery is inserted correctly.



The life of your hearing instrument's battery will vary depending on length of daily use and listening environments. When the hearing aid battery goes dead, remove it immediately and replace with a new battery. Batteries may be purchased through your hearing healthcare professional.

When your hearing instruments are not in use, open the battery compartment to prevent excessive battery drain and to ventilate the instruments.

Important: *Always discard used batteries. Small batteries can be harmful if swallowed. Keep batteries out of the reach of pets and small children. In case of ingestion, contact your physician or call the National Button Battery Hotline collect at 202.625.3333.*

Telephone Use

Unless your hearing loss is severe enough to require the use of an amplified telephone, there is an option available on most hearing aids to eliminate the whistling, or feedback, that may be encountered as a telephone is placed next to a hearing aid. This option is called a Telecoil.

If you DO have a Telecoil

Use your toggle switch to change to the "Telecoil" position. You must have a hearing aid compatible telephone to work with a Telecoil. Your hearing healthcare provider can advise you on how to find out whether your phone is compatible.

If you do NOT have a Telecoil:

If there is a "whistling" sound, simply move the telephone slightly away from the ear.

Cell Phone Compatibility

Some hearing aid users have reported a buzzing sound in their hearing aids when they are using cell phones, indicating that the cell phone and hearing aid may not be compatible. According to the ANSI C63.19 standard (ANSI C63.19-2006 American National Standard Methods of Measurement of Compatibility Between Wireless Communications Devices and Hearing Aids), the compatibility of a particular hearing aid and cell phone can be predicted by adding the rating for the hearing aid immunity to the rating for the cell phone emissions. For example, the sum of a hearing aid rating of 2 (M2/T2) and a telephone rating of 3 (M3/T3) would result in a combined rating of 5. Any combined rating that equals at least 5 would provide "normal use;" a combined rating of 6 or greater would indicate "excellent performance."

The immunity of this hearing aid is at least M2/T2. The equipment performance measurements, categories and system classifications are based upon the best information available but cannot guarantee that all users will be satisfied.

NOTE: *The performance of individual hearing aids may vary with individual cell phones. Therefore, please try this hearing aid with your cell phone or, if you are purchasing a new phone, be sure to try it with your hearing aid prior to purchase. For additional guidance, please ask your cell phone provider for the booklet entitled "Hearing Aid Compatibility with Digital Wireless Cell Phones."*

Adjusting the Volume

Place your finger on the wheel and rotate it forward (towards your nose) to increase the volume.

Rotate the wheel backwards (towards the back of your head) to decrease the volume.

Cleaning Your Hearing Instrument

You should inspect your hearing instruments daily and clean them periodically to control wax buildup. To clean, simply wipe the surface with a soft, dry cloth or use cleaning wipes specifically for hearing instruments.

Do not allow water to enter any openings on the instruments.

Inspect the tip of the ear canal portion of the hearing instruments for any wax buildup. Using only the wax tool supplied, carefully remove any wax from around the sound outlet as directed by your hearing healthcare provider. Be very careful not to push wax into the opening or to insert the tool deeply into the opening.

Cleaning Your Hearing Instrument

The wearing of an earmold or in-the-ear hearing instruments may accelerate the accumulation of ear wax. This can easily be removed by your physician or in some cases your hearing healthcare provider. You should not use cotton swabs to clean your ears, due to the risk of pushing the wax deeper into the ear canal.

The materials used in your hearing instruments are of the highest quality available, and the likelihood of an allergic reaction is slight. If a skin irritation occurs, including redness, swelling or itching where the hearing aid makes contact with your ear - contact your physician and hearing healthcare provider.

Caring for Your Instrument

Hearing aids are exposed to moisture in the form of humidity and perspiration while in your ears. The daily use of a dehumidifying system or drying kit approved by your hearing healthcare provider is recommended. These accessories may be purchased from your hearing healthcare provider.

The most common causes of hearing aid repair are wax and moisture. If you are encountering recurrent repairs, please discuss the various wax/moisture prevention systems and accessories with your hearing healthcare provider.

Extended Storage

If it is necessary to store your hearing instruments for an extended amount of time, follow the steps below for proper storage.

1. Remove the batteries.
2. Remove any wax buildup.
3. Place the instruments in a cool, dry place. Either in their original container or a hearing aid dehumidifier is recommended.

The Do's and Dont's of Hearing Instruments

Do store the instruments overnight in a hearing aid dehumidifier to absorb moisture (Be sure and remove batteries first).

Do always carry spare batteries.

Do dry the battery contacts with a cotton swab in cases of humid weather or heavy perspiration.

Do safely store hearing instruments when not in use. Keep them out of the reach of small children and pets.

Don't leave your hearing instruments on the radiator, near a stove, in a sunny window, in a car or in any other hot place. Excessive heat and cold can damage the instruments.

Don't wear the hearing instruments when using a hair dryer at home or when at the beauty salon, spa or barber.

Don't apply hair spray when wearing your hearing instruments. It may damage the microphone and erode the faceplate.

Don't wear the instruments while taking a bath, shower, spa treatment or while swimming.

Don't store the instruments in the bathroom during a shower or near a steam room as moisture created by the steam may cause damage.

Warranty Information

Please see your hearing healthcare provider regarding your specific warranty information. Our guarantee covers any defect in material or faulty production, except cords and batteries. We will, at our option, repair or replace the instruments or any defective part, at our expense, within the limits of the warranty.

This warranty will become invalid if your hearing instruments have been tampered with or altered in any way, or if the serial number has been altered, effaced, or removed, or if any repairs have been made without authorization.

This warranty does not include a promise of specific hearing improvement because that depends upon your audiologic needs, proper fitting and adjustments by your hearing healthcare provider and proper use by the owner.

Any implied warranty shall be limited to one year from date of purchase. The only entity or individual who is authorized to perform the warranty obligations is the manufacturer.

All in-house warranty work performed by the manufacturer will be done within a reasonable period of time. If warranty work is performed and there appears to be a defect, malfunction, or failure to conform to the warranty, the obligations will be met if brought to our attention within one month from the date of the original warranty work.

This warranty is specifically limited to the scope set forth herein, and may not in any way be expanded by the other oral or written representations.

Returning for Service

You should visit your hearing healthcare provider if your hearing instruments are in need of service. If you no longer have a hearing healthcare provider and are in need of one, you may contact Audina and we will assist you in finding one in your area. If you are unable to visit a hearing healthcare provider, you may send your hearing aids directly to the manufacturer.

Before planning to return your instruments for repair, please refer to the troubleshooting chart on the proceeding page. Many times these solutions can prevent the time and cost associated with a repair.

If you need to send your instruments directly to the manufacturer, carefully pack your instruments in a sturdy box and send it to:

Audina® Hearing Instruments, Inc.
P. O. Box 180877
Casselberry, FL 32718-0877

You must include your current phone number and home address. There is a $40 charge to cover shipping and handling. Please include a check, payable to Audina Hearing Instruments, Inc. for said amount when sending your instruments for service. It is recommended to purchase shipping insurance when sending your hearing aids to the manufacturer.

If your hearing instruments are out of warranty, we will assist you in finding a hearing healthcare provider in your local area.

Troubleshooting Guidelines

No Sound:

- Ensure that hearing instruments are turned on.
- Replace the battery.
- Ensure that there is no wax or debris in the sound channel.

Not loud enough:

- Replace the battery.
- Ensure that there is no wax or debris in the sound channel.
- Ensure that hearing instruments are completely seated in the ear canal.
- Excessive earwax can prevent sound from entering the ear. Contact your physician or hearing healthcare provider.
- Changes in your hearing loss can affect how you hear with a hearing aid. Contact your hearing healthcare provider if you suspect any change in your hearing status.

Whistling / Feedback:

- Ensure that hearing instruments are completely seated in the ear canal.
- Excessive earwax, improperly sized instruments, and large vents can cause whistling. Contact your hearing healthcare provider.

Sound is Distorted or Unclear:

- Replace the battery.
- Hearing instruments may have a poor battery contact or may not be programmed properly. Return to your hearing healthcare provider.

Intermittent Sound:

- Replace the battery.
- Ensure that there is no wax or debris in the sound channel.
- Hearing instruments may have a poor battery contact. Return to your hearing healthcare provider.

Too Much Bass:

- Ensure vent is clear.
- Hearing instruments may need reprogramming. Return to your hearing healthcare provider.

Static Noise:

- Replace the battery.
- If you are near an electromagnetic field (i.e. near a computer screen or fluorescent lights), step away to see if the static clears.
- Hearing instruments may have a poor battery contact. Return to your hearing healthcare provider.
- A buzzing sound is normal when in Telecoil mode. Make sure your toggle switch is in the proper position.

Binaural Amplification

Hearing loss can be improved, in most cases, through medicine, surgery, or with hearing instruments. Hearing aids are the most commonly chosen option.

Today's hearing instruments are quite small in size and feature sophisticated electronics that accommodate a wide variety of hearing losses. Nature gave us two ears for a reason and since the majority of people who have a hearing loss are affected in both ears, two (binaural) hearing instruments are typically recommended. Two hearing instruments should provide these benefits:

- Better overall sound quality when compared to one hearing instrument ("stereo sound").
- The ability to locate a sound source in noisy and normal listening conditions.
- Increased ability to understand speech in noisy and normal listening conditions.
- A sensation of balanced hearing.

Talk with your hearing healthcare provider about what is best for you. Even the best technology can't do the work of two ears!

Children with Hearing Loss

In addition to seeing a physician for a medical evaluation, a child with a hearing loss should be directed to an audiologist for evaluation and rehabilitation since hearing loss may cause problems in language development and the education and social growth of a child. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of a child with hearing loss.

Notice to Prospective Users

Good health practice requires that a person with a hearing loss have a medical exam by a licensed physician, who specializes in diseases of the ear, before purchasing a hearing instrument.

Licensed physicians who specialize in the diseases of the ear are often referred to as Otologists, Otolaryngologists (ear, nose and throat – ENT) or Otorhinolaryngologists.

The purpose of a medical evaluation is to assure that all medically treatable conditions that may affect hearing are identified and treated before the hearing instrument is purchased. Following the medical evaluation, the physician will give you a written statement that your hearing loss has been medically evaluated and that you may be considered a candidate for a hearing instrument.

A hearing specialist will conduct an evaluation to assess your ability to hear. The hearing instrument evaluation will enable the hearing specialist to select and fit a hearing instrument to your individual needs. Federal law restricts the sale of hearing instruments to those individuals who have obtained a medical evaluation from a licensed physician.

Federal law allows a fully informed adult to sign a waiver statement declining the medical evaluation for religious or personal beliefs that preclude consultation with a physician. The exercise of such a waiver is not in your best interest, and its use is strongly discouraged.

Notice for Hearing Healthcare Providers

Hearing healthcare providers should advise a prospective hearing instrument user to consult promptly with a licensed physician (preferably an ear specialist) before dispensing a hearing instrument if the hearing healthcare provider determines through inquiry, actual observation, or review of any other available information, that the prospective user has any of the following conditions:

- Visible congenital or traumatic deformity of the ear.
- History of active drainage from the ear within the previous 90 days.
- History of sudden or rapidly progressive hearing loss within the previous 90 days.
- Acute or chronic dizziness.
- Unilateral hearing loss of sudden or recent onset within the previous 90 days.

- Audiometric air-bone gap greater than or equal to 15 decibels at 500 hertz (Hz), 1000 Hz, and 2000 Hz.
- Visible evidence of significant cerumen (wax) accumulation or a foreign body in the ear canal.
- Pain or discomfort in the ear.

Special care should be exercised in selecting and fitting a hearing instrument which has a maximum sound pressure level in excess of 132 dB (decibels) because there may be a risk of damaging the remaining hearing of the hearing instrument user.

Owner Information

MODEL: _____

SERIAL NUMBERS:

LEFT: _____

RIGHT: _____

PURCHASE DATE: _____

BATTERY SIZE: _____

ORIGINAL WARRANTY EXPIRES: _____

HEARING HEALTH CARE PROVIDER: _____

_____

_____

Notes