Screening flag raised — further evaluation recommended

Screening result

Patient name	Demo TB patient
Age / sex	48 / Female
Patient reference	DEMO-PDF
Internal patient ID	#1
Screening date (UTC)	2026-11-03 16:17:31
Screening module	Tuberculosis (chest X-ray)
Modality routing	Chest X-ray — selected manually and confirmed by the gatekeeper model (1.00)
Image quality	PASSED BASIC CHECKS
Model score	86.35% (uncalibrated model score)
Decision threshold	25.75% (± band 21.26%–30.81%)
Model version	tb-aaa0b6fc
Processing time	366 ms

The model score is the model's raw output, not a calibrated probability of disease.

Image quality measurements

Dimensions	512 x 512 px
Sharpness (Laplacian variance)	26.277
Exposure (1st / 99th percentile)	0.0 / 171.0 (0-255)

Technical checks are separate from disease score. Passing does not establish diagnostic quality.

Grad-CAM visual explanation

Grad-CAM not available for this report

Suggested next steps — Priority referral

- Refer to a qualified clinician (Medical Officer) for clinical assessment.
- Arrange sputum testing, preferably a WHO-recommended molecular test (NAAT such as CBNAAT/Truenat), as per national TB programme guidelines.
- The chest X-ray should be reviewed by a radiologist or clinician.
- This tool does not recommend medicines or treatment. Management decisions belong to the treating clinician.

Questionnaire findings (not part of the image model)

- No questionnaire answers recorded.

Limitations

- The model was trained on a small public dataset and has not been clinically validated.
- The model score is not a calibrated probability of disease.
- A result without a screening flag does not exclude disease.
- Grad-CAM highlights model attention only and is not a lesion map.

Disclaimer

This is an experimental AI screening aid, not a medical diagnosis. Results must be reviewed by a qualified healthcare professional and confirmed through appropriate clinical testing.

Reviewed by (clinician name & signature): _____