

GOVERNMENT OF INDIA - LEGAL METROLOGY

Verification Certificate (OIML R-76)

Inspector ID: LMO-101

Shop/Entity Name: yya

Standard Weight Applied: 1000.0 g

Machine Reading: 998.0 g

Detected Error: 2.0 g

Result: APPROVED & STAMPED

