

Surface Coating Solutions
Certificate of Analysis

COA Number:	COA-2023-06-002
Product:	Sensor Cartridges NEXUS CK MB
Product Code:	CART-CKMB-001
Lot Number:	CART-2023-06-002
Manufacturing Date:	2023-06-20
Expiry Date:	2024-12-20
Quantity:	50,000 units
Customer:	MedTech Diagnostics Inc
Customer PO:	PO-MT-2023-0485

Test Results

Test Parameter	Method	Specification	Result	Status
Antibody Density	UV-Vis 280 nm	2.0 - 2.5 mg/mL	2.4 mg/mL	PASS
Binding Capacity	AM-CART-003	>= 95%	98.5%	PASS
Sterility	Ph. Eur. 2.6.1	No growth (14 days)	No growth	PASS
Endotoxin	LAL kinetic	< 0.5 EU/mL	0.08 EU/mL	PASS
Weight per cartridge	Gravimetric	1.8 - 2.2 g	2.05 g	PASS
Visual Appearance	Visual	No defects	Conforms	PASS
Cap Colour	Visual	Red	Red	PASS

Note: Antibody density of 2.4 mg/mL is within the accepted specification range (2.0 - 2.5 mg/mL) but is at the upper end of the range. Previous lot CART-2023-05-001 had an antibody density of 2.1 mg/mL. This lot-to-lot variation is within normal manufacturing process capability. Binding capacity of 98.5% exceeds the minimum requirement of 95%.

Lot Comparison

Parameter	CART-2023-05-001	CART-2023-06-002	Specification
Antibody Density	2.1 mg/mL	2.4 mg/mL	2.0 - 2.5 mg/mL

Binding Capacity	96.2%	98.5%	>= 95%
Endotoxin	0.12 EU/mL	0.08 EU/mL	< 0.5 EU/mL

Weight per cartridge	2.01 g	2.05 g	1.8 - 2.2 g
----------------------	--------	--------	-------------

Manufacturing Information
Manufactured under ISO 13485:2016 quality management system. Production environment: ISO Class 7 cleanroom.
Antibody source: Monoclonal anti-CK MB (clone 3H6), mouse origin. Coating process: Passive adsorption, 18-hour incubation at 4 C. Post-coating blocking: BSA 1% solution. Antibody master lot: MAB-2023-02 (same as previous lot CART-2023-05-001).

Storage Conditions
Store at 2-8 C. Protect from light. Do not freeze. Use within shelf life. After opening, use within 7 days.

Certification
This is to certify that lot CART-2023-06-002 has been manufactured, tested, and released in accordance with the approved specifications and Surface Coating Solutions quality procedures. All test results meet the acceptance criteria.

QC Manager

Dr. A. Lawrence

Date: 2023-06-23

Production Manager

M. Gibson

Date: 2023-06-23

