

BioCoating Solutions

Certificate of Analysis

COA Number:	COA-2023-05-001
Product:	SPR Cones NEXUS CK MB
Product Code:	SPR-CKMB-001
Lot Number:	SPR-2023-05-001
Manufacturing Date:	2023-05-15
Expiry Date:	2024-11-15
Quantity:	50,000 units
Customer:	MedTech Diagnostics SA
Customer PO:	PO-BM-2023-0420

Test Results

Test Parameter	Method	Specification	Result	Status
Antibody Density	UV-Vis 280 nm	2.0 - 2.5 mg/mL	2.1 mg/mL	PASS
Binding Capacity	AM-SPR-003	>= 95%	96.2%	PASS
Sterility	Ph. Eur. 2.6.1	No growth (14 days)	No growth	PASS
Endotoxin	LAL kinetic	< 0.5 EU/mL	0.12 EU/mL	PASS
Weight per cone	Gravimetric	1.8 - 2.2 g	2.01 g	PASS
Visual Appearance	Visual	No defects	Conforms	PASS
Cap Colour	Visual	Red	Red	PASS

Manufacturing Information

Manufactured under ISO 13485:2016 quality management system. Production environment: ISO Class 7 cleanroom. Antibody source: Monoclonal anti-CK MB (clone 3H6), mouse origin. Coating process: Passive adsorption, 18-hour incubation at 4 C. Post-coating blocking: BSA 1% solution.

Storage Conditions

Store at 2-8 C. Protect from light. Do not freeze. Use within shelf life. After opening, use within 7 days.

Certification

This is to certify that lot SPR-2023-05-001 has been manufactured, tested, and released in accordance with the approved specifications and BioCoating Solutions quality procedures. All test results meet the acceptance criteria.

QC Manager

Dr. A. Lefebvre

Date: 2023-05-18

Production Manager

M. Girard

Date: 2023-05-18