



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005 USA

Phone: +1 425-643-9866
Fax: +1 425-643-9954
E-mail: info@Sea-BirdScientific.com

SHALLOW SEAFET V2 CALIBRATION SHEET:

SeaFET Serial Number: PHS-2146

DuraFET Serial Number:

pH Calibration Date: 27-May-20

External Reference Coefficients:

K0	-1.328181e+00
K2	-1.002588e-03

Internal Reference Coefficients:

KDF0	-1.337108e+00
KDF2	-1.209260e-03

Thermistor Coefficients:

TDFA0	-1.542604e-03
TDFA1	5.451845e-04
TDFA2	-1.988318e-05
TDFA3	4.501564e-07