MRI-READY SYSTEMS BY ABBOTT

For the most up to date information, please refer to the applicable Abbott MRI-Ready Systems Manual at https://medical.abbott/manuals.

3T MRI-Ready Systems

Insertable Cardiac Monitor (ICM) Systems*§

Implanted Device	RF Transmit Conditions	Scan Region
Assert-IQ™ ICM DM5500, DM5300, DM5000 **Jot Dx™ ICM** DM4500 **Confirm Rx™ ICM** DM3500	First Level Controlled Operating Mode or Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)	Full-body

Pacemaker Systems*§

Implanted Device	Implanted Lead/Length (cm)	RF Transmit Conditions	Scan Region
AVEIR™ Leadless Pacemaker LSP112V LSP202V LSP201A	**Leadless**	First Level Controlled Operating Mode or Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)	Full-body
Zenex™ MRI Pacemaker PM1282, PM2282 **Zenus™ MRI Pacemaker** PM1182, PM2182 **Assurity MRI™ Pacemaker** PM1272, PM2272 **Endurity MRI™ Pacemaker** PM1172, PM2172	**UltiPace™ Pacing Lead†** LPA1231-46, 52, 58, 65 **Tendril™ STS Pacing Lead** 2088TC-46, 52, 58 **IsoFlex™ Optim™ Pacing Lead** 1944-46, 52 1948-52, 58	Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)†	Full-body

Implantable Cardioverter Defibrillator (ICD) Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted High Voltage Lead/Length (cm)	RF Transmit Conditions	Scan Region
Neutrino™ NxT VR DF4 ICD CDVRA600Q **Neutrino™ NxT DR DF4 ICD** CDDRA600Q **Gallant™ VR DF4 ICD** CDVRA500Q **Gallant™ DR DF4 ICD** CDDRA500Q **Entrant™ VR DF4 ICD** CDVRA300Q **Entrant™ DR DF4 ICD** CDDRA300Q	**UltiPace™ Pacing Lead** LPA1231-46, 52 **Tendril™ STS Pacing Lead** 2088TC-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65	Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)	Full-body
Neutrino™ NxT VR DF-1 ICD CDVRA600T **Neutrino™ NxT DR DF-1 ICD** CDDRA600T **Gallant™ VR DF-1 ICD** CDVRA500T **Gallant™ DR DF-1 ICD** CDDRA500T **Entrant™ VR DF-1 ICD** CDVRA300T **Entrant™ DR DF-1 ICD** CDDRA300T	**UltiPace™ Pacing Lead** LPA1231-46, 52 **Tendril™ STS Pacing Lead** 2088TC-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120-65 7122-60, 65 **Optisure™ Defibrillation Lead** LDA210-65		

**For additional information about specific MR Conditional systems, including MR Conditions of Use, warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the applicable MRI-Ready Systems Manual at medical.abbott/manuals.*

§ Not all device models are available in all countries.

† MC-2 RF Transmit Condition only applicable for UltiPace Pacing Lead configurations. Please refer to MRI-Ready Systems manual for specific device and lead combinations.

MRI-READY SYSTEMS BY ABBOTT

3T MRI-Ready Systems

Cardiac Resynchronization Therapy Pacemaker (CRT-P) Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted Left Ventricular Lead/Length (cm)	RF Transmit Conditions	Scan Region
Quadra Allure™ CRT-P PM3542 **Quadra Allure MP™ CRT-P** PM3562	**Tendril™ STS Pacing Lead** 2088TC-46, 52, 58 **UltiPace™ Pacing Lead** LPA1231-46, 52, 58, 65 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52 1948-52, 58	**Quartet™ LV Lead** 1456Q-86 1457Q-86 1458Q-86 1458QL-86	Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)	Full-body

Quadripolar-LV CRT-D Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted High Voltage Lead/Length (cm)	Left Ventricular (LV) Lead/Length (cm)	RF Transmit Conditions	Scan Region
Gallant™ HF (DF4, LV IS4) CRT-D CDHFA500Q	**Tendril™ STS Pacing Lead** 2088TC-46, 52 **UltiPace™ Pacing Lead** LPA1231-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65	**Quartet™ LV Lead** 1456Q-75 1456Q-86 1457Q-75 1457Q-86 1458Q-75 1458Q-86 1458Q-92 1458QL-75 1458QL-86	Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)	Full-body
Entrant™ HF (DF4, LV IS4) CRT-D CDHFA300Q					
Neutrino™ NxT HF (DF4, LV IS4) CRT-D CDHFA600Q					
Gallant™ HF (DF-1, LV IS4) CRT-D CDHFA500T		**Durata™ Defibrillation Lead** 7120-65 7122-60, 65 **Optisure™ Defibrillation Lead** LDA210-65			
Entrant™ HF (DF-1, LV IS4) CRT-D CDHFA300T					
Neutrino™ NxT HF (DF-1, LV IS4) CRT-D CDHFA600T					

Bipolar-LV CRT-D Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted High Voltage Lead/Length (cm)	RF Transmit Conditions	Scan Region
Gallant™ HF (DF4, LV IS-1) CRT-D CDHFA500D	**Tendril™ STS Pacing Lead** 2088TC-46, 52, 58 **UltiPace™ Pacing Lead** LPA1231-46, 52, 58, 65 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65	Normal Operating Mode Integrated Whole Body RF Transmit coil with RF excitation: – Circularly Polarized (CP), or – Multichannel-2 (MC-2)	Full-body
Entrant™ HF (DF4, LV IS-1) CRT-D CDHFA300D				
Neutrino™ NxT HF (DF4, LV IS-1) CRT-D CDHFA600D				
Gallant™ HF (DF-1, LV IS-1) CRT-D CDHFA500B		**Durata™ Defibrillation Lead** 7120- 65 7122-60, 65 **Optisure™ Defibrillation Lead** LDA210- 65		
Entrant™ HF (DF-1, LV IS-1) CRT-D CDHFA300B				
Neutrino™ NxT HF (DF-1, LV IS-1) CRT-D CDHFA600B				

**For additional information about specific MR Conditional systems, including MR Conditions of Use, warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the applicable MRI-Ready Systems Manual at medical.abbott/manuals.*

§ Not all device models are available in all countries.

MRI-READY SYSTEMS BY ABBOTT

1.5T MRI-Ready Systems

Insertable Cardiac Monitor (ICM) Systems*§

Implanted Device	RF Transmit Conditions	Scan Region
Assert-IQ™ ICM DM5500, DM5300, DM5000	First Level Controlled Operating Mode or Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body
Jot Dx™ ICM DM4500 **Confirm Rx™ ICM** DM3500	Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body

Pacemaker Systems*§

Implanted Device	Implanted Lead/Length (cm)	RF Transmit Conditions	Scan Region
AVEIR™ Leadless Pacemaker LSP112V LSP202V LSP201A	**Leadless**	First Level Controlled Operating Mode or Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body
Zenex™ MRI Pacemaker PM1282, PM2282 **Zenus™ MRI Pacemaker** PM1182, PM2182 **Assurity MRI™ Pacemaker** PM1272, PM2272 **Endurity MRI™ Pacemaker** PM1172, PM2172 **Endurity™ Pacemaker** PM1162, PM2162 **Endurity™ Core Pacemaker** PM1140, PM1152, PM2140, PM2152	**UltiPace™ Pacing Lead** LPA1231-46, 52, 58, 65 **Tendril MRI™ Pacing Lead†** LPA1200M-46, 52, 58 **Tendril™ STS Pacing Lead** 2088TC-46, 52, 58 **IsoFlex™ Optim™ Pacing Lead** 1944-46, 52 1948-52, 58	Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body

**For additional information about specific MR Conditional systems, including MR Conditions of Use, warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the applicable MRI-Ready Systems Manual at medical.abbott/manuals.*

§ Not all device models are available in all countries.

† First Level Controlled Operating Mode available for Zenex MRI, Zenus MRI, Assurity MRI, and Endurity MRI pacemakers with specific Tendril™ MRI Pacing Lead combinations. Please refer to MRI-Ready Systems Manual for specific device and lead combinations.

MRI-READY SYSTEMS BY ABBOTT

1.5T MRI-Ready Systems

Implantable Cardioverter Defibrillator (ICD) Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted High Voltage Lead/Length (cm)	RF Transmit Conditions	Scan Region
Neutrino™ NxT VR DF4 ICD CDVRA600Q **Neutrino™ NxT DR DF4 ICD** CDDRA600Q **Gallant™ VR DF4 ICD** CDVRA500Q **Gallant™ DR DF4 ICD** CDDRA500Q **Entrant™ VR DF4 ICD** CDVRA300Q **Entrant™ DR DF4 ICD** CDDRA300Q	**UltiPace™ Pacing Lead** LPA1231-46, 52 **Tendril MRI™ Pacing Lead** LPA1200M-46, 52 **Tendril™ STS Pacing Lead** 2088TC-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65	Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body
Neutrino™ NxT VR DF-1 ICD CDVRA600T **Neutrino™ NxT DR DF-1 ICD** CDDRA600T **Gallant™ VR DF-1 ICD** CDVRA500T **Gallant™ DR DF-1 ICD** CDDRA500T **Entrant™ VR DF-1 ICD** CDVRA300T **Entrant™ DR DF-1 ICD** CDDRA300T	**UltiPace™ Pacing Lead** LPA1231-46, 52 **Tendril™ STS Pacing Lead** 2088TC-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120-65 7122-60, 65 **Optisure™ Defibrillation Lead** LDA210-65		
Ellipse™ VR DF4 ICD CD1377-36Q CD1377-36QC **Ellipse™ DR DF4 ICD** CD2377-36Q** CD2377-36QC **Fortify Assura™ VR DF4 ICD** CD1359-40Q CD1359-40QC **Fortify Assura™ DR DF4 ICD** CD2359-40Q CD2359-40QC	**Tendril MRI™ Pacing Lead** LPA1200M-46, 52, 58 **Tendril™ STS Pacing Lead** 2088TC-46, 52 **IsoFlex™ Optim™ Pacing Lead** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65		

Cardiac Resynchronization Therapy Pacemaker (CRT-P) Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Left Ventricular (LV) Lead/Length (cm)	RF Transmit Conditions	Scan Region
Quadra Allure MP™ CRT-P PM3562 **Quadra Allure™ CRT-P** PM3542	**Tendril™ STS Pacing Lead** 2088TC-46, 52, 58 **IsoFlex™ Optim™ Pacing Lead** 1944-46, 52 1948-52, 58 **UltiPace™ Pacing Lead** LPA1231-46, 52, 58, 65	**Quartet™ LV Lead** 1456Q-86 1457Q-86 1458Q-86 1458QL-86	Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body

**For additional information about specific MR Conditional systems, including MR Conditions of Use, warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the applicable MRI-Ready Systems Manual at medical.abbott/manuals.*

§ Not all device models are available in all countries.

MRI-READY SYSTEMS BY ABBOTT

1.5T MRI-Ready Systems

Quadripolar-LV CRT-D Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted High Voltage Lead/Length (cm)	Left Ventricular (LV) Lead/ Length (cm)	RF Transmit Conditions	Scan Region
Gallant™ HF (DF4, LV IS4) CRT-D CDHFA500Q **Entrant™ HF (DF4, LV IS4) CRT-D** CDHFA300Q **Neutrino™ NxT HF (DF4, LV IS4) CRT-D** CDHFA600Q	**Tendril™ STS Pacing Lead** 2088TC-46, 52 **Tendril™ MRI** LPA1200M-46, 52 **UltiPace™ Pacing Lead** LPA1231-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65	**Quartet™ LV Lead** 1456Q-75 1456Q-86 1457Q-75 1457Q-86 1458Q-75 1458Q-86 1458Q-92 1458QL-75 1458QL-86	Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body
Gallant™ HF (DF-1, LV IS4) CRT-D CDHFA500T **Entrant™ HF (DF-1, LV IS4) CRT-D** CDHFA300T **Neutrino™ NxT HF (DF-1, LV IS4) CRT-D** CDHFA600T	**Tendril™ STS Pacing Lead** 2088TC-46, 52 **UltiPace™ Pacing Lead** LPA1231-46, 52 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120- 65 7122-60, 65 **Optisure™ Defibrillation Lead** LDA210- 65			

Bipolar-LV CRT-D Systems*§

Implanted Device	Implanted Low Voltage Lead/Length (cm)	Implanted High Voltage Lead/Length (cm)	RF Transmit Conditions	Scan Region
Gallant™ HF (DF4, LV IS-1) CRT-D CDHFA500D **Entrant™ HF (DF4, LV IS-1) CRT-D** CDHFA300D **Neutrino™ NxT HF (DF4, LV IS-1) CRT-D** CDHFA600D	**Tendril™ STS Pacing Lead** 2088TC-46, 52, 58 **UltiPace™ Pacing Lead** LPA1231-46, 52, 58, 65 **IsoFlex™ Optim™ Pacing Leads** 1944-46, 52	**Durata™ Defibrillation Lead** 7120Q-58, 65 7122Q-58, 65 **Optisure™ Defibrillation Lead** LDA210Q-58, 65 LDA220Q-58, 65	Normal Operating Mode Integrated Whole Body RF Transmit coil -or- Detachable RF Transmit-Receive coils (Head, Lower Extremity, or Upper Extremity) with RF excitation: – Circularly Polarized (CP)	Full-body
Gallant™ HF (DF-1, LV IS-1) CRT-D CDHFA500B **Entrant™ HF (DF-1, LV IS-1) CRT-D** CDHFA300B **Neutrino™ NxT HF (DF-1, LV IS-1) CRT-D** CDHFA600B		**Durata™ Defibrillation Lead** 7120- 65 7122-60, 65 **Optisure™ Defibrillation Lead** LDA210- 65		

*For additional information about specific MR Conditional systems, including MR Conditions of Use, warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the applicable MRI-Ready Systems Manual at medical.abbott/manuals.

§ Not all device models are available in all countries.

MRI-READY SYSTEMS BY ABBOTT

3T MRI Scan Parameters*§

When performing a 3T MRI scan on a patient with an Abbott MR Conditional system, the following scan parameters must be followed:

SCAN PARAMETERS	SETTING
Static Magnetic Field Strength (B_0) and Type of Nuclei	3 Tesla/128 MHz excitation frequency (hydrogen atom only)
Magnet Type and Static Magnetic Field (B_0) Orientation	Cylindrical-bore magnet, horizontal field orientation
Maximum Spatial Field Gradient (SFG)	24 T/m (2400 Gauss/cm) for Jot Dx™ and Confirm Rx™ ICM Systems 30 T/m (3000 Gauss/cm) for all other MR Conditional Systems
Maximum Gradient Slew Rate per Axis	200 T/m/s
RF Transmit Conditions	Refer to 3T MRI-Ready System tables above
Scan Region/Patient Landmarking Criteria	Any landmark is acceptable. Full-body scans are allowed.

1.5T MRI Scanner Parameters*§

When performing a 1.5T MRI scan on a patient with an Abbott MR Conditional system, the following scan parameters must be followed:

SCAN PARAMETERS	SETTING
Static Magnetic Field Strength (B_0) and Type of Nuclei	1.5 Tesla/64 MHz excitation frequency (hydrogen atom only)
Magnet Type and Static Magnetic Field (B_0) Orientation	Cylindrical-bore magnet, horizontal field orientation
Maximum Spatial Field Gradient (SFG)	24 T/m (2400 Gauss/cm) for Jot Dx™ and Confirm Rx™ ICM Systems 30 T/m (3000 Gauss/cm) for all other MR Conditional Systems
Maximum Gradient Slew Rate per Axis	200 T/m/s
RF Transmit Conditions	Refer to 1.5T MRI-Ready System tables above
Scan Region/Patient Landmarking Criteria	Any landmark is acceptable. Full-body scans are allowed.

**For additional information about specific MR Conditional systems, including MR Conditions of Use, warnings, precautions, adverse conditions to MRI scanning and potential adverse events, please refer to the applicable MRI-Ready Systems Manual at medical.abbott/manuals.*

§ Not all device models are available in all countries.

NOT SURE ABOUT A SPECIFIC MR CONDITIONAL SYSTEM?

Check out our MRI-Ready System Manuals online.

Brief Summary: This product is intended for use by or under the direction of a Physician. Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use.

™ Indicates a trademark of the Abbott group of companies.
† Indicates a third-party trademark, which is property of its respective owner.

Abbott
15900 Valley View Court, Sylmar, CA 91342
Tel: +1 818 362 6822
Abbott.com

© 2026 Abbott. All Rights Reserved.
MAT-2407586 v3.0 | Item approved for OUS use.
Products not approved in all geographies.

Abbott