

AURA | 16 kHz / 24-bit / 32-bit slots | NUCLEO-F411RE CN7/CN10


Bare INMP441 + PCM5102A + MAX97220A; selected replacement jack SJ-3523-SMT-TR

NUCLEO CN1 USB supply: JP5 pins 1-2 (U5V), JP1 open for enumerating host; 300 mA board+extensions max

Blockers: exact purchased modules, encoder footprint, MAX land pattern, power/current and listening-level validation



































MAX97220A pseudo-differential: RIN=47k, RF=4.7k, gain −0.1 V/V

SW6 = manual mute, open before attaching earbuds; PA8 controls enable through SW6



Place C1 at microphone; C2/C3 at DVDD; C7/C14 at CPVDD; C15/C16 at AVDD





Architecture schematic; see hardware/README.md for blockers

Sheet: /
File: AURA.kicad_sch

Title: AURA Hearing Assistance Hardware

Size: A2 | Date: | **Rev: A**

KiCad E.D.A. 9.0.8 | Id: 1/1