MZ Biolabs
2102 N Country Club Rd
Tucson, AZ 85716
contact@mzbiolabs.com
www.mzbiolabs.com

# Endotoxin Quantitative Kinetic Turbidimetric Assay Report of Analysis

## Product Information

**Client:** Evolve Peptides
**Product Tested:** CJC-1295 (No DAC) 5mg
**Lot Number:** 2947015836
**Vials Tested:** 2
**Test Performed:** PYROSTAR ES-F/Plate Kinetic Turbidimetric Assay
**Test Date:** 01/21/2026

## Reagent Information

**Lysate Product #:** WPEPK4-50015
**Lysate Lot #:** 122412A
**Lysate Exp:** 12/01/2026
**Lysate Sensitivity:** 0.01 – 10 EU/mL
**Control Standard Endotoxin (CSE) Product #:** WPEPK4-50015
**CSE Lot #:** G-07USA
**CSE Exp:** 09/01/2029

## KTA Test Results

| Solution | Endotoxin Concentration/Added to | Concentration (EU/ml) / % Recovery |
|---|---|---|
| Sample/Product Solution | None (EU/mL) | **A**: <0.079, **B**: <0.079<br>**Average:** <0.079 |
| Positive Product Control (PPC) | 1.0 EU/mL in Product | **A**: 0.718, **B**: 0.683<br>**Percent Recovery:** 70.067% |
| Negative Control | None (EU/mL) | **A**: <0.079, **B**: <0.079 |

**Interpretation:** This test is considered valid when: the positive product control % recovery is between 50% and 200%, and both negative control concentrations do not exceed the limit of the blank value required by the lysate reagent employed (0.079 EU/ml). When sample concentrations match the negative control, samples are considered free of endotoxin.

Sample resuspended in 2 ml endotoxin free water for testing. **Average EU/ml in test solution: <0.079**

## TOTAL Average Endotoxin EU in Sample: NEGATIVE

**Analyst Signature:** 