


a What each validation target can establish

Validation target	Strongest claim it supports	Does not establish on its own
Known motion (phantom, STRAUS, MRXCAT2.0)	representation capacity and technical error in the simulated regime	in-vivo accuracy
Paired DENSE or tagging	material-sensitive agreement under matched definitions	focal location or extent, unless measured
Scan-rescan	precision, not accuracy	bias
LGE or tissue characterisation	biological association, not motion error	pointwise displacement error
Outcome	prognostic association, not clinical utility	decision impact

b A staged programme for regional or focal claims



c Equal-mean test

