

	Anatomical	Rest
Manufacturer	Siemens	Siemens
Model	TrioTim	TrioTim
Headcoil	8 Chan	8 Chan
Field Strength	3T	3T
Sequence	3D MPRAGE	EPI
Flip Angle [Deg]	9.0	90
Inversion Time (TI) [ms]	900	-
Echo Time (TE) [ms]	2.52	30
Repetition Time (TR) [ms]	1900.0	2000
Bandwidth per Voxel (Readout) [Hz]	170	2232
Parallel Acquisition	GRAPPA x 2	-
Partial Fourier	Off	-
Number of Slices	176	32
Slice Orientation	Saggital	Axial
Slice Phase Encoding Direction	Anterior to Posterior	Anterior to Posterior
Slice Acquisition Order	Sequential Ascending	Interleaved Ascending
Slice Thickness [mm]	1.0	3.0
Slice Gap [mm]	0.5	0.99
Field of View [mm]	250	220
Acquisition Matrix	256x246	64x64
Slice In-Place Resolution [mm ²]	1.0x1.0	3.4x3.4
Number of Measurements	-	300
Acquisition Time [min:sec]	4:18	10:04
Fat Suppression	None	Yes
Number of Directions	-	-
Number of B Zeros	-	-
B Value(s) [s/mm ²]	-	-
Averages	-	-