A new angle on light-sheet microscopy and real-time image processing

Bálint Balázs



Pázmány Péter Catholic University

Faculty of Information Technology and Bionics
Roska Tamás Doctoral School of Sciences and Technology

Advisor: Balázs Rózsa, M.D., Ph.D.

Theses of the Ph.D. Dissertation Budapest, 2017

1 Introduction

Imaging techniques such as microscopy are one of the most extensively used tools in medical and biological research. The reason for this is simple: visualizing something invisible to the naked eye is an extremely powerful way to gain insight to its inner workings. Our brain has evolved to receive and process a multitude of signals from various sensors, and arguably the most powerful of these is vision.

As a branch of optics, microscopy (from ancient Greek mikros, "small" and skopein, "to see") is based on observing the interactions of light with an object of interest, such as a cell. To be able to see these interactions, the optics of the microscope magnifies the image of the sample, which can be recorded on a suitable device. For the first microscopes in the 17th century, this was just an eye at the end of the ocular, and the recording is a drawing of the observed image [1].

Microscopy is a truly multidisciplinary field: even in its simplest form, just using a single lens, the principles of physics are applied to gain a deeper understanding of biology and nature. Today, microscopy encompasses most of natural sciences and builds on various technological advancements. While physics and biology are still in the main focus, chemistry (fluorescent molecules), engineering (automation) and computer science (image analysis) are all integrated in a modern microscopy environment.

2 Challenges in three dimensional imaging of live specimens

Challenges in image acquisition

Live imaging is indispensable to understand the processes during embryonic development. In an ideal setting, the ultimate microscope would be able to record a continuous, three dimensional (3D), multicolor dataset of any biological process of interest with the highest possible resolution. Due to several limitations in physics and biology this is not possible. Therefore, a compromise is necessary. The diffractive nature of light, the lifetime of fluorescent probes and the photosensitivity of biological specimens all require microscopy to be able to adapt to answer the question at hand. In order to acquire useful data one has to choose a tradeoff between spatial and temporal resolution, and signal contrast, while making sure the biology is not affected by the imaging process itself (Figure 1a) [2].

Light-sheet fluorescence microscopy (LSFM), also called single plane illumination microscopy (SPIM) [3], is a relatively new addition to the arsenal of tools that comprise light microscopy methods, and is especially suitable for live imaging of embryonic samples over extended periods of time [4–7]. It is also easily adapted to the sample, allowing to image a large variety of specimens, from entire organs [8], to the subcellular processes occurring inside cultured cells [9].

Even though light-sheet microscopy offers intrinsic optical sectioning, when using a single detection objective, its axial resolution is worse than the lateral, similar to all other microscopy methods. To achieve isotropic 3D resolution, multiple views from different directions can

be combined [10]. This is most commonly achieved by embedding the sample in an aqueous gel which offers unrestricted view from multiple directions, and a stiff environment to position the sample [11]. Gel embedding, however is not always possible. Delicate samples, such as mouse embryos, require a very specific environment for proper development [12], which makes sample rotation impossible.




Figure 1: Challenges in microscopy. (a) Tradeoffs in fluorescence microscopy for live imaging. When optimizing the imaging conditions (red dot), a tradeoff has to be made between resolution, contrast, and imaging speed, while avoiding photodamage. Adapted from [2]. (b) Experiment sizes and data rate of different imaging modalities. Comparison of single-plane illumination microscopy (SPIM, red), high-content screening (light blue), single molecule localization microscopy (SMLM, orange) and confocal microscopy (blue) by typical experiment size and data production rate.

Challenges in image processing

When using any kind of microscopy in research, image processing is a crucial part of the workflow. This is especially true for light-sheet microscopy, since it is capable of imaging the same specimen for multiple

days, producing immense amounts of data. A single overnight experiment of *Drosophila* development (which is a very typical use-case for light-sheet microscopy) can produce multiple terabytes of data.

Apart from light-sheet microscopy, many other microscopy modalities are also suffering from this problem. Methods, such as high content screening [13–15], where tens of thousands of different genotypes are imaged generating millions of images; and single molecule localization microscopy (SMLM) [16–18], where just a single plane of a sample is imaged hundreds of thousands of times to acquire super-resolved images.

Not only these methods are capable of generating data extremely fast, but with the sustained high data rate a single experiment can easily reach multiples of terabytes (Figure 1b). Handling this amount of data can quickly become the bottleneck for many discoveries, which is a more and more common issue in biological research [19–21].

3 New scientific results

This work tackles the previously outlined challenges in light-sheet microscopy: high resolution live imaging of delicate samples, such as mouse embryos, and real-time image processing and compression of large light-sheet datasets.

Thesis I. I have designed and constructed a new light-sheet microscope suitable for high resolution imaging of delicate samples. A novel arrangement of two high numerical aperture objectives in 120 degrees combined with a tilted light-sheet allows for near isotropic resolution while increasing light collection efficiency by a factor of two.

Corresponding publications: [J1], [J2], [J3]

Dual Mouse-SPIM is a novel design for symmetric light-sheet microscopy. The use of high NA objectives in 120° not only increases volumetric resolution compared to the conventional 90° setup, but due to the larger detection angle, light collection efficiency is doubled. This is especially beneficial for delicate, light-sensitive specimens, such as mouse embryos, since phototoxic effects are reduced while the contrast is preserved.


I designed the optical path, the layout, and the custom mechanical components, and constructed the microscope. As part of the microscope I have designed a custom beam splitter unit that allows the use of a single galvanometric scanner to generate light-sheets for both objectives. I have also designed a custom detection merging unit that allows the use of a single camera for both detection views.

I have characterized the optical properties of the microscope, measured the illumination profile and point spread function. With a 3.6 μm thick light-sheet, a 95 μm field of view is evenly illuminated. Dual view imaging of bead samples revealed a lateral resolution of 314 nm, and axial resolution of 496 nm. This is a 2.67× improvement compared to the axial resolution of a single detection objective lens. I have also demonstrated the imaging capabilities of the microscope on *Drosophila melanogaster* embryos and mouse zygotes.

Thesis II. I have developed a GPU-based image processing pipeline for multi-view light-sheet microscopy that enables real-time fusion of opposing views.

Corresponding publications: [C1], [C2], [C3]

I have developed a GPU-based image preprocessing pipeline, which






Figure 2: Dual Mouse-SPIM concept. (a) Objective configuration. Both objectives can be used sequentially for both illumination and detection. (b) Combined PSF of the two views. Average of 12 fluorescent beads. Scale bar, 1 µm.

integrates directly to our universal microscope control software in Lab-VIEW. The pipeline currently supports background subtraction and background masking, furthermore it is capable of fusing opposing views of the same plane faster than real-time. I have shown that it is possible to reduce the registration of opposing camera views from a 3D alignment to a 2D alignment without any negative effects in image quality and resolution. This massively reduces the necessary computing resources, and allows the use of CUDA textures for faster than real-time image fusion. Processing speed of this implementation is $138\,\mathrm{fps}$, a $18.3\times$ increase compared to a single threaded CPU implementation.

Thesis Group III. Real-time image compression.

Thesis III/1. I have developed a new image compression algorithm that enables noise dependent lossy compression of light microscopy images, and can reach a compression ratio of 100 fold while




Figure 3: Real-time image processing pipeline for multi-view light-sheet microscopy.

preserving the results of downstream data analysis steps. A fast CUDA implementation allows for real-time image compression of high-speed microscopy images.

Corresponding publications: [J4], [C1], [C2], [C3]

Since many high-speed microscopy methods generate immense amounts of data, easily reaching terabytes per experiment, image compression is especially important to efficiently deal with such datasets. Existing compression methods suitable for microscopy images are not able to deal with the high data rate of modern sCMOS cameras ($\sim 800\,\mathrm{MB/s}$).

I developed a GPU-based parallel image compression algorithm called B³D, capable of over 1 GB/s throughput, allowing live image compression. To further reduce the data size, I developed a noise dependent lossy compression that only modifies the data in a deterministic manner. The allowed differences for each pixel can be specified as a proportion of the inherent image noise, accounting for photon shot noise and camera readout noise. Due to the use of pixel prediction, the

subjective image quality is higher than for other methods that simply quantize the square root of the images.

Thesis III/2. I have shown that within noise level compression does not significantly affect the results of most commonly used image processing tasks, and it allows a $3.32\times$ average increase in compression ratio compared to lossless mode.

Corresponding publications: [J4], [C1], [C2], [C3]

As data integrity in microscopy is paramount for drawing the right conclusions from the experiments, using a lossy compression algorithm might be controversial. I have shown that the within noise level (WNL) mode of $\rm B^3D$ does not significantly affect the results of several commonly used image analysis tasks. For light-sheet microscopy data I have shown that WNL compression introduces less variation to the image than the photon shot noise. When segmenting nuclei of Drosophila embryos and membranes of Phallusia embryos, the overlap of the segmented regions of uncompressed and WNL compressed datasets were 99.6% and 94.5% respectively, while compression ratios were 19.83 for Drosophila and 40.01 for Phallusia embryos.

For single molecule localization microscopy data I have shown that WNL compression only introduces 4% increase in localization uncertainty, while the average compression ratio is increased from 1.44 (lossless) to 4.96 (WNL). I have also shown that change in localization error due to the compression does not depend on the SNR of the input images.




Figure 4: Influence of noise dependent lossy compression on 3D nucleus segmentation. A Drosophila melanogaster embryo expressing H2Av-mCherry nuclear marker was imaged in MuVi-SPIM [11], and 3D nucleus segmentation was performed. (a). To visualize segmentation mismatch, the results of the uncompressed (green) and compressed (magenta) datasets are overlaid in a single image (b, c; overlap in white). For all compression levels the segmentation overlap score was calculated and is plotted in (d) along with the achieved compression ratios.

4 Application of the results

Both the new Dual Mouse-SPIM microscope and the GPU-based image processing and compression pipeline have direct applications in lightsheet imaging of embryonic development.

Multiple potential collaborators indicated their interest in using the Dual Mouse-SPIM for their studies in mouse embryonic development. In the context of the research of symmetry breaking events in the pre-implantation and early post-implantation stages, this system can be used for imaging larger specimens from multiple direction, which is not possible on previous microscopes, and could allow to observe previously unknown mechanisms. Another possible application is investigating chromosome missesgregation mechanisms in the first few divisions during embryonic development. The increased axial resolution of this system will allow to track each individual chromosome during the division process, which is not possible on current microscopes due to the insufficient axial resolution.

The GPU-based image processing pipeline, especially the 2D fusion of opposing views is already being used on our lab's workhorse microscope, the MuVi-SPIM. Being able to fuse the two views of the opposing objectives during imaging not only results in considerable storage space savings, but significantly speeds up the data analysis as well.

The image compression algorithm, B³D, although was developed with light-sheet microscopy in mind, has a more wide-spread use-case. Any kind of high-speed, high-throughput light-microscopy experiment can benefit from the massive data reduction offered by the within-noise-level mode. Since the compression can also be done immediately during imaging, not only the storage requirements, but the data bandwidth is reduced as well, which renders the use of high performance RAID arrays and 10 Gbit networks unnecessary, further reducing costs. Due to the similarly high decompression speed, reading the data is also accelerated, which can be beneficial for data browsing and 3D rendering applications. Several companies of different fields already expressed their interest in the compression library, including Bitplane AG (3D data analysis and visualization), Luxendo GmbH (light-sheet microscopy), and Hamamatsu Photonics K.K. (camera and sensor manufacturing).

The author's publications

[J1] Gustavo de Medeiros, Bálint Balázs, and Lars Hufnagel. "Light-sheet imaging of mammalian development". Seminars in Cell & Developmental Biology. Mammalian development 55 (July 2016), pp. 148–155. DOI: 10.1016/j.semcdb.2015.11.001 (cit. on p. 5).

- [J2] Petr Strnad, Stefan Gunther, Judith Reichmann, Uros Krzic, Balint Balazs, Gustavo de Medeiros, Nils Norlin, Takashi Hiiragi, Lars Hufnagel, and Jan Ellenberg. "Inverted light-sheet microscope for imaging mouse pre-implantation development". Nature Methods 13.2 (Feb. 2016), pp. 139–142. DOI: 10.1038/nmeth. 3690 (cit. on p. 5).
- [J3] Patrick Hoyer, Gustavo de Medeiros, Bálint Balázs, Nils Norlin, Christina Besir, Janina Hanne, Hans-Georg Kräusslich, Johann Engelhardt, Steffen J. Sahl, Stefan W. Hell, and Lars Hufnagel. "Breaking the diffraction limit of light-sheet fluorescence microscopy by RESOLFT". Proceedings of the National Academy of Sciences 113.13 (Mar. 2016), pp. 3442–3446. DOI: 10.1073/pnas.1522292113 (cit. on p. 5).
- [J4] Balint Balazs, Joran Deschamps, Marvin Albert, Jonas Ries, and Lars Hufnagel. "A real-time compression library for microscopy images". bioRxiv (July 2017), p. 164624. DOI: 10.1101/164624 (cit. on pp. 7, 8).

The author's others publications

[J5] Zoltán Jakus, Edina Simon, Bálint Balázs, and Attila Mócsai. "Genetic deficiency of Syk protects mice from autoantibody-induced arthritis". Arthritis and Rheumatism 62.7 (July 2010), pp. 1899–1910. DOI: 10.1002/art.27438.

- [J6] Balázs Győrffy, Zsombor Benke, András Lánczky, Bálint Balázs, Zoltán Szállási, József Timár, and Reinhold Schäfer. "RecurrenceOnline: an online analysis tool to determine breast cancer recurrence and hormone receptor status using microarray data". Breast Cancer Research and Treatment (July 2011). DOI: 10.1007/s10549-011-1676-y.
- [J7] Weiwei Shi, Balint Balazs, Balazs Györffy, Tingting Jiang, W. Fraser Symmans, Christos Hatzis, and Lajos Pusztai. "Combined analysis of gene expression, DNA copy number, and mutation profiling data to display biological process anomalies in individual breast cancers". Breast Cancer Research and Treatment 144.3 (Mar. 2014), pp. 561–568. DOI: 10.1007/s10549-014-2904-z.

The author's conference presentations

- [C1] Bálint Balázs, Marvin Albert, and Lars Hufnagel. "GPU-based image processing for multiview microscopy data". Light Sheet Fluorescence Microscopy International Conference. Sheffield, UK, Sept. 2016 (cit. on pp. 5, 7, 8).
- [C2] Bálint Balázs, Marvin Albert, and Lars Hufnagel. "GPU-based image processing for multi-view microscopy data". Focus on Microscopy. Taipei, Taiwan, Mar. 2016 (cit. on pp. 5, 7, 8).
- [C3] Bálint Balázs, Marvin Albert, and Lars Hufnagel. "GPU-based image processing for multi-view microscopy data". Focus on Microscopy. Bordeaux, France, Apr. 2017 (cit. on pp. 5, 7, 8).

References cited in the thesis

- [1] Robert Hooke. Micrographia: or Some Physiological Descriptions of Minute Bodies Made by Magnifying Glasses. With Observations and Inquiries Thereupon. London: J. Martyn and J. Allestry, 1665 (cit. on p. 1).
- [2] P. Philippe Laissue, Rana A. Alghamdi, Pavel Tomancak, Emmanuel G. Reynaud, and Hari Shroff. "Assessing phototoxicity in live fluorescence imaging". *Nature Methods* 14.7 (July 2017), pp. 657–661. DOI: 10.1038/nmeth.4344 (cit. on pp. 2, 3).
- [3] Jan Huisken, Jim Swoger, Filippo Del Bene, Joachim Wittbrodt, and Ernst H. K Stelzer. "Optical Sectioning Deep Inside Live Embryos by Selective Plane Illumination Microscopy". *Science* 305.5686 (2004), pp. 1007–1009 (cit. on p. 2).
- [4] Philipp J Keller and Ernst HK Stelzer. "Quantitative in vivo imaging of entire embryos with Digital Scanned Laser Light Sheet Fluorescence Microscopy". Current Opinion in Neurobiology 18.6 (Dec. 2008). review, pp. 624–632. DOI: 10.1016/j.conb.2009.03.008 (cit. on p. 2).
- [5] Jan Huisken and Didier Y. R. Stainier. "Selective plane illumination microscopy techniques in developmental biology". *Development* 136.12 (June 2009), pp. 1963–1975. DOI: 10.1242/dev. 022426 (cit. on p. 2).
- [6] Michael Weber and Jan Huisken. "Light sheet microscopy for real-time developmental biology". Current Opinion in Genetics & Development 21.5 (2011), pp. 566-572. DOI: 10.1016/j.gde. 2011.09.009 (cit. on p. 2).

- [7] Raju Tomer, Khaled Khairy, and Philipp J Keller. "Shedding light on the system: studying embryonic development with light sheet microscopy". Current Opinion in Genetics & Development. Developmental mechanisms, patterning and evolution 21.5 (Oct. 2011). Review from 2011, pp. 558–565. DOI: 10.1016/j.gde. 2011.07.003 (cit. on p. 2).
- [8] Hans-Ulrich Dodt, Ulrich Leischner, Anja Schierloh, Nina Jährling, Christoph Peter Mauch, Katrin Deininger, Jan Michael Deussing, Matthias Eder, Walter Zieglgänsberger, and Klaus Becker. "Ultramicroscopy: three-dimensional visualization of neuronal networks in the whole mouse brain". Nature Methods 4.4 (Mar. 2007), pp. 331–336. DOI: 10.1038/nmeth1036 (cit. on p. 2).
- [9] Bi-Chang Chen et al. "Lattice light-sheet microscopy: Imaging molecules to embryos at high spatiotemporal resolution". *Science* 346.6208 (Oct. 2014), p. 1257998. DOI: 10.1126/science. 1257998 (cit. on p. 2).
- [10] Stephan Preibisch, Fernando Amat, Evangelia Stamataki, Mihail Sarov, Robert H. Singer, Eugene Myers, and Pavel Tomancak. "Efficient Bayesian-based multiview deconvolution". *Nature Methods* 11.6 (June 2014), pp. 645–648. DOI: 10.1038/nmeth. 2929 (cit. on p. 3).
- [11] Uros Krzic, Stefan Gunther, Timothy E. Saunders, Sebastian J. Streichan, and Lars Hufnagel. "Multiview light-sheet microscope for rapid in toto imaging". *Nature Methods* 9.7 (July 2012). MuVi-SPIM, pp. 730–733. DOI: 10.1038/nmeth.2064 (cit. on pp. 3, 9).

- [12] Adam S. Doherty and Richard M. Schultz. "Culture of Preimplantation Mouse Embryos". Developmental Biology Protocols. Methods in Molecular BiologyTM. DOI: 10.1385/1-59259-685-1:47. Humana Press, 2000, pp. 47–52 (cit. on p. 3).
- [13] Anne E. Carpenter and David M. Sabatini. "Systematic genome-wide screens of gene function". *Nature Reviews Genetics* 5.1 (Jan. 2004), pp. 11–22. DOI: 10.1038/nrg1248 (cit. on p. 4).
- [14] Christophe J. Echeverri and Norbert Perrimon. "High-throughput RNAi screening in cultured cells: a user's guide". Nature Reviews Genetics 7.5 (May 2006), pp. 373–384. DOI: 10.1038/nrg1836 (cit. on p. 4).
- [15] Rainer Pepperkok and Jan Ellenberg. "High-throughput fluorescence microscopy for systems biology". Nature Reviews Molecular Cell Biology 7.9 (Sept. 2006), pp. 690–696. DOI: 10.1038/nrm1979 (cit. on p. 4).
- [16] Eric Betzig, George H. Patterson, Rachid Sougrat, O. Wolf Lindwasser, Scott Olenych, Juan S. Bonifacino, Michael W. Davidson, Jennifer Lippincott-Schwartz, and Harald F. Hess. "Imaging Intracellular Fluorescent Proteins at Nanometer Resolution". Science 313.5793 (Sept. 2006), pp. 1642–1645. DOI: 10.1126/science.1127344 (cit. on p. 4).
- [17] Samuel T. Hess, Thanu P. K. Girirajan, and Michael D. Mason. "Ultra-High Resolution Imaging by Fluorescence Photoactivation Localization Microscopy". *Biophysical Journal* 91.11 (Dec. 2006), pp. 4258–4272. DOI: 10.1529/biophysj.106.091116 (cit. on p. 4).

- [18] Michael J. Rust, Mark Bates, and Xiaowei Zhuang. "Sub-diffraction-limit imaging by stochastic optical reconstruction microscopy (STORM)". *Nature Methods* 3.10 (Oct. 2006), pp. 793–796. DOI: 10.1038/nmeth929 (cit. on p. 4).
- [19] Roy Wollman and Nico Stuurman. "High throughput microscopy: from raw images to discoveries". *Journal of Cell Science* 120.21 (Nov. 2007), pp. 3715–3722. DOI: 10.1242/jcs.013623 (cit. on p. 4).
- [20] Emmanuel G. Reynaud, Jan Peychl, Jan Huisken, and Pavel Tomancak. "Guide to light-sheet microscopy for adventurous biologists". *Nature Methods* 12.1 (Jan. 2015), pp. 30–34. DOI: 10.1038/nmeth.3222 (cit. on p. 4).
- [21] Jeffrey M. Perkel. "The struggle with image glut". Nature 533.7601 (Apr. 2016), pp. 131–132. DOI: 10.1038/533131a (cit. on p. 4).