Reported Performance of Published DME Classification Systems and the Present 2D OCT Model



Not a head-to-head comparison: every row uses a different dataset, population, device, label definition and evaluation protocol.
Tang 2021 'DME' includes retinal thickening and hard exudates in diabetic eyes; this project's 'DME vs Normal' uses Kermany folder labels. Square markers = volume-level results (different evaluation unit; context only).
The x-axis starts at 0.80 to resolve the published intervals; no statistical test
across studies was performed. Volume-level rows show primary datasets only; all external results are listed in Table 3.
Li 2019 CIs are printed values, narrower than binomial for its n (Table 3 notes); DMENet (Singh & Gorantla 2020) uses fundus photographs and is shown as a different-modality reference only.
This project's official-test AUROC has no meaningful CI (perfect separation).